

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency	CORAL SPRINGS PD	Date	01/27/2023	Type	Agency Inspection		
Inspector	SCHAUT,RICHARD,J	Time	14:59:42	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202106B	0.08 Gas Lot #	06021080A2
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/22/2023	Expiration	04/05/2023
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/27/2023 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency CORAL SPRINGS PD **Date** 02/17/2023 **Type** Agency Inspection
Inspector SCHAUT,RICHARD,J **Time** 14:50:04 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06021080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/17/2023 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency CORAL SPRINGS PD **Date** 03/30/2023 **Type** Agency Inspection
Inspector SCHAUT,RICHARD,J **Time** 14:26:50 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06021080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/04/2023 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency	CORAL SPRINGS PD	Date	04/07/2023	Type	Agency Inspection		
Inspector	SCHAUT,RICHARD,J	Time	11:20:51	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	26422080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	11/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/07/2023 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency CORAL SPRINGS PD **Date** 05/30/2023 **Type** Agency Inspection
Inspector SCHAUT,RICHARD,J **Time** 14:43:02 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2023 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency CORAL SPRINGS PD **Date** 06/06/2023 **Type** Agency Inspection
Inspector SCHAUT,RICHARD,J **Time** 14:42:24 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: RESIDUAL MA FROM 2ND.

Data Download Date/Time 06/06/2023 14:51

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency	CORAL SPRINGS PD	Date	07/26/2023	Type	Agency Inspection		
Inspector	SCHAUT,RICHARD,J	Time	15:20:10	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	26422080A1
Expiration	03/28/2025	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	11/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/26/2023 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency CORAL SPRINGS PD
Inspector SCHAUT,RICHARD,J

Date 08/18/2023
Time 09:52:27

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/18/2023 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency CORAL SPRINGS PD **Date** 09/22/2023 **Type** Agency Inspection
Inspector SCHAUT,RICHARD,J **Time** 15:09:33 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/22/2023 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency CORAL SPRINGS PD **Date** 10/10/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:29:05 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/18/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.077
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.077
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.077
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0010
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 10/10/2023 14:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency	CORAL SPRINGS PD	Date	10/19/2023	Type	Agency Inspection		
Inspector	SCHAUT,RICHARD,J	Time	14:56:10	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	26422080A1
Expiration	03/28/2025	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	11/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.08 g/210L Test # (Repeat) 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance. .08 SIM LID NOT SEATED CORR

Data Download Date/Time 10/19/2023 15:07

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency CORAL SPRINGS PD **Date** 11/21/2023 **Type** Agency Inspection
Inspector SCHAUT,RICHARD,J **Time** 16:12:17 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 26422080A1
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/28/2023 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency CORAL SPRINGS PD
Inspector WHISNANT,JASON,C

Date 12/19/2023
Time 10:14:57

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 26422080A1
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/19/2023 10:24

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

Agency ST PETERSBURG PD **Date** 02/06/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 13:49:37 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0007 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.203
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.203
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.204
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.052
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.204
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.052
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.204
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0011
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/06/2023 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

Agency	ST PETERSBURG PD	Date	02/23/2023	Type	Agency Inspection		
Inspector	MCCULLOUGH,SEAN,J	Time	09:32:35	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	08121080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	05/05/2023
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/23/2023 10:54

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency NASSAU CO SO
Inspector KUTZ, THOMAS, E

Date 01/24/2023
Time 08:08:22

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/24/2023 23:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency NASSAU CO SO	Date 02/16/2023	Type Agency Inspection	
Inspector KUTZ,THOMAS,E	Time 08:36:23	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 26422080A1
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 11/05/2024
Number of Simulators 4	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/21/2023 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency NASSAU CO SO **Date** 03/19/2023 **Type** Agency Inspection
Inspector HIGGINBOTHAM,BRIAN,K **Time** 20:59:12 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/19/2023 22:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency	NASSAU CO SO	Date	04/25/2023	Type	Agency Inspection		
Inspector	HIGGINBOTHAM,BRIAN,K	Time	01:19:08	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	26422080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	11/05/2024
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/25/2023 01:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency	NASSAU CO SO	Date	05/23/2023	Type	Agency Inspection		
Inspector	HIGGINBOTHAM,BRIAN,K	Time	20:49:25	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	26422080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	11/05/2024
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/23/2023 21:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency NASSAU CO SO **Date** 06/26/2023 **Type** Agency Inspection
Inspector HIGGINBOTHAM,BRIAN,K **Time** 23:51:22 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/27/2023 00:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency	NASSAU CO SO	Date	07/27/2023	Type	Agency Inspection		
Inspector	HIGGINBOTHAM,BRIAN,K	Time	18:43:10	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	26422080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	11/05/2024
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/27/2023 18:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency NASSAU CO SO	Date 08/23/2023	Type Agency Inspection	
Inspector HIGGINBOTHAM, BRIAN, K	Time 01:12:16	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 26422080A1
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 11/05/2024
Number of Simulators 4	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:SIM RETURN HOSE NOT FULLY SEALED/CORRECTED SIM HOSE

Data Download Date/Time 08/23/2023 01:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency NASSAU CO SO
Inspector KUTZ, THOMAS, E

Date 09/23/2023
Time 11:13:41

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/23/2023 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency	NASSAU CO SO	Date	10/22/2023	Type	Agency Inspection		
Inspector	HIGGINBOTHAM,BRIAN,K	Time	21:28:25	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	26422080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	11/05/2024
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/22/2023 21:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency NASSAU CO SO **Date** 11/21/2023 **Type** Agency Inspection
Inspector HIGGINBOTHAM,BRIAN,K **Time** 19:45:53 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A:DID NOT READ SLOPE NOT MET . Non-compliance:DID NOT READ SLOPE NOT MET.

Data Download Date/Time 11/21/2023 20:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency NASSAU COUNTY SO **Date** 12/07/2023 **Type** Department Inspection
Inspector SIDDOWAY,BENJAMIN,W **Time** 13:47:55 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.077
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.077
		0.20 g/210L Test # 5	0.196
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.077
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

A F / M A: No Sample Provided.

Data Download Date/Time 12/07/2023 13:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency	NASSAU COUNTY SO	Date	12/25/2023	Type	Agency Inspection		
Inspector	HIGGINBOTHAM,BRIAN,K	Time	22:32:33	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202307A	0.20 Lot #	202206C	0.08 Gas Lot #	26422080A1
Expiration	03/28/2025	Expiration	07/11/2025	Expiration	06/15/2024	Expiration	11/05/2024
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/25/2023 23:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency	POLK COUNTY SO	Date	09/19/2023	Type	Agency Inspection
Inspector	SOTO,ISRAEL,	Time	09:58:11	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 09/21/2023 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency POLK COUNTY SO	Date 09/21/2023	Type Department Inspection	
Inspector SOTO,ISRAEL,	Time 10:13:11	Software 8100.27	
0.05 Lot # 202303K	0.08 Lot # 202303L	0.20 Lot # 202304C	0.08 Gas Lot # 06723080A5
Expiration 03/29/2025	Expiration 03/29/2025	Expiration 04/05/2025	Expiration 04/05/2025
Number of Simulators 5	Average Standard Deviation 0.0003	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0007
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 09/21/2023 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD
Inspector STRUNK,JOHN,D

Date 01/23/2023
Time 09:34:48

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1283438
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 02/25/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/23/2023 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD Date 02/16/2023 Type Agency Inspection
Inspector STRUNK,JOHN,D Time 09:12:08 Software 8100.27

0.05 Lot # 202112A 0.08 Lot # 202206B 0.20 Lot # 202206C 0.08 Gas Lot # 1497672
Expiration 12/07/2023 Expiration 06/14/2024 Expiration 06/15/2024 Expiration 05/20/2024

Number of Simulators 5 Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/16/2023 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD **Date** 04/05/2023 **Type** Agency Inspection
Inspector FARQUHARSON,KEVIN,D **Time** 11:55:13 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted.

Data Download Date/Time 05/31/2023 21:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD **Date** 04/05/2023 **Type** Agency Inspection
Inspector FARQUHARSON,KEVIN,D **Time** 12:53:46 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1497672
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/20/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2023 21:40

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD **Date** 05/31/2023 **Type** Agency Inspection
Inspector FARQUHARSON,KEVIN,D **Time** 21:33:16 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 201893
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/20/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2023 21:41

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD **Date** 06/29/2023 **Type** Agency Inspection
Inspector FORTEZA,CRYSTAL,L **Time** 14:32:06 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1497672
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/20/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/29/2023 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD **Date** 07/26/2023 **Type** Agency Inspection
Inspector HAINES, DAVID, K **Time** 12:30:05 **Software** 8100.27

0.05 Lot # 20211A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1497672
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/20/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/26/2023 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD	Date 08/08/2023	Type Agency Inspection	
Inspector HAINES, DAVID, K	Time 10:36:10	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 1497672
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 05/20/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Sequence Aborted.

Data Download Date/Time 08/08/2023 10:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD	Date 09/02/2023	Type Agency Inspection	
Inspector HAINES, DAVID, K	Time 09:31:59	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 1497672
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 05/20/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

TAKING OUT OF SERVICE FOR ANNUAL INSPECTION

Data Download Date/Time 09/02/2023 09:48

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD **Date** 09/06/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:39:25 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.203
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.203
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 09/06/2023 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD	Date 09/12/2023	Type Agency Inspection	
Inspector HAINES, DAVID, K	Time 12:05:56	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 1497672
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 05/20/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 09/12/2023 12:12

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD	Date 10/31/2023	Type Agency Inspection	
Inspector HAINES, DAVID, K	Time 09:43:41	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 1497672
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 05/20/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/31/2023 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD	Date 11/20/2023	Type Agency Inspection	
Inspector HAINES, DAVID, K	Time 13:57:11	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 1497672
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 05/20/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/17/2023 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency DAVIE PD **Date** 12/26/2023 **Type** Agency Inspection
Inspector HAINES, DAVID, K **Time** 10:21:56 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1497672
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/20/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/26/2023 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD	Date 01/25/2023	Type Agency Inspection	
Inspector ALMEIDA,CAMILO,J	Time 09:09:39	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202106B	0.08 Gas Lot # 11821080A2
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/22/2023	Expiration 06/05/2023
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/25/2023 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD **Date** 02/27/2023 **Type** Agency Inspection
Inspector ALMEIDA,CAMILO,J **Time** 10:42:42 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/27/2023 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD	Date 03/21/2023	Type Agency Inspection	
Inspector ALMEIDA,CAMILO,J	Time 12:42:59	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 11821080A2
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 06/05/2023
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/21/2023 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD	Date 04/19/2023	Type Agency Inspection	
Inspector ALMEIDA,CAMILO,J	Time 13:02:12	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 11821080A2
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 06/05/2023
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance. - SEE HANDWRITTEN NOTE

Data Download Date/Time 04/20/2023 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD **Date** 05/30/2023 **Type** Agency Inspection
Inspector ALMEIDA,CAMILO,J **Time** 10:54:41 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06723080A5
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2023 11:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD	Date 06/20/2023	Type Agency Inspection	
Inspector ALMEIDA,CAMILO,J	Time 09:14:04	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 06723080A5
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 04/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/20/2023 09:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD	Date 06/29/2023	Type Department Inspection	
Inspector SOTO,ISRAEL,	Time 14:36:51	Software 8100.27	
0.05 Lot # 202303K	0.08 Lot # 202303L	0.20 Lot # 202304C	0.08 Gas Lot # AG229803
Expiration 03/29/2025	Expiration 03/29/2025	Expiration 04/05/2025	Expiration 10/25/2024
Number of Simulators 5	Average Standard Deviation 0.0002	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.196
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.196
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 06/29/2023 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD	Date 07/31/2023	Type Agency Inspection	
Inspector ALMEIDA,CAMILO,J	Time 09:54:09	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 06723080A5
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 04/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/31/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD	Date 08/21/2023	Type Agency Inspection	
Inspector ALMEIDA,CAMILO,J	Time 09:30:21	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 06723080A5
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 04/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/21/2023 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD	Date 09/25/2023	Type Agency Inspection	
Inspector ALMEIDA,CAMILO,J	Time 12:42:54	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 06723080A5
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 04/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/25/2023 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD	Date 10/30/2023	Type Agency Inspection	
Inspector ALMEIDA,CAMILO,J	Time 09:24:01	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 06723080A5
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 04/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/30/2023 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD **Date** 11/28/2023 **Type** Agency Inspection
Inspector ALMEIDA,CAMILO,J **Time** 15:06:25 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06723080A5
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/29/2023 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency LAKELAND PD **Date** 12/18/2023 **Type** Agency Inspection
Inspector ALMEIDA,CAMILO,J **Time** 11:58:58 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06723080A5
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/18/2023 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

Agency POLK CSO	Date 09/19/2023	Type Agency Inspection
Inspector SOTO,ISRAEL,	Time 09:58:33	Software 8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance Yes
-----------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 09/21/2023 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

Agency POLK CSO	Date 09/21/2023	Type Department Inspection	
Inspector SOTO,ISRAEL,	Time 10:28:45	Software 8100.27	
0.05 Lot # 202303K	0.08 Lot # 202303L	0.20 Lot # 202304C	0.08 Gas Lot # 06723080A5
Expiration 03/29/2025	Expiration 03/29/2025	Expiration 04/05/2025	Expiration 04/05/2025
Number of Simulators 5	Average Standard Deviation 0.0004	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 09/21/2023 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency FERNANDINA BEACH PD **Date** 02/24/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 10:57:27 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0011
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/24/2023 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency FERNANDINA BEACH PD **Date** 03/13/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 06:55:54 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22722080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/13/2023 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency	FERNANDINA BEACH PD	Date	04/23/2023	Type	Agency Inspection		
Inspector	SMITH,JASON,E	Time	16:01:19	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	22722080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	10/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/24/2023 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency	FERNANDINA BEACH PD	Date	05/11/2023	Type	Agency Inspection		
Inspector	SMITH,JASON,E	Time	17:54:35	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	22722080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	10/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/11/2023 18:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency FERNANDINA BEACH PD **Date** 06/27/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 21:19:34 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22722080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/27/2023 21:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency	FERNANDINA BEACH PD	Date	07/23/2023	Type	Agency Inspection		
Inspector	SMITH,JASON,E	Time	18:47:08	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	22722080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	10/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/23/2023 18:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency	FERNANDINA BEACH PD	Date	08/26/2023	Type	Agency Inspection		
Inspector	SMITH,JASON,E	Time	04:44:19	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	22722080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	10/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.05 g/210L Test # (Repeat) 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.

Data Download Date/Time 08/26/2023 04:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency	FERNANDINA BEACH PD	Date	09/25/2023	Type	Agency Inspection		
Inspector	SMITH,JASON,E	Time	00:57:09	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	22722080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	10/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 09/25/2023 01:05

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency	FERNANDINA BEACH PD	Date	10/17/2023	Type	Agency Inspection		
Inspector	SMITH,JASON,E	Time	19:06:05	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	22722080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	10/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/17/2023 19:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency	FERNANDINA BEACH PD	Date	11/13/2023	Type	Agency Inspection		
Inspector	SMITH,JASON,E	Time	18:10:14	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202307A	0.20 Lot #	202206C	0.08 Gas Lot #	22722080A1
Expiration	03/28/2025	Expiration	07/11/2025	Expiration	06/15/2024	Expiration	10/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/13/2023 18:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

Agency FERNANDINA BEACH PD **Date** 12/18/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 16:33:05 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 22722080A1
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/18/2023 16:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD
Inspector WOMACK,ERIC,

Date 01/30/2023
Time 15:14:38

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08121080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2023

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2023 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD
Inspector WOMACK,ERIC,

Date 02/28/2023
Time 08:50:42

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08121080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2023

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 02/28/2023 09:43

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD
Inspector WOMACK, ERIC,

Date 04/10/2023
Time 15:39:21

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08121080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2023

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/10/2023 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD
Inspector WOMACK,ERIC,

Date 05/08/2023
Time 13:32:54

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/08/2023 13:40

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD
Inspector WOMACK, ERIC,

Date 06/01/2023
Time 09:29:07

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/01/2023 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD
Inspector WOMACK,ERIC,

Date 07/06/2023
Time 14:58:33

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/06/2023 15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD
Inspector WOMACK, ERIC,

Date 08/03/2023
Time 11:07:55

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/03/2023 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD
Inspector WOMACK,ERIC,

Date 09/06/2023
Time 11:21:57

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/06/2023 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD
Inspector WOMACK,ERIC,

Date 10/05/2023
Time 11:33:50

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:MA NOT READ.

Data Download Date/Time 10/05/2023 11:43

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD **Date** 10/12/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 09:36:40 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 10/12/2023 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD
Inspector WOMACK,ERIC,V

Date 10/25/2023
Time 12:16:06

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/25/2023 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD
Inspector WOMACK,ERIC,

Date 11/27/2023
Time 09:14:19

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/27/2023 09:25

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency WINTER HAVEN PD
Inspector WOMACK,ERIC,

Date 12/28/2023
Time 10:53:49

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.190
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.190
		0.20 g/210L Test # (Repeat) 3	0.191
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 12/28/2023 11:04

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency GULFPORT PD **Date** 01/24/2023 **Type** Agency Inspection
Inspector MAROTTA, MICHAEL, S **Time** 23:42:20 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 22521080A1
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/24/2023 23:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency	GULFPORT PD	Date	02/27/2023	Type	Agency Inspection		
Inspector	MAROTTA,MICHAEL,S	Time	01:41:27	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	22521080A1
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	10/05/2023
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2023 01:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency	GULFPORT PD	Date	03/28/2023	Type	Agency Inspection		
Inspector	MAROTTA,MICHAEL,S	Time	00:24:14	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	22521080A1
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	10/05/2023
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/28/2023 03:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency GULFPORT PD **Date** 04/24/2023 **Type** Agency Inspection
Inspector MAROTTA, MICHAEL, S **Time** 00:05:27 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 22521080A1
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/24/2023 18:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency	GULFPORT PD	Date	05/31/2023	Type	Agency Inspection		
Inspector	MAROTTA,MICHAEL,S	Time	20:23:38	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	22521080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	10/05/2023
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/08/2023 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency	GULFPORT PD	Date	06/20/2023	Type	Agency Inspection		
Inspector	MAROTTA,MICHAEL,S	Time	01:47:18	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	22521080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	10/05/2023
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:SIM LEAK.TIGHTENED SIM LID

Data Download Date/Time 06/20/2023 01:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency GULFPORT PD **Date** 07/26/2023 **Type** Agency Inspection
Inspector MAROTTA, MICHAEL, S **Time** 22:13:29 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22521080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/26/2023 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency	GULFPORT PD	Date	08/29/2023	Type	Agency Inspection		
Inspector	MAROTTA,MICHAEL,S	Time	03:44:21	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	22521080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	10/05/2023
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/29/2023 03:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency GULFPORT PD	Date 09/11/2023	Type Department Inspection	
Inspector SOTO,ISRAEL,	Time 11:07:45	Software 8100.27	
0.05 Lot # 202303K	0.08 Lot # 202303L	0.20 Lot # 202304C	0.08 Gas Lot # 06723080A5
Expiration 03/29/2025	Expiration 03/29/2025	Expiration 04/05/2025	Expiration 04/05/2025
Number of Simulators 5	Average Standard Deviation 0.0007	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.053
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.053
		0.08 g/210L Test # 4	0.083
		0.20 g/210L Test # 4	0.204
		0.08 g/210L Dry Gas Std Test # 4	0.082
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.053
		0.08 g/210L Test # 5	0.084
		0.20 g/210L Test # 5	0.204
		0.08 g/210L Dry Gas Std Test # 5	0.082
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.053
0.08 g/210L Test # 6	0.084
0.20 g/210L Test # 6	0.204
0.08 g/210L Dry Gas Std Test # 6	0.083
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.054
0.08 g/210L Test # 7	0.084
0.20 g/210L Test # 7	0.204
0.08 g/210L Dry Gas Std Test # 7	0.083
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.054
0.08 g/210L Test # 8	0.084
0.20 g/210L Test # 8	0.204
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.054
0.08 g/210L Test # 9	0.084
0.20 g/210L Test # 9	0.205
0.08 g/210L Dry Gas Std Test # 9	0.083
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.054
0.08 g/210L Test # 10	0.084
0.20 g/210L Test # 10	0.204
0.08 g/210L Dry Gas Std Test # 10	0.083
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

Data Download Date/Time 09/11/2023 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency GULFPORT PD **Date** 09/16/2023 **Type** Agency Inspection
Inspector MAROTTA, MICHAEL, S **Time** 03:12:18 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 231043
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/16/2023 03:20

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency GULFPORT PD **Date** 10/27/2023 **Type** Agency Inspection
Inspector MAROTTA, MICHAEL, S **Time** 03:30:57 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 231043
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/27/2023 19:39

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency GULFPORT PD **Date** 11/29/2023 **Type** Agency Inspection
Inspector MAROTTA, MICHAEL, S **Time** 23:30:27 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 231043
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/29/2023 23:38

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

Agency GULFPORT PD **Date** 12/28/2023 **Type** Agency Inspection
Inspector MAROTTA, MICHAEL, S **Time** 02:19:01 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 231043
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:SIM LEAK. TIGHTENED SIM LID.

Data Download Date/Time 12/28/2023 02:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 01/04/2023
Time 12:11:51

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail, A. FRAZIER PRESENT.

Data Download Date/Time 01/12/2023 03:14

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 02/01/2023
Time 12:43:43

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

BSO A. FRAIZER WAS PRESENT.

Data Download Date/Time 02/01/2023 14:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 03/11/2023
Time 12:34:29

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.FRAZIER WAS PRESENT

Data Download Date/Time 03/11/2023 14:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 04/12/2023
Time 11:01:06

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

BSO A. FRAZIER WAS PRESENT.

Data Download Date/Time 04/12/2023 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 05/03/2023
Time 11:44:06

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 12/17/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A. FRAZIER

Data Download Date/Time 05/03/2023 11:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD **Date** 05/09/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:18:20 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 05/09/2023 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 05/18/2023
Time 14:56:14

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402290677
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/18/2023 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 06/07/2023
Time 12:35:24

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range Exceeded. TOO MUCH MOUTH ALCOHOL.. 00: Ambient Fail. TURN FAN ON CLEARED AIR.. 05: RFI Detect. UNKNOWN.SOURCE.. BSO A. FRAZIER WAS PRESENT.

Data Download Date/Time 06/14/2023 07:36

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 07/05/2023
Time 10:17:52

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

BSO A. FRAZIER WAS PRESENT.

Data Download Date/Time 07/05/2023 10:29

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 08/02/2023
Time 11:44:51

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

BSO A. FRAZIER WAS PRESENT.

Data Download Date/Time 08/02/2023 11:54

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 09/06/2023
Time 12:45:51

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

BSO A. FRAZIER WAS PRESENT.

Data Download Date/Time 09/06/2023 12:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 10/04/2023
Time 12:08:00

Type Agency Inspection
Software 8100.27

0.05 Lot # 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

BSO A. FRAZIER WAS PRESENT.

Data Download Date/Time 10/04/2023 12:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 11/08/2023
Time 12:24:30

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

BSO A. FRAZIER WAS PRESENT.

Data Download Date/Time 11/08/2023 12:51

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

Agency HOLLYWOOD PD
Inspector DERIET, RAFAEL,

Date 12/13/2023
Time 12:38:09

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

BSO A. FRAZIER WAS PRESENT.

Data Download Date/Time 12/13/2023 12:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency NEW PORT RICHEY PD **Date** 01/07/2023 **Type** Agency Inspection
Inspector HAZIME, CPL. A., **Time** 13:25:46 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Data Download Date/Time 01/07/2023 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency	NEW PORT RICHEY PD	Date	01/07/2023	Type	Agency Inspection		
Inspector	HAZIME,CPL. A.,	Time	14:31:58	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	06022080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	04/05/2024
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED INSTRUMENT, CHECKED FOR CRACKS, WIPED DOWN

Data Download Date/Time 01/07/2023 15:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency	NEW PORT RICHEY PD	Date	02/19/2023	Type	Agency Inspection		
Inspector	HAZIME,CPL A.,	Time	09:30:51	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	06022080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	04/05/2024
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 02/19/2023 11:48

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency NEW PORT RICHEY PD **Date** 03/05/2023 **Type** Agency Inspection
Inspector HAZIME,A, **Time** 08:39:47 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Range ExceedeD TOO MUCH MOUTH ALCOHOL. Non-compliance: .

Data Download Date/Time 03/05/2023 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency NEW PORT RICHEY PD	Date 03/05/2023	Type Agency Inspection	
Inspector HAZIME,A,	Time 09:52:40	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 06022080A3
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 04/05/2024
Number of Simulators 4	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED, WIPED DOWN INSTRUMENT - ALL OK

Data Download Date/Time 03/05/2023 10:10

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency	NEW PORT RICHEY PD	Date	04/06/2023	Type	Agency Inspection		
Inspector	HAZIME,CPL A.,	Time	06:02:36	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	06022080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	04/05/2024
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED INTOX, WIPED DOWN

Data Download Date/Time 04/10/2023 18:34

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency	NEW PORT RICHEY PD	Date	05/27/2023	Type	Agency Inspection		
Inspector	ENGEL,JASON,S	Time	08:47:10	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	06022080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	04/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

CHECKED TUBING

Data Download Date/Time 06/29/2023 12:23

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency NEW PORT RICHEY PD **Date** 06/29/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 14:17:12 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.196
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 06/29/2023 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency TARPON SPRINGS PD **Date** 01/31/2023 **Type** Agency Inspection
Inspector GASSEN,STEVE, **Time** 07:49:48 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 00521080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AREA

Data Download Date/Time 02/20/2023 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency TARPON SPRINGS PD **Date** 02/28/2023 **Type** Agency Inspection
Inspector GASSEN,STEVE, **Time** 07:51:16 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 01923080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AREA

Data Download Date/Time 03/01/2023 09:21

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency TARPON SPRINGS PD **Date** 03/03/2023 **Type** Agency Inspection
Inspector GASSEN,STEVE, **Time** 09:41:30 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 01923080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: No Sample Provided. HAD TO LEAVE ROOM FOR PHONE

Data Download Date/Time 03/03/2023 10:00

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency TARPON SPRINGS PD **Date** 04/28/2023 **Type** Agency Inspection
Inspector GASSEN,STEVE, **Time** 14:41:46 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 01923080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail ALCOHOL BOTTLE LEFT OPEN.

Data Download Date/Time 04/28/2023 14:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency TARPON SPRINGS PD **Date** 05/30/2023 **Type** Agency Inspection
Inspector GASSEN,STEVE, **Time** 10:10:19 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 01923080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AREA CLEANED

Data Download Date/Time 05/30/2023 10:22

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency TARPON SPRINGS PD **Date** 06/07/2023 **Type** Agency Inspection
Inspector GASSEN,STEVE, **Time** 15:37:55 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 01923080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 06/07/2023 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency TARPON SPRINGS PD **Date** 07/31/2023 **Type** Agency Inspection
Inspector GASSEN,STEVE, **Time** 09:20:16 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** **0.08 Gas Lot #**
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.073
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.049
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test # 2	0.076
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 3	0.049
0.08 g/210L Test: OK	No	0.08 g/210L Test # 3	0.077
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

08: RFI Detect/REPLACED GASKET ON SIMULATOR, RFI Detect. Non-compliance:RFI ERRORS REPEATED INSPECTION.

Data Download Date/Time 07/31/2023 10:02

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency	TARPON SPRINGS PD	Date	07/31/2023	Type	Agency Inspection		
Inspector	GASSEN,STEVE,	Time	09:56:18	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202206C	0.08 Gas Lot #	01923080A3
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/15/2024	Expiration	02/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AREA

Data Download Date/Time 07/31/2023 10:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency TARPON SPRINGS PD **Date** 08/03/2023 **Type** Agency Inspection
Inspector GASSEN,STEVE, **Time** 14:07:42 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 01923080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 02/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.08 g/210L Test # (Repeat) 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance SEALS CHECKED.

Data Download Date/Time 08/03/2023 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency	TARPON SPRINGS PD	Date	08/09/2023	Type	Agency Inspection		
Inspector	GASSEN,STEVE,	Time	15:09:19	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202206C	0.08 Gas Lot #	01923080A3
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/15/2024	Expiration	02/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

05: RFI Detect.

Data Download Date/Time 08/15/2023 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency TARPON SPRINGS PD **Date** 08/18/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 12:37:50 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 06723080A5
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0010 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.203
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.204
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.083
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.076
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.204
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0013
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0015

Remarks:

Data Download Date/Time 08/18/2023 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency TARPON SPRINGS PD **Date** 09/13/2023 **Type** Agency Inspection
Inspector GASSEN,STEVE, **Time** 17:36:31 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 01923080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AREA IN SERVICE AFTER DEPARTMENT INSPECTION

Data Download Date/Time 09/13/2023 17:43

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency TARPON SPRINGS PD **Date** 10/31/2023 **Type** Agency Inspection
Inspector GASSEN,STEVE, **Time** 08:21:44 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 01923080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AREA CLEANED

Data Download Date/Time 11/01/2023 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency	TARPON SPRINGS PD	Date	11/01/2023	Type	Agency Inspection		
Inspector	GASSEN,STEVE,	Time	09:08:31	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	01923080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	02/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AREA CLEANED

Data Download Date/Time 11/01/2023 09:19

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency TARPON SPRINGS PD **Date** 12/29/2023 **Type** Agency Inspection
Inspector GASSEN,STEVE, **Time** 07:43:54 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 01923080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AREA

Data Download Date/Time 12/29/2023 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO **Date** 01/20/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 23:16:57 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG115904
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/08/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/20/2023 23:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO **Date** 02/10/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL,X **Time** 10:42:25 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2023

Number of Simulators 5 **Average Standard Deviation** 0.0007 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.203
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0014
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/10/2023 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 02/18/2023
Time 17:07:34

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG115904
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/08/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/22/2023 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 03/22/2023
Time 10:17:17

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG115904
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/08/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/23/2023 21:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO **Date** 04/19/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 10:23:00 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG115904
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/08/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/19/2023 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO **Date** 05/10/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 14:12:07 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG115904
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/08/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/10/2023 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO **Date** 06/14/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 12:38:50 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG229803
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/14/2023 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 07/05/2023
Time 16:01:26

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG229803
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/05/2023 16:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 08/16/2023
Time 08:04:37

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG229803
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/16/2023 08:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 09/06/2023
Time 14:47:50

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG229803
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/06/2023 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

Agency PINELLAS COUNTY SO
Inspector DEANE,M,

Date 10/18/2023
Time 14:20:09

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG229803
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/18/2023 14:30

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency CITY OF MIRAMAR POL **Date** 01/04/2023 **Type** Agency Inspection
Inspector HOLLOWAY,DEVARIOUS,J **Time** 12:15:07 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1273831
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/23/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

A FRAZIER PRESENT

Data Download Date/Time 01/04/2023 13:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency CITY OF MIRAMAR POL **Date** 02/01/2023 **Type** Agency Inspection
Inspector HOLLOWAY,DEVARIOUS,J **Time** 12:46:28 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

J. GUZMAN IN TRAINING.

Data Download Date/Time 02/01/2023 13:43

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency CITY OF MIRAMAR POL **Date** 03/11/2023 **Type** Agency Inspection
Inspector HOLLOWAY,DEVARIOUS,J **Time** 12:37:43 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/11/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A FRAZIER PRESENT

Data Download Date/Time 04/12/2023 10:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency CITY OF MIRAMAR POL **Date** 04/12/2023 **Type** Agency Inspection
Inspector HOLLOWAY,DEVARIOUS,J **Time** 11:04:44 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:TOO MUCH MOUTH ALCOHOL.

Data Download Date/Time 04/12/2023 11:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency CITY OF MIRAMAR POL **Date** 05/03/2023 **Type** Agency Inspection
Inspector HOLLOWAY,DEVARIOUS,J **Time** 11:03:59 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A. FRAZIER WAS PRESENT.

Data Download Date/Time 05/03/2023 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency CITY OF MIRAMAR POL **Date** 06/07/2023 **Type** Agency Inspection
Inspector HOLLOWAY,DEVARIOUS,J **Time** 11:05:35 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: RFI Detect. UNKNOWN SOURCE., RFI Detect. Non-compliance:RFI..

Data Download Date/Time 06/07/2023 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency	CITY OF MIRAMAR POL	Date	06/07/2023	Type	Agency Inspection		
Inspector	HOLLOWAY,DEVARIOUS,J	Time	11:53:14	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	402477283
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/24/2025
Number of Simulators	5	Compliance	No				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # 1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	No	Alcohol Free Test # 2	
Alcohol Free Test: OK	Yes	Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test: OK	No	Interferent Detect Test # 2	INT
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

00: RFI Detect UNKNOWN SOURCE. Int Det: RFI Detect. 08: RFI Detect UNKNOWN SOURCE, RFI Detect. Non-compliance:RFI DETECT UNKNOWN SOURCE..

Data Download Date/Time 06/07/2023 12:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency CITY OF MIRAMAR POL **Date** 06/09/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 15:04:47 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # (Repeat) 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.076
		Interferent Detect Test # 4	INT
		Alcohol Free Test # (Repeat) 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.077
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Alcohol Free Test # (Repeat) 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # (Repeat) 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.077
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # (Repeat) 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # (Repeat) 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # (Repeat) 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

00: Ambient Fail. UNKNOWN. ALLOWED TO AIR OUT BEFORE REPEAT.

Data Download Date/Time 06/09/2023 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency	CITY OF MIRAMAR POL	Date	07/05/2023	Type	Agency Inspection		
Inspector	HOLLOWAY,DEVARIOUS,J	Time	10:58:52	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	402477283
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/24/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A FRAZIER PRESENT

Data Download Date/Time 07/05/2023 11:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency CITY OF MIRAMAR POL **Date** 08/02/2023 **Type** Agency Inspection
Inspector HOLLOWAY,DEVARIOUS,J **Time** 12:31:52 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

BSO A. FRAZIER WAS PRESENT.

Data Download Date/Time 09/13/2023 11:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

Agency CITY OF MIRAMAR POL **Date** 09/06/2023 **Type** Agency Inspection
Inspector HOLLOWAY,DEVARIOUS,J **Time** 13:16:29 **Software** 8100.27

0.05 Lot # 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/06/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

BSO

Data Download Date/Time 09/13/2023 11:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency CLEARWATER PD **Date** 01/20/2023 **Type** Agency Inspection
Inspector HURT,JONATHAN,R **Time** 00:50:36 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 07622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 01/20/2023 01:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency	CLEARWATER PD	Date	02/28/2023	Type	Agency Inspection		
Inspector	BUIS,JUSTIN,M	Time	07:01:59	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202106B	0.08 Gas Lot #	33922080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/22/2023	Expiration	01/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

M/A 2021-D ACETONE 2021-C

Data Download Date/Time 02/28/2023 10:54

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency CLEARWATER PD	Date 03/27/2023	Type Agency Inspection	
Inspector BUIS,JUSTIN,M	Time 07:53:44	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202106B	0.08 Gas Lot # 33922080A1
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/22/2023	Expiration 01/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

M/A 2021-D

Data Download Date/Time 04/11/2023 08:10

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency	CLEARWATER PD	Date	05/23/2023	Type	Agency Inspection
Inspector	SOTO,ISRAEL,	Time	08:11:45	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 05/23/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency CLEARWATER PD **Date** 05/23/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 10:10:55 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 05/23/2023 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency CLEARWATER PD **Date** 06/30/2023 **Type** Agency Inspection
Inspector BUIS,JUSTIN, **Time** 20:25:26 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

M/A 2021-D ACETONE 2022-B

Data Download Date/Time 06/30/2023 20:32

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency CLEARWATER PD **Date** 07/29/2023 **Type** Agency Inspection
Inspector BUIS,JUSTIN, **Time** 22:55:26 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

MA

Data Download Date/Time 08/01/2023 02:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency CLEARWATER PD **Date** 08/24/2023 **Type** Agency Inspection
Inspector BUIS,JUSTIN, **Time** 16:12:46 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

ACETONE 2022-B M/A 2021-D

Data Download Date/Time 08/24/2023 16:23

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency	CLEARWATER PD	Date	09/27/2023	Type	Agency Inspection		
Inspector	BUIS,JUSTIN,M	Time	18:43:16	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	08223080A4
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

M/A 2021-D ACETONE 2022-B

Data Download Date/Time 09/27/2023 18:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency	CLEARWATER PD	Date	10/27/2023	Type	Agency Inspection		
Inspector	HURT,JONATHAN,R	Time	14:08:18	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	08223080A
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.ACETONE LOT #2022-B, M/A 2022-A

Data Download Date/Time 10/27/2023 14:18

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency CLEARWATER PD	Date 11/29/2023	Type Agency Inspection	
Inspector BUIS,JUSTIN,	Time 12:14:01	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 08223080A4
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 06/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

ACETONE 2022-B M/A 2022-A

Data Download Date/Time 11/29/2023 12:32

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency CLEARWATER PD **Date** 12/29/2023 **Type** Agency Inspection
Inspector BUIS,JUSTIN, **Time** 21:18:32 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A4
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

M/A 2022-A ACETONE 2022-B

Data Download Date/Time 12/29/2023 21:29

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency JACKSONVILLE SO **Date** 01/12/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 11:31:29 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 07622080A3
Expiration 12/07/2023 **Expiration** 08/07/2023 **Expiration** 06/22/2023 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

PLACED IN EVID USE AT PTDF 01/12/2023

Data Download Date/Time 01/12/2023 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency	JACKSONVILLE SO	Date	02/01/2023	Type	Agency Inspection		
Inspector	LIGHTSEY,ROBERT,D	Time	14:37:20	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	07622080A3
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2023 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency JACKSONVILLE SO **Date** 02/15/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 11:34:46 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 07622080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

RMVED FROM EVED USE AT PTDF ON 2-15-2023

Data Download Date/Time 02/15/2023 11:43

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency JACKSONVILLE SO **Date** 03/25/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 06:48:47 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 07622080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

PLCED INTO EVED USE AT PTDF ON 3-25-2023

Data Download Date/Time 03/27/2023 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency JACKSONVILLE SO **Date** 04/21/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 10:53:00 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 01923080A4
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/21/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency JACKSONVILLE SO **Date** 05/24/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 13:48:14 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 01923080A4
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/24/2023 14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency	JACKSONVILLE SO	Date	06/14/2023	Type	Agency Inspection		
Inspector	LIGHTSEY,ROBERT,D	Time	14:11:56	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	07622080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

RMVED FROM EVED USE AT PTDF 6-14-2023

Data Download Date/Time 06/14/2023 14:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency	JACKSONVILLE SO	Date	07/18/2023	Type	Agency Inspection		
Inspector	LIGHTSEY,ROBERT,D	Time	10:54:31	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	07622080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

PLACED INTO EVID USE 07/18/2023 @ PTDF

Data Download Date/Time 07/18/2023 11:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency	JACKSONVILLE SO	Date	08/17/2023	Type	Agency Inspection		
Inspector	LIGHTSEY,ROBERT,D	Time	13:26:53	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	07622080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/17/2023 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency	JACKSONVILLE SO	Date	09/19/2023	Type	Agency Inspection		
Inspector	LIGHTSEY,ROBERT,D	Time	11:39:23	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	07622080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

REMOVED FOR DEPARTMENT INSPECTION FDLE

Data Download Date/Time 09/19/2023 11:50

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency	JACKSONVILLE SO	Date	09/26/2023	Type	Department Inspection		
Inspector	SIDDOWAY,BENJAMIN,W	Time	14:40:19	Software	8100.27		
0.05 Lot #	202303K	0.08 Lot #	202303L	0.20 Lot #	202304C	0.08 Gas Lot #	06723080A5
Expiration	03/29/2025	Expiration	03/29/2025	Expiration	04/05/2025	Expiration	04/05/2025
Number of Simulators	5	Average Standard Deviation	0.0007	Compliance	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.082
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.052
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.052
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.083
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.052
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.084
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.052
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.084
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.052
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.084
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0014

Remarks:

Data Download Date/Time 09/26/2023 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency	JACKSONVILLE SO	Date	10/24/2023	Type	Agency Inspection		
Inspector	LIGHTSEY,ROBERT,D	Time	11:46:02	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	08223080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

PLCED INTO EVED USE AT PDF 10-23-2023

Data Download Date/Time 10/24/2023 12:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

Agency JACKSONVILLE SO **Date** 11/17/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 12:21:06 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/17/2023 12:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

Agency FL HIGHWAY PATROL **Date** 03/20/2023 **Type** Agency Inspection
Inspector LAMONT,GREGORY,S **Time** 13:50:04 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/04/2023 12:06

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

Agency FL HIGHWAY PATROL **Date** 05/03/2023 **Type** Agency Inspection
Inspector LAMONT,GREGORY,S **Time** 13:23:41 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/04/2023 12:06

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency CLAY COUNTY SO **Date** 05/23/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 15:49:44 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.077
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.077
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.077
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.077
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 05/23/2023 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency	CLAY COUNTY SO	Date	07/22/2023	Type	Agency Inspection		
Inspector	PATRONE,RICHARD,L	Time	14:34:52	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	07622080A2
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSP RETURN TO SERVICE

Data Download Date/Time 07/28/2023 01:22

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency CLAY COUNTY SO **Date** 07/28/2023 **Type** Agency Inspection
Inspector HARWOOD, CHARLES, M **Time** 01:14:30 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Improper Sample.

Data Download Date/Time 07/28/2023 01:22

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency	CLAY COUNTY SO	Date	08/26/2023	Type	Agency Inspection		
Inspector	JOHNSON,MATTHEW,P	Time	11:13:23	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	07622080A2
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

RFI WAS RESULT OF AGENCY RADIO. MONTHLY AI COMPLETED

Data Download Date/Time 08/26/2023 11:22

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency CLAY COUNTY SO **Date** 09/24/2023 **Type** Agency Inspection
Inspector JOHNSON,MATTHEW,P **Time** 14:12:16 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 09/24/2023 14:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency CLAY COUNTY SO **Date** 10/23/2023 **Type** Agency Inspection
Inspector PATRONE,RICHARD,L **Time** 07:19:19 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 10/23/2023 07:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency CLAY COUNTY SO **Date** 11/26/2023 **Type** Agency Inspection
Inspector PATRONE,RICHARD,L **Time** 13:12:42 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 11/26/2023 13:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency	CLAY COUNTY SO	Date	12/25/2023	Type	Agency Inspection		
Inspector	JOHNSON,MATTHEW,P	Time	13:21:56	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	07622080A2
Expiration	03/28/2025	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 12/25/2023 15:28

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

Agency DADE CITY PD **Date** 05/02/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 11:24:27 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.077
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.076
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.077
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.076
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.076
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.076
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.076
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 05/02/2023 11:28

Inspection Test Electronic Data

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

Agency DADE CITY PD **Date** 06/14/2023 **Type** Agency Inspection
Inspector BALDWIN, DEAN, **Time** 14:02:11 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 25721080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/14/2023 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

Agency DADE CITY PD	Date 07/13/2023	Type Agency Inspection	
Inspector BALDWIN, DEAN,	Time 14:16:10	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 25721080A2
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 11/05/2023
Number of Simulators 4	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/13/2023 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

Agency DADE CITY PD	Date 08/19/2023	Type Agency Inspection	
Inspector BALDWIN, DEAN,	Time 06:48:18	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 25721080A2
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 11/05/2023
Number of Simulators 4	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/19/2023 07:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

Agency DADE CITY PD	Date 09/18/2023	Type Agency Inspection	
Inspector BALDWIN, DEAN,	Time 16:30:02	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 25721080A2
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 11/05/2023
Number of Simulators 4	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/24/2023 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

Agency DADE CITY PD	Date 10/16/2023	Type Agency Inspection	
Inspector BALDWIN, DEAN,	Time 14:16:22	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 25721080A2
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 11/05/2023
Number of Simulators 4	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:USED MOUTH ALCOHOL ON FIRST TEST.

Data Download Date/Time 10/16/2023 14:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

Agency DADE CITY PD **Date** 11/28/2023 **Type** Agency Inspection
Inspector BALDWIN, DEAN, **Time** 16:42:35 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 14323080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/28/2023 16:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

Agency DADE CITY PD **Date** 12/25/2023 **Type** Agency Inspection
Inspector DEL VALLE,M, **Time** 11:14:27 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 14323080A3
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

NO VISIBLE DAMAGE TO INTOX

Data Download Date/Time 12/25/2023 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO **Date** 01/16/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 09:19:38 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: Sequence Aborted . Non-compliance: .

Data Download Date/Time 01/16/2023 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 01/16/2023
Time 10:25:44

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2023 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 02/07/2023
Time 03:50:12

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/07/2023 03:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO **Date** 02/07/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 12:35:25 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

A F / M A: DID MA FIRST.

Data Download Date/Time 02/07/2023 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 02/07/2023
Time 19:10:05

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/07/2023 19:10

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 03/27/2023
Time 13:34:33

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/27/2023 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 04/29/2023
Time 16:20:55

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance FORGOT TO CONNECT HOSE.

Data Download Date/Time 05/03/2023 17:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 05/08/2023
Time 10:26:37

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/08/2023 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 06/30/2023
Time 18:37:06

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/30/2023 18:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 07/11/2023
Time 15:37:11

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/11/2023 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 08/28/2023
Time 14:07:01

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2023 14:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO **Date** 09/16/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 10:58:19 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:REAPLY MOUTH ALC SOLUTION.

Data Download Date/Time 09/16/2023 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 10/13/2023
Time 16:25:09

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/13/2023 16:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 11/27/2023
Time 11:26:05

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/27/2023 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 12/18/2023
Time 13:39:15

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202307C **0.08 Gas Lot #** 08622080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 07/12/2025 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/18/2023 13:47

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD
Inspector STARNES,AILISA,J

Date 01/30/2023
Time 12:37:33

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 13921080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/05/2023

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

WIPE DOWN INSTRUMENT EVERYTHING APPEARS GOOD

Data Download Date/Time 01/30/2023 12:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD **Date** 02/06/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 14:00:04 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/06/2023 14:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD
Inspector STARNES,AILISA,J

Date 02/28/2023
Time 14:58:26

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 13921080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/05/2023

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.075
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2023 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD
Inspector STARNES,AILISA,J

Date 03/28/2023
Time 09:58:12

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 13921080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/05/2023

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.075
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/28/2023 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD
Inspector STARNES,AILISA,J

Date 04/27/2023
Time 13:11:59

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 06023080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/05/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/27/2023 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD
Inspector STARNES,AILISA,J

Date 05/26/2023
Time 11:01:53

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 06023080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/05/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/26/2023 11:08

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD **Date** 06/20/2023 **Type** Agency Inspection
Inspector STARNES,AILISA,J **Time** 13:02:58 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 06023080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/20/2023 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD
Inspector STARNES,AILISA,J

Date 07/28/2023
Time 14:49:29

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 06023080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

WIPED DONE MACHINE EVERYTHING APPEARS FINE

Data Download Date/Time 07/28/2023 14:57

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency	ZEPHYRHILLS PD	Date	08/31/2023	Type	Agency Inspection		
Inspector	STARNES,AILISA,J	Time	09:21:19	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	06023080A2
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	04/25/2025
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 08/31/2023 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD
Inspector STARNES,AILISA,J

Date 09/27/2023
Time 11:10:21

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06023080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

WIPE DOWN MACHINE LOOKS GOOD

Data Download Date/Time 09/27/2023 11:19

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD **Date** 11/02/2023 **Type** Agency Inspection
Inspector SOTO,ISRAEL, **Time** 13:49:24 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 11/03/2023 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD **Date** 11/03/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 09:07:51 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 11/03/2023 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD
Inspector STARNES,AILISA,J

Date 11/16/2023
Time 09:24:04

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06023080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/16/2023 09:33

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency ZEPHYRHILLS PD
Inspector STARNES,AILISA,J

Date 12/18/2023
Time 14:37:29

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06023080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.165
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.166
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.168
		0.20 g/210L Test # (Repeat) 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 12/18/2023 14:47

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency HARDEE COUNTY SO **Date** 02/03/2023 **Type** Agency Inspection
Inspector CARTWRIGHT JR,GARY,W **Time** 14:39:32 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.194
		0.20 g/210L Test # (Repeat) 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 02/06/2023 09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency HARDEE COUNTY SO **Date** 03/31/2023 **Type** Agency Inspection
Inspector CARTWRIGHT JR,GARY,W **Time** 13:41:24 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/31/2023 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency HARDEE COUNTY SO **Date** 04/30/2023 **Type** Agency Inspection
Inspector CARTWRIGHT JR,GARY,W **Time** 12:08:02 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** **0.08 Gas Lot #**
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.110
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.198
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.047
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.104
Alcohol Free Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.090
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test # 3	0.047
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.101
		0.08 g/210L Test # (Repeat) 3	0.088
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside WILL CHANGE SOLUTION AMD RESTART TEST Non-compliance: .

Data Download Date/Time 05/04/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency HARDEE COUNTY SO **Date** 04/30/2023 **Type** Agency Inspection
Inspector CARTWRIGHT JR,GARY,W **Time** 13:19:53 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/04/2023 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency HARDEE COUNTY SO **Date** 05/31/2023 **Type** Agency Inspection
Inspector CARTWRIGHT JR,GARY,W **Time** 17:46:27 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 06/02/2023 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency HARDEE COUNTY SO **Date** 06/30/2023 **Type** Agency Inspection
Inspector CARTWRIGHT JR,GARY,W **Time** 09:38:27 **Software** 8100.27

0.05 Lot # 20211A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/30/2023 09:43

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency HARDEE COUNTY SO **Date** 07/28/2023 **Type** Agency Inspection
Inspector CARWRIGHT JR,GARY,W **Time** 13:57:48 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/28/2023 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency HARDEE COUNTY SO **Date** 08/31/2023 **Type** Agency Inspection
Inspector ALBRITTON,GARRETT,M **Time** 21:31:30 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

SKIPPED MOUTH ALCOHOL

Data Download Date/Time 08/31/2023 21:37

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency HARDEE COUNTY SO **Date** 09/27/2023 **Type** Agency Inspection
Inspector ALBRITTON,GARRETT,M **Time** 23:34:13 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/27/2023 23:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency	HARDEE COUNTY SO	Date	10/12/2023	Type	Department Inspection
Inspector	GUTSCHOW,TAYLOR,D	Time	15:11:32	Software	8100.27

0.05 Lot #	202201C	0.08 Lot #	AG223802	0.20 Lot #		0.08 Gas Lot #	
Expiration	01/11/2024	Expiration	08/26/2024	Expiration		Expiration	

Number of Simulators	5	Average Standard Deviation		Compliance	No
-----------------------------	---	-----------------------------------	--	-------------------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.000
Barometric Pressure Sensor Check: OK	Yes	0.08 g/210L Test # (Repeat) 1	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.050
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.000
Alcohol Free Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.000
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test # 3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test # 3	0.050
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.000
		0.08 g/210L Test # (Repeat) 3	0.000
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.000
		0.08 g/210L Test # (Repeat) 4	0.000
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.000
		0.08 g/210L Test # (Repeat) 5	0.000
		Interferent Detect Test # 5	INT
		Alcohol Free Test # 6	0.000
		0.05 g/210L Test # 6	0.049
		0.08 g/210L Test # 6	0.000
		0.08 g/210L Test # (Repeat) 6	0.000
		Interferent Detect Test # 6	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.000
0.08 g/210L Test # (Repeat) 7	0.000
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.000
0.08 g/210L Test # (Repeat) 8	0.000
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.000
0.08 g/210L Test # (Repeat) 9	0.000
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.000
0.08 g/210L Test # (Repeat) 10	0.000
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev) (Repeat)	0.0000

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Data Download Date/Time 10/12/2023 15:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency	HARDEE COUNTY SO	Date	10/13/2023	Type	Department Inspection		
Inspector	GUTSCHOW,TAYLOR,D	Time	15:00:22	Software	8100.27		
0.05 Lot #	202201C	0.08 Lot #	202201D	0.20 Lot #	202201E	0.08 Gas Lot #	01923080A3
Expiration	01/11/2024	Expiration	01/18/2024	Expiration	01/18/2024	Expiration	02/05/2025
Number of Simulators	5	Average Standard Deviation	0.0003	Compliance	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 10/13/2023 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency	ST PETERSBURG PD	Date	01/10/2023	Type	Department Inspection		
Inspector	GUTSCHOW,TAYLOR,D	Time	15:19:29	Software	8100.27		
0.05 Lot #	202201C	0.08 Lot #	202201D	0.20 Lot #	202201E	0.08 Gas Lot #	00521080A2
Expiration	01/11/2024	Expiration	01/18/2024	Expiration	02/05/2023	Expiration	02/05/2023
Number of Simulators	5	Average Standard Deviation	0.0004	Compliance	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.047
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0010
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/10/2023 15:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency	ST PETERSBURG PD	Date	02/23/2023	Type	Agency Inspection		
Inspector	MCCULLOUGH,SEAN,J	Time	09:34:42	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	08622080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/23/2023 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency	ST PETERSBURG PD	Date	03/29/2023	Type	Agency Inspection		
Inspector	MCCULLOUGH,SEAN,J	Time	08:45:34	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	08622080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/29/2023 09:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency	ST PETERSBURG PD	Date	04/27/2023	Type	Agency Inspection		
Inspector	MCCULLOUGH,SEAN,J	Time	14:06:37	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	08622080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 04/27/2023 15:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency	ST PETERSBURG PD	Date	05/30/2023	Type	Agency Inspection		
Inspector	MCCULLOUGH,SEAN,J	Time	10:19:26	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	08622080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 05/30/2023 13:33

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency	ST PETERSBURG PD	Date	06/21/2023	Type	Agency Inspection		
Inspector	MCCULLOUGH,SEAN,J	Time	09:10:26	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	06023080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	04/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 06/21/2023 09:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency	ST PETERSBURG PD	Date	07/27/2023	Type	Agency Inspection		
Inspector	MCCULLOUGH,SEAN,J	Time	09:03:18	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	06023080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	04/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 07/28/2023 12:03

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency	ST PETERSBURG PD	Date	08/31/2023	Type	Agency Inspection		
Inspector	MCCULLOUGH,SEAN,J	Time	08:26:33	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	08622080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 09/05/2023 08:30

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency ST PETERSBURG PD
Inspector LENAHAAN,RYAN,P

Date 09/27/2023
Time 08:33:05

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 09/27/2023 08:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency	ST PETERSBURG PD	Date	10/25/2023	Type	Agency Inspection		
Inspector	MCCULLOUGH,SEAN,J	Time	09:27:18	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	08622080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 10/25/2023 10:00

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency ST PETERSBURG PD
Inspector LENAHAAN,RYAN,P

Date 11/29/2023
Time 16:07:39

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06023080A1
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 11/30/2023 09:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency ST PETERSBURG PD	Date 12/28/2023	Type Agency Inspection	
Inspector MCCULLOUGH,SEAN,J	Time 16:28:21	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 06023080A1
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 04/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.000
		0.08 g/210L Test # (Repeat) 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 01/03/2024 08:30

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency LARGO PD **Date** 01/11/2023 **Type** Agency Inspection
Inspector THORNSBROUGH,DREW,A **Time** 11:20:31 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 23621080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/11/2023 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency LARGO PD **Date** 02/01/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:38:38 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 00521080A2
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 02/05/2023

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/01/2023 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency LARGO PD **Date** 02/14/2023 **Type** Agency Inspection
Inspector HASTINGS, KENNETH, **Time** 13:47:03 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 23621080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/14/2023 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency LARGO PD **Date** 03/22/2023 **Type** Agency Inspection
Inspector THORNSBROUGH,DREW,A **Time** 13:19:34 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 23621080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 03/22/2023 13:29

Inspection Test Electronic Data

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency LARGO PD **Date** 05/15/2023 **Type** Agency Inspection
Inspector THORNSBROUGH,DREW,A **Time** 13:43:45 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 23621080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/15/2023 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency LARGO PD **Date** 06/20/2023 **Type** Agency Inspection
Inspector THORNSBROUGH,DREW,A **Time** 12:25:35 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 23621080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/20/2023 12:34

Inspection Test Electronic Data

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Inspection Test Electronic Data

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency LARGO PD **Date** 09/18/2023 **Type** Agency Inspection
Inspector HASTINGS, KENNETH, **Time** 12:06:55 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/18/2023 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency LARGO PD **Date** 10/18/2023 **Type** Agency Inspection
Inspector THORNSBROUGH,DREW,A **Time** 13:37:08 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/18/2023 13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency LARGO PD Date 11/07/2023 Type Agency Inspection
Inspector HASTINGS, KENNETH, Time 14:24:16 Software 8100.27

0.05 Lot # 202303J 0.08 Lot # 202206B 0.20 Lot # 202206C 0.08 Gas Lot # 15922080A4
Expiration 03/28/2025 Expiration 06/14/2024 Expiration 06/15/2024 Expiration 08/05/2024

Number of Simulators 5 Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/07/2023 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency LARGO PD **Date** 12/25/2023 **Type** Agency Inspection
Inspector THORNSBROUGH,DREW,A **Time** 10:55:03 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/25/2023 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 01/10/2023	Type Agency Inspection
Inspector COLE,KEVIN,T	Time 10:53:48	Software 8100.27

0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot #
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration

Number of Simulators 5	Compliance No
-------------------------------	----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.131
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.187
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.134
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.190
0.20 g/210L Test: OK	No	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.144
		0.20 g/210L Test # (Repeat) 3	0.193
		Interferent Detect Test # 3	INT

Remarks:

20: SIMULATOR POWER LOST, WAIT 30 MINUTES, ATTEMPT NEW INSPSIMULATOR LOST POWER. WAIT 30 MIN, RE-ATTEMPT.

Data Download Date/Time 01/10/2023 12:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 01/10/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 11:27:57	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 15922080A3
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 08/05/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/10/2023 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 02/01/2023	Type Department Inspection	
Inspector GUTSCHOW,TAYLOR,D	Time 13:48:37	Software 8100.27	
0.05 Lot # 202201C	0.08 Lot # 202201D	0.20 Lot # 202201E	0.08 Gas Lot # 00521080A2
Expiration 01/11/2024	Expiration 01/18/2024	Expiration 01/18/2024	Expiration 02/05/2023
Number of Simulators 5	Average Standard Deviation 0.0003	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.077
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.076
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.076
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.076
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.077
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/01/2023 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 02/14/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 08:59:53	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 15922080A3
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 08/05/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 02/14/2023 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 03/14/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 10:11:06	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 15922080A3
Expiration 12/07/2023	Expiration 06/12/2023	Expiration 06/15/2024	Expiration 08/05/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 03/14/2023 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 04/11/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 10:31:27	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 15922080A3
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 08/05/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 04/11/2023 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 05/18/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 09:50:28	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 26422080A1
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 11/05/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 05/18/2023 12:01

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 06/13/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 10:19:05	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 26422080A1
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 11/05/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 06/13/2023 11:47

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 07/13/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 10:14:25	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 26422080A1
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 11/05/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 07/13/2023 10:43

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 08/15/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 09:47:59	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 26422080A1
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 11/05/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 08/15/2023 10:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 09/12/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 10:46:28	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 08223080A2
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 05/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 09/12/2023 11:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 10/10/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 10:00:33	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 08223080A2
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 05/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/10/2023 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 11/14/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 09:57:49	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 08223080A2
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 05/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 11/14/2023 10:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency POLK CSO	Date 12/05/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 09:40:00	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 08223080A2
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 05/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 12/05/2023 10:13

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 01/16/2023
Time 10:04:06

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: No Sample Provided.

Data Download Date/Time 01/16/2023 10:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency PLANTATION PD **Date** 01/10/2023 **Type** Agency Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 06:52:13 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

Data Download Date/Time 01/11/2023 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency PLANTATION PD **Date** 01/11/2023 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 13:24:38 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 00521080A2
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 02/05/2023

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/11/2023 13:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency PLANTATION PD Date 05/09/2023 Type Agency Inspection
Inspector FORTEZA,CRYSTAL,L Time 19:44:10 Software 8100.27

0.05 Lot # 202112A 0.08 Lot # 202206B 0.20 Lot # 202206C 0.08 Gas Lot #
Expiration 12/07/2023 Expiration 06/14/2024 Expiration 06/15/2024 Expiration

Number of Simulators 4 Compliance No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.011
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.072
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Test # (Repeat) 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test # 1	0.133
Interferent Detect Test: Interferent Detect	Yes	0.20 g/210L Test # (Repeat) 1	0.188
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.041
0.08 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.047
0.20 g/210L Test: OK	No	0.08 g/210L Test # 2	0.077
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # (Repeat) 2	0.077
		0.20 g/210L Test # 2	0.173
		0.20 g/210L Test # (Repeat) 2	0.192
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.05 g/210L Test # (Repeat) 3	0.047
		0.08 g/210L Test # 3	0.077
		0.08 g/210L Test # (Repeat) 3	0.077
		0.20 g/210L Test # 3	0.190
		0.20 g/210L Test # (Repeat) 3	0.193
		Interferent Detect Test # 3	INT

Remarks:

05.CONTROL OUTSIDE TOLERANCE - TIGHTENED TUBES. 08: Control Outside Tolerance - TIGHTENED TUBES. 20: Control Outside Tolerance - TIGHTEN TUBES, Control Outs. Non-compliance: .20 OUTSIDE TOLERANCE.

Data Download Date/Time 05/31/2023 22:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency PLANTATION PD **Date** 05/31/2023 **Type** Agency Inspection
Inspector FARQUHARSON,KEVIN,D **Time** 21:39:12 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 223083
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2023 22:50

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency PLANTATION PD Date 06/29/2023 Type Agency Inspection
Inspector FORTEZA,CRYSTAL,L Time 10:32:37 Software 8100.27

0.05 Lot # 202112A 0.08 Lot # 202206B 0.20 Lot # 202206C 0.08 Gas Lot # 223083
Expiration 12/07/2023 Expiration 06/14/2024 Expiration 06/15/2024 Expiration 04/05/2025

Number of Simulators 4 Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/03/2023 12:36

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency PLANTATION PD **Date** 07/11/2023 **Type** Agency Inspection
Inspector FORTEZA,CRYSTAL,L **Time** 11:33:33 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 223083
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/11/2023 11:53

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency PLANTATION PD
Inspector MARLEY,KEVIN,G

Date 08/10/2023
Time 10:52:14

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06723080A5
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/10/2023 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency PLANTATION PD
Inspector MARLEY,KEVIN,G

Date 09/28/2023
Time 11:42:10

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06723080A5
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/28/2023 12:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency PLANTATION PD
Inspector MARLEY,KEVIN,G

Date 10/31/2023
Time 12:08:59

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06723080A5
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/31/2023 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency PLANTATION PD
Inspector MARLEY,KEVIN,G

Date 11/30/2023
Time 10:41:31

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 06723080A5
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/05/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/30/2023 10:51

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

Agency PLANTATION PD
Inspector MARLEY,KEVIN,G

Date 12/21/2023
Time 16:40:14

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 06723080A5
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/05/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 12/21/2023 16:51

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency	JACKSONVILLE SO	Date	01/12/2023	Type	Agency Inspection		
Inspector	LIGHTSEY,ROBERT,D	Time	14:27:05	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	07622080A3
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

REMOVED FROM EVID USE 1/12/2023

Data Download Date/Time 01/12/2023 15:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency JACKSONVILLE SO **Date** 02/15/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 10:07:20 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 07622080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

PLCED INTO EVED USE AT PTDF ON 2-15-2023

Data Download Date/Time 02/15/2023 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency JACKSONVILLE SO **Date** 03/02/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 07:36:34 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 07622080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/02/2023 07:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency JACKSONVILLE SO **Date** 03/26/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 09:15:04 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 07622080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

REMOVED FROM EVED USE AT PTDF 3-26-2023

Data Download Date/Time 03/27/2023 13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency	JACKSONVILLE SO	Date	04/25/2023	Type	Agency Inspection		
Inspector	LIGHTSEY,ROBERT,D	Time	09:59:37	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	07622080A3
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

PLCED INTO EVED USE AT M/P N/S 4/25/2023

Data Download Date/Time 04/25/2023 10:22

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency	JACKSONVILLE SO	Date	05/19/2023	Type	Agency Inspection		
Inspector	LIGHTSEY,ROBERT,D	Time	18:16:50	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	07622080A3
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

REMOVED FROM EVED USE AT NAS M/P ON 05-19-2023

Data Download Date/Time 05/19/2023 18:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency	JACKSONVILLE SO	Date	06/14/2023	Type	Agency Inspection		
Inspector	LIGHTSEY,ROBERT,D	Time	10:50:21	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	07622080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

PLCED IN EVED USE AT PTDF 6-14-2023

Data Download Date/Time 06/14/2023 11:01

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency JACKSONVILLE SO **Date** 07/18/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 14:35:57 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

REMOVED FROM EVID USE @ PTDF 7/18/2023

Data Download Date/Time 07/18/2023 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency JACKSONVILLE SO **Date** 08/30/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 15:49:02 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 07622080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

PLCED INTO EVED USE AT PTDF 8-30-2023

Data Download Date/Time 08/30/2023 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency	JACKSONVILLE SO	Date	09/20/2023	Type	Agency Inspection		
Inspector	LIGHTSEY,ROBERT,D	Time	12:41:30	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	07622080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	05/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/20/2023 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency	JACKSONVILLE SO	Date	10/25/2023	Type	Agency Inspection		
Inspector	LIGHTSEY,ROBERT,D	Time	11:58:03	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	08223080A3
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/25/2023 12:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency JACKSONVILLE SO **Date** 11/17/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 12:29:19 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/17/2023 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency JACKSONVILLE SO **Date** 12/15/2023 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 14:51:57 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.085
		Interferent Detect Test # 3	INT

Remarks:

RMVD FRM EVED USE AT PDF FOR DEPT INSPC ON 12-15-2023

Data Download Date/Time 12/15/2023 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Agency JACKSONVILLE SO **Date** 12/22/2023 **Type** Department Inspection
Inspector SIDDOWAY,BENJAMIN,W **Time** 12:21:44 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.047
		0.08 g/210L Test # 4	0.076
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.046
		0.08 g/210L Test # 5	0.076
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.047
0.08 g/210L Test # 6	0.076
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.047
0.08 g/210L Test # 7	0.075
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.047
0.08 g/210L Test # 8	0.076
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.047
0.08 g/210L Test # 9	0.076
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.047
0.08 g/210L Test # 10	0.076
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 12/22/2023 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency	FL HIGHWAY PATROL	Date	01/31/2023	Type	Agency Inspection		
Inspector	GONZALEZ,CHRISTOPHER,M	Time	01:34:23	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	1433249
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	01/08/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED, NO CRACK. IN COMPLIANCE

Data Download Date/Time 02/02/2023 16:30

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency FL HIGHWAY PATROL **Date** 02/28/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ, ANTHONY, N **Time** 13:44:54 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS/ IN COMPLIANCE

Data Download Date/Time 02/28/2023 15:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency FL HIGHWAY PATROL **Date** 03/30/2023 **Type** Agency Inspection
Inspector GONZALEZ,CHRISTOPHER,M **Time** 16:07:57 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED CASE NO CRACKS, IN COMPLIANCE

Data Download Date/Time 03/31/2023 14:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency FL HIGHWAY PATROL **Date** 04/28/2023 **Type** Agency Inspection
Inspector GONZALEZ,CHRISTOPHER,M **Time** 19:16:44 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

CHECKED CASE AND WIPED DOWN. IN COMPLIANCE.

Data Download Date/Time 05/03/2023 00:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency FL HIGHWAY PATROL **Date** 05/30/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ, ANTHONY, N **Time** 00:09:48 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS IN COMPLIANCE

Data Download Date/Time 05/30/2023 00:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency FL HIGHWAY PATROL **Date** 06/22/2023 **Type** Agency Inspection
Inspector GONZALEZ,CHRISTOPHER,M **Time** 17:46:22 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:NOT ENOUGH ALCOHOL.

Data Download Date/Time 06/22/2023 17:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency FL HIGHWAY PATROL **Date** 07/31/2023 **Type** Agency Inspection
Inspector EARLYWINE,GARRETT,V **Time** 01:21:45 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS/ IN COMPLIANCE

Data Download Date/Time 07/31/2023 01:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency FL HIGHWAY PATROL **Date** 08/24/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ, ANTHONY, N **Time** 23:12:54 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS/ IN COMPLIANCE

Data Download Date/Time 08/25/2023 00:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency FL HIGHWAY PATROL	Date 09/11/2023	Type Department Inspection	
Inspector SOTO,ISRAEL,	Time 10:57:09	Software 8100.27	
0.05 Lot # 202303K	0.08 Lot # 202303L	0.20 Lot # 202304C	0.08 Gas Lot # 06723080A5
Expiration 03/29/2025	Expiration 03/29/2025	Expiration 04/05/2025	Expiration 04/05/2025
Number of Simulators 5	Average Standard Deviation 0.0004	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.077
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.077
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.076
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.076
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.076
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.076
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 09/11/2023 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency FL HIGHWAY PATROL **Date** 01/04/2023 **Type** Agency Inspection
Inspector RUBEL, NEIL, F **Time** 11:47:56 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1482048
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/21/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 01/09/2023 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency FL HIGHWAY PATROL **Date** 01/09/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 09:42:39 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG113403
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 05/14/2023

Number of Simulators 5 **Average Standard Deviation** 0.0000 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 01/09/2023 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency	FL HIGHWAY PATROL	Date	01/12/2023	Type	Agency Inspection		
Inspector	BARGE,SUSAN,M	Time	13:10:16	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	1482048
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	04/21/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 01/12/2023 14:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency	FL HIGHWAY PATROL	Date	02/15/2023	Type	Agency Inspection		
Inspector	BARGE,SUSAN,M	Time	14:04:51	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	1482048
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	04/21/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 02/15/2023 15:26

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency FL HIGHWAY PATROL
Inspector RUBEL, NEIL, F

Date 03/06/2023
Time 09:31:23

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1482048
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/21/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 03/06/2023 10:41

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency	FL HIGHWAY PATROL	Date	04/15/2023	Type	Agency Inspection
Inspector	LITTLE,NICHOLAS,M	Time	14:47:16	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	5	Compliance	No
-----------------------------	---	-------------------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	Yes		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Time-Date changed. A F / M A: . NO MOUTH ALCOHOL.

Data Download Date/Time 04/17/2023 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency	FL HIGHWAY PATROL	Date	04/15/2023	Type	Agency Inspection		
Inspector	LITTLE,NICHOLAS,M	Time	15:27:16	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202106B	0.08 Gas Lot #	1482048
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/22/2023	Expiration	04/21/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 04/17/2023 14:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency	FL HIGHWAY PATROL	Date	05/01/2023	Type	Agency Inspection		
Inspector	RUBEL,NEIL,F	Time	14:57:23	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	1482048
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	04/21/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 05/01/2023 16:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency FL HIGHWAY PATROL **Date** 06/20/2023 **Type** Agency Inspection
Inspector LITTLE,NICHOLAS,M **Time** 15:00:45 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1482048
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/21/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 07/06/2023 21:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency FL HIGHWAY PATROL **Date** 07/10/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 09:49:03 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0010
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 07/10/2023 10:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency	FL HIGHWAY PATROL	Date	08/13/2023	Type	Agency Inspection		
Inspector	RUBEL,NEIL,F	Time	10:24:26	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	2402328118
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	01/20/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 08/13/2023 11:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency FL HIGHWAY PATROL
Inspector RUBEL, NEIL, F

Date 09/27/2023
Time 10:47:01

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 2402328118
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/20/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 09/27/2023 11:57

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency FL HIGHWAY PATROL	Date 10/23/2023	Type Agency Inspection	
Inspector ABRAHAM,ERIC,D	Time 15:36:43	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 2402328118
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 01/20/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.20 g/210L Test # (Repeat) 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside TEST WAS REPEATED AFTER SIM WAS AT TEMP

Data Download Date/Time 10/23/2023 16:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency	FL HIGHWAY PATROL	Date	11/22/2023	Type	Agency Inspection		
Inspector	ABRAHAM,ERIC,D	Time	20:16:42	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	2402328118
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	01/20/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/25/2023 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency FL HIGHWAY PATROL
Inspector RUBEL, NEIL, F

Date 12/16/2023
Time 12:39:40

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 2402328118
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/20/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 12/16/2023 13:50

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 01/30/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 00:47:39 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 12/07/2023 **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.033
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # (Repeat) 2	0.034
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test # 3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test # (Repeat) 3	0.034
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance BAD SOLUTION, Control Outside. COLD SIMULATOR

Data Download Date/Time 01/30/2023 00:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 01/30/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 02:54:34 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.038
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.038
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.039
		0.05 g/210L Test # (Repeat) 3	0.045
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.

Data Download Date/Time 02/03/2023 04:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 02/09/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 02:10:44 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.033
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.034
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.034
		0.05 g/210L Test # (Repeat) 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance LOOSE SIM CAP.

Data Download Date/Time 02/09/2023 02:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 02/17/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 16:45:05 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/17/2023 16:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 03/03/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 19:58:17 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/03/2023 21:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 04/25/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 03:55:50 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/25/2023 04:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 05/09/2023 **Type** Agency Inspection
Inspector DRACH,STEFAN,G **Time** 13:32:16 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2023 12:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency	FL HIGHWAY PATROL	Date	05/10/2023	Type	Agency Inspection		
Inspector	DRACH,STEFAN,	Time	09:56:09	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	402357111
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	02/10/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2023 12:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency	FL HIGHWAY PATROL	Date	05/10/2023	Type	Agency Inspection		
Inspector	DRACH,STEFAN,	Time	10:29:11	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	402357111
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	02/10/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2023 12:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency	FL HIGHWAY PATROL	Date	05/10/2023	Type	Agency Inspection		
Inspector	DRACH,STEFAN,	Time	11:04:19	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	402357111
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	02/10/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2023 12:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency	FL HIGHWAY PATROL	Date	05/10/2023	Type	Agency Inspection		
Inspector	DRACH,STEFAN,	Time	11:36:38	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	402357111
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	02/10/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2023 12:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 05/31/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 04:31:09 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/31/2023 12:23

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 06/30/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 20:41:24 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/30/2023 20:56

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 07/14/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 19:03:55 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/14/2023 20:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 08/31/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 10:27:25 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/31/2023 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 09/29/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 12:59:10 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/11/2023 18:15

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 10/20/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 21:22:09 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/20/2023 21:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 11/30/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 19:02:42 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/30/2023 19:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

Agency FL HIGHWAY PATROL **Date** 12/27/2023 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 17:45:53 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/27/2023 17:56

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency FL HIGHWAY PATROL **Date** 01/31/2023 **Type** Agency Inspection
Inspector GONZALEZ,CHRISTOPHER,M **Time** 01:42:24 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 02422080A1
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED, NO CRACKS. IN COMPLIANCE.

Data Download Date/Time 02/01/2023 13:30

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency	FL HIGHWAY PATROL	Date	02/28/2023	Type	Agency Inspection		
Inspector	DOBOSIEWICZ,ANTHONY,N	Time	12:28:01	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	02422080A1
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	02/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO ISSUES/ IN COMPLIANCE

Data Download Date/Time 02/28/2023 14:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency	FL HIGHWAY PATROL	Date	03/30/2023	Type	Agency Inspection		
Inspector	GONZALEZ,CHRISTOPHER,M	Time	19:25:19	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202106B	0.08 Gas Lot #	02422080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/22/2023	Expiration	02/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED CASE NO CRACKS, IN COMPLIANCE

Data Download Date/Time 03/31/2023 14:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency	FL HIGHWAY PATROL	Date	04/19/2023	Type	Agency Inspection		
Inspector	GONZALEZ,CHRISTOPHER,M	Time	06:45:10	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	2422080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	02/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

CHECKED CASE AND WIPED DOWN. IN COMPLIANCE.

Data Download Date/Time 05/03/2023 02:43

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency FL HIGHWAY PATROL **Date** 05/29/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ, ANTHONY, N **Time** 04:12:09 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 02422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS IN COMPLIANCE

Data Download Date/Time 05/29/2023 04:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency FL HIGHWAY PATROL **Date** 05/29/2023 **Type** Agency Inspection
Inspector GONZALEZ,CHRISTOPHER,M **Time** 07:02:37 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 2422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED AND WIPED DOWN

Data Download Date/Time 06/22/2023 18:07

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency	FL HIGHWAY PATROL	Date	07/20/2023	Type	Agency Inspection
Inspector	SIDDOWAY,BENJAMIN,W	Time	09:48:26	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 07/21/2023 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency FL HIGHWAY PATROL	Date 07/21/2023	Type Department Inspection	
Inspector SOTO,ISRAEL,	Time 15:34:43	Software 8100.27	
0.05 Lot # 202201C	0.08 Lot # 202303L	0.20 Lot # 202201E	0.08 Gas Lot # 06723080A5
Expiration 01/11/2024	Expiration 03/29/2025	Expiration 01/18/2024	Expiration 04/05/2025
Number of Simulators 5	Average Standard Deviation 0.0009	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.047
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.046
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0010
0.08 g/210L Test (Std Dev)	0.0007
0.20 g/210L Test (Std Dev)	0.0015
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 07/21/2023 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency FL HIGHWAY PATROL **Date** 08/24/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ, ANTHONY, N **Time** 23:42:45 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 02422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS/ IN COMPLIANCE

Data Download Date/Time 08/25/2023 00:19

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency FL HIGHWAY PATROL **Date** 09/05/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ, ANTHONY, N **Time** 09:28:52 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 02422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS/ IN COMPLIANCE

Data Download Date/Time 09/05/2023 10:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency	FL HIGHWAY PATROL	Date	10/30/2023	Type	Agency Inspection		
Inspector	GONZALEZ,CHRISTOPHER,M	Time	17:38:46	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	2422080A1
Expiration	03/28/2025	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	02/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED OK.

Data Download Date/Time 10/30/2023 17:48

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency	FL HIGHWAY PATROL	Date	11/29/2023	Type	Agency Inspection		
Inspector	DOBOSIEWICZ,ANTHONY,N	Time	05:01:01	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	02422080A1
Expiration	03/28/2025	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	02/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS/ IN COMPLIANCE

Data Download Date/Time 11/29/2023 05:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency	FL HIGHWAY PATROL	Date	12/30/2023	Type	Agency Inspection		
Inspector	DOBOSIEWICZ,ANTHONY,N	Time	02:39:56	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	02422080A1
Expiration	03/28/2025	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	02/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS/ IN COMPLIANCE

Data Download Date/Time 12/30/2023 02:57

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Agency DELAND P.D.
Inspector WALTER,CRAIG,K

Date 01/31/2023
Time 18:51:47

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433248
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/06/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 02/06/2023 14:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Agency DELAND P.D. **Date** 02/27/2023 **Type** Agency Inspection
Inspector WALTER,CRAIG,K **Time** 18:46:33 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433248
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/06/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 03/03/2023 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Agency DELAND P.D. **Date** 03/30/2023 **Type** Agency Inspection
Inspector WALTER,CRAIG,K **Time** 15:53:57 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433248
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/06/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 04/04/2023 13:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Agency	DELAND P.D.	Date	04/30/2023	Type	Agency Inspection		
Inspector	WALTER,CRAIG,K	Time	21:31:23	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	1433248
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	01/06/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 05/05/2023 06:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Agency	DELAND P.D.	Date	06/30/2023	Type	Agency Inspection		
Inspector	WALTER,CRAIG,K	Time	15:07:12	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	1433248
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	01/06/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 06/30/2023 15:18

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Agency DELAND P.D.	Date 07/31/2023	Type Agency Inspection	
Inspector WALTER,CRAIG,K	Time 18:39:09	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202108D	0.20 Lot # 202206C	0.08 Gas Lot # 1433248
Expiration 03/28/2025	Expiration 08/17/2023	Expiration 06/15/2024	Expiration 01/06/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 07/31/2023 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Agency DELAND P.D.	Date 08/14/2023	Type Department Inspection	
Inspector SOTO,ISRAEL,	Time 12:27:43	Software 8100.27	
0.05 Lot # 202303K	0.08 Lot # 202303L	0.20 Lot # 202304C	0.08 Gas Lot # 06723080A5
Expiration 03/29/2025	Expiration 03/29/2025	Expiration 04/05/2025	Expiration 04/05/2025
Number of Simulators 5	Average Standard Deviation 0.0005	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.047
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.047
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.047
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.047
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0009
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 08/14/2023 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

Agency	FL HIGHWAY PATROL	Date	01/24/2023	Type	Agency Inspection		
Inspector	HENING,FRED,C	Time	11:51:21	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	33321080A4
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	01/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.132
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.102
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.098
		0.20 g/210L Test # (Repeat) 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance SIM NOT SECURE.

Data Download Date/Time 02/08/2023 10:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

Agency FL HIGHWAY PATROL **Date** 02/09/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 13:32:20 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2023

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.047
		0.08 g/210L Test # 4	0.076
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.046
		0.08 g/210L Test # 5	0.077
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.047
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.046
0.08 g/210L Test # 7	0.077
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.047
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.047
0.08 g/210L Test # 9	0.077
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.047
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/09/2023 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

Agency	FL HIGHWAY PATROL	Date	02/14/2023	Type	Agency Inspection		
Inspector	HENING,FRED,C	Time	15:05:33	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	33321080A4
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	01/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

00: RFI Detect PHONE TOO CLOSE

Data Download Date/Time 02/14/2023 15:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

Agency	FL HIGHWAY PATROL	Date	02/21/2023	Type	Agency Inspection		
Inspector	HENING,FRED,C	Time	14:30:24	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	33321080A4
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	01/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 02/22/2023 15:05

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

Agency FL HIGHWAY PATROL **Date** 03/24/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 13:05:48 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 03/24/2023 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency	EGLIN AFB	Date	01/23/2023	Type	Agency Inspection		
Inspector	TUSKES,ARTHUR,A	Time	06:44:23	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	1283438
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	02/25/2023
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/23/2023 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency EGLIN AFB	Date 03/06/2023	Type Agency Inspection
Inspector SOTO,ISRAEL,	Time 09:21:30	Software 8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance Yes
-----------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 03/06/2023 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency EGLIN AFB **Date** 03/06/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 14:02:26 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.077
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/06/2023 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency EGLIN AFB	Date 04/13/2023	Type Agency Inspection	
Inspector TUSKES,ARTHUR,A	Time 06:49:01	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 402635303
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 01/25/2026
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/13/2023 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency EGLIN AFB	Date 05/23/2023	Type Agency Inspection	
Inspector TUSKES,ARTHUR,A	Time 06:57:00	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 402635303
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 01/25/2026
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/23/2023 08:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency EGLIN AFB	Date 06/20/2023	Type Agency Inspection	
Inspector TUSKES,ARTHUR,A	Time 06:42:29	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 402635303
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 01/25/2026
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/20/2023 07:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency EGLIN AFB
Inspector CLIFFORD,ALAIN,R

Date 07/17/2023
Time 06:44:23

Type Agency Inspection
Software 8100.27

0.05 Lot # 20211128 **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402635303
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/17/2023 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency EGLIN AFB	Date 07/21/2023	Type Department Inspection	
Inspector SOTO,ISRAEL,	Time 09:48:41	Software 8100.27	
0.05 Lot # 202201C	0.08 Lot # 202303L	0.20 Lot # 202201E	0.08 Gas Lot # 06723080A5
Expiration 01/11/2024	Expiration 03/29/2025	Expiration 01/18/2024	Expiration 04/05/2025
Number of Simulators 5	Average Standard Deviation 0.0004	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 07/21/2023 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency EGLIN AFB	Date 08/28/2023	Type Agency Inspection	
Inspector CLIFFORD,ALAIN,R	Time 06:36:51	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 202206B
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 06/14/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.000
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 08/28/2023 08:18

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency EGLIN AFB	Date 08/28/2023	Type Agency Inspection	
Inspector CLIFFORD,ALAIN,R	Time 07:13:23	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 402635303
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 01/25/2026
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2023 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency EGLIN AFB	Date 09/26/2023	Type Agency Inspection	
Inspector CLIFFORD,ALAIN,R	Time 07:20:16	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 402635303
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 01/25/2026
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/26/2023 08:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency EGLIN AFB **Date** 10/31/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 09:13:34 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 10/31/2023 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency EGLIN AFB	Date 12/21/2023	Type Agency Inspection	
Inspector CLIFFORD,ALAIN,R	Time 07:37:40	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 402635303
Expiration 03/28/2025	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 01/25/2026
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/21/2023 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 01/16/2023
Time 10:29:56

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: No Sample Provided.

Data Download Date/Time 01/16/2023 10:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 02/07/2023
Time 03:54:29

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/07/2023 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO **Date** 02/07/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 12:48:25 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.076
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.076
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

Data Download Date/Time 02/07/2023 12:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 02/07/2023
Time 18:57:11

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.05 g/210L Test # (Repeat) 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance TIGHTEN SIMULATOR.

Data Download Date/Time 02/07/2023 18:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 03/27/2023
Time 13:29:51

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/27/2023 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 04/29/2023
Time 16:12:47

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/03/2023 17:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 05/08/2023
Time 10:30:16

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/08/2023 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 06/30/2023
Time 18:43:24

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/11/2023 15:38

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 07/11/2023
Time 15:31:04

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/11/2023 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 08/28/2023
Time 13:58:50

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2023 14:06

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 09/16/2023
Time 10:50:19

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.045
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.045
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.05 g/210L Test # (Repeat) 3	0.045
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance TIGHTENED SIM/ADD WARM UP.

Data Download Date/Time 09/16/2023 10:57

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 10/13/2023
Time 16:16:10

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.05 g/210L Test # (Repeat) 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance MORE WARMUP TIME IN SIM .

Data Download Date/Time 10/13/2023 16:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO **Date** 11/27/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 11:23:41 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 12/07/2023 **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.045
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # (Repeat) 2	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test # 3	0.045
0.08 g/210L Test: OK	No	0.05 g/210L Test # (Repeat) 3	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance . Non-compliance: .

Data Download Date/Time 11/27/2023 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO **Date** 12/06/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 14:28:30 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 12/06/2023 14:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency PASCO COUNTY SO
Inspector NIXON,BARRY,

Date 12/18/2023
Time 13:32:55

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202307C **0.08 Gas Lot #** 08622080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 07/12/2025 **Expiration** 06/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance FORGOT TO CONNECT HOSE.

Data Download Date/Time 12/18/2023 13:43

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency	FL HIGHWAY PATROL	Date	01/25/2023	Type	Agency Inspection		
Inspector	BARGE,SUSAN,M	Time	10:07:52	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	402477284
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	06/27/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 01/25/2023 11:26

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency FL HIGHWAY PATROL **Date** 01/26/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 13:40:57 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.203
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 01/26/2023 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency FL HIGHWAY PATROL **Date** 02/08/2023 **Type** Agency Inspection
Inspector BARGE,SUSAN,M **Time** 15:16:57 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402477284
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 02/08/2023 16:51

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency	FL HIGHWAY PATROL	Date	03/07/2023	Type	Agency Inspection		
Inspector	RUBEL,NEIL,F	Time	12:39:56	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	2402477284
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/27/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 03/07/2023 13:50

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency	FL HIGHWAY PATROL	Date	04/04/2023	Type	Agency Inspection		
Inspector	RUBEL,NEIL,F	Time	03:24:18	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202106B	0.08 Gas Lot #	2402477284
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/22/2023	Expiration	06/27/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 04/10/2023 09:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency FL HIGHWAY PATROL **Date** 04/11/2023 **Type** Agency Inspection
Inspector SOTO,ISRAEL, **Time** 07:09:58 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

CHECKING FORM UPLOAD

Data Download Date/Time 04/11/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency FL HIGHWAY PATROL **Date** 04/11/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 09:43:10 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.046
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.047
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.046
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.046
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.046
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.046
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.046
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 04/11/2023 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency	FL HIGHWAY PATROL	Date	05/01/2023	Type	Agency Inspection		
Inspector	RUBEL,NEIL,F	Time	10:43:16	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	2402477284
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/27/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 05/01/2023 11:56

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency	FL HIGHWAY PATROL	Date	06/21/2023	Type	Agency Inspection		
Inspector	LITTLE,NICHOLAS,M	Time	12:18:49	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	402477284
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/27/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 06/21/2023 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency FL HIGHWAY PATROL **Date** 07/05/2023 **Type** Agency Inspection
Inspector RUBEL, NEIL, F **Time** 10:20:12 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.045
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.05 g/210L Test # (Repeat) 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance/SEALED SIM HEAD.

Data Download Date/Time 07/05/2023 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency FL HIGHWAY PATROL
Inspector RUBEL, NEIL, F

Date 08/13/2023
Time 10:13:09

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 08/13/2023 11:33

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency FL HIGHWAY PATROL
Inspector RUBEL, NEIL, F

Date 09/21/2023
Time 07:19:03

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.079
		0.05 g/210L Test # (Repeat) 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.INCORRECT SIM ATTACHED.

Data Download Date/Time 09/21/2023 08:30

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency FL HIGHWAY PATROL
Inspector RUBEL, NEIL, F

Date 10/17/2023
Time 23:58:58

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 10/18/2023 01:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency	FL HIGHWAY PATROL	Date	11/28/2023	Type	Agency Inspection		
Inspector	RUBEL,NEIL,F	Time	23:32:29	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	2402843437
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	09/08/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 11/29/2023 00:47

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency	FL HIGHWAY PATROL	Date	12/02/2023	Type	Agency Inspection		
Inspector	RUBEL,NEIL,F	Time	11:32:19	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	2402843437
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	09/08/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 12/02/2023 12:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency FL HIGHWAY PATROL **Date** 03/08/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 12:12:28 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201D **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 03/08/2023 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency	FL HIGHWAY PATROL	Date	04/28/2023	Type	Agency Inspection		
Inspector	BUDAY,JOSHUA,M	Time	11:26:24	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	212654
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	01/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

UNIT CLEAN/FUNCTIONAL AT TIME OF INSPECTION

Data Download Date/Time 04/28/2023 12:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency	FL HIGHWAY PATROL	Date	05/09/2023	Type	Agency Inspection		
Inspector	BUDAY,JOSHUA,M	Time	12:26:26	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106B	0.08 Gas Lot #	212654
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	01/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

UNIT CLEAN/FUNCTIONAL AT TIME OF INSPECTION TEST

Data Download Date/Time 05/09/2023 13:01

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency FL HIGHWAY PATROL **Date** 06/29/2023 **Type** Agency Inspection
Inspector BUDAY,JOSHUA,M **Time** 16:45:26 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 212654
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/02/2023 18:21

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency FL HIGHWAY PATROL Date 07/02/2023 Type Agency Inspection
Inspector BUDAY,JOSHUA,M Time 18:01:57 Software 8100.27

0.05 Lot # 202112A 0.08 Lot # 202206B 0.20 Lot # 202206C 0.08 Gas Lot # 212654
Expiration 12/07/2023 Expiration 06/14/2024 Expiration 06/15/2024 Expiration 01/05/2024

Number of Simulators 5 Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.074
		0.08 g/210L Test # (Repeat) 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AMB/PUR - AREA CLR, OK 2ND .08 - SIM TIGHTENED, OK 2ND ATP

Data Download Date/Time 07/02/2023 18:21

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency FL HIGHWAY PATROL **Date** 08/16/2023 **Type** Agency Inspection
Inspector RIVERA, VICTOR, J **Time** 11:40:29 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 212654
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

CLEAN NO ISSUES

Data Download Date/Time 08/16/2023 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency FL HIGHWAY PATROL **Date** 10/03/2023 **Type** Agency Inspection
Inspector RIVERA, VICTOR, J **Time** 08:45:17 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 212654
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

UNIT CLEAN / FUNCTIONAL

Data Download Date/Time 10/03/2023 08:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency	FL HIGHWAY PATROL	Date	11/22/2023	Type	Agency Inspection		
Inspector	BUDAY,JOSHUA,M	Time	11:54:44	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	212654
Expiration	03/28/2025	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	01/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

UNIT CLEAN/FUNCTIONAL AT TIME OF INSPECTION

Data Download Date/Time 11/22/2023 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency	FL HIGHWAY PATROL	Date	12/28/2023	Type	Agency Inspection		
Inspector	BUDAY,JOSHUA,M	Time	09:47:02	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	402872586
Expiration	03/28/2025	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	10/12/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.191
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail AREA CLEARED/VENT. INCREASED. ALL TESTS OK.

Data Download Date/Time 12/28/2023 09:56

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 01/31/2023
Time 22:26:42

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1356262
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/03/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2023 18:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 02/01/2023
Time 01:40:22

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1356262
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/03/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2023 18:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 03/31/2023
Time 22:48:25

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1356262
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/03/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 03/31/2023 22:55

Inspection Test Electronic Data

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 04/01/2023
Time 01:36:16

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 1356262 **0.20 Lot #** 202106B **0.08 Gas Lot #** 1356262
Expiration 12/07/2023 **Expiration** 08/03/2023 **Expiration** 06/22/2023 **Expiration** 08/03/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 04/01/2023 01:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Agency FHP TROOP E MIAMI **Date** 05/31/2023 **Type** Agency Inspection
Inspector LORENZO, YOSNIEL, **Time** 23:36:26 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1356262
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 08/03/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 07/01/2023 00:54

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 07/01/2023
Time 00:41:02

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1356262
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/03/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 07/01/2023 00:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 08/01/2023
Time 00:26:51

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1356262
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 08/03/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.084
		Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 08/01/2023 00:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 08/31/2023
Time 23:39:45

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1356262
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/03/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Improper Sample.MOUTHPIECE NOT PROPERLY INSTALLE

Data Download Date/Time 08/31/2023 23:47

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 09/30/2023
Time 19:54:34

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/24/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 09/30/2023 20:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 10/31/2023
Time 15:13:31

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/24/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/31/2023 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Agency FHP TROOP E MIAMI
Inspector BRANDT, FABIAN, E

Date 11/29/2023
Time 19:08:34

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/24/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

TOO MUCH MOUTH ALCOHOL.

Data Download Date/Time 11/29/2023 20:06

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Agency	FL HIGHWAY PATROL	Date	12/05/2023	Type	Department Inspection		
Inspector	GUTSCHOW,TAYLOR,D	Time	15:32:05	Software	8100.27		
0.05 Lot #	202303K	0.08 Lot #	202303L	0.20 Lot #	202304C	0.08 Gas Lot #	01923080A3
Expiration	03/29/2025	Expiration	03/29/2025	Expiration	04/05/2025	Expiration	02/05/2025
Number of Simulators	5	Average Standard Deviation	0.0003	Compliance	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.082
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 12/05/2023 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 01/31/2023
Time 22:20:05

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/15/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2023 18:33

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 02/01/2023
Time 01:33:51

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/15/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2023 18:33

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 03/31/2023
Time 22:30:24

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/15/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail MOUTH PIECE ATTACHED.

Data Download Date/Time 03/31/2023 22:41

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Inspection Test Electronic Data

COBRA ©1997-2014 CMI, Inc.

Inspection Test Electronic Data

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 05/31/2023
Time 23:28:57

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/15/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI Detect.RADION ON NEAR

Data Download Date/Time 05/31/2023 23:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

Agency FHP TROOP E MIAMI
Inspector LORENZO, YOSNIEL,

Date 07/01/2023
Time 00:34:49

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/15/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 07/01/2023 00:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001122

Agency FHP TROOP E MIAMI **Date** 08/31/2023 **Type** Agency Inspection
Inspector LORENZO, YOSNIEL, **Time** 23:08:07 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** **0.08 Gas Lot #**
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	INT
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.079
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # (Repeat) 2	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test # 2	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test # 3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test # 3	0.049
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI Detect TROOPER PHONE NEAR. 08: Interferent Detect SPILL CLEANED, Interferent Detect. Non-compliance UNK.

Data Download Date/Time 10/04/2023 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency FHP TROOP D ORLANDO
Inspector RIVERA, VICTOR, J

Date 01/18/2023
Time 23:21:53

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402498227
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/22/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

CLEAN NO ISSUES

Data Download Date/Time 01/20/2023 12:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency FHP TROOP D ORLANDO **Date** 02/18/2023 **Type** Agency Inspection
Inspector JEAN BAPTISTE,ALBENS, **Time** 23:31:02 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402477284
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 02/22/2023 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency	FHP TROOP D ORLANDO	Date	03/15/2023	Type	Agency Inspection		
Inspector	RIVERA,VICTOR,J	Time	12:52:50	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202108D	0.20 Lot #	202106N	0.08 Gas Lot #	402477284
Expiration	12/07/2023	Expiration	08/17/2023	Expiration	06/22/2023	Expiration	06/27/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

CLEAN NO ISSUES

Data Download Date/Time 03/16/2023 07:03

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency FHP TROOP D ORLANDO **Date** 04/22/2023 **Type** Agency Inspection
Inspector JEAN BAPTISTE,ALBENS, **Time** 17:36:03 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402477284
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI Detect.

Data Download Date/Time 04/25/2023 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency FHP TROOP D ORLANDO **Date** 05/30/2023 **Type** Agency Inspection
Inspector JEAN BAPTISTE,ALBENS, **Time** 10:31:34 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402477284
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:MORE MOUTH ALCOHOL.

Data Download Date/Time 05/30/2023 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency FHP TROOP D ORLANDO **Date** 06/28/2023 **Type** Agency Inspection
Inspector JEAN BAPTISTE,ALBENS, **Time** 23:49:32 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

UNIT CLEAN NO ISSUE

Data Download Date/Time 09/14/2023 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency FHP TROOP D ORLANDO **Date** 09/14/2023 **Type** Agency Inspection
Inspector SIDDOWAY,BENJAMIN,W **Time** 08:55:04 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

COMPLIANCE NOT DETERMINED, BYPASS FOR DI

Data Download Date/Time 09/14/2023 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency	FHP TROOP D ORLANDO	Date	09/14/2023	Type	Department Inspection
Inspector	SIDDOWAY,BENJAMIN,W	Time	09:52:01	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Average Standard Deviation	Compliance	Yes
-----------------------------	-----------------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	No		
Barometric Pressure Sensor Check: OK	No		
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Data Download Date/Time 09/14/2023 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency FHP TROOP D ORLANDO **Date** 09/14/2023 **Type** Department Inspection
Inspector SIDDOWAY,BENJAMIN,W **Time** 13:30:12 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 06723080A5
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 09/14/2023 13:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency	FHP TROOP D ORLANDO	Date	09/27/2023	Type	Agency Inspection		
Inspector	RIVERA,VICTOR,J	Time	08:26:27	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	402477284
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	06/27/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: RFI Detect.

Data Download Date/Time 10/03/2023 07:58

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency FHP TROOP D ORLANDO
Inspector RIVERA, VICTOR, J

Date 10/03/2023
Time 08:56:13

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

UNIT CLEAN / FUNCTIONAL

Data Download Date/Time 10/03/2023 09:07

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency	FHP TROOP D ORLANDO	Date	12/29/2023	Type	Agency Inspection		
Inspector	BUDAY,JOSHUA,M	Time	12:44:06	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	402477284
Expiration	03/28/2025	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	06/27/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail AREA CHECKED AIRFLOW INC. ALL TESTS OK. . ALL TESTS OK. UNIT CLEAN/FUNCTIONAL.

Data Download Date/Time 12/29/2023 12:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO
Inspector COLE,KEVIN,T

Date 01/10/2023
Time 11:03:43

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/10/2023 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO **Date** 02/17/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 10:29:29 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.203
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.047
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.047
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.047
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0010
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/17/2023 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO
Inspector COLE,KEVIN,T

Date 03/14/2023
Time 10:08:23

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/12/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 03/14/2023 10:32

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO
Inspector COLE,KEVIN,T

Date 04/11/2023
Time 10:28:41

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 04/11/2023 10:59

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO
Inspector COLE,KEVIN,T

Date 05/18/2023
Time 09:47:34

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.084
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 05/18/2023 12:04

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO	Date 06/13/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 10:23:02	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 26422080A1
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 11/05/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 06/13/2023 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO
Inspector COLE,KEVIN,T

Date 06/13/2023
Time 11:27:03

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 06/13/2023 11:51

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO
Inspector COLE,KEVIN,T

Date 07/13/2023
Time 10:17:57

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 07/13/2023 10:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO
Inspector COLE,KEVIN,T

Date 08/15/2023
Time 09:52:10

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 08/15/2023 10:19

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO
Inspector COLE,KEVIN,T

Date 09/12/2023
Time 10:51:33

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 09/12/2023 11:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO
Inspector COLE,KEVIN,T

Date 10/10/2023
Time 10:07:55

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.030
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: . 08: Control Outside Tolerance.

Data Download Date/Time 10/10/2023 11:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO
Inspector COLE,KEVIN,T

Date 10/10/2023
Time 10:39:37

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 26422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2024

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/10/2023 11:09

Printed: 24 January 2024
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO	Date 11/14/2023	Type Agency Inspection	
Inspector COLE,KEVIN,T	Time 10:02:29	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 26422080A1
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 11/05/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 11/14/2023 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency POLK COUNTY SO
Inspector COLE,KEVIN,T

Date 12/05/2023
Time 09:44:19

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/05/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 12/05/2023 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency PUTNAM COUNTY SO
Inspector SMITH,BRIAN,K

Date 01/19/2023
Time 13:05:26

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 28021080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

M/A SOLUTION STILL IN OPERATORS MOUTH.

Data Download Date/Time 01/23/2023 08:36

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency PUTNAM COUNTY SO
Inspector SMITH,BRIAN,K

Date 02/16/2023
Time 09:24:08

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 28021080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/16/2023 09:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency PUTNAM COUNTY SO
Inspector SMITH,BRIAN,K

Date 03/16/2023
Time 12:32:30

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 28021080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/16/2023 13:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency PUTNAM COUNTY SO
Inspector SMITH,BRIAN,K

Date 04/28/2023
Time 10:03:25

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 28021080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/28/2023 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency PUTNAM COUNTY SO
Inspector SMITH,BRIAN,K

Date 05/04/2023
Time 14:02:31

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 28021080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 05/04/2023 14:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency PUTNAM COUNTY SO
Inspector SMITH,BRIAN,K

Date 06/09/2023
Time 08:30:06

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 28021080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/09/2023 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency PUTNAM COUNTY SO
Inspector SMITH,BRIAN,K

Date 07/28/2023
Time 12:04:27

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 28021080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/28/2023 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency PUTNAM COUNTY SO **Date** 08/29/2023 **Type** Agency Inspection
Inspector SMITH,BRIAN,K **Time** 10:23:13 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 28021080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

BOOKING SCANNER OPERATING.

Data Download Date/Time 08/29/2023 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency PUTNAM COUNTY SO **Date** 09/29/2023 **Type** Agency Inspection
Inspector SMITH,BRIAN,K **Time** 10:18:19 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 28021080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/29/2023 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency PUTNAM COUNTY SO	Date 10/30/2023	Type Agency Inspection	
Inspector SMITH,BRIAN,K	Time 11:47:16	Software 8100.27	
0.05 Lot # 202112A	0.08 Lot # 202206B	0.20 Lot # 202206C	0.08 Gas Lot # 28021080A2
Expiration 12/07/2023	Expiration 06/14/2024	Expiration 06/15/2024	Expiration 12/05/2023
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.187
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test # (Repeat) 1	0.197
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test # 1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.20 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test # 2	0.195
		0.20 g/210L Test # (Repeat) 2	0.197
		0.08 g/210L Dry Gas Std Test # 2	0.077
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.075
		0.08 g/210L Test # (Repeat) 3	0.075
		0.20 g/210L Test # 3	0.197
		0.20 g/210L Test # (Repeat) 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

LOOSE FITTING ON HOSE. LOOSE FITTING.

Data Download Date/Time 10/30/2023 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency PUTNAM COUNTY SO **Date** 11/02/2023 **Type** Department Inspection
Inspector SIDDOWAY,BENJAMIN,W **Time** 14:06:56 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 11/02/2023 14:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency	PUTNAM COUNTY SO	Date	11/27/2023	Type	Agency Inspection		
Inspector	SMITH,BRIAN,K	Time	10:05:04	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	26422080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	11/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.20 g/210L Test # (Repeat) 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

LOOSE CONNECTION.

Data Download Date/Time 11/27/2023 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

Agency PUTNAM COUNTY SO	Date 12/21/2023	Type Agency Inspection	
Inspector SMITH,BRIAN,K	Time 09:27:48	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 26422080A1
Expiration 03/28/2025	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 11/05/2024
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.185
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.20 g/210L Test # (Repeat) 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

LOOSE FITTING.

Data Download Date/Time 12/21/2023 09:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

Agency	PUTNAM COUNTY SO	Date	02/21/2023	Type	Agency Inspection
Inspector	NICODEMO,PHIL,	Time	09:29:41	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED

Data Download Date/Time 02/21/2023 12:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

Agency PUTNAM COUNTY SO **Date** 02/21/2023 **Type** Department Inspection
Inspector NICODEMO, PHIL, **Time** 12:19:41 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/21/2023 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

Agency	PUTNAM COUNTY SO	Date	04/28/2023	Type	Agency Inspection		
Inspector	SMITH,BRIAN,K	Time	10:36:28	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	26422080A1
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	11/05/2024
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.201
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.20 g/210L Test # (Repeat) 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

LOOSE FITTING.

Data Download Date/Time 04/28/2023 10:59

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O.
Inspector STONE,JIMMIE,N

Date 01/19/2023
Time 12:52:59

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/25/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/23/2023 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O. **Date** 02/02/2023 **Type** Agency Inspection
Inspector STONE,JIMMIE,N **Time** 12:15:20 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/02/2023 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O. **Date** 03/02/2023 **Type** Agency Inspection
Inspector STONE,JIMMIE,N **Time** 12:49:20 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/02/2023 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O. **Date** 03/07/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 10:45:11 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.077
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.076
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.076
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.075
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.076
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.076
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0011
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/07/2023 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O.
Inspector STONE,JIMMIE,N

Date 04/21/2023
Time 15:49:43

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/21/2023 16:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O. **Date** 05/30/2023 **Type** Agency Inspection
Inspector STONE,JIMMIE,N **Time** 13:05:26 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2023 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O. **Date** 06/13/2023 **Type** Agency Inspection
Inspector STONE,JIMMIE,N **Time** 11:42:15 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/13/2023 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O. **Date** 07/19/2023 **Type** Agency Inspection
Inspector STONE,JIMMIE,N **Time** 13:22:31 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/19/2023 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O. **Date** 08/23/2023 **Type** Agency Inspection
Inspector STONE,JIMMIE,N **Time** 11:59:34 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/23/2023 12:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O. **Date** 09/12/2023 **Type** Agency Inspection
Inspector STONE,JIMMIE,N **Time** 13:03:20 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/12/2023 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O.
Inspector STONE,JIMMIE,N

Date 10/19/2023
Time 10:56:33

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.075
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/19/2023 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

Agency VOLUSIA COUNTY S.O.
Inspector STONE,JIMMIE,N

Date 11/16/2023
Time 11:16:36

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/16/2023 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency VOLUSIA COUNTY SO
Inspector STONE,JIMMIE,N

Date 01/19/2023
Time 13:00:28

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402419647
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/14/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # (Repeat) 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

08: RFI Detect.

Data Download Date/Time 01/20/2023 11:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency VOLUSIA COUNTY SO
Inspector STONE,JIMMIE,N

Date 02/02/2023
Time 12:23:52

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance-CONNECTED DRY GAS .

Data Download Date/Time 02/02/2023 13:04

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency VOLUSIA COUNTY SO
Inspector STONE,JIMMIE,N

Date 03/02/2023
Time 12:58:04

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/02/2023 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency VOLUSIA COUNTY SO
Inspector STONE,JIMMIE,N

Date 04/21/2023
Time 15:43:24

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/21/2023 16:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency VOLUSIA COUNTY SO
Inspector STONE,JIMMIE,N

Date 05/30/2023
Time 13:12:43

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2023 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency	VOLUSIA COUNTY SO	Date	06/06/2023	Type	Department Inspection		
Inspector	NICODEMO,PHIL,	Time	11:56:19	Software	8100.27		
0.05 Lot #	202201C	0.08 Lot #	202201D	0.20 Lot #	202201E	0.08 Gas Lot #	06723080A5
Expiration	01/11/2024	Expiration	01/18/2024	Expiration	01/18/2024	Expiration	04/05/2025
Number of Simulators	5	Average Standard Deviation	0.0004	Compliance	Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 06/06/2023 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency VOLUSIA COUNTY SO
Inspector STONE,JIMMIE,N

Date 07/19/2023
Time 13:25:29

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/19/2023 13:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency VOLUSIA COUNTY SO
Inspector STONE,JIMMIE,N

Date 08/23/2023
Time 12:00:27

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/23/2023 12:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency VOLUSIA COUNTY SO
Inspector STONE,JIMMIE,N

Date 09/12/2023
Time 13:12:23

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

FORGOT TO CONNECT DRY GAS.

Data Download Date/Time 09/12/2023 13:18

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency VOLUSIA COUNTY SO
Inspector STONE,JIMMIE,N

Date 10/19/2023
Time 10:49:21

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/19/2023 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency VOLUSIA COUNTY SO
Inspector STONE,JIMMIE,N

Date 11/16/2023
Time 11:25:35

Type Agency Inspection
Software 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/16/2023 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency	VOLUSIA COUNTY SO	Date	12/05/2023	Type	Agency Inspection		
Inspector	MORGAN,MICHAEL,B	Time	09:47:07	Software	8100.27		
0.05 Lot #	202112A	0.08 Lot #	202206B	0.20 Lot #	202206C	0.08 Gas Lot #	402419647
Expiration	12/07/2023	Expiration	06/14/2024	Expiration	06/15/2024	Expiration	04/14/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/05/2023 10:05