# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 01/31/2023 **Type** Agency Inspection
**Inspector** GONZALEZ,CHRISTOPHER,M **Time** 02:03:07 **Software** 8100.27

**0.05 Lot #** 2021112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 24229080A1
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CASE CHECKED, NO CRACKS. IN COMPLIANCE.

**Data Download Date/Time** 02/02/2023 15:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** DOBOSIEWICZ,ANTHONY,N **Time** 13:31:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 24229080A1
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance FORGOT TO CONNECT DRY GAS . CASE CHECKED NO CRACKS/ IN COMPLIANCE

**Data Download Date/Time** 02/28/2023 14:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 03/30/2023 **Type** Agency Inspection
**Inspector** GONZALEZ,CHRISTOPHER,M **Time** 16:03:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 24229080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 02/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INSPECTED CASE NO CRACKS, IN COMPLIANCE

**Data Download Date/Time** 03/31/2023 16:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 04/28/2023 **Type** Agency Inspection
**Inspector** GONZALEZ,CHRISTOPHER,M **Time** 17:55:02 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 24229080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CHECKED CASE AND WIPED DOWN. IN COMPLIANCE.

**Data Download Date/Time** 05/03/2023 02:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 05/17/2023 **Type** Agency Inspection
**Inspector** MONTGOMERY,KENNETH,R **Time** 14:59:04 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/17/2023 15:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 06/01/2023 **Type** Agency Inspection
**Inspector** MONTGOMERY,KENNETH,R **Time** 13:37:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/01/2023 13:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 06/18/2023 **Type** Agency Inspection
**Inspector** MONTGOMERY,KENNETH,R **Time** 15:51:00 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/22/2023 12:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 06/22/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 11:49:26 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.081 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.081 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0007 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 06/22/2023 12:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 06/23/2023 **Type** Agency Inspection
**Inspector** MONTGOMERY,KENNETH,R **Time** 18:43:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #**
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.178 |
| Alcohol Free Subject Test: 0.000 | Yes | Interferent Detect Test # 1 | INT |
| Mouth Alcohol Test: Slope Not Met | Yes | Alcohol Free Test # 2 | 0.000 |
| Interferent Detect Test: Interferent Detect | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Diagnostic Check (Post-Inspection): OK | No | 0.08 g/210L Test # 2 | 0.079 |
| Alcohol Free Test: OK | Yes | 0.20 g/210L Test # 2 | 0.183 |
| 0.05 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| 0.20 g/210L Test: OK | No | 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Dry Gas Std Test: OK | No | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.185 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance, Ambient Fail. SIMULATOR FAILED TO REACH TEMPERTURE.

**Data Download Date/Time** 06/23/2023 19:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 06/23/2023 **Type** Agency Inspection
**Inspector** MONTGOMERY,KENNETH,R **Time** 19:15:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/23/2023 19:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 07/18/2023 **Type** Agency Inspection
**Inspector** MONTGOMERY,KENNETH,R **Time** 15:13:39 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.05 g/210L Test # (Repeat) 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.05 g/210L Test # (Repeat) 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: RFI Detect.

**Data Download Date/Time** 07/18/2023 15:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 08/14/2023 **Type** Agency Inspection
**Inspector** MONTGOMERY,KENNETH,R **Time** 15:18:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/14/2023 15:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 09/07/2023 **Type** Agency Inspection
**Inspector** MONTGOMERY,KENNETH,R **Time** 16:27:13 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/07/2023 16:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 10/19/2023 **Type** Agency Inspection
**Inspector** MONTGOMERY,KENNETH,R **Time** 19:23:13 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: RFI Detect.

**Data Download Date/Time** 10/19/2023 19:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 11/12/2023 **Type** Agency Inspection
**Inspector** MONTGOMERY,KENNETH,R **Time** 13:31:59 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

DRY GAS NOT CONNECTED.

**Data Download Date/Time** 11/12/2023 13:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

**Agency** FHP TROOP C **Date** 12/19/2023 **Type** Agency Inspection
**Inspector** MONTGOMERY,KENNETH,R **Time** 19:32:04 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/19/2023 19:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOLMES BEACH PD | **Date** | 01/03/2023 | **Type** | Agency Inspection |
| **Inspector** | DESANTIS,ADAM,B | **Time** | 12:13:26 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402419647 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/14/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT WAS CHECKED FOR DAMAGES. NONE WERE OBSERVED.

**Data Download Date/Time** 01/06/2023 12:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOLMES BEACH PD | **Date** | 02/14/2023 | **Type** | Agency Inspection |
| **Inspector** | HURT,ALEXANDER,C | **Time** | 15:10:51 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | Yes | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

00:INCORRECT SIMULATOR ATTACHED.

**Data Download Date/Time** 02/14/2023 16:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

**Agency** HOLMES BEACH PD **Date** 02/14/2023 **Type** Agency Inspection
**Inspector** HURT,ALEXANDER,C **Time** 15:56:10 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

EXTERIOR CASE INSPECTED AND NO SIGNS OF DAMAGE OBSERVED.

**Data Download Date/Time** 02/14/2023 16:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

**Agency** HOLMES BEACH PD **Date** 03/14/2023 **Type** Agency Inspection
**Inspector** DESANTIS,ADAM,B **Time** 09:18:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT WAS INSPECTED FOR DAMAGES. NONE WERE DISCOVERED.

**Data Download Date/Time** 03/14/2023 09:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOLMES BEACH PD | **Date** | 04/10/2023 | **Type** | Agency Inspection |
| **Inspector** | DESANTIS,ADAM,B | **Time** | 15:32:27 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402419647 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/14/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT WAS INSPECTED FOR DAMAGES. NONE WERE FOUND

**Data Download Date/Time** 04/10/2023 15:55

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOLMES BEACH PD | **Date** | 05/22/2023 | **Type** | Agency Inspection |
| **Inspector** | HURT,ALEXANDER,C | **Time** | 15:29:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402419647 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/14/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

EXTERIOR CHECKED NO DAMAGE OBSERVED

**Data Download Date/Time** 05/22/2023 16:51

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

**Agency** HOLMES BEACH PD **Date** 06/08/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:47:11 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 06/08/2023 12:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOLMES BEACH PD | **Date** | 06/14/2023 | **Type** | Agency Inspection |
| **Inspector** | HURT,ALEXANDER,C | **Time** | 23:51:34 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402419647 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/14/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NO NEW DAMAGE OBSERVED AFTER TRAVELING

**Data Download Date/Time** 06/15/2023 09:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

**Agency** HOLMES BEACH PD **Date** 07/13/2023 **Type** Agency Inspection
**Inspector** DESANTIS,ADAM,B **Time** 08:48:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 40241964
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT WAS INSPECTED FOR DAMAGES. NONE WERE OBSERVED.

**Data Download Date/Time** 07/13/2023 08:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

**Agency** HOLMES BEACH PD **Date** 08/15/2023 **Type** Agency Inspection
**Inspector** DESANTIS,ADAM,B **Time** 10:34:51 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 40241964
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT WAS CHECKED AND NO DAMAGES WERE OBSERVED.

**Data Download Date/Time** 08/15/2023 10:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOLMES BEACH PD | **Date** | 09/11/2023 | **Type** | Agency Inspection |
| **Inspector** | HURT,ALEXANDER,C | **Time** | 18:16:10 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 40241964 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/14/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT INSPECTED AND WAS FREE OF DAMAGE

**Data Download Date/Time** 09/12/2023 16:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HOLMES BEACH PD | **Date** | 10/14/2023 | **Type** | Agency Inspection |
| **Inspector** | DESANTIS,ADAM,B | **Time** | 10:50:27 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402419647 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/14/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT WAS CHECK FOR DAMAGES. NONE WERE DISCOVERED.

**Data Download Date/Time** 10/18/2023 09:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

**Agency** HOLMES BEACH PD **Date** 11/16/2023 **Type** Agency Inspection
**Inspector** DESANTIS,ADAM,B **Time** 10:16:37 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT INSPECTED. NO DAMAGES FOUND

**Data Download Date/Time** 11/21/2023 10:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

**Agency** MONROE COUNTY SO **Date** 01/11/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 13:37:23 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

SPARE INTOX

**Data Download Date/Time** 01/11/2023 13:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

**Agency** MONROE COUNTY SO **Date** 02/16/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 12:11:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

SPARE INTOX

**Data Download Date/Time** 02/16/2023 12:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

**Agency** MONROE COUNTY SO **Date** 03/06/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 10:37:53 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

POST-EVIDENTIARY AI COMPLETED. BEING SENT FOR DEPT INSP.

**Data Download Date/Time** 03/06/2023 10:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

**Agency** MONROE COUNTY SO **Date** 04/07/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:36:16 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 04/07/2023 12:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

**Agency** MONROE COUNTY SO **Date** 05/03/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 13:45:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PRE-EVIDENTIARY AI. CAME BACK FROM ANNUAL DEPT INSP

**Data Download Date/Time** 05/03/2023 14:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

**Agency** MONROE COUNTY SO **Date** 06/07/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 10:58:53 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402477284
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KW INTOX

**Data Download Date/Time** 06/08/2023 09:53

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

**Agency** MONROE COUNTY SO **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 11:06:19 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/05/2023 11:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

**Agency** MONROE COUNTY SO **Date** 08/02/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 10:15:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KW SPARE INTOX

**Data Download Date/Time** 08/02/2023 10:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

**Agency** MONROE COUNTY SO **Date** 09/14/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 11:49:48 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

SPARE INTOX

**Data Download Date/Time** 09/14/2023 11:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

**Agency** MONROE COUNTY SO **Date** 10/04/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 12:10:08 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

SPARE INTOX

**Data Download Date/Time** 10/04/2023 12:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

**Agency** MONROE COUNTY SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 10:23:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

SPARE INTOX

**Data Download Date/Time** 11/08/2023 11:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

**Agency** MONROE COUNTY SO **Date** 12/20/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 13:30:00 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

SPARE INTOX

**Data Download Date/Time** 12/20/2023 13:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 01/09/2023 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 09:32:50 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402419647
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

DEP CLINE PRESENT

**Data Download Date/Time** 01/09/2023 09:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 02/06/2023 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM,D **Time** 14:28:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402419647
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | No |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/06/2023 15:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 02/06/2023 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 15:09:16 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402419647
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/06/2023 15:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 03/13/2023 **Type** Agency Inspection
**Inspector** CLINE,GLENN, **Time** 10:00:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402419647
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/13/2023 10:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 04/10/2023 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM,D **Time** 14:12:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402419647
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/10/2023 14:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 05/15/2023 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM,D **Time** 09:45:04 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402419647
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/15/2023 10:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 06/12/2023 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 13:46:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402419647
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/14/2024 **Expiration** 04/14/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | Yes | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 06/12/2023 14:01

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 06/23/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 11:51:33 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 06/23/2023 11:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 07/20/2023 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM,D **Time** 15:31:48 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/20/2023 16:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 08/14/2023 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM,D **Time** 13:23:16 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 03/28/2025 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/14/2023 13:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 09/18/2023 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM,D **Time** 09:26:01 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/18/2023 09:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 10/09/2023 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 09:13:27 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/09/2023 09:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 11/06/2023 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 17:46:32 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.018 |
| 0.05 g/210L Test # (Repeat) 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.019 |
| 0.05 g/210L Test # (Repeat) 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.022 |
| 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time** 11/06/2023 09:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 11/08/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 13:28:10 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

TEST OF FORMS RELOAD. AI NOT CONDUCTED.

**Data Download Date/Time** 11/08/2023 13:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

**Agency** MANATEE COUNTY SO **Date** 11/16/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:12:12 **Software** 8100.27

**0.05 Lot #** 202303K **Expiration** 03/29/2025
**0.08 Lot #** 202303L **Expiration** 03/29/2025
**0.20 Lot #** 202304C **Expiration** 04/05/2025
**0.08 Gas Lot #** AG223802 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 11/16/2023 12:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 01/09/2023 | **Type** | Agency Inspection |
| **Inspector** | LEDFORD,TAYLOR,M | **Time** | 09:53:16 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202101C | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | 402357111 |
| **Expiration** | 01/12/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/09/2023 10:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

**Agency** FHP TROOP F **Date** 02/06/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR,M **Time** 10:03:02 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.000 |
| 0.20 g/210L Test # (Repeat) 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.000 |
| 0.20 g/210L Test # (Repeat) 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.000 |
| 0.20 g/210L Test # (Repeat) 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

DID NOT CONNECT.

**Data Download Date/Time** 02/06/2023 10:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

**Agency** FHP TROOP F **Date** 03/13/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR, **Time** 10:21:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

DRY GAS TUBE NOT ATTACHED FOR DGS

**Data Download Date/Time** 03/13/2023 10:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 04/10/2023 | **Type** | Agency Inspection |
| **Inspector** | LEDFORD,TAYLOR, | **Time** | 10:05:35 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | 402357111 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 04/10/2023 10:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

**Agency** FHP TROOP F **Date** 05/15/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR, **Time** 09:34:41 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/15/2023 10:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

**Agency** FHP TROOP F **Date** 06/12/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR,M **Time** 09:32:55 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/14/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/12/2023 10:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

**Agency** FHP TROOP F **Date** 07/20/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR,M **Time** 10:00:56 **Software** 8100.27

**0.05 Lot #** 2023J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: RFI Detect.

**Data Download Date/Time** 07/20/2023 11:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

**Agency** FL HIGHWAY PATROL **Date** 08/08/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:57:28 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

| | |
|---|---|
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 5 | 0.079 |
| Interferent Detect Test # 5 | INT |
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.047 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.047 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.077 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.047 |
| 0.08 g/210L Test # 10 | 0.077 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10 | 0.079 |
| Interferent Detect Test # 10 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

| | |
|---|---|
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0332 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat) | 0.0004 |

**Remarks:**

08: Control Outside Tolerance. DID NOT ATTACH DGS HOSE..

**Data Download Date/Time** 08/08/2023 14:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

**Agency** KENNEDY SPACE CENTER **Date** 01/23/2023 **Type** Agency Inspection
**Inspector** HOFFMANN,JOHN,K **Time** 21:47:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/23/2023 22:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | KENNEDY SPACE CENTER | **Date** | 10/02/2023 | **Type** | Agency Inspection |
| **Inspector** | SIDDOWAY,BENJAMIN,W | **Time** | 08:50:52 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

COMPLIANCE NOT DETERMINED, BYPASS TO OPERATE

**Data Download Date/Time** 10/02/2023 12:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

**Agency** KENNEDY SPACE CENTER **Date** 10/02/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 11:59:38 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 06723080A5
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

| | |
|---|---|
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 5 | 0.079 |
| Interferent Detect Test # 5 | INT |
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10 | 0.079 |
| Interferent Detect Test # 10 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

| | |
|---|---|
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0334 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat) | 0.0004 |

**Remarks:**

08: Control Outside Tolerance INLET UNPLUGGED.

**Data Download Date/Time** 10/02/2023 12:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

**Agency** KENNEDY SPACE CENTER **Date** 11/29/2023 **Type** Agency Inspection
**Inspector** TORRES,ISMAEL,J **Time** 11:43:26 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 14323020A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/29/2023 11:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FHP **Date** 01/04/2023 **Type** Agency Inspection
**Inspector** RUBEL,NEIL,F **Time** 11:51:09 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1327124
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 05/14/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 01/06/2023 14:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FL HIGHWAY PATROL **Date** 01/09/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 08:54:48 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG113403
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 05/14/2023

**Number of Simulators** 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 01/09/2023 11:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FL HIGHWAY PATROL **Date** 01/12/2023 **Type** Agency Inspection
**Inspector** BARGE,SUSAN,M **Time** 13:04:26 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1327124
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 05/14/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.044 |
| 0.05 g/210L Test # (Repeat) 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.074 |
| 0.08 g/210L Test # (Repeat) 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.190 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.05 g/210L Test # (Repeat) 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.08 g/210L Test # (Repeat) 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.08 g/210L Test # (Repeat) 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance. 08: Control Outside Toleran e.OUTSIDE CONTROL TOLERANCE .05 .08 TIGHTEN LID

**Data Download Date/Time** 01/12/2023 14:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FL HIGHWAY PATROL **Date** 02/16/2023 **Type** Agency Inspection
**Inspector** BARGE,SUSAN,M **Time** 11:57:00 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1327124
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 05/14/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 02/16/2023 13:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FL HIGHWAY PATROL **Date** 03/06/2023 **Type** Agency Inspection
**Inspector** RUBEL,NEIL,F **Time** 09:26:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1327124
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/14/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 03/06/2023 10:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FL HIGHWAY PATROL **Date** 04/04/2023 **Type** Agency Inspection
**Inspector** RUBEL,NEIL,F **Time** 03:25:14 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1327124
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 05/14/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 04/07/2023 23:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FL HIGHWAY PATROL **Date** 05/10/2023 **Type** Agency Inspection
**Inspector** RUBEL,NEIL,F **Time** 13:27:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 2402477284
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 05/10/2023 14:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FL HIGHWAY PATROL **Date** 06/21/2023 **Type** Agency Inspection
**Inspector** RUBEL,NEIL,F **Time** 07:27:13 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 2402477284
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 06/21/2023 08:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FL HIGHWAY PATROL **Date** 07/03/2023 **Type** Agency Inspection
**Inspector** RUBEL,NEIL,F **Time** 18:31:45 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 2402477284
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 07/03/2023 19:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FL HIGHWAY PATROL **Date** 08/13/2023 **Type** Agency Inspection
**Inspector** RUBEL,NEIL,F **Time** 10:27:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 2402477284
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 08/13/2023 11:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FL HIGHWAY PATROL **Date** 09/21/2023 **Type** Agency Inspection
**Inspector** RUBEL,NEIL,F **Time** 07:31:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 2402477284
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 09/21/2023 08:53

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FL HIGHWAY PATROL **Date** 10/23/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 13:09:59 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # (Repeat) 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # (Repeat) 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # (Repeat) 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # (Repeat) 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # (Repeat) 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # (Repeat) 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # (Repeat) 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # (Repeat) 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # (Repeat) 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # (Repeat) 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) (Repeat) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

08: Ambient Fail CLEARED AND AIRED OUT AREA WITH FAN.

**Data Download Date/Time** 10/23/2023 14:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 11/19/2023 | **Type** | Agency Inspection |
| **Inspector** | RUBEL,NEIL,F | **Time** | 14:17:55 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402477284 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/27/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 11/19/2023 15:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

**Agency** FL HIGHWAY PATROL **Date** 12/11/2023 **Type** Agency Inspection
**Inspector** RUBEL,NEIL,F **Time** 15:00:44 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 12/11/2023 16:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 06/13/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 12:32:39 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE

**Data Download Date/Time** 06/14/2023 11:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

**Agency** FL HIGHWAY PATROL **Date** 06/14/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 11:52:44 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.047 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 06/14/2023 11:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

**Agency** FL HIGHWAY PATROL **Date** 07/20/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR,M **Time** 10:12:52 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/20/2023 12:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

**Agency** FL HIGHWAY PATROL **Date** 08/14/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR,M **Time** 10:25:55 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/14/2023 10:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

**Agency** FL HIGHWAY PATROL **Date** 09/18/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR,M **Time** 09:43:58 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/18/2023 09:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

**Agency** FL HIGHWAY PATROL **Date** 10/09/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR,M **Time** 09:39:20 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/09/2023 09:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

**Agency** FL HIGHWAY PATROL **Date** 11/06/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR,M **Time** 10:07:11 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/06/2023 10:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

**Agency** FL HIGHWAY PATROL **Date** 12/04/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR, **Time** 09:49:32 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/04/2023 09:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

**Agency** FHP TROOP L **Date** 01/05/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 00:02:26 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/19/2023 10:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

**Agency** FHP TROOP L **Date** 02/23/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 12:30:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/23/2023 12:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

**Agency** FHP TROOP L **Date** 03/29/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 11:51:46 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.075 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/29/2023 11:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

**Agency** FHP TROOP L **Date** 04/18/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 14:43:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/18/2023 14:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

**Agency** FHP TROOP L **Date** 05/30/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 12:31:08 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: LOSE HOSE.

**Data Download Date/Time** 05/30/2023 12:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

**Agency** FHP TROOP L **Date** 06/28/2023 **Type** Agency Inspection
**Inspector** JONES,CORTLAND,T **Time** 14:40:12 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2023 14:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP L | **Date** | 08/07/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 09:53:11 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

**Data Download Date/Time** 08/08/2023 13:01

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

**Agency** FL HIGHWAY PATROL **Date** 08/08/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:54:53 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 08/08/2023 13:01

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

**Agency** FL HIGHWAY PATROL **Date** 08/25/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 12:44:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/25/2023 12:51

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

**Agency** FL HIGHWAY PATROL **Date** 09/27/2023 **Type** Agency Inspection
**Inspector** JONES,CORTLAND,T **Time** 11:01:08 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2023 11:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

**Agency** FL HIGHWAY PATROL **Date** 10/26/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 09:44:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/26/2023 09:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

**Agency** FL HIGHWAY PATROL **Date** 11/29/2023 **Type** Agency Inspection
**Inspector** JONES,CORTLAND,T **Time** 10:11:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/29/2023 10:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006762

**Agency** FL HIGHWAY PATROL **Date** 12/15/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 16:53:23 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/15/2023 17:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

**Agency** FHP TROOP G **Date** 03/24/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 13:39:12 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.047 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.076 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.047 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.047 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.047 |
| 0.08 g/210L Test # 9 | 0.077 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.047 |
| 0.08 g/210L Test # 10 | 0.077 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 03/24/2023 13:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP G | **Date** | 04/26/2023 | **Type** | Agency Inspection |
| **Inspector** | FARLEY,JOSEPH,E | **Time** | 11:40:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | | **0.20 Lot #** | | **0.08 Gas Lot #** | |
| **Expiration** | 12/07/2023 | **Expiration** | | **Expiration** | | **Expiration** | |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | INT |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # 1 | INT |
| Alcohol Free Subject Test: 0.000 | Yes | Alcohol Free Test # 2 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.05 g/210L Test # 2 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 2 | INT |
| Diagnostic Check (Post-Inspection): OK | No | Alcohol Free Test # 3 | 0.000 |
| Alcohol Free Test: OK | Yes | Interferent Detect Test # 3 | INT |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

05: Interferent Detect, SOLUTION BEING CHANGED Non-compliance: .

**Data Download Date/Time** 05/05/2023 09:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

**Agency** FHP TROOP G **Date** 08/08/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 10:16:58 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 08/08/2023 10:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP G | **Date** | 10/19/2023 | **Type** | Agency Inspection |
| **Inspector** | SMITH,MATTHEW,M | **Time** | 20:51:09 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | Interferent Detect Test # 1 | 0.000 |
| Minimum Sample Volume Check: OK | | Alcohol Free Test # 2 | 0.000 |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # 2 | 0.000 |
| Alcohol Free Subject Test: 0.000 | Yes | Alcohol Free Test # 3 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 3 | 0.000 |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | Yes | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

Int Det: .

**Data Download Date/Time** 10/19/2023 21:52

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

**Agency** FHP TROOP G **Date** 10/19/2023 **Type** Agency Inspection
**Inspector** SMITH,MATTHEW,M **Time** 21:35:33 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.05 g/210L Test # (Repeat) 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.05 g/210L Test # (Repeat) 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | |
| 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: RFI Detect.

**Data Download Date/Time** 10/19/2023 21:52

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 12/04/2023 | **Type** | Agency Inspection |
| **Inspector** | SIDDOWAY,BENJAMIN,W | **Time** | 09:26:08 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

COMPLIANCE NOT DETERMINED, BYPASS AI TO OPERATE

**Data Download Date/Time** 12/04/2023 15:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

**Agency** FL HIGHWAY PATROL **Date** 12/04/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 15:00:37 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0008 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

**Remarks:**

**Data Download Date/Time** 12/04/2023 15:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

**Agency** FHP TROOP G **Date** 01/25/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 19:49:48 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.20 g/210L Test # (Repeat) 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.20 g/210L Test # (Repeat) 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | |
| 0.20 g/210L Test # (Repeat) 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

20: RFI Detect. RADIO ON

**Data Download Date/Time** 01/27/2023 08:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

**Agency** FHP TROOP G **Date** 02/15/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 09:15:14 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/15/2023 10:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

**Agency** FHP TROOP G **Date** 03/28/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 12:05:02 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/28/2023 12:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP G | **Date** | 04/26/2023 | **Type** | Agency Inspection |
| **Inspector** | FARLEY,JOSEPH,E | **Time** | 13:32:00 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402623828 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/26/2023 15:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

**Agency** FHP TROOP G **Date** 05/17/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 13:53:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/17/2023 15:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

**Agency** FHP TROOP G **Date** 06/01/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 12:52:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/01/2023 13:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

**Agency** FHP TROOP G **Date** 07/18/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 14:18:34 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/18/2023 14:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

**Agency** FHP TROOP G **Date** 08/29/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 20:47:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.05 g/210L Test # (Repeat) 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.05 g/210L Test # (Repeat) 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # (Repeat) 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: RFI Detect.RADIO ON

**Data Download Date/Time** 09/05/2023 11:55

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

**Agency** FHP TROOP G **Date** 09/07/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 16:25:06 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/07/2023 16:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP G | **Date** | 10/19/2023 | **Type** | Agency Inspection |
| **Inspector** | FARLEY,JOSEPH,E | **Time** | 19:19:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402623828 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/19/2023 19:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

**Agency** FHP TROOP G **Date** 11/12/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 13:25:12 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: RFI Detect. RADIO ON DURING TEST

**Data Download Date/Time** 11/13/2023 11:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

**Agency** FL HIGHWAY PATROL **Date** 11/13/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 11:05:26 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 11/13/2023 11:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 01/24/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 09:13:38 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1408275 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 11/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INSPECTED. NO MARKS OR CRACKS.

**Data Download Date/Time** 01/26/2023 15:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

**Agency** FHP TROOP F **Date** 01/26/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:40:21 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 00521080A2
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 02/05/2023

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 01/26/2023 15:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 02/13/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 11:32:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1408275 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 11/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

VISUALLY INSPECTED. NO MARKS OR CRACKS.

**Data Download Date/Time** 02/13/2023 13:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 03/08/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 09:24:13 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1408275 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 11/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NO MARKS OR CRACKS.

**Data Download Date/Time** 03/08/2023 12:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 04/04/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 12:22:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1408275 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 11/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

DETECTED ALCOHOL AND NOT SLOPE NOT MET.

**Data Download Date/Time** 04/05/2023 10:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 05/22/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 11:40:11 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | Interferent Detect Test # 1 | 0.000 |
| Minimum Sample Volume Check: OK | | Alcohol Free Test # 2 | 0.000 |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # 2 | 0.000 |
| Alcohol Free Subject Test: 0.000 | Yes | Alcohol Free Test # 3 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 3 | 0.000 |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | Yes | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

SIMULATOR NOT HEATED ENOUGH. Non-compliance: .

**Data Download Date/Time** 05/25/2023 10:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

**Agency** FHP TROOP F **Date** 05/23/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 11:14:33 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1408275
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/25/2023 10:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 06/27/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 09:40:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1408275 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 11/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NO MARKS OR CRACKS.

**Data Download Date/Time** 06/29/2023 09:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 07/28/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 09:21:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1408275 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 11/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

VISUALLY INSPECTED. NO MARKS OR CRACKS.

**Data Download Date/Time** 08/01/2023 10:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 08/01/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 09:26:04 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1408275 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 11/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

VISUALLY INSPECTED. NO MARKS OR CRACKS.

**Data Download Date/Time** 08/01/2023 10:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 09/26/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 14:30:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1408275 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 11/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/11/2023 13:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 10/11/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 12:21:51 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202307A | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1408275 |
| **Expiration** | 03/28/2025 | **Expiration** | 07/11/2025 | **Expiration** | 06/15/2024 | **Expiration** | 11/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.190 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/11/2023 13:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 11/22/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 10:07:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202307A | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024028434 |
| **Expiration** | 03/28/2025 | **Expiration** | 07/11/2025 | **Expiration** | 06/15/2024 | **Expiration** | 09/08/2026 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/05/2023 22:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 12/05/2023 | **Type** | Agency Inspection |
| **Inspector** | BLAKE,DAVID,A | **Time** | 04:24:04 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

GETTING DEVICE OUT OF DISABLED MODE FOR STABILITY TEST

**Data Download Date/Time** 12/05/2023 22:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

**Agency** FHP TROOP F **Date** 12/05/2023 **Type** Agency Inspection
**Inspector** BLAKE,DAVID,A **Time** 05:24:42 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024028434
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NO CRACKS OR DAMAGE

**Data Download Date/Time** 12/05/2023 22:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

**Agency** FL HIGHWAY PATROL **Date** 03/13/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR, **Time** 10:11:04 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | 0.000 |
| Interferent Detect Test # (Repeat) 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Interferent Detect Test # (Repeat) 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

Int Det:NOT ATTATCHED.

**Data Download Date/Time** 03/13/2023 10:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

**Agency** FL HIGHWAY PATROL **Date** 04/10/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR, **Time** 09:53:33 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/10/2023 10:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

**Agency** FL HIGHWAY PATROL **Date** 05/15/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR, **Time** 09:27:50 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

DEP CLINE PRESENT

**Data Download Date/Time** 05/15/2023 10:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

**Agency** FL HIGHWAY PATROL **Date** 06/09/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 15:10:03 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 06/09/2023 15:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

**Agency** FL HIGHWAY PATROL **Date** 07/20/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR,M **Time** 10:15:29 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/20/2023 11:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 08/14/2023 | **Type** | Agency Inspection |
| **Inspector** | LEDFORD,TAYLOR,M | **Time** | 10:16:33 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402357111 |
| **Expiration** | 03/28/2025 | **Expiration** | 08/17/2023 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/14/2023 10:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

**Agency** FL HIGHWAY PATROL **Date** 09/18/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR,M **Time** 09:33:39 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/18/2023 09:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

**Agency** FL HIGHWAY PATROL **Date** 10/09/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR, **Time** 09:29:57 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 10/09/2023 09:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

**Agency** FL HIGHWAY PATROL **Date** 11/06/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR,M **Time** 10:11:54 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/06/2023 10:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

**Agency** FL HIGHWAY PATROL **Date** 12/04/2023 **Type** Agency Inspection
**Inspector** LEDFORD,TAYLOR,M **Time** 09:52:30 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/04/2023 10:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 01/27/2023 | **Type** | Agency Inspection |
| **Inspector** | MARCUM,JOSEPH,K | **Time** | 08:58:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | 1355293 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 07/29/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 01/27/2023 15:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 02/16/2023 | **Type** | Agency Inspection |
| **Inspector** | BARGE,SUSAN,M | **Time** | 12:04:41 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | 1355293 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 07/29/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 02/16/2023 13:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

**Agency** FHP **Date** 02/21/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 10:43:45 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.076 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 02/21/2023 11:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

**Agency** FHP **Date** 03/30/2023 **Type** Agency Inspection
**Inspector** WALLEN,JUSTIN,R **Time** 10:31:00 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1355293
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 07/29/2023

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

DID NOT CONNECT A/F IN A/F TESTNon-compliance: .

**Data Download Date/Time** 03/30/2023 14:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 03/30/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 11:12:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1355293 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/15/2024 | **Expiration** | 07/29/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 03/30/2023 14:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 04/21/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 13:45:22 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1355293 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 07/29/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 04/21/2023 16:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 05/16/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 11:06:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1355293 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 07/29/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 05/16/2023 14:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 06/14/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 10:04:10 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 06/14/2023 13:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 07/19/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 11:56:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | No | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

MOUTH ALCOHOL PRESENT IN A/F ERROR IN CLEARING MOUTH OF ALCNon-compliance:MOUTH NOT CLEAR ALCOHOL IN A/F.

**Data Download Date/Time** 07/19/2023 15:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

**Agency** FHP **Date** 07/19/2023 **Type** Agency Inspection
**Inspector** WALLEN,JUSTIN,R **Time** 12:35:25 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024023571
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | No |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: .

**Data Download Date/Time** 07/19/2023 15:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 07/19/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 12:47:31 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A: Range Exceeded.

**Data Download Date/Time** 07/19/2023 15:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 07/19/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 13:21:32 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 07/19/2023 15:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 08/28/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 11:14:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.180 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.168 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.182 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.179 |
| 0.20 g/210L Test: OK | No | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.189 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.177 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 08/28/2023 13:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 08/28/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 12:29:06 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 08/28/2023 13:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 09/25/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 08:56:44 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 09/25/2023 10:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 10/24/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 08:53:55 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 10/24/2023 10:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 11/21/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 11:47:45 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 11/21/2023 13:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 12/04/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 09:56:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/14/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 12/04/2023 11:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006769

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 12/19/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 09:14:50 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

**Data Download Date/Time** 12/19/2023 13:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006769

**Agency** FL HIGHWAY PATROL **Date** 12/19/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:19:18 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 01923080A3
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 02/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.082 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.082 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006769

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 12/19/2023 13:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 01/24/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 09:14:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INSPECTED. NO MARKS OR CRACKS.

**Data Download Date/Time** 01/26/2023 14:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

**Agency** FHP TROOP F **Date** 01/26/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:27:44 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 00521080A2
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 02/05/2023

**Number of Simulators** 5 **Average Standard Deviation** 0.0000 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 01/26/2023 14:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

**Agency** FHP TROOP F **Date** 02/13/2023 **Type** Agency Inspection
**Inspector** TERUZ,GABRIEL,S **Time** 11:36:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024023571
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

FIRST DGST TESTS DID NOT HAVE GAS PLUGGED IN.

**Data Download Date/Time** 02/13/2023 13:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

**Agency** FHP TROOP F **Date** 03/08/2023 **Type** Agency Inspection
**Inspector** TERUZ,GABRIEL,S **Time** 09:23:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024023571
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NO MARKS OR CRACKS.

**Data Download Date/Time** 03/08/2023 12:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

**Agency** FHP TROOP F **Date** 04/04/2023 **Type** Agency Inspection
**Inspector** TERUZ,GABRIEL,S **Time** 12:16:52 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024023571
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.219 |
| 0.20 g/210L Test # (Repeat) 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.216 |
| 0.20 g/210L Test # (Repeat) 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.212 |
| 0.20 g/210L Test # (Repeat) 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CONTACTED FDLE. SWAPPED SIMULATORS. READINGS CORRECTED.

**Data Download Date/Time** 04/05/2023 10:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 05/22/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 11:40:47 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | No |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | No |
| 0.08 g/210L Test: OK | No |
| 0.20 g/210L Test: OK | No |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test # 3 | 0.000 |

**Remarks:**

SIMULATOR NOT HEATED ENOUGH.

**Data Download Date/Time** 05/25/2023 10:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

**Agency** FHP TROOP F **Date** 05/23/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 13:58:56 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 3024023571
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # (Repeat) 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # (Repeat) 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # (Repeat) 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

MISTAKENLY USED 0.08 SIM. FOR 0.20 TEST.REPLACED&TESTPASSED.

**Data Download Date/Time** 05/25/2023 10:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 06/27/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 09:42:38 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.053 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NO CRACKS OR MARKS.

**Data Download Date/Time** 06/29/2023 09:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

**Agency** FHP TROOP F **Date** 07/28/2023 **Type** Agency Inspection
**Inspector** TERUZ,GABRIEL,S **Time** 09:24:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024023571
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

VISUALLY INSPECTED. NO MARKS OR CRACKS.

**Data Download Date/Time** 08/01/2023 10:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 08/01/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 09:24:58 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

VISUALLY INSPECTED. NO MARKS OR CRACKS.

**Data Download Date/Time** 08/01/2023 10:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 09/26/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 13:42:39 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | No |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | No |
| 0.08 g/210L Test: OK | No |
| 0.20 g/210L Test: OK | No |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test # 3 | 0.000 |

**Remarks:**

**Data Download Date/Time** 10/11/2023 13:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

**Agency** FHP TROOP F **Date** 09/26/2023 **Type** Agency Inspection
**Inspector** TERUZ,GABRIEL,S **Time** 14:29:30 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024023571
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Range Exceeded. 00: Ambient Fail.

**Data Download Date/Time** 10/11/2023 13:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 10/11/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 12:19:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202307A | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 03/28/2025 | **Expiration** | 07/11/2025 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/11/2023 13:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 11/22/2023 | **Type** | Agency Inspection |
| **Inspector** | TERUZ,GABRIEL,S | **Time** | 10:08:47 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202307A | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 3024023571 |
| **Expiration** | 03/28/2025 | **Expiration** | 07/11/2025 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/05/2023 22:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP F | **Date** | 12/05/2023 | **Type** | Agency Inspection |
| **Inspector** | BLAKE,DAVID,A | **Time** | 04:23:23 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

GETTING INTOX OUT OF DISABLED MODE FOR STABILITY TESTNon-compliance: .

**Data Download Date/Time** 12/05/2023 22:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

**Agency** FHP TROOP F **Date** 12/05/2023 **Type** Agency Inspection
**Inspector** BLAKE,DAVID,A **Time** 05:25:28 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024023571
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NO CRACKS OR DAMAGE

**Data Download Date/Time** 12/05/2023 22:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

**Agency** FL HIGHWAY PATROL **Date** 03/28/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 11:48:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/28/2023 12:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

**Agency** FL HIGHWAY PATROL **Date** 04/13/2023 **Type** Agency Inspection
**Inspector** MILLS,SCOTT,T **Time** 20:54:09 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.05 g/210L Test # (Repeat) 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test # (Repeat) 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.05 g/210L Test # (Repeat) 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: RFI Detect.

**Data Download Date/Time** 04/18/2023 09:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

**Agency** FL HIGHWAY PATROL **Date** 05/30/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 23:42:24 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2023 13:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 06/01/2023 | **Type** | Agency Inspection |
| **Inspector** | FARLEY,JOSEPH,E | **Time** | 12:33:06 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | INT |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | Yes | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A: Ambient FailAmbient Fail. 00: Interferent Detect. CLEARED AREA, REPEATEDNon-compliance: .

**Data Download Date/Time** 09/27/2023 13:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 06/07/2023 | **Type** | Agency Inspection |
| **Inspector** | NICODEMO,PHIL, | **Time** | 14:20:25 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED

**Data Download Date/Time** 09/27/2023 13:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

**Agency** FL HIGHWAY PATROL **Date** 12/22/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 10:31:37 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 12/22/2023 11:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

**Agency** FHP TROOP L **Date** 01/19/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 10:20:25 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/19/2023 10:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

**Agency** FHP TROOP L **Date** 02/23/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 12:22:51 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/23/2023 12:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

**Agency** FHP TROOP L **Date** 03/29/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 11:54:30 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 06021080A3
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 04/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/29/2023 12:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

**Agency** FHP TROOP L **Date** 04/18/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 14:56:27 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08:LOOSE DRY GAS HOSE.

**Data Download Date/Time** 04/18/2023 15:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP L | **Date** | 05/30/2023 | **Type** | Agency Inspection |
| **Inspector** | LIMB,HUNTER, | **Time** | 12:16:18 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1474747 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/12/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/30/2023 12:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

**Agency** FHP TROOP L **Date** 06/28/2023 **Type** Agency Inspection
**Inspector** JONES,CORTLAND,T **Time** 14:24:55 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | 0.000 |
| Interferent Detect Test # (Repeat) 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | 0.000 |
| Interferent Detect Test # (Repeat) 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | 0.000 |
| Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

LOOSE HOSE ON SIM.

**Data Download Date/Time** 06/28/2023 14:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

**Agency** FHP TROOP L **Date** 07/26/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 11:30:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/26/2023 11:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

**Agency** FHP TROOP L **Date** 08/25/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 11:25:12 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/25/2023 11:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

**Agency** FL HIGHWAY PATROL **Date** 09/07/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:21:08 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 09/07/2023 14:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

**Agency** FL HIGHWAY PATROL **Date** 09/27/2023 **Type** Agency Inspection
**Inspector** JONES,CORTLAND,T **Time** 11:02:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2023 11:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

**Agency** FL HIGHWAY PATROL **Date** 10/26/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 09:35:38 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 23621080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/26/2023 09:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

**Agency** FL HIGHWAY PATROL **Date** 11/29/2023 **Type** Agency Inspection
**Inspector** JONES,CORTLAND,T **Time** 10:32:33 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # (Repeat) 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.08 g/210L Test # (Repeat) 1 | 0.082 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # (Repeat) 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.08 g/210L Test # (Repeat) 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.074 |
| 0.08 g/210L Test # (Repeat) 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: Ambient Fail. 08:LOOSE HOSE

**Data Download Date/Time** 11/29/2023 10:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006772

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 12/15/2023 | **Type** | Agency Inspection |
| **Inspector** | LIMB,HUNTER, | **Time** | 17:00:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1474747 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/12/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/15/2023 17:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

**Agency** MIAMI-DADE PD **Date** 01/05/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 04:41:32 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 01/05/2023 08:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

**Agency** MIAMI-DADE PD **Date** 02/16/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 03:43:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.082 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 02/16/2023 07:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

**Agency** MIAMI-DADE PD **Date** 03/02/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 06:08:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 03/02/2023 07:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

**Agency** MIAMI-DADE PD **Date** 04/07/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 11:23:28 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0007 |

**Remarks:**

**Data Download Date/Time** 04/07/2023 11:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

**Agency** MIAMI-DADE PD **Date** 05/11/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 04:24:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI G PACHECO PRESENT

**Data Download Date/Time** 05/11/2023 07:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

**Agency** MIAMI-DADE PD **Date** 06/02/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 04:17:13 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 06/02/2023 08:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

**Agency** MIAMI-DADE PD **Date** 07/06/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 02:08:48 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI G PACHECO PRESENT

**Data Download Date/Time** 07/06/2023 02:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

**Agency** MIAMI-DADE PD **Date** 08/17/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 04:02:00 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 08/17/2023 04:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

**Agency** MIAMI-DADE PD **Date** 09/01/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 03:00:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.052 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/01/2023 03:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

**Agency** MIAMI-DADE PD **Date** 10/05/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 01:15:55 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 10/05/2023 02:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

**Agency** MIAMI-DADE PD **Date** 11/03/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 04:46:41 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202106B **0.20 Lot #** 202106C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 11/03/2023 05:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

**Agency** MIAMI-DADE PD **Date** 12/01/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 04:02:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 12/01/2023 04:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

**Agency** BROWARD COUNTY SO **Date** 01/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 14:20:13 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 01/04/2023 14:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

**Agency** BROWARD COUNTY SO **Date** 02/01/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:26:35 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING

**Data Download Date/Time** 02/01/2023 14:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

**Agency** BROWARD COUNTY SO **Date** 03/11/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:53:48 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 03/11/2023 11:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

**Agency** BROWARD COUNTY SO **Date** 04/12/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:21:05 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

**Remarks:**

**Data Download Date/Time** 04/12/2023 12:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

**Agency** BROWARD COUNTY SO **Date** 05/03/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 13:59:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 05/03/2023 14:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

**Agency** BROWARD COUNTY SO **Date** 06/07/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 13:32:10 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 06/07/2023 14:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BROWARD COUNTY SO | **Date** | 07/05/2023 | **Type** | Agency Inspection |
| **Inspector** | FRAZIER,ANAYA,S | **Time** | 09:39:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402477283 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/24/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/05/2023 10:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

**Agency** BROWARD COUNTY SO **Date** 08/02/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:52:18 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | No |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 08/02/2023 12:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

**Agency** BROWARD COUNTY SO **Date** 09/06/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:07:07 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 09/06/2023 11:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

**Agency** BROWARD COUNTY SO **Date** 10/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:23:04 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 10/04/2023 10:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

**Agency** BROWARD COUNTY SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:26:38 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING

**Data Download Date/Time** 11/08/2023 11:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

**Agency** BROWARD COUNTY SO **Date** 12/13/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:38:45 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 202206C
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/15/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/13/2023 12:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

**Agency** BROWARD COUNTY SO **Date** 01/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 13:07:15 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.044 |
| 0.05 g/210L Test # (Repeat) 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test # (Repeat) 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance.JGUZMAN IN TRAINING

**Data Download Date/Time** 01/04/2023 13:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

**Agency** BROWARD COUNTY SO **Date** 02/01/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:14:40 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.084 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING

**Data Download Date/Time** 02/01/2023 13:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

**Agency** BROWARD COUNTY SO **Date** 07/17/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:58:05 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 07/17/2023 13:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

**Agency** BROWARD COUNTY SO **Date** 08/02/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:08:15 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J.GUZMAN IN TRAINING.

**Data Download Date/Time** 08/02/2023 11:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

**Agency** BROWARD COUNTY SO **Date** 09/06/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:59:06 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 09/06/2023 12:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

**Agency** BROWARD COUNTY SO **Date** 10/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:14:49 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING.

**Data Download Date/Time** 10/04/2023 10:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

**Agency** BROWARD COUNTY SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:26:53 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING.

**Data Download Date/Time** 11/08/2023 11:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

**Agency** BROWARD COUNTY SO **Date** 12/13/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:34:06 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/13/2023 13:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

**Agency** TALLAHASSEE PD **Date** 01/24/2023 **Type** Agency Inspection
**Inspector** HARRIS,JAMES,J **Time** 08:51:27 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 34520080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 02/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/24/2023 10:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | TALLAHASSEE PD | **Date** | 02/02/2023 | **Type** | Agency Inspection |
| **Inspector** | HARRIS,JAMES,J | **Time** | 08:07:21 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | 33922080A1 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/22/2023 | **Expiration** | 01/05/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/02/2023 08:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

**Agency** TALLAHASSEE PD **Date** 03/24/2023 **Type** Agency Inspection
**Inspector** HARRIS,JAMES,J **Time** 14:59:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.084 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.075 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/27/2023 10:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | TALLAHASSEE PD | **Date** | 04/25/2023 | **Type** | Agency Inspection |
| **Inspector** | HARRIS,JAMES,J | **Time** | 08:14:34 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 33922080A1 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/05/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.084 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/25/2023 09:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

**Agency** TALLAHASSEE PD **Date** 05/02/2023 **Type** Agency Inspection
**Inspector** HARRIS,JAMES,J **Time** 12:42:26 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/02/2023 13:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

**Agency** TALLAHASSEE PD **Date** 05/10/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 11:50:53 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 05/10/2023 11:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

**Agency** TALLAHASSEE PD **Date** 06/01/2023 **Type** Agency Inspection
**Inspector** HARRIS,JAMES,J **Time** 08:11:42 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/01/2023 12:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

**Agency** TALLAHASSEE PD **Date** 07/20/2023 **Type** Agency Inspection
**Inspector** HARRIS,JAMES,J **Time** 14:18:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/20/2023 14:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

**Agency** TALLAHASSEE PD **Date** 08/25/2023 **Type** Agency Inspection
**Inspector** HARRIS,JAMES,J **Time** 04:33:52 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/25/2023 05:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

**Agency** TALLAHASSEE PD **Date** 09/14/2023 **Type** Agency Inspection
**Inspector** HARRIS,JAMES,J **Time** 08:09:05 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/14/2023 08:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

**Agency** TALLAHASSEE PD **Date** 10/30/2023 **Type** Agency Inspection
**Inspector** HARRIS,JAMES,J **Time** 09:12:19 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/30/2023 09:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

**Agency** TALLAHASSEE PD **Date** 11/01/2023 **Type** Agency Inspection
**Inspector** HARRIS,JAMES,J **Time** 08:30:04 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/01/2023 08:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

**Agency** TALLAHASSEE PD **Date** 12/22/2023 **Type** Agency Inspection
**Inspector** HARRIS,JAMES,J **Time** 12:03:27 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.084 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.084 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/22/2023 12:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

**Agency** BAY COUNTY SO **Date** 01/23/2023 **Type** Agency Inspection
**Inspector** MILLER,BRIAN,D **Time** 14:40:36 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** AG121002
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/29/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/24/2023 13:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BAY COUNTY SO | **Date** | 02/27/2023 | **Type** | Agency Inspection |
| **Inspector** | MILLER,BRIAN,D | **Time** | 11:52:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | AG121002 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 07/29/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/27/2023 13:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

**Agency** BAY COUNTY SO **Date** 03/29/2023 **Type** Agency Inspection
**Inspector** MILLER,BRIAN,D **Time** 09:22:13 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** AG121002
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/29/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/29/2023 13:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

**Agency** BAY COUNTY SO **Date** 04/30/2023 **Type** Agency Inspection
**Inspector** MILLER,BRIAN,D **Time** 15:35:06 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** AG121002
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/29/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/01/2023 13:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BAY COUNTY SO | **Date** | 05/15/2023 | **Type** | Agency Inspection |
| **Inspector** | SOTO,ISRAEL, | **Time** | 08:30:22 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

CHECKING FORMS UPLOAD

**Data Download Date/Time** 05/15/2023 11:23

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

**Agency** BAY COUNTY SO **Date** 05/15/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 10:08:25 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006937

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 05/15/2023 11:23

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

**Agency** FORT LAUDERDALE PD **Date** 01/09/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 21:34:49 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

OFC CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 01/17/2023 20:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

**Agency** FORT LAUDERDALE PD **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 20:22:00 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC J CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 02/28/2023 20:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

**Agency** FORT LAUDERDALE PD **Date** 03/27/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 21:34:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

OFC CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 04/04/2023 20:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

**Agency** FORT LAUDERDALE PD **Date** 04/25/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 20:32:36 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 04/25/2023 20:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

**Agency** FORT LAUDERDALE PD **Date** 05/23/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 22:46:07 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance DRY GAS HOSE NOT ATTACHED

**Data Download Date/Time** 05/23/2023 23:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

**Agency** FORT LAUDERDALE PD **Date** 06/27/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 23:00:38 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 06/27/2023 23:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

**Agency** FORT LAUDERDALE PD **Date** 08/29/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 22:08:06 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC J CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 08/29/2023 22:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

**Agency** FORT LAUDERDALE PD **Date** 09/19/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 21:26:14 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC J CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 09/19/2023 21:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

**Agency** FORT LAUDERDALE PD **Date** 10/24/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 22:22:33 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

OFC J CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 10/24/2023 22:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FORT LAUDERDALE PD | **Date** | 12/01/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 10:10:45 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

**Data Download Date/Time** 12/01/2023 14:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

**Agency** FORT LAUDERDALE PD **Date** 12/01/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:34:56 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 01923080A3
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 02/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 12/01/2023 14:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

**Agency** FORT LAUDERDALE PD **Date** 12/26/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 23:14:42 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 12/27/2023 20:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

**Agency** FORT LAUDERDALE PD **Date** 06/08/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 11:54:32 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 06/08/2023 12:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

**Agency** FORT LAUDERDALE PD **Date** 06/27/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 23:04:45 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC J CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 06/27/2023 23:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

**Agency** FORT LAUDERDALE PD **Date** 07/30/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 20:53:09 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance DRY GAS HOSE NOT SEATED.

**Data Download Date/Time** 07/30/2023 21:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

**Agency** FORT LAUDERDALE PD **Date** 08/29/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 22:00:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC J CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 08/29/2023 22:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

**Agency** FORT LAUDERDALE PD **Date** 09/19/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 20:54:25 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFFICER CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 09/19/2023 21:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

**Agency** FORT LAUDERDALE PD **Date** 10/24/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 22:25:24 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: No Sample Provided. TIMED OUT

**Data Download Date/Time** 10/24/2023 22:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

**Agency** FORT LAUDERDALE PD **Date** 11/28/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 22:00:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 11/28/2023 22:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

**Agency** FORT LAUDERDALE PD **Date** 12/26/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 23:06:09 **Software** 8100.27

**0.05 Lot #** 20303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 202206B
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/14/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.000 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.000 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance, LOOSE HOSE . LOOSE HOSE

**Data Download Date/Time** 12/27/2023 20:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 01/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:39:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 01/04/2023 10:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 02/01/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 15:06:00 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINNG

**Data Download Date/Time** 02/01/2023 15:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 02/16/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:23:00 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 02/16/2023 14:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 03/11/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 14:09:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 03/13/2023 13:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 04/12/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:15:03 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/12/2023 10:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 05/03/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:54:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 05/03/2023 12:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 06/07/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:43:41 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 06/07/2023 11:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:36:06 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 07/05/2023 11:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 08/02/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:13:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 08/02/2023 10:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BROWARD COUNTY SO | **Date** | 08/02/2023 | **Type** | Agency Inspection |
| **Inspector** | FRAZIER,ANAYA,S | **Time** | 10:20:11 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A: Sequence Aborted. ALREADY DID INSPECTION. ACCIDE. ALREADY DID INSPECTION, DID NOT SWITCH. ACCIDENTAL.

**Data Download Date/Time** 08/02/2023 10:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 09/06/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 13:12:10 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 09/06/2023 13:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 10/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:03:18 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 10/04/2023 11:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:22:58 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 11/08/2023 10:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

**Agency** BROWARD COUNTY SO **Date** 12/13/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:37:20 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient Fail.CLEARED ROOM AIR W/FAN.

**Data Download Date/Time** 12/13/2023 11:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

**Agency** FORT LAUDERDALE PD **Date** 01/09/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 21:41:46 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC J CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 01/14/2023 14:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

**Agency** FORT LAUDERDALE PD **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 20:22:45 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

OFC J CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 03/05/2023 11:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

**Agency** FORT LAUDERDALE PD **Date** 03/27/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 21:38:55 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 04/10/2023 15:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

**Agency** FORT LAUDERDALE PD **Date** 04/25/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 20:25:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 04/25/2023 20:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

**Agency** FORT LAUDERDALE PD **Date** 05/23/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 22:48:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFFICER CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 05/27/2023 19:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

**Agency** FORT LAUDERDALE PD **Date** 06/27/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 22:54:33 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC J CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 06/27/2023 23:01

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

**Agency** FORT LAUDERDALE PD **Date** 07/30/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 20:39:24 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/30/2023 20:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

**Agency** FORT LAUDERDALE PD **Date** 08/29/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 22:13:40 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 20206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

OFC J CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 08/29/2023 22:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

**Agency** FORT LAUDERDALE PD **Date** 09/19/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 20:41:40 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC J CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 09/19/2023 20:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

**Agency** FORT LAUDERDALE PD **Date** 10/24/2023 **Type** Agency Inspection
**Inspector** STONER\,NATHAN,C **Time** 22:30:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

OFC J CARTER PRESNET AND OBSERVED

**Data Download Date/Time** 10/24/2023 22:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FORT LAUDERDALE PD | **Date** | 12/01/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 11:52:29 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

**Data Download Date/Time** 12/18/2023 13:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

**Agency** FORT LAUDERDALE PD **Date** 12/19/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 11:58:56 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 01923080A3
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 02/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 12/19/2023 12:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

**Agency** FORT LAUDERDALE PD **Date** 12/26/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 22:59:14 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 12/26/2023 23:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

**Agency** BROWARD COUNTY SO **Date** 01/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 14:28:46 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 01/04/2023 14:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

**Agency** BROWARD COUNTY SO **Date** 02/01/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 13:35:26 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING

**Data Download Date/Time** 02/01/2023 14:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

**Agency** BROWARD COUNTY SO **Date** 02/23/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:07:46 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 02/23/2023 14:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

**Agency** BROWARD COUNTY SO **Date** 05/03/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:37:18 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 05/03/2023 12:52

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

**Agency** BROWARD COUNTY SO **Date** 06/07/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:50:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.05 g/210L Test # (Repeat) 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # (Repeat) 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # (Repeat) 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: RFI Detect. 05: RFI JGUZMAN IN TRAINING

**Data Download Date/Time** 06/07/2023 14:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

**Agency** BROWARD COUNTY SO **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:21:20 **Software** 8100.27

**0.05 Lot #** 202-112A **0.08 Lot #** 202206B **0.20 Lot #** 202-206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 07/05/2023 10:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

**Agency** BROWARD COUNTY SO **Date** 08/02/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:40:50 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 08/02/2023 12:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

**Agency** BROWARD COUNTY SO **Date** 09/06/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:17:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 09/06/2023 11:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

**Agency** BROWARD COUNTY SO **Date** 10/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:24:13 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
**Expiration** 03/28/2025 **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | Interferent Detect Test # 1 | INT |
| Barometric Pressure Sensor Check: OK | | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 2 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 3 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 3 | 0.049 |
| Alcohol Free Test: OK | Yes | Interferent Detect Test # 3 | INT |
| 0.05 g/210L Test: OK | Yes | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

J GUZMAN IN TRAINING

**Data Download Date/Time** 10/04/2023 12:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

**Agency** BROWARD COUNTY SO **Date** 10/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:03:10 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING

**Data Download Date/Time** 10/04/2023 12:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

**Agency** BROWARD COUNTY SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:27:51 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING.

**Data Download Date/Time** 11/08/2023 11:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

**Agency** BROWARD COUNTY SO **Date** 12/13/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:25:59 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A:BLEW MOUTH ALCOHOL AT WRONG TIME..

**Data Download Date/Time** 12/13/2023 11:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

**Agency** FT. LAUDERDALE PD **Date** 01/10/2023 **Type** Department Inspection
**Inspector** REYES-RIVERA,DAVID,E **Time** 10:08:44 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 00521080A2 **0.20 Lot #** 202201E **0.08 Gas Lot #** 00521080A2
**Expiration** 01/11/2024 **Expiration** 02/05/2023 **Expiration** 01/18/2024 **Expiration** 02/05/2023

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.000 |
| 0.08 g/210L Test # (Repeat) 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.000 |
| 0.08 g/210L Test # (Repeat) 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.000 |
| 0.08 g/210L Test # (Repeat) 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.000 |
| 0.08 g/210L Test # (Repeat) 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

| | |
|---|---|
| 0.08 g/210L Test # 5 | 0.000 |
| 0.08 g/210L Test # (Repeat) 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.000 |
| 0.08 g/210L Test # (Repeat) 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.000 |
| 0.08 g/210L Test # (Repeat) 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.000 |
| 0.08 g/210L Test # (Repeat) 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.000 |
| 0.08 g/210L Test # (Repeat) 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.000 |
| 0.08 g/210L Test # (Repeat) 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

| | |
|---|---|
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) (Repeat) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

08: Control Outside Tolerance USED DRY GAS.

**Data Download Date/Time** 01/10/2023 10:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

**Agency** FORT LAUDERDALE PD **Date** 08/21/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 11:59:29 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 08/21/2023 12:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

**Agency** FORT LAUDERDALE PD **Date** 09/19/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 20:48:29 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFFICER CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 09/19/2023 21:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

**Agency** FORT LAUDERDALE PD **Date** 10/24/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 22:36:33 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: No Sample Provided.TIMED OUT

**Data Download Date/Time** 10/24/2023 22:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

**Agency** FORT LAUDERDALE PD **Date** 11/28/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 21:56:55 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.000 |
| 0.05 g/210L Test # (Repeat) 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.000 |
| 0.05 g/210L Test # (Repeat) 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.000 |
| 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance SIM LID LOOSE.

**Data Download Date/Time** 11/28/2023 22:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

**Agency** FORT LAUDERDALE PD **Date** 12/26/2023 **Type** Agency Inspection
**Inspector** STONER,NATHAN,C **Time** 23:19:12 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

OFC CARTER PRESENT AND OBSERVED

**Data Download Date/Time** 12/27/2023 20:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 01/11/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 09:14:36 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 01/11/2023 10:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 02/07/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 09:13:52 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 02/07/2023 09:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 03/09/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 08:28:25 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 03/09/2023 08:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 04/07/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 15:26:40 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 04/11/2023 09:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 05/02/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 14:16:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 05/02/2023 15:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 06/01/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 14:49:40 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 06/02/2023 09:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 07/13/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 11:34:50 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 07/13/2023 13:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 08/17/2023 **Type** Agency Inspection
**Inspector** LANDAS,CARTER,J **Time** 11:32:03 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 08/17/2023 12:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 09/04/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 14:56:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 09/04/2023 15:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 09/06/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:06:19 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.051 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.051 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 09/06/2023 14:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 09/12/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 09:25:08 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 09/12/2023 09:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 10/05/2023 **Type** Agency Inspection
**Inspector** LANDAS,CARTER,J **Time** 11:32:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 10/05/2023 11:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 11/11/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 07:49:35 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 11/11/2023 08:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

**Agency** COLLIER COUNTY SO **Date** 12/05/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 10:25:37 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 12/05/2023 12:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

**Agency** COLLIER COUNTY SO **Date** 01/10/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 14:28:12 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 01/10/2023 15:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

**Agency** COLLIER COUNTY SO **Date** 02/06/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 15:54:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 02/07/2023 09:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

**Agency** COLLIER COUNTY SO **Date** 03/07/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 11:25:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 03/07/2023 11:51

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

**Agency** COLLIER COUNTY SO **Date** 04/06/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 16:23:04 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 04/07/2023 09:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

**Agency** COLLIER COUNTY SO **Date** 05/04/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 08:48:45 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 05/04/2023 09:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

**Agency** COLLIER COUNTY SO **Date** 05/09/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:44:38 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 05/09/2023 15:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | COLLIER COUNTY SO | **Date** | 09/07/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 11:45:12 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

**Data Download Date/Time** 09/08/2023 16:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

**Agency** COLLIER COUNTY SO **Date** 09/08/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 15:59:19 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 09/08/2023 16:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

**Agency** COLLIER COUNTY SO **Date** 09/12/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 13:55:40 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 09/12/2023 14:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

**Agency** COLLIER COUNTY SO **Date** 10/06/2023 **Type** Agency Inspection
**Inspector** LANDAS,CARTER,J **Time** 13:49:19 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 10/06/2023 14:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

**Agency** COLLIER COUNTY SO **Date** 11/10/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 07:45:40 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 11/10/2023 08:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

**Agency** COLLIER COUNTY SO **Date** 12/04/2023 **Type** Agency Inspection
**Inspector** GONZALES,YVETTE,R **Time** 14:38:39 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED AT NJC

**Data Download Date/Time** 12/04/2023 15:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 01/29/2023 **Type** Agency Inspection
**Inspector** DEMPSTER,PAUL,R **Time** 21:22:10 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 01/30/2023 09:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** KERN,MICHAEL,P **Time** 20:20:51 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 02/28/2023 20:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 03/31/2023 **Type** Agency Inspection
**Inspector** KERN,MICHAEL,P **Time** 22:31:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 04/03/2023 15:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | CHARLOTTE COUNTY SO | **Date** | 05/03/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 10:27:03 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

**Data Download Date/Time** 05/04/2023 14:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 05/04/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:24:26 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 05/04/2023 14:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 05/20/2023 **Type** Agency Inspection
**Inspector** DEMPSTER,PAUL,R **Time** 20:51:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 05/22/2023 08:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 06/29/2023 **Type** Agency Inspection
**Inspector** KERN,MICHAEL,P **Time** 07:26:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
**Expiration** 12/07/2023 **Expiration** 06/24/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 06/29/2023 07:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 07/31/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:09:31 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 07/31/2023 14:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 08/22/2023 **Type** Agency Inspection
**Inspector** DEMPSTER,PAUL,R **Time** 23:13:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 08/22/2023 23:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 09/24/2023 **Type** Agency Inspection
**Inspector** CHANDLER,VANESSA,N **Time** 20:36:48 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 09/24/2023 20:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 10/25/2023 **Type** Agency Inspection
**Inspector** DEMPSTER,PAUL,R **Time** 20:20:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 10/25/2023 20:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 11/20/2023 **Type** Agency Inspection
**Inspector** CHANDLER,VANESSA,N **Time** 00:36:54 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 11/20/2023 00:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 12/15/2023 **Type** Agency Inspection
**Inspector** DEMPSTER,PAUL,R **Time** 22:42:43 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 12/15/2023 22:51

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

**Agency** CHARLOTTE COUNTY SO **Date** 12/15/2023 **Type** Agency Inspection
**Inspector** DEMPSTER,PAUL,R **Time** 23:22:49 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 12/15/2023 23:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

**Agency** MIAMI-DADE PD **Date** 01/05/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 02:17:40 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 01/05/2023 07:01

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

**Agency** MIAMI-DADE PD **Date** 02/09/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 03:51:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 02/09/2023 04:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

**Agency** MIAMI-DADE PD **Date** 03/09/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 03:12:34 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 03/09/2023 03:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

**Agency** MIAMI-DADE PD **Date** 04/06/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 03:55:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 04/06/2023 08:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

**Agency** MIAMI-DADE PD **Date** 05/12/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 04:41:40 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.067 |
| 0.08 g/210L Test # (Repeat) 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.062 |
| 0.08 g/210L Test # (Repeat) 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.068 |
| 0.08 g/210L Test # (Repeat) 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance/POSS SIM ISSUE.

**Data Download Date/Time** 05/12/2023 07:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

**Agency** MIAMI-DADE PD **Date** 05/30/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:54:54 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG115904
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 06/08/2023

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.082 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.082 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.082 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.082 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 05/30/2023 14:01

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI-DADE PD | **Date** | 06/09/2023 | **Type** | Agency Inspection |
| **Inspector** | CLOSIUS,RICHARD,L | **Time** | 04:27:08 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 33321080A2 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/05/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/09/2023 08:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

**Agency** MIAMI-DADE PD **Date** 07/14/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 01:05:06 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI G PACHECO PRESENT

**Data Download Date/Time** 07/14/2023 01:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

**Agency** MIAMI-DADE PD **Date** 08/17/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 01:55:35 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 08/17/2023 02:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

**Agency** MIAMI-DADE PD **Date** 09/01/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 05:24:59 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/01/2023 05:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

**Agency** MIAMI-DADE PD **Date** 10/05/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 01:11:59 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 10/05/2023 01:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

**Agency** MIAMI-DADE PD **Date** 11/03/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 02:01:48 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202106B **0.20 Lot #** 202106C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.074 |
| 0.08 g/210L Test # (Repeat) 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.08 g/210L Test # (Repeat) 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.08 g/210L Test # (Repeat) 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Cont Outside TOL/LOOSE CONN/AI R. MYRTIL PRESENT

**Data Download Date/Time** 11/03/2023 02:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

**Agency** MIAMI-DADE PD **Date** 12/01/2023 **Type** Agency Inspection
**Inspector** CLOSIUS,RICHARD,L **Time** 04:07:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI R MYRTIL PRESENT

**Data Download Date/Time** 12/01/2023 04:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

**Agency** BROWARD COUNTY S.O. **Date** 01/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 09:57:42 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 40277283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 01/04/2023 10:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

**Agency** BROWARD COUNTY S.O. **Date** 02/01/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 14:50:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: .J GUZMAN IN TRAINING

**Data Download Date/Time** 02/01/2023 15:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

**Agency** BROWARD COUNTY S.O. **Date** 03/11/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:41:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 03/13/2023 13:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

**Agency** BROWARD COUNTY S.O. **Date** 04/12/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:13:48 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.047 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.047 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.047 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.077 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.077 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 04/12/2023 13:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

**Agency** BROWARD COUNTY S.O. **Date** 05/03/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:26:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 05/03/2023 12:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

**Agency** BROWARD COUNTY S.O. **Date** 06/07/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:33:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 06/07/2023 12:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

**Agency** BROWARD COUNTY S.O. **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:54:07 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/07/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 07/05/2023 11:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

**Agency** BROWARD COUNTY S.O. **Date** 08/02/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:47:09 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 08/02/2023 11:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

**Agency** BROWARD COUNTY S.O. **Date** 09/06/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:25:39 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/06/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 09/06/2023 11:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

**Agency** BROWARD COUNTY S.O. **Date** 10/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:45:56 **Software** 8100.27

**0.05 Lot #** 202206B **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 06/14/2024 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 10/04/2023 11:55

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

**Agency** BROWARD COUNTY S.O. **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:48:49 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN PRESENT

**Data Download Date/Time** 11/08/2023 12:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

**Agency** BROWARD COUNTY S.O. **Date** 12/13/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:38:48 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/13/2023 12:52

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

**Agency** BROWARD COUNTY S.O. **Date** 01/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:03:24 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 01/04/2023 10:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

**Agency** BROWARD COUNTY S.O. **Date** 02/01/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:22:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 02/01/2023 12:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

**Agency** BROWARD COUNTY S.O. **Date** 03/11/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:01:25 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 03/11/2023 11:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

**Agency** BROWARD COUNTY S.O. **Date** 04/12/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:05:33 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 04/12/2023 13:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

**Agency** BROWARD COUNTY S.O. **Date** 05/03/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 13:13:50 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 05/03/2023 13:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

**Agency** BROWARD COUNTY S.O. **Date** 06/07/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 14:16:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 06/07/2023 14:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

**Agency** BROWARD COUNTY S.O. **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 09:33:00 **Software** 8100.27

**0.05 Lot #** 202-112A **0.08 Lot #** 202-206B **0.20 Lot #** 202-206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/05/2023 09:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

**Agency** BROWARD COUNTY S.O. **Date** 08/02/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:02:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING

**Data Download Date/Time** 08/02/2023 11:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

**Agency** BROWARD COUNTY S.O. **Date** 09/06/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:04:25 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Range Exceeded.TOO MUCH MOUTH ALCOHOL.

**Data Download Date/Time** 09/06/2023 12:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

**Agency** BROWARD COUNTY S.O. **Date** 10/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:29:52 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 10/04/2023 10:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

**Agency** BROWARD COUNTY S.O. **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:18:00 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 11/08/2023 12:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007108

**Agency** BROWARD COUNTY S.O. **Date** 12/13/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:07:39 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/13/2023 12:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

**Agency** HILLSBOROUGH CO SO **Date** 01/10/2023 **Type** Agency Inspection
**Inspector** DENDY,HALEY,M **Time** 10:49:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 01/10/2023 10:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

**Agency** HILLSBOROUGH CO SO **Date** 02/07/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 12:57:10 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI H.DENDY #260868 PRESENT FOR INSPECTION.

**Data Download Date/Time** 02/07/2023 13:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

**Agency** HILLSBOROUGH CO SO **Date** 03/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 08:51:46 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI H. DENDY #260868 PRESENT FOR INSPECTION.

**Data Download Date/Time** 03/02/2023 09:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

**Agency** HILLSBOROUGH CO SO **Date** 04/10/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:43:04 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 04/10/2023 14:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

**Agency** HILLSBOROUGH CO SO **Date** 05/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:24:27 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 05/02/2023 10:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

**Agency** HILLSBOROUGH CO SO **Date** 06/05/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:40:29 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 06/05/2023 12:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

**Agency** HILLSBOROUGH CO SO **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 12:04:42 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 07/05/2023 12:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

**Agency** HILLSBOROUGH CO SO **Date** 08/01/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:38:12 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 08/01/2023 11:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

**Agency** HILLSBOROUGH CO SO **Date** 09/05/2023 **Type** Agency Inspection
**Inspector** DAIL,JULIE,A **Time** 11:04:37 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 09/05/2023 11:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 01/06/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 11:22:37 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG113403
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 05/14/2023

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 01/06/2023 12:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 01/27/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 13:45:10 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1346146
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 07/07/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/31/2023 09:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 02/21/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 10:57:19 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1346146
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 07/07/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 02/23/2023 09:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 03/09/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 08:30:03 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1346146
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 07/07/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/09/2023 10:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 04/25/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 06:53:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1408275
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/25/2023 09:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 05/30/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 11:33:29 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1408275
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/30/2023 14:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 06/27/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 13:50:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1408275
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2023 08:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 07/30/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 16:24:25 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1408275
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/31/2023 13:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 08/30/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 12:24:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1408275
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/30/2023 14:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 09/27/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 15:01:29 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1408275
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2023 16:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 10/25/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 12:26:45 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 222152
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/27/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/25/2023 15:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 11/28/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 12:51:36 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 222152
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/27/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | No |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Range Exceeded.

**Data Download Date/Time** 11/28/2023 15:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 11/28/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 13:59:12 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 222152
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/27/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/28/2023 15:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

**Agency** PENSACOLA P.D. **Date** 12/21/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 12:12:47 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 222152
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/27/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/21/2023 14:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 02/03/2023 **Type** Agency Inspection
**Inspector** CARTWRIGHT JR,GARY,W **Time** 14:48:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.045 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/06/2023 09:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 03/31/2023 **Type** Agency Inspection
**Inspector** CARTWRIGHT JR,GARY,W **Time** 15:34:56 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/03/2023 11:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 04/30/2023 **Type** Agency Inspection
**Inspector** CARTWRIGHT JR,GARY,W **Time** 12:45:21 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/04/2023 13:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 05/31/2023 **Type** Agency Inspection
**Inspector** CARTWRIGHT JR,GARY,W **Time** 16:40:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.074 |
| 0.08 g/210L Test # (Repeat) 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.045 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.08 g/210L Test # (Repeat) 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.045 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.08 g/210L Test # (Repeat) 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 06/02/2023 10:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 06/30/2023 **Type** Agency Inspection
**Inspector** CARTWRIGHT JR,GARY,W **Time** 09:39:59 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/30/2023 09:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 07/28/2023 **Type** Agency Inspection
**Inspector** CARTWRIGHT JR,GARY,W **Time** 13:57:35 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/28/2023 14:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 08/31/2023 **Type** Agency Inspection
**Inspector** ALBRITTON,GARRETT,M **Time** 21:29:45 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/31/2023 21:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 09/19/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:23:39 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 09/19/2023 12:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 09/27/2023 **Type** Agency Inspection
**Inspector** ALBRITTON,GARRETT,M **Time** 23:38:35 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2023 23:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 10/22/2023 **Type** Agency Inspection
**Inspector** ALBRITTON,GARRETT,M **Time** 21:19:19 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/22/2023 21:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 11/05/2023 **Type** Agency Inspection
**Inspector** ALBRITTON,GARRETT,M **Time** 21:32:54 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #**
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | No |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.183 |
| 0.20 g/210L Test # (Repeat) 1 | 0.183 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.20 g/210L Test # 2 | 0.184 |
| 0.20 g/210L Test # (Repeat) 2 | 0.184 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.184 |
| 0.20 g/210L Test # (Repeat) 3 | 0.184 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: DID NOT MEET APPROPRIATE LEVELS .

**Data Download Date/Time** 12/05/2023 10:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 11/05/2023 **Type** Agency Inspection
**Inspector** ALBRITTON,GARRETT,M **Time** 22:12:27 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #**
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | No |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.183 |
| 0.20 g/210L Test # (Repeat) 1 | 0.188 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.186 |
| 0.20 g/210L Test # (Repeat) 2 | 0.186 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.186 |
| 0.20 g/210L Test # (Repeat) 3 | 0.187 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 12/05/2023 10:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

**Agency** HARDEE COUNTY SO **Date** 12/05/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:50:59 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 01923080A3
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 02/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 12/05/2023 15:01

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 01/10/2023 **Type** Agency Inspection
**Inspector** DENDY,HALEY,M **Time** 10:55:09 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 01/10/2023 11:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 02/07/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:45:40 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** **0.08 Gas Lot #**
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.107 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.074 |
| Alcohol Free Subject Test: 0.000 | Yes | Interferent Detect Test # 1 | INT |
| Mouth Alcohol Test: Slope Not Met | Yes | Alcohol Free Test # 2 | 0.000 |
| Interferent Detect Test: Interferent Detect | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Diagnostic Check (Post-Inspection): OK | No | 0.08 g/210L Test # 2 | 0.101 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.074 |
| 0.05 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| 0.20 g/210L Test: OK | No | 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Dry Gas Std Test: OK | No | 0.08 g/210L Test # 3 | 0.096 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.074 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance SIM ISSUE-SIM REPLACED . SIM ISSUE-SIM REPLACED, 0.080 WB READING LOWNon-compliance: 0.080 WB OUT OF TOLERANCE.

**Data Download Date/Time** 02/07/2023 10:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 02/07/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 10:30:30 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI H.DENDY #260868 PRESENT FOR INSPECTION.

**Data Download Date/Time** 02/07/2023 10:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 03/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 10:36:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI H. DENDY #260868 PRESENT FOR INSPECTION.

**Data Download Date/Time** 03/02/2023 10:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 04/10/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:34:30 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 04/10/2023 10:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 04/13/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 10:45:34 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.077 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.077 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 04/13/2023 10:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 06/05/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:23:56 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 06/05/2023 10:01

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 12:20:12 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance. DGS HOSE NOT CONNECTED TEST1. AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 07/05/2023 12:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 08/01/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:45:24 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 08/01/2023 11:53

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 09/05/2023 **Type** Agency Inspection
**Inspector** DAIL,JULIE,A **Time** 11:00:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 09/05/2023 11:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 10/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:07:04 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/02/2023 11:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 10:59:15 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 11/08/2023 11:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

**Agency** HILLSBOROUGH CO SO **Date** 12/01/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 14:24:01 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A4
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.191 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: No Sample Provided.NO AF IN TIME LIMIT.

**Data Download Date/Time** 12/01/2023 14:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

**Agency** FFWCC **Date** 01/30/2023 **Type** Agency Inspection
**Inspector** CARROLL,MICHAEL,J **Time** 11:44:56 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1380119
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 09/22/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/30/2023 11:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

**Agency** FFWCC **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** CARTAYA,GUILLERMO, **Time** 21:29:52 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
**Expiration** 12/07/2023 **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.000 |
| Minimum Sample Volume Check: OK | | Interferent Detect Test # 1 | INT |
| Barometric Pressure Sensor Check: OK | | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.05 g/210L Test # 2 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 2 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 3 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 3 | 0.000 |
| Alcohol Free Test: OK | Yes | Interferent Detect Test # 3 | INT |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

05: Control Outside Tolerance. OUTSIDE INT.

**Data Download Date/Time** 03/15/2023 19:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

**Agency** FFWCC **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** CARTAYA,GUILLERMO, **Time** 22:09:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1380119
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 09/22/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/15/2023 19:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

**Agency** FFWCC **Date** 03/31/2023 **Type** Agency Inspection
**Inspector** CARTAYA,GUILLERMO, **Time** 13:26:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1380119
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/22/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/16/2023 10:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

**Agency** FFWCC **Date** 05/16/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 12:25:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 2021O8D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1380119
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 09/22/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI OK

**Data Download Date/Time** 05/16/2023 12:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 06/20/2023 | **Type** | Agency Inspection |
| **Inspector** | MARROQUIN,RICHARD,J | **Time** | 13:12:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1380119 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 09/22/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/20/2023 13:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

**Agency** FFWCC **Date** 07/24/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 15:14:50 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1380119
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/22/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/24/2023 15:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

**Agency** FFWCC **Date** 08/08/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 10:30:03 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

**Remarks:**

**Data Download Date/Time** 08/08/2023 10:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

**Agency** FFWCC **Date** 08/28/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 13:55:30 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1380119
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/22/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/15/2023 14:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

**Agency** FFWCC **Date** 09/15/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 15:46:27 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG222203
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/15/2023 16:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 10/31/2023 | **Type** | Agency Inspection |
| **Inspector** | MARROQUIN,RICHARD,J | **Time** | 11:07:58 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | AG222203 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 08/10/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 10/31/2023 14:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

**Agency** FFWCC **Date** 11/01/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 13:30:26 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** AG222203 **0.20 Lot #** **0.08 Gas Lot #**
**Expiration** 03/28/2025 **Expiration** 08/10/2024 **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.049 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.000 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Test # (Repeat) 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 2 | 0.044 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.000 |
| 0.08 g/210L Test: OK | No | 0.08 g/210L Test # (Repeat) 2 | 0.000 |
| 0.20 g/210L Test: OK | No | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.044 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.000 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.000 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 11/01/2023 15:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 11/01/2023 | **Type** | Agency Inspection |
| **Inspector** | MARROQUIN,RICHARD,J | **Time** | 14:16:24 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | AG222203 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2022 | **Expiration** | 06/15/2024 | **Expiration** | 08/10/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

2ND INSPECTION - DRY GAS WAS NOT GOING THROUGH

**Data Download Date/Time** 11/01/2023 15:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

**Agency** FFWCC **Date** 12/29/2023 **Type** Agency Inspection
**Inspector** CARROLL,MICHAEL,J **Time** 15:05:13 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** AG222203
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 08/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/29/2023 15:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

**Agency** FFWCC **Date** 01/30/2023 **Type** Agency Inspection
**Inspector** CARROLL,MICHAEL,J **Time** 11:49:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402498227
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/30/2023 11:55

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

**Agency** FFWCC **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** CARTAYA,GUILLERMO, **Time** 21:24:14 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
**Expiration** 12/07/2023 **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.000 |
| Minimum Sample Volume Check: OK | | Interferent Detect Test # 1 | INT |
| Barometric Pressure Sensor Check: OK | | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.05 g/210L Test # 2 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 2 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 3 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 3 | 0.000 |
| Alcohol Free Test: OK | Yes | Interferent Detect Test # 3 | INT |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

05: Control Outside Tolerance. OUTSIDE INT.

**Data Download Date/Time** 03/15/2023 19:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 02/28/2023 | **Type** | Agency Inspection |
| **Inspector** | CARTAYA,GUILLERMO, | **Time** | 22:05:47 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402498227 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/15/2024 | **Expiration** | 07/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/15/2023 19:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

**Agency** FFWCC **Date** 03/31/2023 **Type** Agency Inspection
**Inspector** CARTAYA,GUILLERMO, **Time** 13:20:19 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402498227
**Expiration** 12/07/2021 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Range Exceeded.

**Data Download Date/Time** 05/16/2023 11:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 05/16/2023 | **Type** | Agency Inspection |
| **Inspector** | MARROQUIN,RICHARD,J | **Time** | 12:31:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202108D | **0.08 Gas Lot #** | 402498227 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 08/17/2023 | **Expiration** | 07/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | Yes | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 05/16/2023 12:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

**Agency** FFWCC **Date** 06/20/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 13:16:25 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402498227
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/20/2023 13:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 07/24/2023 | **Type** | Agency Inspection |
| **Inspector** | MARROQUIN,RICHARD,J | **Time** | 14:46:54 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | Interferent Detect Test # 1 | 0.000 |
| Minimum Sample Volume Check: OK | | Interferent Detect Test # (Repeat) 1 | 0.000 |
| Barometric Pressure Sensor Check: OK | | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Subject Test: 0.000 | Yes | Interferent Detect Test # 2 | INT |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # (Repeat) 2 | 0.000 |
| Interferent Detect Test: Interferent Detect | No | Alcohol Free Test # 3 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | Interferent Detect Test # 3 | INT |
| Alcohol Free Test: OK | Yes | Interferent Detect Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

Int Det: , . Non-compliance: NOT DETECTING INT.

**Data Download Date/Time** 08/08/2023 15:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 08/04/2023 | **Type** | Agency Inspection |
| **Inspector** | SIDDOWAY,BENJAMIN,W | **Time** | 09:28:03 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

**Data Download Date/Time** 08/08/2023 15:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

**Agency** FFWCC **Date** 08/08/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 15:20:59 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 08/08/2023 15:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

**Agency** FFWCC **Date** 08/28/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 13:50:24 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402498227
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/15/2023 14:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

**Agency** FFWCC **Date** 09/15/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 15:41:45 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402498227
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/15/2023 15:52

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

**Agency** FFWCC **Date** 10/31/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 11:42:43 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402498227
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 10/31/2023 14:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

**Agency** FFWCC **Date** 11/01/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 12:51:09 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402498227
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/01/2023 14:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 12/29/2023 | **Type** | Agency Inspection |
| **Inspector** | CARROLL,MICHAEL,J | **Time** | 15:09:42 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202307A | **0.20 Lot #** | 202307C | **0.08 Gas Lot #** | 402498227 |
| **Expiration** | 03/28/2025 | **Expiration** | 07/11/2025 | **Expiration** | 07/12/2025 | **Expiration** | 07/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/29/2023 15:23

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

**Agency** MONROE COUNTY SO **Date** 01/11/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 13:26:55 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024022906
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KW INTOX

**Data Download Date/Time** 01/11/2023 13:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

**Agency** MONROE COUNTY SO **Date** 02/16/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 12:02:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024022906
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KW INTOX

**Data Download Date/Time** 02/16/2023 12:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

**Agency** MONROE COUNTY SO **Date** 03/06/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 10:28:12 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024022906
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KW INTOX. DRYGAS LOT 302-402290677

**Data Download Date/Time** 03/06/2023 10:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

**Agency** MONROE COUNTY SO **Date** 04/04/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 12:59:36 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW INTOX

**Data Download Date/Time** 04/04/2023 13:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

**Agency** MONROE COUNTY SO **Date** 05/03/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 13:37:55 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

POST-EVIDENTIARY AI. BEING SENT TO FDLE FOR ANNUAL DEPT ISP

**Data Download Date/Time** 05/03/2023 13:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

**Agency** MONROE COUNTY SO **Date** 05/18/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:35:35 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 05/18/2023 14:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

**Agency** MONROE COUNTY SO **Date** 06/07/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 15:45:27 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PRE-EVIDENTIARY AI. CAME BACK FROM ANNUAL DEPT INSPECTION.

**Data Download Date/Time** 06/08/2023 12:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 07/06/2023 | **Type** | Agency Inspection |
| **Inspector** | KEITH,JASON,T | **Time** | 12:54:06 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402357111 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/15/2024 | **Expiration** | 02/10/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/06/2023 13:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

**Agency** MONROE COUNTY SO **Date** 08/02/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 10:21:14 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KW INTOX

**Data Download Date/Time** 08/02/2023 10:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

**Agency** MONROE COUNTY SO **Date** 09/14/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 11:44:15 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KW INTOX

**Data Download Date/Time** 09/14/2023 11:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

**Agency** MONROE COUNTY SO **Date** 10/04/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 12:04:42 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KW INTOX

**Data Download Date/Time** 10/04/2023 12:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 11/08/2023 | **Type** | Agency Inspection |
| **Inspector** | KEITH,JASON,T | **Time** | 10:16:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402558895 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/03/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KW INTOX

**Data Download Date/Time** 11/08/2023 11:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

**Agency** MONROE COUNTY SO **Date** 12/20/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 12:54:01 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KW INTOX

**Data Download Date/Time** 12/20/2023 13:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

**Agency** FISH & WILDLIFE CC **Date** 01/23/2023 **Type** Agency Inspection
**Inspector** CARTAYA,GUILLERMO, **Time** 11:10:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/15/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/23/2023 12:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FISH & WILDLIFE CC | **Date** | 03/06/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 10:15:43 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT.

**Data Download Date/Time** 03/06/2023 13:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

**Agency** FISH & WILDLIFE CC **Date** 03/06/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:06:19 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 03/06/2023 13:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

**Agency** FISH & WILDLIFE CC **Date** 03/31/2023 **Type** Agency Inspection
**Inspector** CARTAYA,GUILLERMO, **Time** 13:21:34 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/23/2023 15:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

**Agency** FISH & WILDLIFE CC **Date** 05/23/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 14:02:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/23/2023 15:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

**Agency** FISH & WILDLIFE CC **Date** 06/28/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 12:49:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2023 13:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

**Agency** FISH & WILDLIFE CC **Date** 07/18/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 12:14:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 07/18/2023 12:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

**Agency** FISH & WILDLIFE CC **Date** 08/29/2023 **Type** Agency Inspection
**Inspector** SAAVEDRA,OSCAR, **Time** 14:16:21 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A:DID NOT HAVE SOLUTION IN MOUTHNo Sample Provided.

**Data Download Date/Time** 08/29/2023 14:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

**Agency** FISH & WILDLIFE CC **Date** 09/25/2023 **Type** Agency Inspection
**Inspector** SAAVEDRA,OSCAR, **Time** 14:31:24 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/26/2023 13:51

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

**Agency** FISH & WILDLIFE CC **Date** 10/24/2023 **Type** Agency Inspection
**Inspector** SAAVEDRA,OSCAR, **Time** 13:35:40 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/24/2023 13:55

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

**Agency** FISH & WILDLIFE CC **Date** 11/27/2023 **Type** Agency Inspection
**Inspector** SAAVEDRA,OSCAR, **Time** 16:24:48 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/04/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/27/2023 16:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

**Agency** FISH & WILDLIFE CC **Date** 12/28/2023 **Type** Agency Inspection
**Inspector** CONROY,JESSICA,P **Time** 13:56:21 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024025903
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 11/04/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/28/2023 14:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

**Agency** FFWCC **Date** 07/10/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 14:14:03 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
**Expiration** 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 07/10/2023 14:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

**Agency** MONROE COUNTY SO **Date** 01/12/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 07:03:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KV INTOX

**Data Download Date/Time** 01/12/2023 07:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

**Agency** MONROE COUNTY SO **Date** 02/15/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 12:54:33 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV INTOX

**Data Download Date/Time** 02/16/2023 09:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 03/07/2023 | **Type** | Agency Inspection |
| **Inspector** | KEITH,JASON,T | **Time** | 09:14:31 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | | **0.08 Gas Lot #** | |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | | **Expiration** | |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.074 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.074 |
| Alcohol Free Subject Test: 0.000 | Yes | Interferent Detect Test # 1 | INT |
| Mouth Alcohol Test: Slope Not Met | Yes | Alcohol Free Test # 2 | 0.000 |
| Interferent Detect Test: Interferent Detect | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Diagnostic Check (Post-Inspection): OK | No | 0.08 g/210L Test # 2 | 0.075 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.074 |
| 0.05 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| 0.20 g/210L Test: OK | No | 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Dry Gas Std Test: OK | No | 0.08 g/210L Test # 3 | 0.075 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.075 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:CONTROL OUTSIDE OF TOL. .

**Data Download Date/Time** 03/07/2023 11:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

**Agency** MONROE COUNTY SO **Date** 03/07/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 11:10:53 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KV INTOX

**Data Download Date/Time** 03/07/2023 11:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

**Agency** MONROE COUNTY SO **Date** 04/05/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 10:21:19 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KV INTOX

**Data Download Date/Time** 04/05/2023 11:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

**Agency** MONROE COUNTY SO **Date** 05/04/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 11:21:47 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV INTOX

**Data Download Date/Time** 05/04/2023 11:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

**Agency** MONROE COUNTY SO **Date** 06/08/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 09:00:58 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KV INTOX

**Data Download Date/Time** 06/08/2023 09:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

**Agency** MONROE COUNTY SO **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 15:08:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

POST-EVIDENTIARY AI. BEING SENT FOR ANNUAL DEPT INSPECTION.

**Data Download Date/Time** 07/05/2023 15:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

**Agency** MONROE COUNTY SO **Date** 07/13/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:09:44 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.047 |
| 0.08 g/210L Test # 4 | 0.076 |
| 0.20 g/210L Test # 4 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.076 |
| 0.20 g/210L Test # 6 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.047 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.077 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.077 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 07/13/2023 14:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

**Agency** MONROE COUNTY SO **Date** 08/02/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 14:25:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PRE-EVIDENTIARY AI. CAME BACK FROM ANNUAL DEPT INSPECTION.

**Data Download Date/Time** 08/02/2023 14:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

**Agency** MONROE COUNTY SO **Date** 09/13/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 11:12:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KV INTOX

**Data Download Date/Time** 09/13/2023 11:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

**Agency** MONROE COUNTY SO **Date** 10/05/2023 **Type** Agency Inspection
**Inspector** KEITH,JASON,T **Time** 11:11:50 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

KV INTOX

**Data Download Date/Time** 10/05/2023 11:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 11/09/2023 | **Type** | Agency Inspection |
| **Inspector** | KEITH,JASON,T | **Time** | 11:12:45 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1433249 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/08/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV INTOX

**Data Download Date/Time** 11/09/2023 11:23

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 12/21/2023 | **Type** | Agency Inspection |
| **Inspector** | KEITH,JASON,T | **Time** | 11:33:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1433249 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/08/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

KV INTOX

**Data Download Date/Time** 12/21/2023 11:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 05/01/2023 | **Type** | Agency Inspection |
| **Inspector** | RATH,ROBERT,S | **Time** | 03:08:30 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASS

**Data Download Date/Time** 05/06/2022 15:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

**Agency** MARION COUNTY SO **Date** 01/23/2023 **Type** Agency Inspection
**Inspector** RATH,ROBERT,S **Time** 15:21:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402498227
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 01/23/2023 15:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

**Agency** MARION COUNTY SO **Date** 02/20/2023 **Type** Agency Inspection
**Inspector** RATH,ROBERT,S **Time** 15:35:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402498227
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 02/20/2023 15:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

**Agency** MARION COUNTY SO **Date** 03/29/2023 **Type** Agency Inspection
**Inspector** RATH,ROBERT,S **Time** 10:28:10 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402370890
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 05/09/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 03/29/2023 10:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 04/26/2023 | **Type** | Agency Inspection |
| **Inspector** | RATH,ROBERT,S | **Time** | 10:38:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402370890 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 05/09/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 04/26/2023 10:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

**Agency** MARION COUNTY SO **Date** 05/24/2023 **Type** Agency Inspection
**Inspector** RATH,ROBERT,S **Time** 11:08:18 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402370890
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/09/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 05/24/2023 11:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

**Agency** MARION COUNTY SO **Date** 06/27/2023 **Type** Agency Inspection
**Inspector** RATH,ROBERT,S **Time** 10:40:52 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402690226
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/09/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 06/27/2023 10:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 07/27/2023 | **Type** | Agency Inspection |
| **Inspector** | RATH,ROBERT,S | **Time** | 02:23:39 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402690226 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 03/09/2026 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

00:AMBIENT.DUI SUSP. WALKED BY DURING TEST.

**Data Download Date/Time** 07/27/2023 02:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

**Agency** MARION COUNTY SO **Date** 08/23/2023 **Type** Agency Inspection
**Inspector** RATH,ROBERT,S **Time** 09:47:37 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402806340
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/26/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 08/23/2023 09:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

**Agency** MARION COUNTY SO **Date** 09/20/2023 **Type** Agency Inspection
**Inspector** RATH,ROBERT,S **Time** 15:13:17 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402806340
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/26/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 09/20/2023 15:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 10/25/2023 | **Type** | Agency Inspection |
| **Inspector** | YOUNG,ROBERT,A | **Time** | 14:27:25 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A: .

**Data Download Date/Time** 10/25/2023 15:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

**Agency** MARION COUNTY SO **Date** 10/25/2023 **Type** Agency Inspection
**Inspector** YOUNG,ROBERT,A **Time** 15:03:13 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402806340
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/26/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/25/2023 15:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

**Agency** MARION COUNTY SO **Date** 11/03/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 15:03:09 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.051 |
| 0.08 g/210L Test # 4 | 0.081 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.051 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 11/03/2023 15:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

**Agency** MARION COUNTY SO **Date** 11/27/2023 **Type** Agency Inspection
**Inspector** YOUNG,ROBERT,A **Time** 15:38:09 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402806340
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/26/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 11/27/2023 15:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 12/20/2023 | **Type** | Agency Inspection |
| **Inspector** | RATH,ROBERT,S | **Time** | 13:39:39 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402872586 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/12/2026 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 12/20/2023 13:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

**Agency** MARION COUNTY SO **Date** 01/23/2023 **Type** Agency Inspection
**Inspector** RATH,ROBERT,S **Time** 15:32:21 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402498227
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 01/23/2023 15:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

**Agency** MARION COUNTY SO **Date** 02/20/2023 **Type** Agency Inspection
**Inspector** RATH,ROBERT,S **Time** 15:31:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402498227
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 02/20/2023 15:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

**Agency** MARION COUNTY SO **Date** 02/28/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 15:13:05 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0013 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0007 |

**Remarks:**

**Data Download Date/Time** 02/28/2023 15:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

**Agency** MARION COUNTY SO **Date** 09/06/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 11:48:26 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 06723080A5
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.070 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

| | |
|---|---|
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 5 | 0.079 |
| Interferent Detect Test # 5 | INT |
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10 | 0.079 |
| Interferent Detect Test # 10 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

| | |
|---|---|
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0028 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat) | 0.0003 |

**Remarks:**

08: Control Outside Tolerance. DGS HOSE NOT ATTACHED.

**Data Download Date/Time** 09/06/2023 11:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

**Agency** MARION COUNTY SO **Date** 09/20/2023 **Type** Agency Inspection
**Inspector** RATH,ROBERT,S **Time** 15:28:02 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402806340
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/26/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 09/20/2023 15:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

**Agency** MARION COUNTY SO **Date** 10/25/2023 **Type** Agency Inspection
**Inspector** YOUNG,ROBERT,A **Time** 15:12:45 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402806340
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/26/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 10/25/2023 15:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

**Agency** MARION COUNTY SO **Date** 11/27/2023 **Type** Agency Inspection
**Inspector** YOUNG,ROBERT,A **Time** 15:46:44 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402806340
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/26/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 11/27/2023 15:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARION COUNTY SO | **Date** | 12/20/2023 | **Type** | Agency Inspection |
| **Inspector** | RATH,ROBERT,S | **Time** | 13:46:43 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402806340 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 07/26/2026 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

WITHIN COMPLIANCE.

**Data Download Date/Time** 12/20/2023 13:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

**Agency** WALTON COUNTY SO **Date** 01/31/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 12:11:56 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402367487
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/28/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/31/2023 13:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

**Agency** WALTON COUNTY SO **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 10:43:36 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402367487
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/28/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/28/2023 11:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

**Agency** WALTON COUNTY SO **Date** 03/31/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 10:50:21 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402367487
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/28/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/04/2023 08:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | WALTON COUNTY SO | **Date** | 05/05/2023 | **Type** | Agency Inspection |
| **Inspector** | SOTO,ISRAEL, | **Time** | 12:50:21 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED INSPECTION TO OPERATE INSTRUMENT

**Data Download Date/Time** 05/08/2023 10:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

**Agency** WALTON COUNTY SO **Date** 05/08/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 09:13:05 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 05/08/2023 10:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

**Agency** WALTON COUNTY SO **Date** 05/31/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 12:05:46 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402367487
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/28/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/31/2023 13:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

**Agency** WALTON COUNTY SO **Date** 06/30/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 09:06:04 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402367487
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/14/2024 **Expiration** 04/28/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/03/2023 14:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

**Agency** WALTON COUNTY SO **Date** 07/27/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 14:29:04 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** WO32302415
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 09/11/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.052 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/27/2023 15:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

**Agency** WALTON COUNTY SO **Date** 08/31/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 13:00:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** WO32302415
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/11/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/31/2023 14:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

**Agency** WALTON COUNTY SO **Date** 09/29/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 09:57:26 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** WO32302415
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/11/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/29/2023 11:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

**Agency** WALTON COUNTY SO **Date** 10/30/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 19:57:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402370891
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/23/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/30/2023 21:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

**Agency** WALTON COUNTY SO **Date** 11/29/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 13:07:30 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402370891
**Expiration** 12/07/2023 **Expiration** 06/14/2023 **Expiration** 06/15/2024 **Expiration** 02/23/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/29/2023 14:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

**Agency** WALTON COUNTY SO **Date** 12/30/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 21:54:34 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402370891
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/23/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/30/2023 23:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007174

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MICCOSUKEE PD | **Date** | 11/03/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 09:48:50 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

**Data Download Date/Time** 11/03/2023 12:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007174

**Agency** MICCOSUKEE PD **Date** 11/03/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:29:52 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG223802
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007174

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 11/03/2023 12:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

**Agency** GROVELAND PD **Date** 01/24/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:53:08 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 00521080A2
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 02/05/2023

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.077 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.077 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 01/24/2023 15:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

**Agency** GROVELAND PD **Date** 08/14/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:55:23 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.077 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.077 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 08/14/2023 13:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | GROVELAND PD | **Date** | 08/28/2023 | **Type** | Agency Inspection |
| **Inspector** | PENA,RUSSEL,S | **Time** | 13:45:55 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 28021080A2 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 12/05/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # (Repeat) 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # (Repeat) 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

Int Det: RFI Detect.

**Data Download Date/Time** 08/28/2023 13:51

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

**Agency** GROVELAND PD **Date** 09/25/2023 **Type** Agency Inspection
**Inspector** PENA,RUSSEL,A **Time** 15:19:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 28021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/25/2023 15:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

**Agency** GROVELAND PD **Date** 10/18/2023 **Type** Agency Inspection
**Inspector** PENA,RUSSEL,S **Time** 15:48:49 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 28021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Ambient Fail. 00: RFI Detect.

**Data Download Date/Time** 10/18/2023 15:55

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | GROVELAND PD | **Date** | 11/13/2023 | **Type** | Agency Inspection |
| **Inspector** | PENA,RUSSEL,S | **Time** | 15:18:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 28021080A2 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 12/05/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/13/2023 15:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | GROVELAND PD | **Date** | 12/04/2023 | **Type** | Agency Inspection |
| **Inspector** | PENA,RUSSEL,S | **Time** | 14:32:41 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 07622080A2 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 05/05/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/04/2023 14:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

**Agency** PENSACOLA PD **Date** 01/27/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 13:09:53 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1346146
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 07/07/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/06/2023 08:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

**Agency** PENSACOLA PD **Date** 02/09/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 12:20:10 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 02/09/2023 13:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PENSACOLA PD | **Date** | 02/21/2023 | **Type** | Agency Inspection |
| **Inspector** | MARTINEZ,HENDRICK,J | **Time** | 11:37:07 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1346146 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/15/2024 | **Expiration** | 07/07/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/23/2023 09:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

**Agency** PENSACOLA PD **Date** 03/09/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 07:40:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1346146
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 07/07/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 03/09/2023 11:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PENSACOLA PD | **Date** | 04/25/2023 | **Type** | Agency Inspection |
| **Inspector** | MARTINEZ,HENDRICK,J | **Time** | 07:40:15 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1408275 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/15/2024 | **Expiration** | 11/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/25/2023 09:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

**Agency** PENSACOLA PD **Date** 05/30/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 10:46:12 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1408275
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/30/2023 14:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

**Agency** PENSACOLA PD **Date** 06/27/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 14:31:58 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1408275
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # (Repeat) 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # (Repeat) 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # (Repeat) 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 06/28/2023 08:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

**Agency** PENSACOLA PD **Date** 07/30/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 15:35:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1408275
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/31/2023 13:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PENSACOLA PD | **Date** | 08/30/2023 | **Type** | Agency Inspection |
| **Inspector** | MARTINEZ,HENDRICK,J | **Time** | 13:08:36 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1408275 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 11/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/30/2023 14:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

**Agency** PENSACOLA PD **Date** 09/27/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 14:17:25 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1408275
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2023 16:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

**Agency** PENSACOLA PD **Date** 10/25/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 13:30:06 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 222152
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/27/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # (Repeat) 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # (Repeat) 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # (Repeat) 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 10/25/2023 15:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

**Agency** PENSACOLA PD **Date** 11/28/2023 **Type** Agency Inspection
**Inspector** MARTINEZ,HENDRICK,J **Time** 11:36:17 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 222152
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/27/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.20 g/210L Test # (Repeat) 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # (Repeat) 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # (Repeat) 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient Fail. 20: RFI Detect.

**Data Download Date/Time** 11/28/2023 15:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PENSACOLA PD | **Date** | 12/21/2023 | **Type** | Agency Inspection |
| **Inspector** | MARTINEZ,HENDRICK,J | **Time** | 13:03:55 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 222152 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/27/2026 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/21/2023 14:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

**Agency** ST JOHNS SO **Date** 01/25/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEXANDER, **Time** 22:35:19 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 220703
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/27/2023 16:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

**Agency** ST JOHNS SO **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEXANDER, **Time** 15:51:10 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #**
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration**

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | No |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # (Repeat) 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # (Repeat) 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # (Repeat) 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance, Control Outside Tolerance.

**Data Download Date/Time** 03/03/2023 13:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

**Agency** ST JOHNS SO **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEXANDER, **Time** 16:28:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 220703
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/03/2023 13:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

**Agency** ST JOHNS SO **Date** 03/29/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEXANDER, **Time** 14:07:02 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 220703
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.054 |
| 0.08 g/210L Test # 1 | 0.082 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.054 |
| 0.08 g/210L Test # 2 | 0.083 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.053 |
| 0.08 g/210L Test # 3 | 0.083 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/31/2023 13:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 04/27/2023 | **Type** | Agency Inspection |
| **Inspector** | ELLIS,ALEX, | **Time** | 22:18:08 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 220703 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/03/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/28/2023 14:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 05/31/2023 | **Type** | Agency Inspection |
| **Inspector** | GREEN,JASON,A | **Time** | 19:58:14 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | Interferent Detect Test # 1 | INT |
| Minimum Sample Volume Check: OK | | Alcohol Free Test # 2 | 0.000 |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # 2 | INT |
| Alcohol Free Subject Test: 0.000 | Yes | Alcohol Free Test # 3 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 3 | INT |
| Interferent Detect Test: Interferent Detect | Yes | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | Yes | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

**Data Download Date/Time** 06/01/2023 16:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 05/31/2023 | **Type** | Agency Inspection |
| **Inspector** | GREEN,JASON,A | **Time** | 20:32:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 220703 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/03/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/01/2023 16:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 06/28/2023 | **Type** | Agency Inspection |
| **Inspector** | GREEN,JASON,A | **Time** | 20:37:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 220703 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/03/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | No |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2023 23:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

**Agency** ST JOHNS SO **Date** 06/28/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 21:13:45 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 220703
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2023 23:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 07/26/2023 | **Type** | Agency Inspection |
| **Inspector** | GREEN,JASON,A | **Time** | 22:14:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 220703 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/15/2024 | **Expiration** | 10/03/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/26/2023 22:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 08/23/2023 | **Type** | Agency Inspection |
| **Inspector** | GREEN,JASON,A | **Time** | 21:48:53 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 220703 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/03/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.191 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/23/2023 22:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

**Agency** ST JOHNS SO **Date** 09/18/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 11:40:58 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 220703
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Interferent Detect Test # (Repeat) 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # (Repeat) 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

Time-Date changed. Int Det: RFI Detect.

**Data Download Date/Time** 09/18/2023 13:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 10/30/2023 | **Type** | Agency Inspection |
| **Inspector** | GREEN,JASON,A | **Time** | 11:47:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 220703 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/03/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.075 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.075 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/01/2023 14:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

**Agency** ST JOHNS SO **Date** 11/01/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 14:34:41 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.052 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.052 |
| 0.08 g/210L Test # 4 | 0.081 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.052 |
| 0.08 g/210L Test # 5 | 0.082 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.052 |
| 0.08 g/210L Test # 7 | 0.082 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.052 |
| 0.08 g/210L Test # 8 | 0.082 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.052 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 11/01/2023 14:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 11/02/2023 | **Type** | Agency Inspection |
| **Inspector** | GREEN,JASON,A | **Time** | 20:34:21 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 221783 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/02/2023 23:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

**Agency** ST JOHNS SO **Date** 12/05/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 23:52:55 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 221783
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.191 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/06/2023 00:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

**Agency** WALTON COUNTY SO **Date** 01/31/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 12:05:48 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402370891
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/23/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/31/2023 13:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

**Agency** WALTON COUNTY SO **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 10:30:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402370891
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/23/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/28/2023 11:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

**Agency** WALTON COUNTY SO **Date** 03/31/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 10:56:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402370891
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/23/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/04/2023 08:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

**Agency** WALTON COUNTY SO **Date** 11/09/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 11:14:59 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 11/09/2023 11:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | WALTON COUNTY SO | **Date** | 11/29/2023 | **Type** | Agency Inspection |
| **Inspector** | MAY,WILLIAM,P | **Time** | 12:32:03 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A: TEST ACCEPTABLE - M/A READ RANGE DID NOT LVL OUT.

**Data Download Date/Time** 11/29/2023 14:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

**Agency** WALTON COUNTY SO **Date** 11/29/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 13:18:18 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402370891
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/23/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | No | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AF TEST OUTSIDE 0.000 DUE TO MA REPEAT

**Data Download Date/Time** 11/29/2023 14:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

**Agency** WALTON COUNTY SO **Date** 12/30/2023 **Type** Agency Inspection
**Inspector** MAY,WILLIAM,P **Time** 22:00:26 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402370891
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/23/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/30/2023 23:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

**Agency** ST JOHNS SO **Date** 01/25/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEXANDER, **Time** 20:36:36 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 211543
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/27/2023 16:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

**Agency** ST JOHNS SO **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEXANDER, **Time** 17:15:25 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 211543
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.073 |
| 0.08 g/210L Test # (Repeat) 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.190 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.08 g/210L Test # (Repeat) 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.08 g/210L Test # (Repeat) 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 03/03/2023 13:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

**Agency** ST JOHNS SO **Date** 03/29/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEXANDER, **Time** 12:52:59 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 211543
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/31/2023 12:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 04/27/2023 | **Type** | Agency Inspection |
| **Inspector** | ELLIS,ALEXANDER, | **Time** | 20:28:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 211543 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/05/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/28/2023 14:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

**Agency** ST JOHNS SO **Date** 05/31/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 18:36:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 221264
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/01/2023 16:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 06/28/2023 | **Type** | Agency Inspection |
| **Inspector** | GREEN,JASON,A | **Time** | 22:20:13 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 221264 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/03/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2023 23:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 07/26/2023 | **Type** | Agency Inspection |
| **Inspector** | GREEN,JASON,A | **Time** | 20:32:04 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** 202112A | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** 12/07/2023 | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | Yes | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | Interferent Detect Test # 1 | 0.000 |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # (Repeat) 1 | INT |
| Alcohol Free Subject Test: 0.000 | Yes | Alcohol Free Test # 2 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | Interferent Detect Test # (Repeat) 2 | INT |
| Alcohol Free Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test: OK | No | Interferent Detect Test # 3 | INT |
| 0.20 g/210L Test: OK | No | Interferent Detect Test # (Repeat) 3 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

Time-Date changed. Int Det: .

**Data Download Date/Time** 07/26/2023 21:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 07/26/2023 | **Type** | Agency Inspection |
| **Inspector** | GREEN,JASON,A | **Time** | 21:11:36 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 221264 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/15/2024 | **Expiration** | 10/03/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/26/2023 21:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

**Agency** ST JOHNS SO **Date** 08/23/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 22:47:15 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 221264
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/23/2023 23:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

**Agency** ST JOHNS SO **Date** 09/18/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 12:47:11 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 221264
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/18/2023 13:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 10/30/2023 | **Type** | Agency Inspection |
| **Inspector** | GREEN,JASON,A | **Time** | 11:55:35 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 221264 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/03/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.075 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 11/01/2023 15:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

**Agency** ST JOHNS SO **Date** 11/01/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 15:01:37 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.051 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.051 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 11/01/2023 15:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

**Agency** ST JOHNS SO **Date** 11/02/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 20:56:00 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 221264
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/02/2023 23:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ST JOHNS SO | **Date** | 12/05/2023 | **Type** | Agency Inspection |
| **Inspector** | GREEN,JASON,A | **Time** | 23:58:47 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 222723 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 08/09/2026 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/06/2023 00:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

**Agency** BROWARD COUNTY SO **Date** 01/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:45:26 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING.

**Data Download Date/Time** 01/04/2023 10:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

**Agency** BROWARD COUNTY SO **Date** 02/01/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:01:24 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING

**Data Download Date/Time** 02/01/2023 14:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

**Agency** BROWARD COUNTY SO **Date** 03/11/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 14:41:14 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Interferent Detect. JGUZMAN IN TRAINING

**Data Download Date/Time** 03/13/2023 14:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

**Agency** BROWARD COUNTY SO **Date** 04/12/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:50:15 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 04/12/2023 11:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

**Agency** BROWARD COUNTY SO **Date** 05/03/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 13:52:34 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 05/03/2023 14:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

**Agency** BROWARD COUNTY SO **Date** 11/08/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:04:19 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG223802
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 11/08/2023 13:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

**Agency** BROWARD COUNTY SO **Date** 12/13/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:57:14 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 12/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/13/2023 12:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 01/31/2023 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 11:40:55 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient Fail.

**Data Download Date/Time** 01/31/2023 12:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 02/25/2023 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 10:02:10 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 02/27/2023 10:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 03/08/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:17:39 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 03/08/2023 14:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 03/22/2023 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 14:38:51 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 03/22/2023 14:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 04/28/2023 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 10:52:21 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 04/28/2023 11:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 05/30/2023 **Type** Agency Inspection
**Inspector** ROCA,LUIS,C **Time** 14:22:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NO ISSUES

**Data Download Date/Time** 05/30/2023 14:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 06/30/2023 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 10:54:01 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient Fail UNKNOWN WHY THERE WAS ALCOHOL IN THE AIR.

**Data Download Date/Time** 06/30/2023 11:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 07/26/2023 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 10:15:01 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 07/26/2023 10:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 08/31/2023 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 10:36:10 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | 0.000 |
| Interferent Detect Test # (Repeat) 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | 0.000 |
| Interferent Detect Test # (Repeat) 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | 0.000 |
| Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

Int Det: .THE SIMULATOR WAS NOT SEALED

**Data Download Date/Time** 08/31/2023 10:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 09/14/2023 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 20:47:05 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 09/14/2023 20:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 10/31/2023 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 09:52:22 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 10/31/2023 10:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 11/30/2023 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 13:41:11 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient FailMOUTH ALCOHOL IN THE AIR.

**Data Download Date/Time** 11/30/2023 13:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

**Agency** HOMESTEAD PD **Date** 12/30/2023 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 10:30:19 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 12/30/2023 10:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 01/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:28:10 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING.

**Data Download Date/Time** 01/04/2023 11:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 02/01/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 14:28:38 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 402477283 **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/24/2025 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.000 |
| 0.20 g/210L Test # (Repeat) 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.000 |
| 0.20 g/210L Test # (Repeat) 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.000 |
| 0.20 g/210L Test # (Repeat) 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: Interferent Detect. 20: Control Outside Tolerance.

**Data Download Date/Time** 02/01/2023 14:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 02/16/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:02:16 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.077 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.077 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 02/16/2023 14:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 03/11/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 14:03:47 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 03/13/2023 14:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 04/12/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:07:14 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/12/2023 10:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 05/03/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:23:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 05/03/2023 10:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 06/07/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:49:46 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 06/07/2023 11:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:20:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 07/05/2023 12:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 08/02/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:20:03 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 08/02/2023 10:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 09/06/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:17:19 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 09/06/2023 12:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 10/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:20:25 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 10/04/2023 10:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:36:08 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

UNKNOWN

**Data Download Date/Time** 11/08/2023 12:53

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

**Agency** BROWARD COUNTY SO **Date** 12/13/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:37:14 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/13/2023 11:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | SUWANNEE COUNTY SO | **Date** | 01/11/2023 | **Type** | Agency Inspection |
| **Inspector** | SOTO,ISRAEL, | **Time** | 13:34:34 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 01/12/2023 12:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

**Agency** SUWANNEE COUNTY SO **Date** 01/12/2023 **Type** Department Inspection
**Inspector** NICODEMO,PHIL, **Time** 12:01:04 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 01/12/2023 12:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

**Agency** SUWANNEE COUNTY SO **Date** 01/19/2023 **Type** Department Inspection
**Inspector** NICODEMO,PHIL, **Time** 15:22:24 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0000 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 01/19/2023 15:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

**Agency** SUWANNEE COUNTY SO **Date** 01/26/2023 **Type** Agency Inspection
**Inspector** DUFFEK,GENEVA,K **Time** 12:14:52 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402469736
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/16/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 01/26/2023 12:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

**Agency** SUWANNEE COUNTY SO **Date** 02/21/2023 **Type** Agency Inspection
**Inspector** DUFFEK,GENEVA,K **Time** 14:15:51 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #**
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| Interferent Detect Test # 1 | 0.000 |
| Interferent Detect Test # (Repeat) 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| Interferent Detect Test # 2 | 0.000 |
| Interferent Detect Test # (Repeat) 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| Interferent Detect Test # 3 | 0.000 |
| Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

Int Det:SIM SEAL NOT SECURE. Non-compliance: .

**Data Download Date/Time** 02/22/2023 14:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

**Agency** SUWANNEE COUNTY SO **Date** 02/21/2023 **Type** Agency Inspection
**Inspector** DUFFEK,GENEVA,K **Time** 15:06:36 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402469736
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/16/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance INLET TUBE NOT SECURE.

**Data Download Date/Time** 02/22/2023 14:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

**Agency** BROWARD COUNTY SO **Date** 01/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 13:48:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 01/04/2023 14:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

**Agency** BROWARD COUNTY SO **Date** 02/01/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:57:49 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 02/01/2023 13:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BROWARD COUNTY SO | **Date** | 04/11/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 11:37:56 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT DONCUTED. BYPASSED TO OPERATE INSTRUMENT.

**Data Download Date/Time** 04/12/2023 13:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

**Agency** BROWARD COUNTY SO **Date** 04/12/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:09:29 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 04/12/2023 13:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

**Agency** BROWARD COUNTY SO **Date** 05/19/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 21:49:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

SGT. Y WASHINGTON PRESENT.

**Data Download Date/Time** 05/22/2023 08:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

**Agency** BROWARD COUNTY SO **Date** 06/07/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 14:48:51 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 06/07/2023 15:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

**Agency** BROWARD COUNTY SO **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:16:50 **Software** 8100.27

**0.05 Lot #** 202-112A **0.08 Lot #** 202-206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 07/05/2023 12:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

**Agency** BROWARD COUNTY SO **Date** 08/02/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:52:48 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.052 |
| 0.08 g/210L Test # 1 | 0.082 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.082 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.052 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 08/02/2023 11:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

**Agency** BROWARD COUNTY SO **Date** 09/06/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:56:53 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | 0.000 |
| Interferent Detect Test # (Repeat) 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | 0.000 |
| Interferent Detect Test # (Repeat) 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | 0.000 |
| Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

Int Det:SIMULATOR NOT ATTACHED SECURELY. J GUZMAN IN TRAINI

**Data Download Date/Time** 09/06/2023 13:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

**Agency** BROWARD COUNTY SO **Date** 10/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:24:02 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 10/04/2023 11:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

**Agency** BROWARD COUNTY SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:33:31 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING.

**Data Download Date/Time** 11/08/2023 11:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007433

**Agency** BROWARD COUNTY SO **Date** 12/13/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:32:26 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/13/2023 11:53

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

**Agency** BROWARD COUNTY SO **Date** 01/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 13:00:38 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 01/04/2023 13:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

**Agency** BROWARD COUNTY SO **Date** 02/01/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 14:12:07 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477283
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/24/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient Fail.JGUZMAN IN TRAINING

**Data Download Date/Time** 02/01/2023 14:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BROWARD COUNTY SO | **Date** | 03/11/2023 | **Type** | Agency Inspection |
| **Inspector** | FRAZIER,ANAYA,S | **Time** | 13:28:16 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402477283 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/24/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 03/13/2023 13:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

**Agency** BROWARD COUNTY SO **Date** 04/12/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:44:08 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 04/12/2023 14:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

**Agency** BROWARD COUNTY SO **Date** 05/03/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:15:27 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.080 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.204 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

08: RFI Detect UNKNOWN SOURCE..

**Data Download Date/Time** 05/03/2023 11:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

**Agency** BROWARD COUNTY SO **Date** 06/07/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 13:20:53 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J.GUZMAN IN TRAINING.

**Data Download Date/Time** 06/07/2023 14:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

**Agency** BROWARD COUNTY SO **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:40:41 **Software** 8100.27

**0.05 Lot #** 202-112A **0.08 Lot #** 202-206B **0.20 Lot #** 202-206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A

**Data Download Date/Time** 07/05/2023 11:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

**Agency** BROWARD COUNTY SO **Date** 08/02/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:56:03 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

J. GUZMAN IN TRAINING.

**Data Download Date/Time** 08/02/2023 11:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

**Agency** BROWARD COUNTY SO **Date** 09/06/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:28:35 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: RFI Detect.UNKNOWN SOURCE.JGUZMAN IN TRAINING

**Data Download Date/Time** 09/06/2023 11:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

**Agency** BROWARD COUNTY SO **Date** 10/04/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 12:27:50 **Software** 8100.27

**0.05 Lot #** 202203J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

J GUZMAN IN TRAINING

**Data Download Date/Time** 10/04/2023 12:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

**Agency** BROWARD COUNTY SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 11:13:54 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

JGUZMAN IN TRAINING

**Data Download Date/Time** 11/08/2023 11:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007434

**Agency** BROWARD COUNTY SO **Date** 12/13/2023 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:42:42 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.203 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/13/2023 11:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

**Agency** BAKER COUNTY SO **Date** 01/10/2023 **Type** Department Inspection
**Inspector** NICODEMO,PHIL, **Time** 11:51:11 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 01/10/2023 11:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

**Agency** BAKER COUNTY SO **Date** 01/12/2023 **Type** Agency Inspection
**Inspector** CALEB COLLINS,DEREK WATSON,/ **Time** 13:38:05 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 06021080A2
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/20/2023 14:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

**Agency** BAKER COUNTY SO **Date** 02/18/2023 **Type** Agency Inspection
**Inspector** CALEB COLLINS,DEREK WATSON,/ **Time** 09:23:09 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 06021080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/22/2023 08:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

**Agency** BAKER COUNTY SO **Date** 03/22/2023 **Type** Agency Inspection
**Inspector** DEREK WATSON,CALEB COLLINS,/ **Time** 14:30:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 06021080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/27/2023 09:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

**Agency** BAKER COUNTY SO **Date** 04/30/2023 **Type** Agency Inspection
**Inspector** COLLINS,CALEB,A **Time** 12:19:56 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/30/2023 12:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

**Agency** BAKER COUNTY SO **Date** 05/28/2023 **Type** Agency Inspection
**Inspector** COLLINS,CALEB,A **Time** 13:51:27 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/01/2023 09:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

**Agency** BAKER COUNTY SO **Date** 06/30/2023 **Type** Agency Inspection
**Inspector** COLLINS,CALEB,A **Time** 13:00:18 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: RFI Detect.

**Data Download Date/Time** 06/30/2023 13:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

**Agency** BAKER COUNTY SO **Date** 07/22/2023 **Type** Agency Inspection
**Inspector** COLLINS,CALEB,A **Time** 12:47:42 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206B **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/14/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/22/2023 12:52

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

**Agency** BAKER COUNTY SO **Date** 08/25/2023 **Type** Agency Inspection
**Inspector** COLLINS',CALEB,A **Time** 14:17:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | 0.046 |
| Interferent Detect Test # (Repeat) 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | 0.046 |
| Interferent Detect Test # (Repeat) 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

Int Det: Sequence Aborted.

**Data Download Date/Time** 08/25/2023 14:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BAKER COUNTY SO | **Date** | 09/26/2023 | **Type** | Agency Inspection |
| **Inspector** | COLLINS,CALEB,A | **Time** | 11:25:35 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | Alcohol Free Test # 2 | 0.000 |
| Minimum Sample Volume Check: OK | | Alcohol Free Test # 3 | 0.000 |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | Yes | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | Yes | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

**Data Download Date/Time** 09/26/2023 12:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

**Agency** BAKER COUNTY SO **Date** 09/26/2023 **Type** Agency Inspection
**Inspector** COLLINS,CALEB,A **Time** 12:09:25 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 292206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/26/2023 12:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BAKER COUNTY SO | **Date** | 10/15/2023 | **Type** | Agency Inspection |
| **Inspector** | COLLINS,CALEB,A | **Time** | 11:55:51 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | | **0.08 Gas Lot #** | |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | | **Expiration** | |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.047 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.050 |
| Alcohol Free Subject Test: 0.000 | Yes | Interferent Detect Test # 1 | INT |
| Mouth Alcohol Test: Slope Not Met | Yes | Alcohol Free Test # 2 | 0.000 |
| Interferent Detect Test: Interferent Detect | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Diagnostic Check (Post-Inspection): OK | No | 0.08 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.052 |
| 0.05 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| 0.20 g/210L Test: OK | No | 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Dry Gas Std Test: OK | No | 0.08 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.036 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Ambient Fail. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 10/15/2023 12:01

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BAKER COUNTY SO | **Date** | 12/12/2023 | **Type** | Agency Inspection |
| **Inspector** | SIDDOWAY,BENJAMIN,W | **Time** | 10:00:38 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

COMPLIANCE NOT DETERMINED, BYPASS TO OPERATE

**Data Download Date/Time** 12/12/2023 12:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

**Agency** BAKER COUNTY SO **Date** 12/12/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 12:39:16 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 12/12/2023 12:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

**Agency** FHP **Date** 01/31/2023 **Type** Agency Inspection
**Inspector** GONZALEZ,CHRISTOPHER,M **Time** 01:35:34 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 03521080A3
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CHECKED CASE NO CRACKS, IN COMPLIANCE

**Data Download Date/Time** 02/02/2023 16:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

**Agency** FHP **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** DOBOSIEWICZ,ANTHONY,N **Time** 13:53:24 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 03521080A3
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: MOUTH PIECE CAME OUT PRIOR TO GETTING TO DECLINE.

**Data Download Date/Time** 02/28/2023 15:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

**Agency** FHP **Date** 03/30/2023 **Type** Agency Inspection
**Inspector** GONZALEZ,CHRISTOPHER,M **Time** 16:03:25 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402649340
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/27/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INSPECTED CASE NO CRACKS, IN COMPLIANCE

**Data Download Date/Time** 03/31/2023 14:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

**Agency** FHP **Date** 04/28/2023 **Type** Agency Inspection
**Inspector** GONZALEZ,CHRISTOPHER,M **Time** 18:12:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202306B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402649340
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/27/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CHECKED CASE AND WIPED DOWN. IN COMPLIANCE.

**Data Download Date/Time** 05/03/2023 02:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

**Agency** FHP **Date** 05/29/2023 **Type** Agency Inspection
**Inspector** DOBOSIEWICZ,ANTHONY,N **Time** 04:19:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402649340
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/27/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CASE CHECKED NO CRACKS IN COMPLIANCE

**Data Download Date/Time** 05/29/2023 05:01

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

**Agency** FHP **Date** 06/22/2023 **Type** Agency Inspection
**Inspector** GONZALEZ,CHRISTOPHER,M **Time** 17:55:13 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402649340
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/27/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CASE CHECKED AND WIPED DOWN

**Data Download Date/Time** 06/22/2023 18:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

**Agency** FHP **Date** 07/21/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 14:52:07 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
**Expiration** 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.047 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.047 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 07/21/2023 14:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

**Agency** FHP **Date** 12/30/2023 **Type** Agency Inspection
**Inspector** DOBOSIEWICZ,ANTHONY,N **Time** 02:41:21 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402649340
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/27/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CASE CHECKED NO CRACKS/ IN COMPLIANCE

**Data Download Date/Time** 12/30/2023 03:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 01/27/2023 | **Type** | Agency Inspection |
| **Inspector** | MARCUM,JOSEPH,K | **Time** | 11:00:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | 03521080A3 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 03/05/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 01/27/2023 16:01

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

**Agency** FHP **Date** 02/06/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 12:29:52 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 02/06/2023 13:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 02/17/2023 | **Type** | Agency Inspection |
| **Inspector** | BARGE,SUSAN,M | **Time** | 16:27:39 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402477284 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/27/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 02/20/2023 14:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 03/30/2023 | **Type** | Agency Inspection |
| **Inspector** | DRANE,RICHARD,B | **Time** | 12:38:21 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402477284 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/15/2024 | **Expiration** | 06/27/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 03/30/2023 14:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 04/21/2023 | **Type** | Agency Inspection |
| **Inspector** | DRANE,RICHARD,B | **Time** | 14:57:32 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402477284 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/27/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 04/21/2023 17:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

**Agency** FHP **Date** 05/16/2023 **Type** Agency Inspection
**Inspector** WALLEN,JUSTIN,R **Time** 12:39:24 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 05/16/2023 14:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 06/14/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 11:27:16 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402477284 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/27/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 06/14/2023 13:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 07/19/2023 | **Type** | Agency Inspection |
| **Inspector** | DRANE,RICHARD,B | **Time** | 14:24:44 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402477284 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/27/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 07/19/2023 15:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 08/28/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 14:03:35 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402477284 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/27/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.067 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 08/28/2023 15:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 09/25/2023 | **Type** | Agency Inspection |
| **Inspector** | DRANE,RICHARD,B | **Time** | 07:18:40 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

DIAGNOSTIC FAILNon-compliance: .

**Data Download Date/Time** 09/25/2023 10:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 09/25/2023 | **Type** | Agency Inspection |
| **Inspector** | DRANE,RICHARD,B | **Time** | 08:01:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402477284 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/27/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 09/25/2023 10:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 10/24/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 10:01:10 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402477284 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/27/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 10/24/2023 11:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 11/21/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 10:59:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402477284 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/27/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: .

**Data Download Date/Time** 11/21/2023 12:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

**Agency** FHP **Date** 12/04/2023 **Type** Agency Inspection
**Inspector** DRANE,RICHARD,B **Time** 08:20:59 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 12/04/2023 10:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

**Agency** FHP TRP B LAKE CITY **Date** 02/13/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 12:30:33 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 02/13/2023 12:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

**Agency** FHP TRP B LAKE CITY **Date** 03/15/2023 **Type** Agency Inspection
**Inspector** LAMONT,GREGORY,S **Time** 12:17:47 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/15/2023 13:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

**Agency** FHP TRP B LAKE CITY **Date** 04/11/2023 **Type** Agency Inspection
**Inspector** MURPHY,JAVON,D **Time** 20:07:13 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 02/24/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/11/2023 20:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

**Agency** FHP TRP B LAKE CITY **Date** 05/25/2023 **Type** Agency Inspection
**Inspector** MURPHY,JAVON,D **Time** 12:12:39 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/24/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/25/2023 12:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

**Agency** FHP TRP B LAKE CITY **Date** 06/30/2023 **Type** Agency Inspection
**Inspector** MURPHY,JAVON,D **Time** 20:01:48 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/24/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/01/2023 17:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

**Agency** FHP TRP B LAKE CITY **Date** 07/29/2023 **Type** Agency Inspection
**Inspector** MURPHY,JAVON,D **Time** 17:37:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/24/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/29/2023 17:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

**Agency** FHP TRP B LAKE CITY **Date** 08/31/2023 **Type** Agency Inspection
**Inspector** MURPHY,JAVON,D **Time** 13:49:53 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/24/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/31/2023 15:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

**Agency** FHP TRP B LAKE CITY **Date** 10/14/2023 **Type** Agency Inspection
**Inspector** MURPHY,JAVON,D **Time** 02:45:58 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/24/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 10/14/2023 03:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

**Agency** FHP TRP B LAKE CITY **Date** 12/02/2023 **Type** Agency Inspection
**Inspector** MURPHY,JAVON,D **Time** 21:02:23 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 02/24/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/03/2023 20:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 10/09/2023 | **Type** | Agency Inspection |
| **Inspector** | SOTO,ISRAEL, | **Time** | 13:12:33 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 10/12/2023 12:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

**Agency** FHP **Date** 10/12/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 12:37:07 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.082 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.082 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.082 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.082 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0007 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 10/12/2023 12:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

**Agency** FHP **Date** 01/31/2023 **Type** Agency Inspection
**Inspector** DOBOSIEWICZ,ANTHONY,N **Time** 01:46:07 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 3521080A3
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CASE CHECKED, NO CRACKS. IN COMPLIANCE.

**Data Download Date/Time** 02/01/2023 10:55

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

**Agency** FHP **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** DOBOSIEWICZ,ANTHONY,N **Time** 12:52:07 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 3521080A3
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

CASE CHECKED NO CRACKS/ IN COMPLIANCE

**Data Download Date/Time** 02/28/2023 14:23

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

**Agency** FHP **Date** 03/30/2023 **Type** Agency Inspection
**Inspector** GONZALEZ,CHRISTOPHER,M **Time** 16:03:45 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 2402498227
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INSPECTED CASE NO CRACKS, IN COMPLIANCE

**Data Download Date/Time** 03/31/2023 15:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

**Agency** FHP **Date** 04/28/2023 **Type** Agency Inspection
**Inspector** GONZALEZ,CHRISTOPHER,M **Time** 18:32:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 2402498227
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CHECKED CASE AND WIPED DOWN. IN COMPLIANCE.

**Data Download Date/Time** 05/03/2023 02:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 05/29/2023 | **Type** | Agency Inspection |
| **Inspector** | DOBOSIEWICZ,ANTHONY,N | **Time** | 03:27:18 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 2402498227 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 07/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed. CASE CHECKED NO CRACKS IN COMPLIANCE

**Data Download Date/Time** 05/29/2023 04:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

**Agency** FHP **Date** 06/22/2023 **Type** Agency Inspection
**Inspector** GONZALEZ,CHRISTOPHER,M **Time** 18:08:27 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 2402498227
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CASE CHECKED AND WIPED DOWN

**Data Download Date/Time** 06/22/2023 18:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

**Agency** FHP **Date** 07/21/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 15:20:38 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
**Expiration** 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.081 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.081 |
| 0.20 g/210L Test # 5 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007444

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0016 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 07/21/2023 15:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

**Agency** FHP **Date** 01/25/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 19:52:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 10/25/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/27/2023 08:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 02/15/2023 | **Type** | Agency Inspection |
| **Inspector** | FARLEY,JOSEPH,E | **Time** | 11:01:56 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402623828 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/15/2023 12:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

**Agency** FHP **Date** 03/28/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 11:42:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/28/2023 12:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 04/26/2023 | **Type** | Agency Inspection |
| **Inspector** | FARLEY,JOSEPH,E | **Time** | 14:21:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402623828 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/22/2025 |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

Non-compliance:VOLTAGE FAIL.

**Data Download Date/Time** 04/26/2023 15:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 04/26/2023 | **Type** | Agency Inspection |
| **Inspector** | FARLEY,JOSEPH,E | **Time** | 15:06:09 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402623828 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/26/2023 15:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 05/17/2023 | **Type** | Agency Inspection |
| **Inspector** | FARLEY,JOSEPH,E | **Time** | 13:46:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402623828 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/17/2023 15:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

**Agency** FHP **Date** 06/01/2023 **Type** Agency Inspection
**Inspector** SMITH,MATTHEW,M **Time** 14:04:45 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024026238
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 06/09/2023 13:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 06/07/2023 | **Type** | Agency Inspection |
| **Inspector** | NICODEMO,PHIL, | **Time** | 14:18:09 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED

**Data Download Date/Time** 06/09/2023 13:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

**Agency** FHP **Date** 06/09/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 12:55:55 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 06/09/2023 13:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 07/07/2023 | **Type** | Agency Inspection |
| **Inspector** | SMITH,MATTHEW,M | **Time** | 15:50:14 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M: .

**Data Download Date/Time** 07/07/2023 17:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

**Agency** FHP **Date** 07/07/2023 **Type** Agency Inspection
**Inspector** SMITH,MATTHEW,M **Time** 16:39:49 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 212654
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/07/2023 17:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

**Agency** FHP **Date** 08/19/2023 **Type** Agency Inspection
**Inspector** SMITH,MATTHEW,M **Time** 09:02:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/19/2023 09:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 08/19/2023 | **Type** | Agency Inspection |
| **Inspector** | SMITH,MATTHEW,M | **Time** | 09:43:26 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | Interferent Detect Test # 1 | INT |
| Minimum Sample Volume Check: OK | | Alcohol Free Test # 2 | 0.000 |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # 2 | INT |
| Alcohol Free Subject Test: 0.000 | Yes | Alcohol Free Test # 3 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 3 | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | Yes | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

Int Det: RFI Detect.

**Data Download Date/Time** 08/19/2023 10:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

**Agency** FHP **Date** 08/19/2023 **Type** Agency Inspection
**Inspector** SMITH,MATTHEW,M **Time** 10:34:04 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/19/2023 10:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

**Agency** FHP **Date** 09/17/2023 **Type** Agency Inspection
**Inspector** SMITH,MATTHEW,M **Time** 12:25:52 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 4026238828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/17/2023 13:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

**Agency** FHP **Date** 10/19/2023 **Type** Agency Inspection
**Inspector** SMITH,MATTHEW,M **Time** 19:37:45 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/19/2023 21:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

**Agency** FHP **Date** 11/17/2023 **Type** Agency Inspection
**Inspector** SMITH,MATTHEW,M **Time** 19:44:27 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 11/17/2023 19:55

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

**Agency** FHP **Date** 12/19/2023 **Type** Agency Inspection
**Inspector** SMITH,MATTHEW,M **Time** 20:52:16 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 3024026238
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 12/19/2023 21:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 01/10/2023 | **Type** | Agency Inspection |
| **Inspector** | HENING,FRED,C | **Time** | 12:48:23 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | 03521080A3 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 03/05/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 01/10/2023 12:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

**Agency** FHP **Date** 01/11/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 12:37:05 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG113403
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 05/14/2023

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 01/11/2023 12:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 01/25/2023 | **Type** | Agency Inspection |
| **Inspector** | MONTGOMERY,KENNETH,R | **Time** | 17:28:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 2021112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | | **0.08 Gas Lot #** | |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | | **Expiration** | |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.049 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.070 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Test # (Repeat) 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.070 |
| 0.08 g/210L Test: OK | No | 0.08 g/210L Test # (Repeat) 2 | 0.000 |
| 0.20 g/210L Test: OK | No | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.070 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.000 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

05: RFI Detect. 08: Control Outside Tolerance, Control Outside Tolerance.

**Data Download Date/Time** 01/27/2023 08:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

**Agency** FHP **Date** 01/25/2023 **Type** Agency Inspection
**Inspector** FARLEY,JOSEPH,E **Time** 18:47:09 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/27/2023 08:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

**Agency** FHP **Date** 02/15/2023 **Type** Agency Inspection
**Inspector** MILLS,SCOTT,T **Time** 14:10:52 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/15/2023 15:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

**Agency** FHP **Date** 03/29/2023 **Type** Agency Inspection
**Inspector** MILLS,SCOTT,T **Time** 10:55:15 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 302-402623
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 04/03/2023 20:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

**Agency** FHP **Date** 04/13/2023 **Type** Agency Inspection
**Inspector** MILLS,SCOTT,T **Time** 20:25:38 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/11/2023 11:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 05/11/2023 | **Type** | Agency Inspection |
| **Inspector** | HENING,FRED,C | **Time** | 10:38:46 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

VOLTAGE FAILURENon-compliance:FAILED DIAGNOSTIC TEST.

**Data Download Date/Time** 05/11/2023 11:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

**Agency** FHP **Date** 05/11/2023 **Type** Agency Inspection
**Inspector** HENING,FRED,C **Time** 11:24:27 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 05/11/2023 11:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 06/07/2023 | **Type** | Agency Inspection |
| **Inspector** | HENING,FRED,C | **Time** | 19:19:20 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402623828 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 06/08/2023 12:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

**Agency** FHP **Date** 07/10/2023 **Type** Agency Inspection
**Inspector** HENING,FRED,C **Time** 11:52:50 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/10/2023 12:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 08/21/2023 | **Type** | Agency Inspection |
| **Inspector** | HENING,FRED,C | **Time** | 15:21:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402623828 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 08/21/2023 15:53

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 09/12/2023 | **Type** | Agency Inspection |
| **Inspector** | HENING,FRED,C | **Time** | 13:11:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402623828 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 09/12/2023 13:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

**Agency** FHP **Date** 10/11/2023 **Type** Agency Inspection
**Inspector** HENING,FRED,C **Time** 13:12:33 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.045 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 10/11/2023 13:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 11/20/2023 | **Type** | Agency Inspection |
| **Inspector** | HENING,FRED,C | **Time** | 14:44:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402623828 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 11/20/2023 15:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 12/19/2023 | **Type** | Agency Inspection |
| **Inspector** | HENING,FRED,C | **Time** | 14:10:09 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402623828 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/22/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 12/19/2023 15:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

**Agency** FHP **Date** 03/15/2023 **Type** Agency Inspection
**Inspector** LAMONT,GREGORY,S **Time** 12:25:52 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: .

**Data Download Date/Time** 03/15/2023 12:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

**Agency** FHP **Date** 04/21/2023 **Type** Agency Inspection
**Inspector** LAMONT,GREGORY,S **Time** 13:18:27 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/04/2023 08:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 05/03/2023 | **Type** | Agency Inspection |
| **Inspector** | LAMONT,GREGORY,S | **Time** | 12:52:16 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1433248 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/06/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/04/2023 08:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

**Agency** FHP **Date** 05/26/2023 **Type** Department Inspection
**Inspector** NICODEMO,PHIL, **Time** 15:28:36 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 05/26/2023 15:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

**Agency** FHP **Date** 08/18/2023 **Type** Agency Inspection
**Inspector** LAMONT,GREGORY,S **Time** 12:07:16 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/18/2023 12:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 09/07/2023 | **Type** | Agency Inspection |
| **Inspector** | LAMONT,GREGORY,S | **Time** | 16:32:58 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1433248 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/06/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/07/2023 16:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 10/17/2023 | **Type** | Agency Inspection |
| **Inspector** | LAMONT,GREGORY,S | **Time** | 11:33:02 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1433248 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/06/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/17/2023 11:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 11/20/2023 | **Type** | Agency Inspection |
| **Inspector** | HERNANDEZ,ALAIN, | **Time** | 10:36:55 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1433248 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/06/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/20/2023 11:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

**Agency** FHP **Date** 12/21/2023 **Type** Agency Inspection
**Inspector** HERNANDEZ,ALAIN, **Time** 11:13:19 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/21/2023 11:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 01/19/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 10:14:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/19/2023 10:53

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 02/23/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 12:36:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/23/2023 12:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 03/29/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 11:48:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/29/2023 11:53

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 04/18/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 14:52:45 **Software** 8100.27

**0.05 Lot #** 20112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08:LOOSE HOSE.

**Data Download Date/Time** 04/18/2023 15:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 05/30/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 12:22:14 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/30/2023 13:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 06/09/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:52:14 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 06/09/2023 14:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 06/28/2023 **Type** Agency Inspection
**Inspector** JONES,CORTLAND,T **Time** 14:33:52 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2023 14:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 07/26/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 11:37:26 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/26/2023 11:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 08/25/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 12:40:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/25/2023 12:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 09/27/2023 **Type** Agency Inspection
**Inspector** JONES,CORTLAND,T **Time** 11:24:37 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/27/2023 11:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 10/26/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 09:41:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/26/2023 09:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 11/29/2023 **Type** Agency Inspection
**Inspector** JONES,CORTLAND,T **Time** 10:06:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/29/2023 10:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007458

**Agency** FHP TROOP L **Date** 12/15/2023 **Type** Agency Inspection
**Inspector** LIMB,HUNTER, **Time** 16:59:06 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1474747
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/12/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/15/2023 17:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 01/18/2023 | **Type** | Agency Inspection |
| **Inspector** | RIVERA,VICTOR,J | **Time** | 22:12:04 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | Interferent Detect Test # 1 | 0.000 |
| Minimum Sample Volume Check: OK | | Alcohol Free Test # 2 | 0.000 |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # 2 | 0.000 |
| Alcohol Free Subject Test: 0.000 | No | Alcohol Free Test # 3 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 3 | 0.000 |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | Yes | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

Int Det:RESIDUAL MOUTH ALCOHOL. Non-compliance:RESIDUAL MOUTH ALCOHOL.

**Data Download Date/Time** 01/20/2023 12:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

**Agency** FHP **Date** 01/18/2023 **Type** Agency Inspection
**Inspector** RIVERA,VICTOR,J **Time** 23:24:56 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402231696
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/20/2023 12:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

**Agency** FHP **Date** 02/18/2023 **Type** Agency Inspection
**Inspector** JEAN BAPTISTE,ALBENS, **Time** 23:35:45 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402231696
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 02/22/2023 13:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 03/02/2023 | **Type** | Agency Inspection |
| **Inspector** | SOTO,ISRAEL, | **Time** | 11:38:39 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 03/03/2023 11:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

**Agency** FHP **Date** 03/03/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 11:07:20 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 03/03/2023 11:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 03/15/2023 | **Type** | Agency Inspection |
| **Inspector** | RIVERA,VICTOR,J | **Time** | 13:00:30 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | 402231696 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 10/07/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

CLEAN NO ISSUES

**Data Download Date/Time** 03/16/2023 07:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

**Agency** FHP **Date** 04/22/2023 **Type** Agency Inspection
**Inspector** JEAN BAPTISTE,ALBENS, **Time** 17:43:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402231696
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

UNIT CLEAN NO ISSUE

**Data Download Date/Time** 04/26/2023 00:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

**Agency** FHP **Date** 05/30/2023 **Type** Agency Inspection
**Inspector** JEAN BAPTISTE,ALBENS, **Time** 10:25:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402231696
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/07/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

UNIT CLEAN NO ISSUE

**Data Download Date/Time** 05/30/2023 12:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 06/28/2023 | **Type** | Agency Inspection |
| **Inspector** | JEAN BAPTISTE,ALBENS, | **Time** | 23:54:07 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402231696 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/07/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

UNIT CLEAN NO ISSUE

**Data Download Date/Time** 09/26/2023 15:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 09/26/2023 | **Type** | Agency Inspection |
| **Inspector** | SIDDOWAY,BENJAMIN,W | **Time** | 10:28:15 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

COMPLIANCE NOT DETERMINED, BYPASS TO OPERATE

**Data Download Date/Time** 09/26/2023 15:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

**Agency** FHP **Date** 09/26/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 15:37:11 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 06723080A5
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0007 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.075 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0018 |

**Remarks:**

**Data Download Date/Time** 09/26/2023 15:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 01/27/2023 | **Type** | Agency Inspection |
| **Inspector** | MARCUM,JOSEPH,K | **Time** | 14:16:24 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | 1482048 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 04/21/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 01/27/2023 16:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

**Agency** FHP **Date** 02/09/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 08:18:48 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 02/09/2023 09:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 02/17/2023 | **Type** | Agency Inspection |
| **Inspector** | BARGE,SUSAN,M | **Time** | 16:22:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1482048 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/21/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 02/17/2023 17:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 03/30/2023 | **Type** | Agency Inspection |
| **Inspector** | DRANE,RICHARD,B | **Time** | 11:44:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1482048 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/15/2024 | **Expiration** | 04/21/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 03/30/2023 14:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 04/21/2023 | **Type** | Agency Inspection |
| **Inspector** | DRANE,RICHARD,B | **Time** | 15:37:21 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1482048 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/21/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 04/21/2023 16:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 05/16/2023 | **Type** | Agency Inspection |
| **Inspector** | DRANE,RICHARD,B | **Time** | 11:13:15 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1482048 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/21/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 05/16/2023 14:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 06/14/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 10:11:53 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 1482048 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/21/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 06/14/2023 13:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007465

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 07/19/2023 | **Type** | Agency Inspection |
| **Inspector** | WALLEN,JUSTIN,R | **Time** | 11:20:16 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | | **0.20 Lot #** | | **0.08 Gas Lot #** | |
| **Expiration** | 12/07/2023 | **Expiration** | | **Expiration** | | **Expiration** | |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | |
| Minimum Sample Volume Check: OK | | Interferent Detect Test # 1 | INT |
| Barometric Pressure Sensor Check: OK | | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Subject Test: 0.000 | Yes | Interferent Detect Test # 2 | INT |
| Mouth Alcohol Test: Slope Not Met | Yes | Alcohol Free Test # 3 | 0.000 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 3 | INT |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | Yes | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

05: Range Exceeded, Ambient Fail. Non-compliance: SIM ERROR.

**Data Download Date/Time** 10/31/2023 13:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

**Agency** BAY COUNTY SO **Date** 01/23/2023 **Type** Agency Inspection
**Inspector** MILLER,BRIAN,D **Time** 11:10:55 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1303107
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/28/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/23/2023 13:53

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

**Agency** BAY COUNTY SO **Date** 02/27/2023 **Type** Agency Inspection
**Inspector** MILLER,BRIAN,D **Time** 10:46:38 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1303107
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/28/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/27/2023 13:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

**Agency** BAY COUNTY SO **Date** 03/29/2023 **Type** Agency Inspection
**Inspector** MILLER,BRIAN,D **Time** 11:19:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** AG121002
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/29/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/29/2023 13:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

**Agency** BAY COUNTY SO **Date** 04/30/2023 **Type** Agency Inspection
**Inspector** MILLER,BRIAN,D **Time** 15:22:50 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** AG121002
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/29/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/01/2023 11:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

**Agency** BAY COUNTY SO **Date** 05/01/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 10:04:59 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 05/01/2023 11:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

**Agency** BAY COUNTY SO **Date** 05/02/2023 **Type** Agency Inspection
**Inspector** MILLER,BRIAN,D **Time** 13:34:42 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** AG121002
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/29/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/02/2023 16:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

**Agency** BAY COUNTY SO **Date** 08/14/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 13:06:04 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 06723080A5
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0012 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 08/14/2023 13:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

**Agency** BAY COUNTY SO **Date** 09/13/2023 **Type** Agency Inspection
**Inspector** MILLER,BRIAN,D **Time** 09:31:06 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402754628
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/30/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/13/2023 11:55

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

**Agency** BAY COUNTY SO **Date** 10/23/2023 **Type** Agency Inspection
**Inspector** MILLER,BRIAN,D **Time** 10:15:16 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402754628
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/30/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/23/2023 11:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

**Agency** BAY COUNTY SO **Date** 11/29/2023 **Type** Agency Inspection
**Inspector** MILLER,BRIAN,D **Time** 14:29:59 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402754628
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/30/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/29/2023 16:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

**Agency** BAY COUNTY SO **Date** 12/28/2023 **Type** Agency Inspection
**Inspector** MILLER,BRIAN,D **Time** 15:32:14 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402754628
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 05/30/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/28/2023 17:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

**Agency** HILLSBOROUGH CO SO **Date** 01/10/2023 **Type** Agency Inspection
**Inspector** DENDY,HALEY,M **Time** 09:25:24 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 01/10/2023 09:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

**Agency** HILLSBOROUGH CO SO **Date** 02/10/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 10:26:55 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 02/10/2023 10:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

**Agency** HILLSBOROUGH CO SO **Date** 06/06/2023 **Type** Department Inspection
**Inspector** NICODEMO,PHIL, **Time** 11:41:39 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.047 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

A F / M A: Range Exceeded.RINSED MOUTH AND REPEATED.

**Data Download Date/Time** 06/06/2023 11:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

**Agency** HILLSBOROUGH CO SO **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:19:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/05/2023 09:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 08/01/2023 | **Type** | Agency Inspection |
| **Inspector** | KNOX,STEPHANIE,R | **Time** | 11:29:29 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | N | **0.08 Gas Lot #** | 29122080A1 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 12/05/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 08/01/2023 11:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

**Agency** HILLSBOROUGH CO SO **Date** 09/05/2023 **Type** Agency Inspection
**Inspector** DAIL,JULIE,A **Time** 13:51:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
**Expiration** 12/07/2023 **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.044 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.044 |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # 1 | INT |
| Alcohol Free Subject Test: 0.000 | Yes | Alcohol Free Test # 2 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.05 g/210L Test # 2 | 0.045 |
| Interferent Detect Test: Interferent Detect | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.044 |
| Diagnostic Check (Post-Inspection): OK | No | Interferent Detect Test # 2 | INT |
| Alcohol Free Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test: OK | No | 0.05 g/210L Test # 3 | 0.045 |
| 0.08 g/210L Test: OK | No | 0.05 g/210L Test # (Repeat) 3 | 0.044 |
| 0.20 g/210L Test: OK | No | Interferent Detect Test # 3 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

05: Control Outside Tolerance SIM TURNED OFF . AI S. KNOX #261252 PRESENT FOR INSPECTION.Non-compliance:0.05 CONTROL OUTSIDE TOLERANCE.

**Data Download Date/Time** 09/05/2023 14:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

**Agency** HILLSBOROUGH CO SO **Date** 09/05/2023 **Type** Agency Inspection
**Inspector** DAIL,JULIE,A **Time** 14:48:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252

**Data Download Date/Time** 09/05/2023 14:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

**Agency** HILLSBOROUGH CO SO **Date** 09/22/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 08:47:02 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.045 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.045 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 09/22/2023 08:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

**Agency** HILLSBOROUGH CO SO **Date** 10/12/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 11:11:49 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 10/12/2023 11:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

**Agency** HILLSBOROUGH CO SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:31:27 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 11/08/2023 09:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

**Agency** HILLSBOROUGH CO SO **Date** 12/01/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:24:49 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 12/01/2023 09:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

**Agency** HILLSBOROUGH CO SO **Date** 01/10/2023 **Type** Agency Inspection
**Inspector** DENDY,HALEY,M **Time** 09:32:26 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 01/10/2023 09:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

**Agency** HILLSBOROUGH CO SO **Date** 02/07/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 12:39:38 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI H.DENDY #260868 PRESENT FOR INSPECTION.

**Data Download Date/Time** 02/07/2023 12:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

**Agency** HILLSBOROUGH CO SO **Date** 03/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 10:26:58 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI H. DENDY #260868 PRRESENT FOR INSPECTION.

**Data Download Date/Time** 03/02/2023 10:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 04/10/2023 | **Type** | Agency Inspection |
| **Inspector** | KNOX,STEPHANIE,R | **Time** | 11:49:27 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 13921080A2 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 07/05/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 04/10/2023 14:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

**Agency** HILLSBOROUGH CO SO **Date** 05/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:21:47 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 05/02/2023 11:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

**Agency** HILLSBOROUGH CO SO **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:32:46 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 07/05/2023 09:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

**Agency** HILLSBOROUGH CO SO **Date** 08/01/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:35:09 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 08/01/2023 11:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

**Agency** HILLSBOROUGH CO SO **Date** 09/05/2023 **Type** Agency Inspection
**Inspector** DAIL,JULIE,A **Time** 11:10:08 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 09/05/2023 11:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

**Agency** HILLSBOROUGH CO SO **Date** 09/12/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 10:38:23 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 06723080A5
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 09/12/2023 10:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

**Agency** HILLSBOROUGH CO SO **Date** 09/22/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 08:29:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 09/22/2023 08:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

**Agency** HILLSBOROUGH CO SO **Date** 10/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 10:51:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 10/02/2023 11:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

**Agency** HILLSBOROUGH CO SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 13:31:45 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 11/08/2023 13:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

**Agency** HILLSBOROUGH CO SO **Date** 12/01/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:32:02 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 12/01/2023 09:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

**Agency** HILLSBOROUGH CO SO **Date** 01/10/2023 **Type** Agency Inspection
**Inspector** DENDY,HALEY,M **Time** 11:41:54 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 01/10/2023 11:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

**Agency** HILLSBOROUGH CO SO **Date** 02/10/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 14:45:44 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 02/10/2023 14:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

**Agency** HILLSBOROUGH CO SO **Date** 03/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 08:45:41 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI H. DENDY #260868 PRESENT FOR INSPECTION.

**Data Download Date/Time** 03/02/2023 08:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

**Agency** HILLSBOROUGH CO SO **Date** 04/10/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:26:46 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 04/10/2023 09:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

**Agency** HILLSBOROUGH CO SO **Date** 05/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:02:59 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 05/02/2023 10:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

**Agency** HILLSBOROUGH CO SO **Date** 06/05/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 08:57:58 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 06/05/2023 09:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

**Agency** HILLSBOROUGH CO SO **Date** 07/05/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:57:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 07/05/2023 12:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

**Agency** HILLSBOROUGH CO SO **Date** 08/01/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:59:05 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 08/01/2023 10:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

**Agency** HILLSBOROUGH CO SO **Date** 09/05/2023 **Type** Agency Inspection
**Inspector** DAIL,JULIE,A **Time** 13:54:35 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 09/05/2023 14:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

**Agency** HILLSBOROUGH CO SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:13:49 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 11/08/2023 09:28

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007481

**Agency** HILLSBOROUGH CO SO **Date** 12/01/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:45:22 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 12/01/2023 09:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

**Agency** HILLSBOROUGH CO SO **Date** 01/10/2023 **Type** Agency Inspection
**Inspector** DENDY,HALEY,M **Time** 09:43:42 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 01/10/2023 09:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

**Agency** HILLSBOROUGH CO SO **Date** 02/07/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 10:43:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI H.DENDY #260868 PRESENT FOR INSPECTION.

**Data Download Date/Time** 02/07/2023 10:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

**Agency** HILLSBOROUGH CO SO **Date** 03/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 10:13:50 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI H. DENDY #260868 PRESENT FOR INSPECTION.

**Data Download Date/Time** 03/02/2023 10:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

**Agency** HILLSBOROUGH CO SO **Date** 04/10/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:36:34 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 04/10/2023 13:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

**Agency** HILLSBOROUGH CO SO **Date** 05/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:29:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 05/02/2023 11:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

**Agency** HILLSBOROUGH CO SO **Date** 06/05/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:13:09 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 06/05/2023 10:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 08/07/2023 | **Type** | Agency Inspection |
| **Inspector** | SIDDOWAY,BENJAMIN,W | **Time** | 08:33:35 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

**Data Download Date/Time** 08/08/2023 15:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

**Agency** HILLSBOROUGH CO SO **Date** 08/08/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 15:37:31 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 08/08/2023 15:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

**Agency** HILLSBOROUGH CO SO **Date** 09/05/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:13:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A4
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 09/05/2023 09:23

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

**Agency** HILLSBOROUGH CO SO **Date** 10/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:00:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A4
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 10/02/2023 11:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

**Agency** HILLSBOROUGH CO SO **Date** 01/10/2023 **Type** Agency Inspection
**Inspector** DENDY,HALEY,M **Time** 10:37:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 01/10/2023 10:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

**Agency** HILLSBOROUGH CO SO **Date** 02/07/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 10:55:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI H.DENDY #260868 PRESENT FOR INSPECTION.

**Data Download Date/Time** 02/07/2023 11:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

**Agency** HILLSBOROUGH CO SO **Date** 03/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 10:32:07 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI H. DENDY #260868 PRESENT FOR INSPECTION.

**Data Download Date/Time** 03/02/2023 10:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 04/10/2023 | **Type** | Agency Inspection |
| **Inspector** | KNOX,STEPHANIE,R | **Time** | 11:31:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 13921080A2 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 07/05/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 04/10/2023 14:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

**Agency** HILLSBOROUGH CO SO **Date** 05/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:40:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 05/02/2023 11:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

**Agency** HILLSBOROUGH CO SO **Date** 07/10/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 10:37:39 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 07/10/2023 10:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

**Agency** HILLSBOROUGH CO SO **Date** 08/01/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:46:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 08/01/2023 09:53

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

**Agency** HILLSBOROUGH CO SO **Date** 09/05/2023 **Type** Agency Inspection
**Inspector** DAIL,JULIE,A **Time** 14:12:32 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.045 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 09/05/2023 14:23

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

**Agency** HILLSBOROUGH CO SO **Date** 10/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 13:03:49 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 10/02/2023 16:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

**Agency** HILLSBOROUGH CO SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 13:44:29 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 11/08/2023 14:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

**Agency** HILLSBOROUGH CO SO **Date** 12/01/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 14:02:21 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: No Sample Provided.NO AF IN TIME LIMIT.

**Data Download Date/Time** 12/01/2023 14:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

**Agency** HILLSBOROUGH CO SO **Date** 01/10/2023 **Type** Agency Inspection
**Inspector** DENDY,HALEY,M **Time** 10:43:16 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI S. KNOX #261252 PRESENT FOR INSPECTION.

**Data Download Date/Time** 01/10/2023 10:51

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

**Agency** HILLSBOROUGH CO SO **Date** 02/07/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 12:44:24 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI H.DENDY #260868 PRESENT FOR INSPECTION.

**Data Download Date/Time** 02/07/2023 12:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

**Agency** HILLSBOROUGH CO SO **Date** 03/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:03:08 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI H. DENDY #260868 PRESENT FOR INSPECTION.

**Data Download Date/Time** 03/02/2023 09:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

**Agency** HILLSBOROUGH CO SO **Date** 03/08/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 12:23:06 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 03/08/2023 12:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

**Agency** HILLSBOROUGH CO SO **Date** 04/10/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:17:04 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 04/10/2023 09:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

**Agency** HILLSBOROUGH CO SO **Date** 05/02/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:33:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI J.DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 05/02/2023 12:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 06/05/2023 | **Type** | Agency Inspection |
| **Inspector** | KNOX,STEPHANIE,R | **Time** | 11:34:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 13921080A2 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 07/05/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J. DAIL #254236 PRESENT FOR INSPECTION.

**Data Download Date/Time** 06/05/2023 12:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

**Agency** HILLSBOROUGH CO SO **Date** 06/15/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 10:37:28 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 06/15/2023 10:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HILLSBOROUGH CO SO | **Date** | 08/07/2023 | **Type** | Agency Inspection |
| **Inspector** | SIDDOWAY,BENJAMIN,W | **Time** | 07:33:47 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

**Data Download Date/Time** 08/08/2023 15:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

**Agency** HILLSBOROUGH CO SO **Date** 08/08/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 15:52:41 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 08/08/2023 15:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

**Agency** HILLSBOROUGH CO SO **Date** 10/31/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 09:58:34 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 10/31/2023 10:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

**Agency** HILLSBOROUGH CO SO **Date** 11/08/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 09:24:09 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPECTION.

**Data Download Date/Time** 11/08/2023 09:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

**Agency** HILLSBOROUGH CO SO **Date** 12/01/2023 **Type** Agency Inspection
**Inspector** KNOX,STEPHANIE,R **Time** 11:32:52 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B. KROME #263389 PRESENT FOR INSPCTION.

**Data Download Date/Time** 12/01/2023 11:51

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

**Agency** ORANGE CITY P.D. **Date** 01/26/2023 **Type** Agency Inspection
**Inspector** HOFFMAN,JOSHUA,J **Time** 04:26:49 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/26/2023 04:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

**Agency** ORANGE CITY P.D. **Date** 02/27/2023 **Type** Agency Inspection
**Inspector** HOFFMAN,JOSHUA,J **Time** 21:07:37 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | 0.028 |
| Interferent Detect Test # (Repeat) 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | 0.028 |
| Interferent Detect Test # (Repeat) 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | 0.029 |
| Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

A F / M A: Interferent Detect. Int Det:WAS NOT DETECTING SOLUTION/GETTING WRONG READING. .

**Data Download Date/Time** 03/02/2023 22:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

**Agency** ORANGE CITY P.D. **Date** 03/22/2023 **Type** Agency Inspection
**Inspector** KIBBE,BRIAN,J **Time** 02:24:37 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/26/2023 21:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

**Agency** ORANGE CITY P.D. **Date** 04/27/2023 **Type** Agency Inspection
**Inspector** KIBBE,BRIAN,J **Time** 11:36:47 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/02/2023 21:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

**Agency** ORANGE CITY P.D. **Date** 05/25/2023 **Type** Agency Inspection
**Inspector** KIBBE,BRIAN,J **Time** 01:09:46 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921989A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/25/2023 01:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

**Agency** ORANGE CITY P.D. **Date** 06/10/2023 **Type** Agency Inspection
**Inspector** HOFFMAN,JOSHUA,J **Time** 16:53:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13921080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/14/2023 04:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

**Agency** ORANGE CITY P.D. **Date** 07/28/2023 **Type** Agency Inspection
**Inspector** HOFFMAN,JOSHUA,J **Time** 05:05:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/28/2023 05:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

**Agency** ORANGE CITY P.D. **Date** 08/25/2023 **Type** Agency Inspection
**Inspector** HOFFMAN,JOSHUA,J **Time** 14:56:34 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/25/2023 15:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

**Agency** ORANGE CITY P.D. **Date** 09/28/2023 **Type** Agency Inspection
**Inspector** KIBBE,BRIAN,J **Time** 12:08:21 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/28/2023 12:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

**Agency** ORANGE CITY P.D. **Date** 10/24/2023 **Type** Agency Inspection
**Inspector** HOFFMAN,JOSHUA,J **Time** 22:03:54 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 2022206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/24/2023 22:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

**Agency** ORANGE CITY P.D. **Date** 11/02/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 14:44:02 **Software** 8100.27

**0.05 Lot #** 202303K **Expiration** 03/29/2025
**0.08 Lot #** 202303L **Expiration** 03/29/2025
**0.20 Lot #** 202304C **Expiration** 04/05/2025
**0.08 Gas Lot #** AG325603 **Expiration** 09/13/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.051 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 11/02/2023 14:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

**Agency** ORANGE CITY P.D. **Date** 12/18/2023 **Type** Agency Inspection
**Inspector** KIBBE,BRIAN,J **Time** 05:22:57 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/18/2023 05:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

**Agency** FFWCC **Date** 01/03/2023 **Type** Agency Inspection
**Inspector** DESANTIS,ADAM,B **Time** 13:21:55 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 31420080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INSTRUMENT WAS CHECKED FOR DAMAGES. NONE WERE OBSERVED

**Data Download Date/Time** 01/06/2023 12:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

**Agency** FFWCC **Date** 05/23/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 13:59:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/23/2023 15:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

**Agency** FFWCC **Date** 06/20/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 13:17:53 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/20/2023 13:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 07/28/2023 | **Type** | Agency Inspection |
| **Inspector** | PULASKI,DAMON,J | **Time** | 12:18:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 15922080A4 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 08/05/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/28/2023 13:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

**Agency** FFWCC **Date** 08/28/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 14:06:19 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/28/2023 14:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 09/26/2023 | **Type** | Agency Inspection |
| **Inspector** | PULASKI,DAMON,J | **Time** | 13:50:53 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 15922080A4 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 08/05/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/26/2023 14:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 10/12/2023 | **Type** | Agency Inspection |
| **Inspector** | PULASKI,DAMON,J | **Time** | 12:45:57 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 15922080A4 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 08/05/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/12/2023 13:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

**Agency** FFWCC **Date** 11/02/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 13:49:26 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/02/2023 14:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

**Agency** FFWCC **Date** 11/03/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 13:36:52 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 11/03/2023 13:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

**Agency** FFWCC **Date** 11/13/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 15:50:12 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/13/2023 16:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007492

**Agency** FFWCC **Date** 12/04/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 12:29:47 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/04/2023 12:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 01/11/2023 | **Type** | Agency Inspection |
| **Inspector** | PULASKI,DAMON,J | **Time** | 14:11:16 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | 15922080A4 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 08/05/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/11/2023 15:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

**Agency** FFWCC **Date** 02/13/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 13:36:53 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 15922080A4
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/13/2023 13:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 03/21/2023 | **Type** | Agency Inspection |
| **Inspector** | PULASKI,DAMON,J | **Time** | 13:23:08 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 15922080A4 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 08/05/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/21/2023 13:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

**Agency** FFWCC **Date** 04/11/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 12:41:52 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/11/2023 13:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

**Agency** FFWCC **Date** 05/05/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 12:58:35 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/05/2023 13:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 06/20/2023 | **Type** | Agency Inspection |
| **Inspector** | PULASKI,DAMON,J | **Time** | 13:30:21 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 15922080A4 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 08/05/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 06/20/2023 13:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

**Agency** FFWCC **Date** 07/28/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 12:29:40 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/28/2023 13:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

**Agency** FFWCC **Date** 08/28/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 14:13:31 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/28/2023 14:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

**Agency** FFWCC **Date** 09/26/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 13:58:42 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/26/2023 14:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 10/12/2023 | **Type** | Agency Inspection |
| **Inspector** | PULASKI,DAMON,J | **Time** | 12:52:58 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 15922080A4 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 08/05/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/12/2023 13:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

**Agency** FFWCC **Date** 11/02/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 13:59:59 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/02/2023 14:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

**Agency** FFWCC **Date** 11/03/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 14:18:31 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.047 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 11/03/2023 14:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 01/14/2023 | **Type** | Agency Inspection |
| **Inspector** | LAWRENCE,JOSHUA,S | **Time** | 21:25:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | AG129201 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 10/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/19/2023 11:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

**Agency** FFWCC **Date** 02/10/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL,X **Time** 16:26:13 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 02/10/2023 16:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 03/04/2023 | **Type** | Agency Inspection |
| **Inspector** | LAWRENCE,JOSHUA,S | **Time** | 14:45:44 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | AG129201 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 10/19/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/04/2023 18:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

**Agency** FFWCC **Date** 04/14/2023 **Type** Agency Inspection
**Inspector** LAWRENCE,JOSHUA,S **Time** 13:58:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG129201
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/14/2023 15:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

**Agency** FFWCC **Date** 05/22/2023 **Type** Agency Inspection
**Inspector** LAWRENCE,JOSHUA,S **Time** 18:12:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.016 |
| 0.05 g/210L Test # (Repeat) 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.017 |
| 0.05 g/210L Test # (Repeat) 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.020 |
| 0.05 g/210L Test # (Repeat) 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance. UNPLUGGED DURING TEST. . OUTSIDE TOLERANCE DUE TO UNPLUGGED DURING TEST. FIXED

**Data Download Date/Time** 05/22/2023 19:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

**Agency** FFWCC **Date** 06/20/2023 **Type** Agency Inspection
**Inspector** LAWRENCE,JOSHUA,S **Time** 11:07:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PASSED

**Data Download Date/Time** 06/20/2023 11:52

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 07/24/2023 | **Type** | Agency Inspection |
| **Inspector** | LAWRENCE,JOSHUA,S | **Time** | 11:16:57 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | AG310302 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/13/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

PASSED

**Data Download Date/Time** 07/24/2023 13:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 08/18/2023 | **Type** | Agency Inspection |
| **Inspector** | LAWRENCE,JOSHUA,S | **Time** | 23:00:30 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | AG310302 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/13/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/21/2023 16:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 09/19/2023 | **Type** | Agency Inspection |
| **Inspector** | LAWRENCE,JOSHUA,S | **Time** | 15:39:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | AG310302 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/13/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/20/2023 10:51

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

**Agency** FFWCC **Date** 10/29/2023 **Type** Agency Inspection
**Inspector** LAWRENCE,JOSHUA,S **Time** 20:20:01 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PASSED

**Data Download Date/Time** 10/30/2023 16:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

**Agency** FFWCC **Date** 11/10/2023 **Type** Agency Inspection
**Inspector** LAWRENCE,JOSHUA,S **Time** 22:19:32 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/11/2023 14:53

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

**Agency** FFWCC **Date** 12/29/2023 **Type** Agency Inspection
**Inspector** LAWRENCE,JOSHUA,S **Time** 15:14:13 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/29/2023 16:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 02/09/2023 | **Type** | Agency Inspection |
| **Inspector** | NICODEMO,PHIL, | **Time** | 11:07:16 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED

**Data Download Date/Time** 02/09/2023 13:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

**Agency** FFWCC **Date** 02/09/2023 **Type** Department Inspection
**Inspector** NICODEMO,PHIL, **Time** 13:34:03 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 02/09/2023 13:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

**Agency** FFWCC **Date** 03/06/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 10:31:50 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** AG222203
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed. A F / M A: No Sample Provided.

**Data Download Date/Time** 03/06/2023 11:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 04/28/2023 | **Type** | Agency Inspection |
| **Inspector** | MARROQUIN,RICHARD,J | **Time** | 11:14:46 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | AG222203 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 08/10/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 04/28/2023 11:51

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

**Agency** FFWCC **Date** 05/16/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 12:16:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** AG222203
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.025 |
| 0.05 g/210L Test # (Repeat) 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.028 |
| 0.05 g/210L Test # (Repeat) 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.032 |
| 0.05 g/210L Test # (Repeat) 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time** 05/16/2023 12:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

**Agency** FFWCC **Date** 06/20/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 13:21:31 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG222203
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

AI OK

**Data Download Date/Time** 06/20/2023 13:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

**Agency** FFWCC **Date** 07/24/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 15:10:49 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG222203
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/24/2023 15:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

**Agency** FFWCC **Date** 08/28/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 13:46:36 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG222203
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI OK

**Data Download Date/Time** 09/15/2023 14:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

**Agency** FFWCC **Date** 09/15/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 16:01:05 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG222203
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/15/2023 16:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

**Agency** FFWCC **Date** 10/31/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 14:04:04 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG222203
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: NO ENOUGH M/A.

**Data Download Date/Time** 10/31/2023 14:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 11/01/2023 | **Type** | Agency Inspection |
| **Inspector** | MARROQUIN,RICHARD,J | **Time** | 12:46:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | AG222203 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 08/10/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/01/2023 14:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 12/20/2023 | **Type** | Agency Inspection |
| **Inspector** | MARROQUIN,RICHARD,J | **Time** | 15:00:01 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | AG222203 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 08/10/2024 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/20/2023 16:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

**Agency** FFWCC **Date** 01/12/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 11:46:52 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 211143
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | No | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: NOT ENOUGH M/A. RESIDUAL M/A.

**Data Download Date/Time** 01/18/2023 08:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

**Agency** FFWCC **Date** 01/12/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 12:20:54 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 211143
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI OK

**Data Download Date/Time** 01/18/2023 08:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

**Agency** FFWCC **Date** 01/18/2023 **Type** Department Inspection
**Inspector** NICODEMO,PHIL, **Time** 13:25:40 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 01/18/2023 13:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 03/06/2023 | **Type** | Agency Inspection |
| **Inspector** | MARROQUIN,RICHARD,J | **Time** | 10:40:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 2021O8D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | 211143 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 08/05/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 03/06/2023 11:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 04/28/2023 | **Type** | Agency Inspection |
| **Inspector** | MARROQUIN,RICHARD,J | **Time** | 11:10:37 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202106B | **0.08 Gas Lot #** | 211143 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/22/2023 | **Expiration** | 08/05/2023 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 04/28/2023 11:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

**Agency** FFWCC **Date** 05/16/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 12:16:58 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 211143
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/16/2023 12:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

**Agency** FFWCC **Date** 06/20/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 12:59:35 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 211143
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/20/2023 13:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 07/24/2023 | **Type** | Agency Inspection |
| **Inspector** | MARROQUIN,RICHARD,J | **Time** | 15:07:07 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | AG310302 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 04/13/2025 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/24/2023 15:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

**Agency** FFWCC **Date** 08/28/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 13:49:58 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/15/2023 14:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

**Agency** FFWCC **Date** 09/15/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 15:53:34 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: RFI Detect.

**Data Download Date/Time** 09/15/2023 16:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

**Agency** FFWCC **Date** 10/31/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 14:00:01 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/31/2023 14:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

**Agency** FFWCC **Date** 11/01/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 12:47:32 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/01/2023 14:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007497

**Agency** FFWCC **Date** 12/20/2023 **Type** Agency Inspection
**Inspector** MARROQUIN,RICHARD,J **Time** 14:49:04 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/20/2023 16:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

**Agency** FFWCC **Date** 01/23/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 18:15:33 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 01/23/2023 19:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 02/03/2023 | **Type** | Agency Inspection |
| **Inspector** | BASFORD,NICOLE,M | **Time** | 18:06:31 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202108D | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 15421080A1 |
| **Expiration** | 12/07/2023 | **Expiration** | 08/17/2023 | **Expiration** | 06/15/2024 | **Expiration** | 08/05/2023 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 02/06/2023 15:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

**Agency** FFWCC **Date** 03/02/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 14:48:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 03/02/2023 16:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

**Agency** FFWCC **Date** 04/24/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 14:11:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/24/2023 15:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 05/15/2023 | **Type** | Agency Inspection |
| **Inspector** | BASFORD,NICOLE,M | **Time** | 14:15:12 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

**Data Download Date/Time** 05/15/2023 16:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

**Agency** FFWCC **Date** 05/15/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 15:09:52 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 05/15/2023 16:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

**Agency** FFWCC **Date** 06/12/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 19:35:36 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 06/16/2023 09:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

**Agency** FFWCC **Date** 07/24/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 18:19:53 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/24/2023 19:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 08/15/2023 | **Type** | Agency Inspection |
| **Inspector** | BASFORD,NICOLE,M | **Time** | 16:42:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402648278 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/27/2026 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | Yes | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 08/15/2023 17:54

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

**Agency** FFWCC **Date** 09/29/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 19:06:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402648278
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/27/2026

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/29/2023 20:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

**Agency** FFWCC **Date** 10/02/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 17:28:35 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402648278
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/27/2026

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/02/2023 18:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

**Agency** FFWCC **Date** 11/15/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 12:28:26 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG310901
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/19/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 11/15/2023 13:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 11/26/2023 | **Type** | Agency Inspection |
| **Inspector** | BASFORD,NICOLE,M | **Time** | 13:16:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402648278 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/27/2026 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/26/2023 14:27

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007498

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 12/22/2023 | **Type** | Agency Inspection |
| **Inspector** | BASFORD,NICOLE,M | **Time** | 16:19:15 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402648278 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 01/27/2026 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/22/2023 17:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

**Agency** FFWCC **Date** 01/23/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 18:10:16 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402469736
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/16/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 01/23/2023 19:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

**Agency** FFWCC **Date** 02/03/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 18:13:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402469736
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/16/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/06/2023 15:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

**Agency** FFWCC **Date** 03/02/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 14:58:26 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402469736
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/16/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 03/02/2023 16:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

**Agency** FFWCC **Date** 04/29/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 17:17:51 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402469736
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/16/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/29/2023 18:23

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

**Agency** FFWCC **Date** 05/01/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 13:54:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402469736
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/16/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/01/2023 15:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

**Agency** FFWCC **Date** 06/12/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 14:52:54 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402469736
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/16/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 06/16/2023 09:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

**Agency** FFWCC **Date** 06/19/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 14:13:50 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
**Expiration** 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 10/25/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # (Repeat) 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # (Repeat) 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

| | |
|---|---|
| Alcohol Free Test # (Repeat) 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # (Repeat) 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # (Repeat) 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # (Repeat) 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # (Repeat) 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

00: Ambient Fail MOUTH ALCOHOL TOO CLOSE.

**Data Download Date/Time** 06/19/2023 15:19

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 07/24/2023 | **Type** | Agency Inspection |
| **Inspector** | BASFORD,NICOLE,M | **Time** | 18:38:24 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402469736 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/16/2025 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/24/2023 19:45

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 08/15/2023 | **Type** | Agency Inspection |
| **Inspector** | BASFORD,NICOLE,M | **Time** | 17:02:23 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402469736 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/16/2025 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 08/15/2023 18:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

**Agency** FFWCC **Date** 09/29/2023 **Type** Agency Inspection
**Inspector** BASFORD,NICOLE,M **Time** 19:17:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402469736
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/16/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/29/2023 20:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 10/02/2023 | **Type** | Agency Inspection |
| **Inspector** | BASFORD,NICOLE,M | **Time** | 17:35:39 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402469736 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/16/2025 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/02/2023 18:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 11/26/2023 | **Type** | Agency Inspection |
| **Inspector** | BASFORD,NICOLE,M | **Time** | 13:11:43 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402469736 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/16/2025 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/26/2023 14:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 12/22/2023 | **Type** | Agency Inspection |
| **Inspector** | BASFORD,NICOLE,M | **Time** | 16:13:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303J | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 402469736 |
| **Expiration** | 03/28/2025 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 06/16/2025 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/22/2023 17:23

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

**Agency** FFWCC **Date** 01/17/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 10:51:47 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/17/2023 12:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FFWCC | **Date** | 02/28/2023 | **Type** | Agency Inspection |
| **Inspector** | GILMORE,RICHARD,L | **Time** | 13:17:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202112A | **0.08 Lot #** | 202206B | **0.20 Lot #** | 202206C | **0.08 Gas Lot #** | 15421080A1 |
| **Expiration** | 12/07/2023 | **Expiration** | 06/14/2024 | **Expiration** | 06/15/2024 | **Expiration** | 08/05/2023 |

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/28/2023 13:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

**Agency** FFWCC **Date** 03/03/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 17:54:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/03/2023 17:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

**Agency** FFWCC **Date** 04/24/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 21:26:30 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.075 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/16/2023 14:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

**Agency** FFWCC **Date** 05/16/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 15:53:06 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CORRECTION TO .20 LOT NUMBER 202206C

**Data Download Date/Time** 05/16/2023 15:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

**Agency** FFWCC **Date** 06/23/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 11:44:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 06/23/2023 11:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

**Agency** FFWCC **Date** 07/24/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 13:33:35 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/24/2023 13:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

**Agency** FFWCC **Date** 08/03/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 11:21:10 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A5
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMP

**Data Download Date/Time** 08/03/2023 11:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

**Agency** FFWCC **Date** 09/30/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 19:46:16 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A5
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 09/30/2023 19:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

**Agency** FFWCC **Date** 10/01/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 13:10:50 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A5
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/01/2023 13:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

**Agency** FFWCC **Date** 11/21/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 12:41:20 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 11/21/2023 12:48

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007500

**Agency** FFWCC **Date** 12/29/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 19:44:26 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A5
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Interferent Detect Test # (Repeat) 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Interferent Detect Test # (Repeat) 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

Int Det: RFI Detect.

**Data Download Date/Time** 12/29/2023 19:53

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

**Agency** FFWCC **Date** 01/17/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 10:47:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/17/2023 12:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

**Agency** FFWCC **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 13:24:07 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/28/2023 13:30

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

**Agency** FFWCC **Date** 03/03/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 18:01:40 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A:RESIDUAL MOUTH ALCOHOL.

**Data Download Date/Time** 03/03/2023 18:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

**Agency** FFWCC **Date** 04/24/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 21:17:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/16/2023 14:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

**Agency** FFWCC **Date** 05/16/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 15:58:26 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206B **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

CORRECTION TO .20 LOT NUMBER 202206C

**Data Download Date/Time** 05/16/2023 16:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

**Agency** FFWCC **Date** 06/23/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 12:02:54 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/23/2023 12:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

**Agency** FFWCC **Date** 07/24/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 13:25:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15421080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/24/2023 13:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

**Agency** FFWCC **Date** 08/03/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 11:15:30 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06723080A5
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMP

**Data Download Date/Time** 08/03/2023 11:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

**Agency** FFWCC **Date** 09/30/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 19:41:32 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06723080A5
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/30/2023 19:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

**Agency** FFWCC **Date** 10/01/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 13:15:57 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06723080A5
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

BUILDING POWER OUTAGE AT START UP

**Data Download Date/Time** 10/01/2023 13:23

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

**Agency** FFWCC **Date** 11/21/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 14:22:46 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.083 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.083 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.083 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.083 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 11/21/2023 14:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

**Agency** FFWCC **Date** 12/29/2023 **Type** Agency Inspection
**Inspector** GILMORE,RICHARD,L **Time** 19:38:01 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06723080A5
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/29/2023 19:49

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 01/23/2023 **Type** Agency Inspection
**Inspector** CARTAYA,GUILLERMO, **Time** 11:14:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/15/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: .

**Data Download Date/Time** 01/23/2023 12:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FISH & WILDLIFE CC | **Date** | 03/07/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 09:59:29 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT.

**Data Download Date/Time** 03/07/2023 14:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 03/07/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:46:29 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.076 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.076 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.076 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.077 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 03/07/2023 14:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 03/31/2023 **Type** Agency Inspection
**Inspector** CARTAYA,GUILLERMO, **Time** 13:25:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/23/2023 12:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 05/23/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 11:45:38 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/23/2023 12:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 06/28/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 13:05:10 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2023 13:32

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 07/18/2023 **Type** Agency Inspection
**Inspector** PULASKI,DAMON,J **Time** 12:02:55 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/18/2023 12:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 08/29/2023 **Type** Agency Inspection
**Inspector** SAAVEDRA,OSCAR, **Time** 13:30:47 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A:DID NOT HAVE SOLUTION IN MOUTH. Non-compliance:NO MOUTH ALCOHOL DURING MOUTH ALCOHOL SAMPLE.

**Data Download Date/Time** 08/29/2023 14:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 08/29/2023 **Type** Agency Inspection
**Inspector** SAAVEDRA,OSCAR, **Time** 14:35:29 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/04/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | No |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A:NOT ENOUGH MOUTH ALCOHOL SOLUTION.

**Data Download Date/Time** 08/29/2023 14:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 08/31/2023 **Type** Agency Inspection
**Inspector** CONROY,JESSICA,P **Time** 14:28:58 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/04/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/31/2023 14:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 09/25/2023 **Type** Agency Inspection
**Inspector** SAAVEDRA,OSCAR, **Time** 14:22:13 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/25/2023 14:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 10/24/2023 **Type** Agency Inspection
**Inspector** SAAVEDRA,OSCAR, **Time** 13:49:56 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/24/2023 14:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 11/27/2023 **Type** Agency Inspection
**Inspector** SAAVEDRA,OSCAR, **Time** 16:34:25 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: Range Exceeded.

**Data Download Date/Time** 11/27/2023 16:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007502

**Agency** FISH & WILDLIFE CC **Date** 12/28/2023 **Type** Agency Inspection
**Inspector** CONROY,JESSICA,P **Time** 14:15:00 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 402558895
**Expiration** 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/28/2023 14:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007503

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | POLK CSO | **Date** | 02/21/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 09:24:53 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT

**Data Download Date/Time** 02/21/2023 13:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007503

**Agency** POLK CSO **Date** 02/21/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:41:48 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007503

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.077 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 02/21/2023 13:47

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007503

**Agency** POLK CSO **Date** 07/14/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 10:18:33 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
**Expiration** 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.080 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007503

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 07/14/2023 10:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | POLK CSO | **Date** | 09/20/2023 | **Type** | Agency Inspection |
| **Inspector** | SOTO,ISRAEL, | **Time** | 08:55:48 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 09/20/2023 12:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

**Agency** POLK CSO **Date** 09/20/2023 **Type** Department Inspection
**Inspector** SOTO,ISRAEL, **Time** 12:37:22 **Software** 8100.27

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202303K | **0.08 Lot #** | 202303L | **0.20 Lot #** | 202304C | **0.08 Gas Lot #** | 06723080A5 |
| **Expiration** | 03/29/2025 | **Expiration** | 03/29/2025 | **Expiration** | 04/05/2025 | **Expiration** | 04/05/2025 |

**Number of Simulators** 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.050 |
| 0.08 g/210L Test # 4 | 0.081 |
| 0.20 g/210L Test # 4 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |
| 0.08 g/210L Test # 5 | 0.081 |
| 0.20 g/210L Test # 5 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.082 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.082 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0006 |
| 0.20 g/210L Test (Std Dev) | 0.0014 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 09/20/2023 12:46

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

**Agency** HIGHLANDS COUNTY SO **Date** 01/18/2023 **Type** Agency Inspection
**Inspector** GLASCO,GREGORY,S **Time** 19:59:58 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/20/2023 23:24

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

**Agency** HIGHLANDS COUNTY SO **Date** 02/24/2023 **Type** Agency Inspection
**Inspector** GLASCO,GREGORY,S **Time** 20:59:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/24/2023 21:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

**Agency** HIGHLANDS COUNTY SO **Date** 03/23/2023 **Type** Agency Inspection
**Inspector** GLASCO,GREGORY,S **Time** 22:56:31 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/28/2023 02:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

**Agency** HIGHLANDS COUNTY SO **Date** 04/12/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 11:46:24 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.08 g/210L Test # (Repeat) 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.08 g/210L Test # (Repeat) 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.08 g/210L Test # (Repeat) 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.08 g/210L Test # (Repeat) 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.050 |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

| | |
|---|---|
| 0.08 g/210L Test # 5 | INT |
| 0.08 g/210L Test # (Repeat) 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # (Repeat) 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # (Repeat) 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # (Repeat) 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # (Repeat) 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # (Repeat) 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) (Repeat) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

08: Interferent Detect. UNKNOWN. AIRED OUT WITH FAN. .

**Data Download Date/Time** 04/12/2023 11:52

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

**Agency** HIGHLANDS COUNTY SO **Date** 05/30/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 08:34:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/30/2023 08:40

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

**Agency** HIGHLANDS COUNTY SO **Date** 06/26/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 07:50:41 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/26/2023 07:56

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

**Agency** HIGHLANDS COUNTY SO **Date** 07/10/2023 **Type** Agency Inspection
**Inspector** GLASCO,GREGORY,S **Time** 06:12:15 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.075 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/10/2023 06:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

**Agency** HIGHLANDS COUNTY SO **Date** 08/28/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 09:02:58 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/28/2023 09:10

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

**Agency** HIGHLANDS COUNTY SO **Date** 09/22/2023 **Type** Agency Inspection
**Inspector** GLASCO,GREGORY,S **Time** 07:43:12 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/22/2023 07:51

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

**Agency** HIGHLANDS COUNTY SO **Date** 10/27/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 09:27:27 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/27/2023 09:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

**Agency** HIGHLANDS COUNTY SO **Date** 11/29/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 08:18:48 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/05/2023 20:03

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | HIGHLANDS COUNTY SO | **Date** | 12/18/2023 | **Type** | Agency Inspection |
| **Inspector** | ABELN,SHON,W | **Time** | 08:37:16 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A: Range Exceeded.

**Data Download Date/Time** 12/19/2023 04:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

**Agency** HIGHLANDS COUNTY SO **Date** 12/18/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 09:28:10 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/19/2023 04:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

**Agency** POLK CSO **Date** 01/09/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 15:35:55 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 00521080A2
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 02/05/2023

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.081 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 01/09/2023 15:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 01/18/2023 **Type** Agency Inspection
**Inspector** GLASCO,GREGORY,S **Time** 19:18:15 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/18/2023 19:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 02/24/2023 **Type** Agency Inspection
**Inspector** GLASCO,GREGORY,S **Time** 20:23:54 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/24/2023 20:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 03/23/2023 **Type** Agency Inspection
**Inspector** GLASCO,GREGORY,S **Time** 22:19:20 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient Fail.

**Data Download Date/Time** 03/24/2023 01:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 04/21/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 07:51:33 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/25/2023 14:38

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 05/30/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 07:45:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/13/2023 13:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 06/14/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 11:25:39 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0008 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.052 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.053 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.053 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.053 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.054 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.055 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.055 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0023 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

**Remarks:**

**Data Download Date/Time** 06/14/2023 11:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 06/26/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 08:24:43 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/26/2023 08:34

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 07/10/2023 **Type** Agency Inspection
**Inspector** GLASCO,GREGORY,S **Time** 06:49:16 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/10/2023 06:59

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 08/28/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 08:24:15 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/28/2023 08:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 09/22/2023 **Type** Agency Inspection
**Inspector** GLASCO,GREGORY,S **Time** 07:10:57 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06023080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/22/2023 07:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 10/27/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 08:47:03 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06023080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/27/2023 08:57

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 11/29/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 07:43:17 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06023080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/29/2023 07:55

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

**Agency** HIGHLANDS COUNTY SO **Date** 12/18/2023 **Type** Agency Inspection
**Inspector** ABELN,SHON,W **Time** 08:23:27 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06023080A2
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/18/2023 08:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY S.O. **Date** 01/25/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 22:25:30 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22521080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/25/2023 22:42

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY S.O. **Date** 02/09/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 18:37:41 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 225108A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: RFI Detect.

**Data Download Date/Time** 02/09/2023 19:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY S.O. **Date** 03/30/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 09:44:13 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/30/2023 10:04

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY S.O. **Date** 04/28/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 10:46:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/28/2023 11:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY S.O. **Date** 05/02/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 21:03:01 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/02/2023 21:37

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY S.O. **Date** 06/28/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 15:50:22 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2023 16:05

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY S.O. **Date** 07/28/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 19:12:12 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/28/2023 19:21

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY S.O. **Date** 08/31/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 11:02:37 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22521080A1
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/31/2023 11:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY SO **Date** 09/19/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:05:20 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 09/19/2023 12:14

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY SO **Date** 09/28/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 22:04:54 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 143230080A
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/28/2023 22:18

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY SO **Date** 10/30/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 20:18:57 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 14323080A2
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/30/2023 20:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY SO **Date** 11/28/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 20:24:38 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 13223080A2
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/28/2023 20:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

**Agency** DESOTO COUNTY SO **Date** 12/07/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 21:04:28 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 14323080A2
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/07/2023 21:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY S.O. **Date** 01/25/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 22:32:44 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22521080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/25/2023 22:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY S.O. **Date** 02/09/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 19:22:16 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22521080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/09/2023 19:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY S.O. **Date** 03/30/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 09:55:29 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22521080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/30/2023 10:06

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY S.O. **Date** 04/28/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 10:53:33 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22521080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/28/2023 11:25

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY S.O. **Date** 05/02/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 21:07:04 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22521080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/02/2023 21:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY S.O. **Date** 06/28/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 15:56:23 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22521080A1
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2023 16:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY S.O. **Date** 07/13/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:44:07 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 07/13/2023 13:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY S.O. **Date** 07/28/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 19:12:09 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22521080A1
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/28/2023 19:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY SO **Date** 08/10/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:44:07 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 08/10/2023 12:51

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY SO **Date** 08/31/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 11:02:04 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 19021080A2
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 08/31/2023 11:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY SO **Date** 09/28/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 21:57:43 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 14323080A2
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/28/2023 22:16

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY SO **Date** 10/30/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 20:16:58 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 14323080A2
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/30/2023 20:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY SO **Date** 11/28/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 20:23:50 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 14323080A2
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/28/2023 20:35

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

**Agency** DESOTO COUNTY SO **Date** 12/07/2023 **Type** Agency Inspection
**Inspector** BELLFLOWER,MICHAEL,D **Time** 21:00:27 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 14323080A2
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/07/2023 21:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 01/25/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEXANDER, **Time** 21:13:14 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 212654
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.191 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.045 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.191 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/27/2023 16:39

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEXANDER, **Time** 12:30:08 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
**Expiration** 12/07/2023 **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.044 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.045 |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # 1 | INT |
| Alcohol Free Subject Test: 0.000 | Yes | Alcohol Free Test # 2 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.05 g/210L Test # 2 | 0.045 |
| Interferent Detect Test: Interferent Detect | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.045 |
| Diagnostic Check (Post-Inspection): OK | No | Interferent Detect Test # 2 | INT |
| Alcohol Free Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # 3 | 0.045 |
| 0.08 g/210L Test: OK | No | 0.05 g/210L Test # (Repeat) 3 | 0.045 |
| 0.20 g/210L Test: OK | No | Interferent Detect Test # 3 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

SOULTION TOO COLD STILL. Non-compliance: .

**Data Download Date/Time** 03/03/2023 13:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 02/28/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEX, **Time** 13:24:25 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 212654
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.189 |
| 0.20 g/210L Test # (Repeat) 1 | 0.190 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.045 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.20 g/210L Test # 2 | 0.190 |
| 0.20 g/210L Test # (Repeat) 2 | 0.190 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.045 |
| 0.08 g/210L Test # 3 | 0.075 |
| 0.20 g/210L Test # 3 | 0.190 |
| 0.20 g/210L Test # (Repeat) 3 | 0.190 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance COLD SOLUTIN.

**Data Download Date/Time** 03/03/2023 13:22

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 03/29/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEXANDER, **Time** 13:49:30 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 211543
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/31/2023 12:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 04/27/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEXANDER, **Time** 21:40:30 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 212654
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/28/2023 14:09

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 05/31/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 19:40:10 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 212654
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.08 g/210L Test # (Repeat) 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # (Repeat) 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # (Repeat) 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: RFI Detect.

**Data Download Date/Time** 06/01/2023 16:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 06/28/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 20:27:28 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 212654
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/28/2023 22:02

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 07/26/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 20:54:53 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 212654
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/05/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/26/2023 21:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 08/23/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 21:57:39 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 221783
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/23/2023 22:07

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 09/18/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 11:46:15 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 221783
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/18/2023 13:43

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 10/30/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 13:15:05 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 221783
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/22/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.20 g/210L Test # 2 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/01/2023 15:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 11/01/2023 **Type** Department Inspection
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 15:05:06 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.049 |
| 0.08 g/210L Test # 4 | 0.079 |
| 0.20 g/210L Test # 4 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.079 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

**Remarks:**

**Data Download Date/Time** 11/01/2023 15:17

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 11/02/2023 **Type** Agency Inspection
**Inspector** ELLIS,ALEXANDER,M **Time** 18:47:12 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 220703
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.072 |
| 0.08 g/210L Test # (Repeat) 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.191 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.074 |
| 0.08 g/210L Test # (Repeat) 2 | 0.075 |
| 0.20 g/210L Test # 2 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.074 |
| 0.08 g/210L Test # (Repeat) 3 | 0.075 |
| 0.20 g/210L Test # 3 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 11/14/2023 19:00

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

**Agency** ST JOHNS COUNTY SO **Date** 12/06/2023 **Type** Agency Inspection
**Inspector** GREEN,JASON,A **Time** 00:12:30 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 220703
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/03/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/06/2023 00:36

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

**Agency** MIAMI PD **Date** 01/18/2023 **Type** Agency Inspection
**Inspector** PASTOR,JOSE,F **Time** 12:12:11 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1506449
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.075 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.045 |
| | | 0.08 g/210L Test # 3 | 0.075 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/18/2023 12:44

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

**Agency** MIAMI PD **Date** 02/16/2023 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 14:04:19 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1506449
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/16/2023 14:08

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 06/30/2023 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 10:44:36 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

**Data Download Date/Time** 06/30/2023 14:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

**Agency** MIAMI PD **Date** 06/30/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 14:14:32 **Software** 8100.27

**0.05 Lot #** 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
**Expiration** 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.047 |
| 0.08 g/210L Test # 4 | 0.077 |
| 0.20 g/210L Test # 4 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.077 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.077 |
| 0.20 g/210L Test # 5 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.077 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.047 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.077 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.077 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.077 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.047 |
| 0.08 g/210L Test # 8 | 0.077 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.077 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.077 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.077 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.077 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.077 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0007 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

**Data Download Date/Time** 06/30/2023 14:20

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

**Agency** MIAMI PD **Date** 12/19/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 12:46:21 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 01923080A3
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 02/05/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.082 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.049 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.082 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.077 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.082 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0007 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 12/19/2023 12:58

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 01/18/2023 **Type** Agency Inspection
**Inspector** PASTOR,JOSE,F **Time** 12:17:03 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1506449
**Expiration** 12/07/2023 **Expiration** 08/12/2023 **Expiration** 06/22/2023 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.042 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.045 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.195 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | INT |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.046 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

05: POWER SHUT OFF TO SIM.

**Data Download Date/Time** 01/18/2023 12:55

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 02/16/2023 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 14:17:19 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202112A **0.20 Lot #** 202106B **0.08 Gas Lot #** 1506449
**Expiration** 12/07/2023 **Expiration** 10/07/2023 **Expiration** 06/22/2023 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/16/2023 14:26

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 03/14/2023 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 16:49:37 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1506449
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.045 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/14/2023 17:01

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 04/13/2023 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 14:08:02 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1506449
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.044 |
| 0.05 g/210L Test # (Repeat) 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.044 |
| 0.05 g/210L Test # (Repeat) 2 | 0.045 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.044 |
| 0.05 g/210L Test # (Repeat) 3 | 0.045 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05:RECONNECTED SIMULATOR

**Data Download Date/Time** 04/13/2023 14:41

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 05/29/2023 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 13:08:34 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1506449
**Expiration** 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.045 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.075 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/31/2023 11:13

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 06/19/2023 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 13:14:19 **Software** 8100.27

**0.05 Lot #** 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1506449
**Expiration** 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/19/2023 13:29

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 07/31/2023 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 13:17:08 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1506449
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/31/2023 13:31

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 08/18/2023 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 16:46:15 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1506449
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/18/2023 16:52

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 09/01/2023 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 14:09:36 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1506449
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/01/2023 14:15

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 10/27/2023 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 12:04:24 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1506449
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/27/2023 12:11

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 11/03/2023 **Type** Department Inspection
**Inspector** GUTSCHOW,TAYLOR,D **Time** 13:00:56 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG223802
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 08/26/2024

**Number of Simulators** 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | Yes |
| Barometric Pressure Sensor Check: OK | Yes |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |
| Alcohol Free Test # 4 | 0.000 |
| 0.05 g/210L Test # 4 | 0.048 |
| 0.08 g/210L Test # 4 | 0.078 |
| 0.20 g/210L Test # 4 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| Interferent Detect Test # 4 | INT |
| Alcohol Free Test # 5 | 0.000 |
| 0.05 g/210L Test # 5 | 0.048 |
| 0.08 g/210L Test # 5 | 0.078 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| Interferent Detect Test # 5 | INT |

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

| | |
|---|---|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.078 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.048 |
| 0.08 g/210L Test # 8 | 0.078 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.048 |
| 0.08 g/210L Test # 9 | 0.078 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.078 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0000 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

**Remarks:**

**Data Download Date/Time** 11/03/2023 13:12

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 11/27/2023 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 09:42:26 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1506449
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 11/27/2023 09:50

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

**Agency** MIAMI PD **Date** 12/18/2023 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 14:23:43 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1506449
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/10/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 12/18/2023 14:33

**Printed: 24 January 2024**

COBRA ©1997-2014 CMI, Inc.