

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

Agency: OKALOOSA COUNTY SO

Date/Time: 02/27/2024 11:28:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/27/2024 13:11

Air Blank	0.000	11:26
Control Test	0.197	11:27
Air Blank	0.000	11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

Agency: FL HIGHWAY PATROL

Date/Time: 02/06/2024 13:03:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/06/2024 15:33

Air Blank	0.000	13:01
Control Test	0.051	13:02
Air Blank	0.000	13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 02/17/2024 15:07:56

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 02/29/2024 12:30

Air Blank	0.000	15:07
ITP Check	0.081	15:07
Air Blank	0.000	15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 02/28/2024 12:38:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2024 13:39

Air Blank	0.000	12:37
Control Test	0.000	12:38
Air Blank	0.000	12:38

Agency: LEESBURG P.D.

Date/Time: 02/28/2024 12:41:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2024 13:39

Air Blank	0.000	12:39
Control Test	0.049	12:40
Air Blank	0.000	12:41

Agency: LEESBURG P.D.

Date/Time: 02/28/2024 12:43:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2024 13:39

Air Blank	0.000	12:42
Control Test	0.079	12:42
Air Blank	0.000	12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

Agency: LEESBURG P.D.

Date/Time: 02/28/2024 12:46:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/28/2024 13:39

Air Blank	0.000	12:45
Control Test	0.198	12:46
Air Blank	0.000	12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency: MARTIN COUNTY SO

Date/Time: 02/27/2024 15:46:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/27/2024 16:51

Air Blank	0.000	15:44
Control Test	0.200	15:45
Air Blank	0.000	15:46

Agency: MARTIN COUNTY SO

Date/Time: 02/27/2024 15:48:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/27/2024 16:51

Air Blank	0.000	15:47
Control Test	0.078	15:48
Air Blank	0.000	15:48

Agency: MARTIN COUNTY SO

Date/Time: 02/27/2024 15:50:46

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/27/2024 16:51

Air Blank	0.000	15:49
Control Test	0.048	15:50
Air Blank	0.000	15:50

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

Agency: MARTIN COUNTY SO

Date/Time: 02/27/2024 15:52:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/27/2024 16:51

Air Blank	0.000	15:51
Control Test	0.000	15:52
Air Blank	0.000	15:52

Agency: MARTIN COUNTY SO

Date/Time: 02/27/2024 15:54:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/27/2024 16:51

Air Blank	0.000	15:53
Control Test	0.000	15:54
Air Blank	0.000	15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency: CALHOUN COUNTY SO

Date/Time: 02/05/2024 12:17:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/05/2024 14:47

Air Blank	0.000	12:16
Control Test	0.076	12:17
Air Blank	0.000	12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

Agency: MIAMI DADE PD

Date/Time: 01/04/2024 01:58:12

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/04/2024 04:33

Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 01/25/2024 20:28:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2024 22:59

Air Blank	0.000	20:27
Control Test	0.000	20:28
Air Blank	0.000	20:28

Agency: PORT ST JOE PD

Date/Time: 01/25/2024 20:31:33

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/25/2024 22:59

Air Blank	0.000	20:30
Control Test	0.081	20:31
Air Blank	0.000	20:31

Agency: PORT ST JOE PD

Date/Time: 01/25/2024 20:49:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2024 23:08

Air Blank	0.000	20:48
Control Test	0.079	20:49
Air Blank	0.000	20:49

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 01/25/2024 20:53:16

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/25/2024 23:08

Air Blank	0.000	20:52
Control Test	0.199	20:52
Air Blank	0.000	20:53

Agency: PORT ST JOE PD

Date/Time: 01/25/2024 20:54:49

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/25/2024 23:08

Air Blank	0.000	20:54
Control Test	0.080	20:54
Air Blank	0.000	20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

Agency: BREVARD COUNTY SO

Date/Time: 01/04/2024 11:52:23

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 01/11/2024 11:10

Air Blank	0.000	11:51
ITP Check	0.078	11:51
Air Blank	0.000	11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

Agency: HENDRY COUNTY SO

Date/Time: 01/25/2024 21:41:54

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/26/2024 13:26

Air Blank	0.000	21:41
Control Test	0.073	21:41
Air Blank	0.000	21:41

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

Agency: WAKULLA COUNTY SO

Date/Time: 02/19/2024 10:51:44

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/19/2024 12:43

Air Blank	0.000	10:50
Control Test	0.000	10:51
Air Blank	0.000	10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

Agency: JACKSONVILLE SO

Date/Time: 02/16/2024 14:22:18

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/23/2024 10:45

Air Blank	0.000	14:21
Control Test	0.076	14:21
Air Blank	0.000	14:22

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency: ZEPHYRHILLS PD

Date/Time: 01/29/2024 10:11:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/29/2024 11:18

Air Blank	0.000	10:10
Control Test	0.048	10:10
Air Blank	0.000	10:11

Agency: ZEPHYRHILLS PD

Date/Time: 01/29/2024 10:13:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/29/2024 11:18

Air Blank	0.000	10:12
Control Test	0.078	10:12
Air Blank	0.000	10:13

Agency: ZEPHYRHILLS PD

Date/Time: 01/29/2024 10:15:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/29/2024 11:18

Air Blank	0.000	10:13
Control Test	0.200	10:14
Air Blank	0.000	10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency: ZEPHYRHILLS PD

Date/Time: 01/29/2024 10:16:41

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/29/2024 11:18

Air Blank	0.000	10:15
Control Test	0.078	10:16
Air Blank	0.000	10:16

Agency: ZEPHYRHILLS PD

Date/Time: 02/26/2024 08:51:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/26/2024 09:47

Air Blank	0.000	08:49
Control Test	0.048	08:50
Air Blank	0.000	08:51

Agency: ZEPHYRHILLS PD

Date/Time: 02/26/2024 08:53:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/26/2024 09:47

Air Blank	0.000	08:51
Control Test	0.077	08:52
Air Blank	0.000	08:53

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency: ZEPHYRHILLS PD

Date/Time: 02/26/2024 08:54:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/26/2024 09:47

Air Blank	0.000	08:53
Control Test	0.200	08:54
Air Blank	0.000	08:54

Agency: ZEPHYRHILLS PD

Date/Time: 02/26/2024 08:56:28

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/26/2024 09:47

Air Blank	0.000	08:55
Control Test	0.079	08:55
Air Blank	0.000	08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

Agency: ST PETERSBURG PD

Date/Time: 02/14/2024 00:46:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2024 09:58

Air Blank	0.000	00:44
Control Test	0.000	00:45
Air Blank	0.000	00:46

Agency: ST PETERSBURG PD

Date/Time: 02/14/2024 00:47:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2024 09:58

Air Blank	0.000	00:46
Control Test	0.079	00:47
Air Blank	0.000	00:47

Agency: ST PETERSBURG PD

Date/Time: 02/14/2024 00:49:53

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2024 09:58

Air Blank	0.000	00:48
Control Test	0.000	00:49
Air Blank	0.000	00:49

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 02/08/2024 08:51:12

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/28/2024 09:51

Air Blank	0.000	08:49
Control Test	0.081	08:49
Air Blank		08:50
Air Blank		08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 01/17/2024 09:14:54

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/17/2024 10:35

Air Blank	0.000	09:14
Control Test	0.081	09:14
Air Blank	0.000	09:14

Agency: POLK CSO

Date/Time: 01/17/2024 09:17:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 10:35

Air Blank	0.000	09:15
Control Test	0.000	09:16
Air Blank	0.000	09:17

Agency: POLK CSO

Date/Time: 01/17/2024 09:18:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 10:35

Air Blank	0.000	09:17
Control Test	0.048	09:18
Air Blank	0.000	09:18

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK CSO

Date/Time: 01/17/2024 09:21:05

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 10:35

Air Blank	0.000	09:19
Control Test	0.079	09:20
Air Blank	0.000	09:21

Agency: POLK CSO

Date/Time: 01/17/2024 09:23:24

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 10:35

Air Blank	0.000	09:22
Control Test	0.193	09:22
Air Blank	0.000	09:23

Agency: POLK COUNTY SO

Date/Time: 02/13/2024 09:08:15

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/13/2024 10:49

Air Blank	0.000	09:07
Control Test	0.082	09:07
Air Blank	0.000	09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 02/13/2024 09:10:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2024 10:49

Air Blank	0.000	09:08
Control Test	0.048	09:09
Air Blank	0.000	09:10

Agency: POLK COUNTY SO

Date/Time: 02/13/2024 09:12:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2024 10:49

Air Blank	0.000	09:10
Control Test	0.000	09:11
Air Blank	0.000	09:12

Agency: POLK COUNTY SO

Date/Time: 02/13/2024 09:13:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2024 10:50

Air Blank	0.000	09:12
Control Test	0.080	09:13
Air Blank	0.000	09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 02/13/2024 09:17:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2024 10:50

Air Blank	0.000	09:16
Control Test	0.192	09:16
Air Blank	0.000	09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

Agency: FL HIGHWAY PATROL

Date/Time: 02/06/2024 14:00:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/06/2024 15:27

Air Blank	0.000	13:59
Control Test	0.196	14:00
Air Blank	0.000	14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency: FL HIGHWAY PATROL

Date/Time: 02/06/2024 14:05:09

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/06/2024 17:22

Air Blank	0.000	14:03
Control Test	0.074	14:04
Air Blank	0.000	14:05

Agency: FL HIGHWAY PATROL

Date/Time: 02/06/2024 14:16:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/06/2024 17:22

Air Blank	0.000	14:15
Control Test	0.075	14:15
Air Blank	0.000	14:16

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 01/17/2024 09:14:50

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 10:42

Air Blank	0.000	09:13
Control Test	0.000	09:14
Air Blank	0.000	09:14

Agency: POLK COUNTY SO

Date/Time: 01/17/2024 09:16:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 10:42

Air Blank	0.000	09:15
Control Test	0.049	09:16
Air Blank	0.000	09:16

Agency: POLK COUNTY SO

Date/Time: 01/17/2024 09:19:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 10:42

Air Blank	0.000	09:17
Control Test	0.082	09:18
Air Blank	0.000	09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 01/17/2024 09:21:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 10:42

Air Blank	0.000	09:19
Control Test	0.198	09:20
Air Blank	0.000	09:21

Agency: POLK COUNTY SO

Date/Time: 01/17/2024 09:22:36

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/17/2024 10:42

Air Blank	0.000	09:21
Control Test	0.081	09:22
Air Blank	0.000	09:22

Agency: POLK COUNTY SO

Date/Time: 02/13/2024 09:07:37

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/13/2024 10:34

Air Blank	0.000	09:06
Control Test	0.081	09:07
Air Blank	0.000	09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 02/13/2024 09:09:39

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2024 10:34

Air Blank	0.000	09:08
Control Test	0.000	09:09
Air Blank	0.000	09:09

Agency: POLK COUNTY SO

Date/Time: 02/13/2024 09:11:30

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2024 10:34

Air Blank	0.000	09:10
Control Test	0.081	09:10
Air Blank	0.000	09:11

Agency: POLK COUNTY SO

Date/Time: 02/13/2024 09:13:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2024 10:34

Air Blank	0.000	09:12
Control Test	0.048	09:13
Air Blank	0.000	09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 02/13/2024 09:15:33

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2024 10:35

Air Blank	0.000	09:14
Control Test	0.195	09:15
Air Blank	0.000	09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 02/13/2024 14:18:07

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2024 15:20

Air Blank	0.000	14:16
Control Test	0.051	14:17
Air Blank	0.000	14:18

Agency: GROVELAND PD

Date/Time: 02/13/2024 14:20:18

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2024 15:20

Air Blank	0.000	14:19
Control Test	0.080	14:19
Air Blank	0.000	14:20

Agency: GROVELAND PD

Date/Time: 02/13/2024 14:22:19

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/13/2024 15:20

Air Blank	0.000	14:21
Control Test	0.200	14:21
Air Blank	0.000	14:22

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 01/24/2024 20:41:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/24/2024 22:02

Air Blank	0.000	20:39
Control Test	0.000	20:40
Air Blank	0.000	20:41

Agency: MARTIN COUNTY SO

Date/Time: 01/24/2024 20:46:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/24/2024 22:02

Air Blank	0.000	20:45
Control Test	0.000	20:46
Air Blank	0.000	20:46

Agency: MARTIN COUNTY SO

Date/Time: 01/24/2024 20:48:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/24/2024 22:02

Air Blank	0.000	20:47
Control Test	0.050	20:48
Air Blank	0.000	20:48

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency: MARTIN COUNTY SO

Date/Time: 01/24/2024 20:50:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/24/2024 22:02

Air Blank	0.000	20:49
Control Test	0.079	20:50
Air Blank	0.000	20:50

Agency: MARTIN COUNTY SO

Date/Time: 01/24/2024 20:52:59

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/24/2024 22:02

Air Blank	0.000	20:51
Control Test	0.200	20:52
Air Blank	0.000	20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

Agency: PANAMA CITY BEACH
PD

Date/Time: 02/05/2024 17:40:49

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/05/2024 19:26

Air Blank	0.000	17:40
Control Test	0.000	17:40
Air Blank	0.000	17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency: MONROE COUNTY SO

Date/Time: 02/05/2024 11:53:57

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/05/2024 14:08

Air Blank	0.000	11:53
Control Test	0.081	11:53
Air Blank	0.000	11:53

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

Agency: ORANGE COUNTY S.O.

Date/Time: 01/01/2024 18:00:45

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/15/2024 10:26

Air Blank	0.000	17:59
Control Test	0.077	18:00
Air Blank	0.000	18:00

Agency: ORANGE COUNTY S.O.

Date/Time: 01/01/2024 18:02:05

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/15/2024 10:26

Air Blank	0.000	18:01
Control Test	0.077	18:01
Air Blank	0.000	18:02

Agency: ORANGE COUNTY SO

Date/Time: 01/29/2024 11:03:20

Standard Type: ITP

Event Source: Local

Data Download Date/Time: 02/07/2024 11:32

Air Blank	0.000	11:02
ITP Check	0.080	11:02
Air Blank	0.000	11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency: INDIAN HARBOUR PD

Date/Time: 01/19/2024 19:11:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/19/2024 20:16

Air Blank	0.000	19:09
Control Test	0.000	19:10
Air Blank	0.000	19:11

Agency: INDIAN HARBOUR PD

Date/Time: 02/24/2024 07:23:02

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/24/2024 08:42

Air Blank	0.000	07:21
Control Test	0.000	07:22
Air Blank	0.000	07:23

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 02/01/2024 02:04:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/01/2024 04:40

Air Blank	0.000	02:03
Control Test	0.000	02:04
Air Blank	0.000	02:04

Agency: MIAMI-DADE PD

Date/Time: 02/01/2024 02:06:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/01/2024 04:40

Air Blank	0.000	02:05
Control Test	0.052	02:06
Air Blank	0.000	02:06

Agency: MIAMI-DADE PD

Date/Time: 02/01/2024 02:08:57

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/01/2024 04:40

Air Blank	0.000	02:07
Control Test	0.082	02:08
Air Blank	0.000	02:08

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 02/01/2024 02:11:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/01/2024 04:40

Air Blank	0.000	02:10
Control Test	0.202	02:10
Air Blank	0.000	02:11

Agency: MIAMI-DADE PD

Date/Time: 02/15/2024 00:21:52

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/01/2024 05:48

Air Blank	0.000	00:20
Control Test	0.007	00:21
Air Blank	0.000	00:21

Agency: MIAMI-DADE PD

Date/Time: 02/15/2024 00:24:04

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/01/2024 05:48

Air Blank	0.000	00:22
Control Test	0.052	00:23
Air Blank	0.000	00:24

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

Agency: MIAMI-DADE PD

Date/Time: 02/15/2024 00:26:08

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/01/2024 05:48

Air Blank	0.000	00:24
Control Test	0.081	00:25
Air Blank	0.000	00:26

Agency: MIAMI-DADE PD

Date/Time: 02/15/2024 00:28:36

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 03/01/2024 05:48

Air Blank	0.000	00:27
Control Test	0.201	00:28
Air Blank	0.000	00:28

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

Agency: BAY COUNTY SO

Date/Time: 01/17/2024 13:55:21

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/29/2024 12:48

Air Blank	0.000	13:54
Control Test	0.082	13:54
Air Blank	0.000	13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency: PINELLAS COUNTY SO

Date/Time: 01/04/2024 23:44:26

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/17/2024 13:46

Air Blank	0.000	23:43
Control Test	0.081	23:43
Air Blank	0.000	23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency: MARTIN COUNTY SO

Date/Time: 01/28/2024 06:50:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/27/2024 16:41

Air Blank	0.000	06:49
Control Test	0.000	06:50
Air Blank	0.000	06:50

Agency: MARTIN COUNTY SO

Date/Time: 01/28/2024 06:52:04

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/27/2024 16:41

Air Blank	0.000	06:51
Control Test	0.080	06:51
Air Blank	0.000	06:52

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 01/18/2024 11:16:32

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 12:08

Air Blank	0.000	11:15
Control Test	0.000	11:16
Air Blank	0.000	11:16

Agency: PINELLAS PARK P.D.

Date/Time: 01/18/2024 11:18:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 12:08

Air Blank	0.000	11:17
Control Test	0.000	11:17
Air Blank	0.000	11:18

Agency: PINELLAS PARK P.D.

Date/Time: 01/18/2024 11:20:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 12:08

Air Blank	0.000	11:19
Control Test	0.047	11:20
Air Blank	0.000	11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 01/18/2024 11:22:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 12:08

Air Blank	0.000	11:21
Control Test	0.076	11:21
Air Blank	0.000	11:22

Agency: PINELLAS PARK P.D.

Date/Time: 01/18/2024 11:24:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 12:08

Air Blank	0.000	11:23
Control Test	0.194	11:23
Air Blank	0.000	11:24

Agency: PINELLAS PARK P.D.

Date/Time: 02/20/2024 14:28:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2024 15:29

Air Blank	0.000	14:27
Control Test	0.000	14:28
Air Blank	0.000	14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 02/20/2024 14:30:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2024 15:29

Air Blank	0.000	14:29
Control Test	0.000	14:29
Air Blank	0.000	14:30

Agency: PINELLAS PARK P.D.

Date/Time: 02/20/2024 14:32:37

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2024 15:29

Air Blank	0.000	14:31
Control Test	0.048	14:32
Air Blank	0.000	14:32

Agency: PINELLAS PARK P.D.

Date/Time: 02/20/2024 14:34:20

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2024 15:29

Air Blank	0.000	14:33
Control Test	0.075	14:33
Air Blank	0.000	14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency: PINELLAS PARK P.D.

Date/Time: 02/20/2024 14:36:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2024 15:29

Air Blank	0.000	14:35
Control Test	0.000	14:35
Air Blank	0.000	14:36

Agency: PINELLAS PARK P.D.

Date/Time: 02/20/2024 14:38:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2024 15:29

Air Blank	0.000	14:37
Control Test	0.000	14:38
Air Blank	0.000	14:38

Agency: PINELLAS PARK P.D.

Date/Time: 02/20/2024 14:42:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/20/2024 15:29

Air Blank	0.000	14:41
Control Test	0.190	14:41
Air Blank	0.000	14:42

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY
SO

Date/Time: 01/08/2024 18:45:52

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/18/2024 22:15

Air Blank	0.000	18:45
Control Test	0.081	18:45
Air Blank	0.000	18:45

Agency: CHARLOTTE COUNTY
SO

Date/Time: 01/18/2024 20:58:03

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/18/2024 22:15

Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:58

Agency: CHARLOTTE COUNTY
SO

Date/Time: 01/18/2024 20:59:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 22:15

Air Blank	0.000	20:58
Control Test	0.000	20:59
Air Blank	0.000	20:59

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY
SO

Date/Time: 01/18/2024 21:02:01

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 22:15

Air Blank	0.000	21:00
Control Test	0.000	21:01
Air Blank	0.000	21:02

Agency: CHARLOTTE COUNTY
SO

Date/Time: 01/18/2024 21:04:26

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 22:15

Air Blank	0.000	21:03
Control Test	0.049	21:03
Air Blank	0.000	21:04

Agency: CHARLOTTE COUNTY
SO

Date/Time: 01/18/2024 21:06:48

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 22:15

Air Blank	0.000	21:05
Control Test	0.078	21:06
Air Blank	0.000	21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY
SO

Date/Time: 01/18/2024 21:09:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 22:15

Air Blank	0.000	21:08
Control Test	0.070	21:08
Air Blank	0.000	21:09

Agency: CHARLOTTE COUNTY
SO

Date/Time: 01/18/2024 21:23:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 22:15

Air Blank	0.000	21:22
Control Test	0.195	21:22
Air Blank	0.000	21:23

Agency: CHARLOTTE COUNTY
SO

Date/Time: 01/18/2024 21:24:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 22:15

Air Blank		21:23
Air Blank	0.000	21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency: CHARLOTTE COUNTY
SO

Date/Time: 01/18/2024 21:26:11

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 22:15

Air Blank	0.000	21:24
Control Test	0.198	21:25
Air Blank	0.000	21:26

Agency: CHARLOTTE COUNTY
SO

Date/Time: 01/18/2024 21:28:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/18/2024 22:15

Air Blank	0.000	21:26
Control Test	0.198	21:27
Air Blank	0.000	21:28

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency: ST PETERSBURG PD

Date/Time: 02/14/2024 00:46:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2024 10:02

Air Blank	0.000	00:44
Control Test	0.049	00:45
Air Blank	0.000	00:46

Agency: ST PETERSBURG PD

Date/Time: 02/14/2024 00:48:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2024 10:02

Air Blank	0.000	00:46
Control Test	0.196	00:47
Air Blank	0.000	00:48

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency: Ocala PD

Date/Time: 01/04/2024 14:21:31

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/11/2024 11:46

Air Blank	0.000	14:20
Control Test	0.079	14:20
Air Blank	0.000	14:21

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency: MIAMI DADE PD

Date/Time: 01/04/2024 01:54:06

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/04/2024 04:23

Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency: MIAMI DADE PD

Date/Time: 01/04/2024 01:58:20

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/04/2024 04:26

Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58

Agency: MIAMI DADE PD

Date/Time: 01/11/2024 05:44:51

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/01/2024 05:01

Air Blank	0.000	05:43
Control Test	0.079	05:44
Air Blank	0.000	05:44

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency: MIAMI DADE PD

Date/Time: 01/11/2024 02:13:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/11/2024 03:36

Air Blank	0.000	02:12
Control Test	0.000	02:13
Air Blank	0.000	02:13

Agency: MIAMI DADE PD

Date/Time: 01/11/2024 02:15:58

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/11/2024 03:36

Air Blank	0.000	02:14
Control Test	0.050	02:15
Air Blank	0.000	02:15

Agency: MIAMI DADE PD

Date/Time: 01/11/2024 02:18:12

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/11/2024 03:36

Air Blank	0.000	02:16
Control Test	0.080	02:17
Air Blank	0.000	02:18

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency: MIAMI DADE PD

Date/Time: 01/11/2024 02:20:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/11/2024 03:36

Air Blank	0.000	02:19
Control Test	0.205	02:19
Air Blank	0.000	02:20

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

Agency: MANATEE COUNTY SO

Date/Time: 02/12/2024 10:30:07

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/12/2024 13:20

Air Blank		10:29
Air Blank	0.000	10:30

Agency: MANATEE COUNTY SO

Date/Time: 02/12/2024 10:31:30

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/12/2024 13:20

Air Blank	0.000	10:30
Control Test	0.055	10:30
Air Blank	0.000	10:31

Agency: MANATEE COUNTY SO

Date/Time: 02/12/2024 10:38:01

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/12/2024 13:20

Air Blank		10:37
Air Blank	0.000	10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

Agency: MANATEE COUNTY SO

Date/Time: 02/12/2024 10:39:01

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/12/2024 13:20

Air Blank		10:38
Air Blank	0.000	10:39

Agency: MANATEE COUNTY SO

Date/Time: 02/12/2024 10:41:17

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/12/2024 13:20

Air Blank	0.000	10:40
Control Test	0.000	10:40
Air Blank		10:40
Air Blank	0.000	10:41

Agency: MANATEE COUNTY SO

Date/Time: 02/12/2024 10:43:03

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/12/2024 13:20

Air Blank	0.000	10:41
Control Test	0.000	10:42
Air Blank		10:42
Air Blank	0.000	10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

Agency: MANATEE COUNTY SO

Date/Time: 02/12/2024 10:44:23

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/12/2024 13:20

Air Blank	0.000	10:43
Control Test	0.000	10:43
Air Blank	0.000	10:44

Agency: MANATEE COUNTY SO

Date/Time: 02/12/2024 10:45:38

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 02/12/2024 13:20

Air Blank	0.000	10:44
Control Test	0.000	10:45
Air Blank	0.000	10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

Agency: MIAMI-DADE PD

Date/Time: 01/04/2024 02:01:43

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/04/2024 04:36

Air Blank	0.000	02:00
Control Test	0.082	02:01
Air Blank	0.000	02:01

Agency: MIAMI-DADE PD

Date/Time: 01/04/2024 02:04:55

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2024 04:36

Air Blank	0.000	02:03
Control Test	0.000	02:04
Air Blank	0.000	02:04

Agency: MIAMI-DADE PD

Date/Time: 01/04/2024 02:06:56

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2024 04:36

Air Blank	0.000	02:05
Control Test	0.050	02:06
Air Blank	0.000	02:06

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

Agency: MIAMI-DADE PD

Date/Time: 01/04/2024 02:09:03

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2024 04:36

Air Blank	0.000	02:07
Control Test	0.058	02:08
Air Blank	0.000	02:09

Agency: MIAMI-DADE PD

Date/Time: 01/04/2024 02:11:35

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2024 04:36

Air Blank	0.000	02:10
Control Test	0.061	02:11
Air Blank	0.000	02:11

Agency: MIAMI-DADE PD

Date/Time: 01/04/2024 02:16:29

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2024 04:36

Air Blank	0.000	02:15
Control Test	0.198	02:15
Air Blank	0.000	02:16

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

Agency: MIAMI-DADE PD

Date/Time: 01/04/2024 02:29:54

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/04/2024 04:36

Air Blank	0.000	02:28
Control Test	0.080	02:29
Air Blank	0.000	02:29

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

Agency: MIAMI-DADE PD

Date/Time: 01/04/2024 02:00:08

Standard Type: Dry Gas

Event Source: Local

Data Download Date/Time: 01/04/2024 04:29

Air Blank	0.000	01:59
Control Test	0.083	01:59
Air Blank	0.000	02:00

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

Agency: GROVELAND PD

Date/Time: 01/08/2024 14:37:06

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/08/2024 15:47

Air Blank	0.000	14:35
Control Test	0.047	14:36
Air Blank	0.000	14:37

Agency: GROVELAND PD

Date/Time: 01/08/2024 14:39:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/08/2024 15:47

Air Blank	0.000	14:37
Control Test	0.077	14:38
Air Blank	0.000	14:39

Agency: GROVELAND PD

Date/Time: 01/08/2024 14:40:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/08/2024 15:47

Air Blank	0.000	14:39
Control Test	0.196	14:40
Air Blank	0.000	14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

Agency: GROVELAND PD

Date/Time: 02/14/2024 09:03:41

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2024 11:11

Air Blank	0.000	09:02
Control Test	0.048	09:03
Air Blank	0.000	09:03

Agency: GROVELAND PD

Date/Time: 02/14/2024 09:05:43

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2024 11:11

Air Blank	0.000	09:04
Control Test	0.078	09:05
Air Blank	0.000	09:05

Agency: GROVELAND PD

Date/Time: 02/14/2024 09:07:27

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/14/2024 11:11

Air Blank	0.000	09:06
Control Test	0.197	09:06
Air Blank	0.000	09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007441

Agency: FHP

Date/Time: 02/12/2024 06:40:15

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/12/2024 11:20

Air Blank	0.000	06:39
Control Test	0.000	06:39
Air Blank	0.000	06:40

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/17/2024 19:27:25

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 21:02

Air Blank	0.000	19:26
Control Test	0.195	19:26
Air Blank	0.000	19:27

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/17/2024 19:29:28

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 21:03

Air Blank	0.000	19:28
Control Test	0.047	19:28
Air Blank	0.000	19:29

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/17/2024 19:31:38

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 21:03

Air Blank	0.000	19:30
Control Test	0.077	19:31
Air Blank	0.000	19:31

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/17/2024 19:33:44

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 21:03

Air Blank	0.000	19:32
Control Test	0.000	19:33
Air Blank	0.000	19:33

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/26/2024 21:12:14

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/26/2024 22:34

Air Blank	0.000	21:11
Control Test	0.077	21:11
Air Blank	0.000	21:12

Florida Department of Law Enforcement

Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/17/2024 19:22:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 20:19

Air Blank	0.000	19:20
Control Test	0.198	19:21
Air Blank	0.000	19:22

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/17/2024 19:24:42

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 20:19

Air Blank	0.000	19:23
Control Test	0.048	19:24
Air Blank	0.000	19:24

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/17/2024 19:27:00

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 20:19

Air Blank	0.000	19:25
Control Test	0.077	19:26
Air Blank	0.000	19:27

Florida Department of Law Enforcement
Alcohol Testing Program

Control Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/17/2024 19:29:13

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 01/17/2024 20:19

Air Blank	0.000	19:27
Control Test	0.000	19:28
Air Blank	0.000	19:29

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/26/2024 21:08:10

Standard Type: Wet Bath

Event Source: Local

Data Download Date/Time: 02/26/2024 21:59

Air Blank	0.000	21:06
Control Test	0.078	21:07
Air Blank	0.000	21:08
