

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

|                                    |                        |                               |
|------------------------------------|------------------------|-------------------------------|
| <b>Agency</b> FDLE                 | <b>Date</b> 02/06/2024 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 10:02:37   | <b>Software</b> 8100.27       |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                       |
|-----------------------------|-----------------------|
| <b>Number of Simulators</b> | <b>Compliance</b> Yes |
|-----------------------------|-----------------------|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

Remarks:

**Data Download Date/Time** 02/06/2024 13:34

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

|                                    |                                          |                                   |                                  |
|------------------------------------|------------------------------------------|-----------------------------------|----------------------------------|
| <b>Agency</b> FDLE                 | <b>Date</b> 02/06/2024                   | <b>Type</b> Department Inspection |                                  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:24:20                     | <b>Software</b> 8100.27           |                                  |
| <b>0.05 Lot #</b> 202303K          | <b>0.08 Lot #</b> 202303L                | <b>0.20 Lot #</b> 202304C         | <b>0.08 Gas Lot #</b> 01923080A3 |
| <b>Expiration</b> 03/29/2025       | <b>Expiration</b> 03/29/2025             | <b>Expiration</b> 04/05/2025      | <b>Expiration</b> 02/05/2025     |
| <b>Number of Simulators</b> 5      | <b>Average Standard Deviation</b> 0.0008 | <b>Compliance</b> Yes             |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000206

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0011 |
| 0.08 g/210L Test (Std Dev)             | 0.0009 |
| 0.20 g/210L Test (Std Dev)             | 0.0011 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 02/06/2024 13:34

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

|                                 |                              |                               |                                 |
|---------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> FDLE              | <b>Date</b> 01/30/2024       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> BETHAM,KEITH,D | <b>Time</b> 22:55:40         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202303J       | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 402623830 |
| <b>Expiration</b> 03/28/2025    | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 12/28/2025    |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.063  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.076  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.064  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.064  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: SIM NOT TIGHTEN PROPERLY . 08: SIM NOT TIGHTEN PROPERLY

**Data Download Date/Time** 01/30/2024 23:27

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

|                                 |                              |                               |                                 |
|---------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> FDLE              | <b>Date</b> 02/24/2024       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> BETHAM,KEITH,D | <b>Time</b> 00:57:10         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202303J       | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 402623830 |
| <b>Expiration</b> 03/28/2025    | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 12/28/2025    |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/06/2024 12:59

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

|                                 |                                          |                                   |                                  |
|---------------------------------|------------------------------------------|-----------------------------------|----------------------------------|
| <b>Agency</b> FDLE              | <b>Date</b> 03/06/2024                   | <b>Type</b> Department Inspection |                                  |
| <b>Inspector</b> NICODEMO,PHIL, | <b>Time</b> 14:42:00                     | <b>Software</b> 8100.27           |                                  |
| <b>0.05 Lot #</b> 202303K       | <b>0.08 Lot #</b> 202303L                | <b>0.20 Lot #</b> 202304C         | <b>0.08 Gas Lot #</b> 06723080A5 |
| <b>Expiration</b> 03/29/2025    | <b>Expiration</b> 03/29/2025             | <b>Expiration</b> 04/05/2025      | <b>Expiration</b> 04/05/2025     |
| <b>Number of Simulators</b> 5   | <b>Average Standard Deviation</b> 0.0002 | <b>Compliance</b> Yes             |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 03/06/2024 14:52

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

**Agency** FDLE  
**Inspector** FARLEY,JOSEPH,E

**Date** 01/10/2024  
**Time** 20:27:13

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402623828  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 10/22/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # 1             | 0.050  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: RFI Detect. RADIO ON

**Data Download Date/Time** 01/25/2024 12:28

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

**Agency** FDLE  
**Inspector** FARLEY,JOSEPH,E

**Date** 01/19/2024  
**Time** 18:47:51

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402623828  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 10/22/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/25/2024 12:28

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

**Agency** FDLE  
**Inspector** FARLEY,JOSEPH,E

**Date** 02/01/2024  
**Time** 16:05:08

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402623828  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 10/22/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/13/2024 10:06

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

**Agency** FDLE  
**Inspector** FARLEY,JOSEPH,E

**Date** 03/03/2024  
**Time** 15:41:29

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402623828  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 10/22/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.000  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.049  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.049  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

05: Control Outside SIM NOT CONNECTED.

**Data Download Date/Time** 03/04/2024 14:51

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

|                                 |                                          |                                   |                                  |
|---------------------------------|------------------------------------------|-----------------------------------|----------------------------------|
| <b>Agency</b> FDLE              | <b>Date</b> 03/19/2024                   | <b>Type</b> Department Inspection |                                  |
| <b>Inspector</b> NICODEMO,PHIL, | <b>Time</b> 15:00:40                     | <b>Software</b> 8100.27           |                                  |
| <b>0.05 Lot #</b> 202303K       | <b>0.08 Lot #</b> 202303L                | <b>0.20 Lot #</b> 202304C         | <b>0.08 Gas Lot #</b> 06723080A5 |
| <b>Expiration</b> 03/29/2025    | <b>Expiration</b> 03/29/2025             | <b>Expiration</b> 04/05/2025      | <b>Expiration</b> 04/05/2025     |
| <b>Number of Simulators</b> 5   | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes             |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 03/19/2024 15:09

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

**Agency** FDLE **Date** 01/31/2024 **Type** Agency Inspection  
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 09:46:31 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

COMPLIANCE NOT DETERMINED, BYPASS INSPECTION TO OPERATE

**Data Download Date/Time** 02/14/2024 09:30

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

|                                 |                                          |                                   |                                  |
|---------------------------------|------------------------------------------|-----------------------------------|----------------------------------|
| <b>Agency</b> FDLE              | <b>Date</b> 02/14/2024                   | <b>Type</b> Department Inspection |                                  |
| <b>Inspector</b> NICODEMO,PHIL, | <b>Time</b> 11:08:43                     | <b>Software</b> 8100.27           |                                  |
| <b>0.05 Lot #</b> 202303K       | <b>0.08 Lot #</b> 202303L                | <b>0.20 Lot #</b> 202304C         | <b>0.08 Gas Lot #</b> 06723080A5 |
| <b>Expiration</b> 03/29/2025    | <b>Expiration</b> 03/29/2025             | <b>Expiration</b> 04/05/2025      | <b>Expiration</b> 04/05/2025     |
| <b>Number of Simulators</b> 5   | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes             |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.052  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.082  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 02/14/2024 11:19

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

|                                    |                        |                               |
|------------------------------------|------------------------|-------------------------------|
| <b>Agency</b> FDLE                 | <b>Date</b> 02/06/2024 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 09:58:19   | <b>Software</b> 8100.27       |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                       |
|-----------------------------|-----------------------|
| <b>Number of Simulators</b> | <b>Compliance</b> Yes |
|-----------------------------|-----------------------|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

**Data Download Date/Time** 02/06/2024 14:04

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

|                                    |                                          |                                   |                                  |
|------------------------------------|------------------------------------------|-----------------------------------|----------------------------------|
| <b>Agency</b> FDLE                 | <b>Date</b> 02/06/2024                   | <b>Type</b> Department Inspection |                                  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:53:17                     | <b>Software</b> 8100.27           |                                  |
| <b>0.05 Lot #</b> 202303K          | <b>0.08 Lot #</b> 202303L                | <b>0.20 Lot #</b> 202304C         | <b>0.08 Gas Lot #</b> 01923080A3 |
| <b>Expiration</b> 03/29/2025       | <b>Expiration</b> 03/29/2025             | <b>Expiration</b> 04/05/2025      | <b>Expiration</b> 02/05/2025     |
| <b>Number of Simulators</b> 5      | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes             |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 02/06/2024 14:04

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

|                                       |                              |                               |                                 |
|---------------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> FDLE                    | <b>Date</b> 01/10/2024       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> HOLLOWAY,DEVARIOUS,J | <b>Time</b> 11:06:36         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202303J             | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 402477284 |
| <b>Expiration</b> 03/28/2025          | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 06/27/2025    |
| <b>Number of Simulators</b> 5         | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A FRAIZER PRESENT

**Data Download Date/Time** 01/10/2024 11:21

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

|                                       |                              |                               |                                 |
|---------------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> FDLE                    | <b>Date</b> 02/07/2024       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> HOLLOWAY,DEVARIOUS,J | <b>Time</b> 11:13:13         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202303J             | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 402477284 |
| <b>Expiration</b> 03/28/2025          | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 06/27/2025    |
| <b>Number of Simulators</b> 5         | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A FRAIZER PRESENT

**Data Download Date/Time** 02/07/2024 11:23

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

|                                    |                                          |                                   |                                  |
|------------------------------------|------------------------------------------|-----------------------------------|----------------------------------|
| <b>Agency</b> FDLE                 | <b>Date</b> 02/13/2024                   | <b>Type</b> Department Inspection |                                  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 12:48:44                     | <b>Software</b> 8100.27           |                                  |
| <b>0.05 Lot #</b> 202303K          | <b>0.08 Lot #</b> 202303L                | <b>0.20 Lot #</b> 202304C         | <b>0.08 Gas Lot #</b> 01923080A3 |
| <b>Expiration</b> 03/29/2025       | <b>Expiration</b> 03/29/2025             | <b>Expiration</b> 04/05/2025      | <b>Expiration</b> 02/05/2025     |
| <b>Number of Simulators</b> 5      | <b>Average Standard Deviation</b> 0.0003 | <b>Compliance</b> Yes             |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 02/13/2024 12:58

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> ALACHUA COUNTY SO  | <b>Date</b> 01/25/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LEVY, JAYSON, I | <b>Time</b> 08:32:32         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 26422080A1 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 11/05/2024     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 01/25/2024 09:56



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> ALACHUA COUNTY SO  | <b>Date</b> 02/07/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LEVY, JAYSON, I | <b>Time</b> 08:29:49         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 26422080A1 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 11/05/2024     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/07/2024 11:20

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

|                                  |                                          |                                   |
|----------------------------------|------------------------------------------|-----------------------------------|
| <b>Agency</b> ALACHUA COUNTY SO  | <b>Date</b> 02/15/2024                   | <b>Type</b> Department Inspection |
| <b>Inspector</b> NICODEMO, PHIL, | <b>Time</b> 12:12:09                     | <b>Software</b> 8100.27           |
| <b>0.05 Lot #</b> 202303K        | <b>0.08 Lot #</b> 202303L                | <b>0.20 Lot #</b> 202304C         |
| <b>0.08 Gas Lot #</b> 06723080A5 |                                          |                                   |
| <b>Expiration</b> 03/29/2025     | <b>Expiration</b> 03/29/2025             | <b>Expiration</b> 04/05/2025      |
| <b>Expiration</b> 04/05/2025     |                                          |                                   |
| <b>Number of Simulators</b> 5    | <b>Average Standard Deviation</b> 0.0003 | <b>Compliance</b> Yes             |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.082  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.082  |
|                                             |        | Interferent Detect Test # 5      | INT    |

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.082  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.082  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.082  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.082  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 02/15/2024 12:21

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> ALACHUA COUNTY SO  | <b>Date</b> 02/23/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LEVY, JAYSON, I | <b>Time</b> 15:24:14         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 26422080A1 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 07/11/2023 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 11/05/2024     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: RFI Detect SMART WATCH TOO CLOSE TO INST.

**Data Download Date/Time** 02/23/2024 15:35

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> ALACHUA COUNTY SO  | <b>Date</b> 03/21/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LEVY, JAYSON, I | <b>Time</b> 07:40:29         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 26422080A1 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 11/05/2024     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 03/21/2024 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | OKALOOSA COUNTY SO | <b>Date</b>       | 01/26/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | NICHOLS,MICHAEL,J  | <b>Time</b>       | 17:26:33   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 402623828  |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 12/22/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 01/26/2024 18:40

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

|                             |                     |                   |            |                   |                   |                       |            |
|-----------------------------|---------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | OKALOOSA COUNTY SO  | <b>Date</b>       | 02/27/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SCHLAGER,JONATHAN,D | <b>Time</b>       | 12:01:12   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J             | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 402663828  |
| <b>Expiration</b>           | 03/28/2025          | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 12/22/2025 |
| <b>Number of Simulators</b> | 5                   | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 02/27/2024 13:11

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

|                             |                     |                   |            |                   |                   |                       |            |
|-----------------------------|---------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | OKALOOSA COUNTY SO  | <b>Date</b>       | 03/22/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SCHLAGER,JONATHAN,D | <b>Time</b>       | 12:03:59   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J             | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 402623828  |
| <b>Expiration</b>           | 03/28/2025          | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 12/22/2025 |
| <b>Number of Simulators</b> | 5                   | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # (Repeat) 1    | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

05: RFI DetecT RADIO IN ROOM.

**Data Download Date/Time** 03/22/2024 13:14

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | VALPARAISO PD      | <b>Date</b>       | 01/15/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | DUMBAUGH,CHARLES,C | <b>Time</b>       | 23:36:32   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402477284  |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 06/27/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/16/2024 00:43

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | VALPARAISO PD      | <b>Date</b>       | 02/11/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | DUMBAUGH,CHARLES,C | <b>Time</b>       | 22:20:38   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402477284  |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 06/27/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/11/2024 23:48

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

**Agency** VALPARAISO PD  
**Inspector** LUNGER,JOSEPH,W

**Date** 03/18/2024  
**Time** 22:48:04

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 402477284  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 06/27/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 03/19/2024 00:00

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | ESCAMBIA COUNTY SO | <b>Date</b>       | 01/16/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SHELLEY,SAMUEL,L   | <b>Time</b>       | 11:55:07   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402690226  |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 03/09/2026 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

Data Download Date/Time 01/16/2024 13:04

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

**Agency** ESCAMBIA COUNTY SO  
**Inspector** SHELLEY,SAMUEL,L

**Date** 02/14/2024  
**Time** 10:45:56

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 402690226  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 03/09/2026

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

**Data Download Date/Time** 02/14/2024 11:55

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | ESCAMBIA COUNTY SO | <b>Date</b>       | 03/12/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SHELLEY,SAMUEL,L   | <b>Time</b>       | 11:02:04   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402690226  |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 03/09/2026 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

Data Download Date/Time 03/12/2024 12:14

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

**Agency** ESCAMBIA COUNTY SO  
**Inspector** SHELLEY,SAMUEL,L

**Date** 01/16/2024  
**Time** 12:35:34

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 402690226  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 03/09/2026

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

**Data Download Date/Time** 01/16/2024 13:43

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

**Agency** ESCAMBIA COUNTY SO  
**Inspector** SHELLEY,SAMUEL,L

**Date** 02/14/2024  
**Time** 09:51:34

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 402690226  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 03/09/2026

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

**Data Download Date/Time** 02/14/2024 11:01

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

|                             |                    |                                   |            |                   |                       |                       |            |
|-----------------------------|--------------------|-----------------------------------|------------|-------------------|-----------------------|-----------------------|------------|
| <b>Agency</b>               | ESCAMBIA COUNTY SO | <b>Date</b>                       | 02/16/2024 | <b>Type</b>       | Department Inspection |                       |            |
| <b>Inspector</b>            | NICODEMO,PHIL,     | <b>Time</b>                       | 11:39:44   | <b>Software</b>   | 8100.27               |                       |            |
| <b>0.05 Lot #</b>           | 202303K            | <b>0.08 Lot #</b>                 | 202303L    | <b>0.20 Lot #</b> | 202304C               | <b>0.08 Gas Lot #</b> | 06723080A5 |
| <b>Expiration</b>           | 03/29/2025         | <b>Expiration</b>                 | 03/29/2025 | <b>Expiration</b> | 04/05/2025            | <b>Expiration</b>     | 04/05/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Average Standard Deviation</b> | 0.0003     | <b>Compliance</b> | Yes                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.082  |
|                                             |        | Interferent Detect Test # 5      | INT    |

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 02/16/2024 12:49

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000749

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | ESCAMBIA COUNTY SO | <b>Date</b>       | 03/12/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SHELLEY,SAMUEL,L   | <b>Time</b>       | 11:51:58   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402690226  |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 03/09/2026 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

Data Download Date/Time 03/12/2024 13:02

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

**Agency** FL HIGHWAY PATROL **Date** 01/27/2024 **Type** Agency Inspection  
**Inspector** RUBEL, NEIL, F **Time** 11:51:01 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1482048  
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/21/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 01/27/2024 13:01

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 02/06/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | ABRAHAM,ERIC,D    | <b>Time</b>       | 14:24:33   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J           | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 1482048    |
| <b>Expiration</b>           | 03/28/2025        | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 04/21/2024 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.084  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 02/06/2024 15:33

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

|                             |                   |                                   |            |                   |                       |                       |            |
|-----------------------------|-------------------|-----------------------------------|------------|-------------------|-----------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>                       | 02/23/2024 | <b>Type</b>       | Department Inspection |                       |            |
| <b>Inspector</b>            | NICODEMO,PHIL,    | <b>Time</b>                       | 10:39:34   | <b>Software</b>   | 8100.27               |                       |            |
| <b>0.05 Lot #</b>           | 202303K           | <b>0.08 Lot #</b>                 | 202303L    | <b>0.20 Lot #</b> | 202304C               | <b>0.08 Gas Lot #</b> | 06723080A5 |
| <b>Expiration</b>           | 03/29/2025        | <b>Expiration</b>                 | 03/29/2025 | <b>Expiration</b> | 04/05/2025            | <b>Expiration</b>     | 04/05/2025 |
| <b>Number of Simulators</b> | 5                 | <b>Average Standard Deviation</b> | 0.0004     | <b>Compliance</b> | Yes                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 02/23/2024 12:19

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

|                                 |                              |                               |                               |
|---------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 03/04/2024       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> RUBEL, NEIL, F | <b>Time</b> 15:03:57         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202303J       | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 1482048 |
| <b>Expiration</b> 03/28/2025    | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 04/21/2024  |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

N/A

Data Download Date/Time 03/04/2024 16:23

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> ALACHUA COUNTY SO  | <b>Date</b> 01/25/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LEVY, JAYSON, I | <b>Time</b> 13:18:09         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 26422080A1 |
| <b>Expiration</b> 03/26/2025     | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 11/05/2024     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 01/25/2024 13:34

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

**Agency** ALACHUA COUNTY SO **Date** 01/30/2024 **Type** Department Inspection  
**Inspector** SIDDOWAY,BENJAMIN,W **Time** 12:30:05 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG310901  
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/19/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 01/30/2024 12:38

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> ALACHUA COUNTY SO  | <b>Date</b> 02/02/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LEVY, JAYSON, I | <b>Time</b> 10:06:47         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 26422080A1 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 11/05/2024     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1          | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                 | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.198  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.081  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.081  |
|                                             |        | Interferent Detect Test # 3          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

#### Remarks:

Int Det: RFI Detect PHONE WAS IN THE AREA AND RANG.

**Data Download Date/Time** 02/02/2024 13:11

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> ALACHUA COUNTY SO  | <b>Date</b> 03/21/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LEVY, JAYSON, I | <b>Time</b> 10:48:14         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 26422080A1 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 11/05/2024     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 03/21/2024 11:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

**Agency** OKALOOSA COUNTY SO  
**Inspector** NICHOLS, MICHAEL, J

**Date** 01/26/2024  
**Time** 14:16:29

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402623828  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 06/15/2024      **Expiration** 12/22/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/26/2024 15:26

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | OKALOOSA COUNTY SO | <b>Date</b>       | 02/26/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | NICHOLS,MICHAEL,J  | <b>Time</b>       | 15:12:34   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402623828  |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 12/22/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 02/26/2024 16:26

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | OKALOOSA COUNTY SO | <b>Date</b>       | 03/26/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | NICHOLS,MICHAEL,J  | <b>Time</b>       | 14:34:55   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402623828  |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 12/22/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 03/26/2024 15:47



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

**Agency** SUWANNEE CO SO  
**Inspector** GRIFFIN,ANGELA,F

**Date** 01/25/2024  
**Time** 23:43:07

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 402469736  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 07/12/2025      **Expiration** 06/16/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 01/25/2024 23:53

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

**Agency** SUWANNEE CO SO  
**Inspector** GRIFFIN, ANGELA, F

**Date** 02/26/2024  
**Time** 23:26:13

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 402469736  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 07/12/2025      **Expiration** 06/16/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail.

**Data Download Date/Time** 02/26/2024 23:37

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

**Agency** SUWANNEE CO SO  
**Inspector** GRIFFIN, ANGELA, F

**Date** 03/29/2024  
**Time** 23:14:38

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 14323080A3  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 07/12/2025      **Expiration** 09/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: .

**Data Download Date/Time** 03/29/2024 23:25

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | HAMILTON COUNTY SO | <b>Date</b>       | 01/31/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | CURRY,WILLIAM,C    | <b>Time</b>       | 01:18:20   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 29122080A1 |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 12/05/2024 |
| <b>Number of Simulators</b> | 4                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 01/31/2024 01:28

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | HAMILTON COUNTY SO | <b>Date</b>       | 02/27/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | CURRY, WILLIAM, C  | <b>Time</b>       | 17:05:30   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 29122080A1 |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 12/05/2024 |
| <b>Number of Simulators</b> | 4                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.191  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.192  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 02/27/2024 17:17

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000770

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | HAMILTON COUNTY SO | <b>Date</b>       | 03/30/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | CURRY,WILLIAM,C    | <b>Time</b>       | 17:46:10   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 29122080A1 |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 12/05/2024 |
| <b>Number of Simulators</b> | 4                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.191  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 03/30/2024 18:03

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

|                             |                      |                   |            |                   |                   |                       |            |
|-----------------------------|----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SANTA ROSA COUNTY SO | <b>Date</b>       | 01/05/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SPENCER, RAYMOND, L  | <b>Time</b>       | 05:09:51   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J              | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 402558895  |
| <b>Expiration</b>           | 03/28/2025           | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 10/03/2025 |
| <b>Number of Simulators</b> | 5                    | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.191  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.192  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 01/08/2024 05:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

|                             |                      |                   |            |                   |                   |                       |            |
|-----------------------------|----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SANTA ROSA COUNTY SO | <b>Date</b>       | 02/05/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SPENCER, RAYMOND, L  | <b>Time</b>       | 18:17:09   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J              | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 402558895  |
| <b>Expiration</b>           | 03/28/2025           | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 10/03/2025 |
| <b>Number of Simulators</b> | 5                    | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/05/2024 19:30

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

**Agency** SANTA ROSA COUNTY SO      **Date** 03/05/2024      **Type** Agency Inspection  
**Inspector** SPENCER, RAYMOND, L      **Time** 21:29:05      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1          | 0.045  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1 | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 1   | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Alcohol Free Test # 2         | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # 2          | 0.048  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.08 g/210L Test # 2          | 0.046  |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # 2   | INT    |
| 0.05 g/210L Test: OK                        | Yes    | Alcohol Free Test # 3         | 0.000  |
| 0.08 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.049  |
| 0.20 g/210L Test: OK                        | No     | 0.08 g/210L Test # 3          | 0.048  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | Interferent Detect Test # 3   | INT    |

#### Remarks:

08: Control Outside Tolerance, RFI Detect. Non-compliance: SEAL WAS NOT TIGHT. RFI DETECTED FROM RADIO .

**Data Download Date/Time** 03/06/2024 00:22

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

|                             |                      |                   |            |                   |                   |                       |            |
|-----------------------------|----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SANTA ROSA COUNTY SO | <b>Date</b>       | 03/05/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SPENCER, RAYMOND, L  | <b>Time</b>       | 23:08:40   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J              | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 402558895  |
| <b>Expiration</b>           | 03/28/2025           | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 10/03/2025 |
| <b>Number of Simulators</b> | 5                    | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.065  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.063  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.065  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: Control Outside Tolerance. SIMULATOR CHANGED AND REHEAT

Data Download Date/Time 03/06/2024 00:22

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

**Agency** LAFAYETTE CO SO  
**Inspector** HENING,FRED,C

**Date** 01/23/2024  
**Time** 11:11:44

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 33922080A1  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 01/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 01/23/2024 11:25

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

**Agency** LAFAYETTE CO SO  
**Inspector** HENING,FRED,C

**Date** 02/21/2024  
**Time** 14:20:34

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 33922080A1  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 06/15/2024      **Expiration** 01/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 02/21/2024 14:33

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> LAFAYETTE CO SO    | <b>Date</b> 03/27/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> HENING, FRED, C | <b>Time</b> 11:20:06         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 33922080A1 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 01/05/2025     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 03/27/2024 11:37

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

**Agency** LAFAYETTE CO SO  
**Inspector** HENING,FRED,C

**Date** 01/23/2024  
**Time** 11:19:32

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 33922080A1  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 01/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.208  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.208  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.207  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 01/23/2024 11:30

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

**Agency** LAFAYETTE COUNTY SO      **Date** 01/26/2024      **Type** Department Inspection  
**Inspector** SIDDOWAY,BENJAMIN,W      **Time** 13:37:10      **Software** 8100.27

**0.05 Lot #** 202303K      **0.08 Lot #** 202303L      **0.20 Lot #** 202304C      **0.08 Gas Lot #** AG229803  
**Expiration** 03/29/2025      **Expiration** 03/29/2025      **Expiration** 04/05/2025      **Expiration** 10/25/2024

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.077  |
| 0.20 g/210L Test # 6                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.077  |
| 0.20 g/210L Test # 8                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.077  |
| 0.20 g/210L Test # 9                   | 0.194  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0006 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 01/26/2024 13:47

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

**Agency** COLUMBIA COUNTY SO      **Date** 01/30/2024      **Type** Agency Inspection  
**Inspector** COLE,DANDRA,M      **Time** 10:37:27      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: RFI Detect.

**Data Download Date/Time** 01/30/2024 11:22

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | COLUMBIA COUNTY SO | <b>Date</b>       | 01/30/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | COLE,DANDRA,M      | <b>Time</b>       | 11:11:37   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 08622080A3 |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 06/05/2024 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/30/2024 11:22

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

|                  |                    |             |            |                 |                   |
|------------------|--------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | COLUMBIA COUNTY SO | <b>Date</b> | 02/27/2024 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | LORTON,CORNELIUS,Q | <b>Time</b> | 17:28:13   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |    |
|-----------------------------|---|-------------------|----|
| <b>Number of Simulators</b> | 5 | <b>Compliance</b> | No |
|-----------------------------|---|-------------------|----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | Yes    |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

00: RFI Detect. Non-compliance:RFI.

**Data Download Date/Time** 02/27/2024 18:17

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

|                  |                    |             |            |                 |                   |
|------------------|--------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | COLUMBIA COUNTY SO | <b>Date</b> | 02/27/2024 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | LORTON,CORNELIUS,Q | <b>Time</b> | 17:30:04   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |    |
|-----------------------------|---|-------------------|----|
| <b>Number of Simulators</b> | 5 | <b>Compliance</b> | No |
|-----------------------------|---|-------------------|----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: RFI Detect. Non-compliance: .

**Data Download Date/Time** 02/27/2024 18:18

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | COLUMBIA COUNTY SO | <b>Date</b>       | 02/27/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LORTON,CORNELIUS,Q | <b>Time</b>       | 18:06:46   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 08622080A3 |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 06/05/2024 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 02/27/2024 18:18

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

**Agency** NAS JACKSONVILLE PD      **Date** 01/26/2024      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 10:20:57      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 07622080A3  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 05/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.045  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/06/2024 12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

**Agency** NAS JACKSONVILLE PD      **Date** 01/31/2024      **Type** Agency Inspection  
**Inspector** POTTER, MICHAEL, R      **Time** 12:14:45      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 08223080A3  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 06/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/06/2024 12:20

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

|                  |                     |             |            |                 |                   |
|------------------|---------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | NAS JACKSONVILLE PD | <b>Date</b> | 02/06/2024 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | SIDDOWAY,BENJAMIN,W | <b>Time</b> | 09:08:38   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

COMPLIANCE NOT DETERMINED, BYPASS INSPECTION TO OPERATE

**Data Download Date/Time** 02/06/2024 12:20

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

**Agency** NAS JACKSONVILLE PD      **Date** 02/06/2024      **Type** Department Inspection  
**Inspector** SIDDOWAY,BENJAMIN,W      **Time** 12:16:17      **Software** 8100.27

**0.05 Lot #** 202303K      **0.08 Lot #** 202303L      **0.20 Lot #** 202304C      **0.08 Gas Lot #** AG310901  
**Expiration** 03/29/2025      **Expiration** 03/29/2025      **Expiration** 04/05/2025      **Expiration** 04/19/2025

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 02/06/2024 12:20

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

|                             |                     |                   |            |                   |                   |                       |            |
|-----------------------------|---------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | NAS JACKSONVILLE PD | <b>Date</b>       | 02/22/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LIGHTSEY,ROBERT,D   | <b>Time</b>       | 12:23:17   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J             | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 08223080A3 |
| <b>Expiration</b>           | 03/28/2025          | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 06/05/2025 |
| <b>Number of Simulators</b> | 5                   | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # (Repeat) 1    | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: RFI Detect RADIO TRAF IN HLWY RERAN, EVED USE NAS JAX

**Data Download Date/Time** 02/22/2024 12:32

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

**Agency** NAS JACKSONVILLE PD      **Date** 03/26/2024      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 15:34:30      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 08223080A3  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 06/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CONTINUED EVED USE AT NAS JAX 3-26-2024

**Data Download Date/Time** 03/26/2024 15:43

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MADISON COUNTY SO  | <b>Date</b> 01/22/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> HENING, FRED, C | <b>Time</b> 12:35:44         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 33922080A1 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 01/05/2025     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.084  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.083  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.084  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 01/22/2024 13:18

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MADISON COUNTY SO  | <b>Date</b> 02/20/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> HENING, FRED, C | <b>Time</b> 12:30:40         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 33922080A1 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 01/05/2025     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.084  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.084  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.084  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 02/20/2024 13:20

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MADISON COUNTY SO  | <b>Date</b> 03/26/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> HENING, FRED, C | <b>Time</b> 12:57:40         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 33922080A1 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 01/05/2025     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 03/26/2024 13:38

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

**Agency** SUWANNEE CO SO  
**Inspector** GRIFFIN,ANGELA,F

**Date** 01/26/2024  
**Time** 02:43:00

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 14323080A3  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 07/12/2025      **Expiration** 09/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail.

**Data Download Date/Time** 01/26/2024 02:53

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

**Agency** SUWANNEE CO SO  
**Inspector** GRIFFIN,ANGELA,F

**Date** 02/27/2024  
**Time** 00:18:22

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 14323080A3  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 07/12/2025      **Expiration** 09/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.052  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 02/27/2024 00:30

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

**Agency** SUWANNEE CO SO  
**Inspector** GRIFFIN,ANGELA,F

**Date** 03/30/2024  
**Time** 00:06:04

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 14323080A3  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 07/12/2025      **Expiration** 09/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 03/30/2024 00:18

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

|                                 |                              |                               |                               |
|---------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> FHP               | <b>Date</b> 01/28/2024       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> RUBEL, NEIL, F | <b>Time</b> 11:38:18         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 202303J       | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 1482048 |
| <b>Expiration</b> 03/28/2025    | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 04/21/2024  |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

N/A

**Data Download Date/Time** 01/28/2024 12:48

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

**Agency** FHP  
**Inspector** ABRAHAM,ERIC,D

**Date** 02/06/2024  
**Time** 14:35:21

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 1482048  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 06/15/2024      **Expiration** 04/21/2024

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/06/2024 15:44

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

|                                  |                                          |                                   |
|----------------------------------|------------------------------------------|-----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL  | <b>Date</b> 02/23/2024                   | <b>Type</b> Department Inspection |
| <b>Inspector</b> NICODEMO,PHIL,  | <b>Time</b> 11:13:40                     | <b>Software</b> 8100.27           |
| <b>0.05 Lot #</b> 202303K        | <b>0.08 Lot #</b> 202303L                | <b>0.20 Lot #</b> 202304C         |
| <b>0.08 Gas Lot #</b> 06723080A5 |                                          |                                   |
| <b>Expiration</b> 03/29/2025     | <b>Expiration</b> 03/29/2025             | <b>Expiration</b> 04/05/2025      |
|                                  |                                          | <b>Expiration</b> 04/05/2025      |
| <b>Number of Simulators</b> 5    | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes             |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 02/23/2024 12:22

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 03/04/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | RUBEL,NEIL,F      | <b>Time</b>       | 15:08:44   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J           | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 1482048    |
| <b>Expiration</b>           | 03/28/2025        | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 04/21/2024 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

N/A

**Data Download Date/Time** 03/04/2024 16:26

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

**Agency** FL HIGHWAY PATROL  
**Inspector** RUBEL, NEIL, F

**Date** 01/04/2024  
**Time** 10:14:54

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402843437  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 09/08/2026

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A : . INSPECTOR ERROR-VOLUME

**Data Download Date/Time** 01/04/2024 11:25

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

**Agency** FL HIGHWAY PATROL **Date** 01/30/2024 **Type** Department Inspection  
**Inspector** NICODEMO,PHIL, **Time** 11:22:26 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG310901  
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/19/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0006 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 01/30/2024 11:31

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

|                                 |                              |                               |                                  |
|---------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> FHP               | <b>Date</b> 01/27/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> RUBEL, NEIL, F | <b>Time</b> 11:45:02         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J       | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 2402328118 |
| <b>Expiration</b> 03/28/2025    | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 01/20/2025     |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 01/27/2024 12:54

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

|                                 |                              |                               |                                  |
|---------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> FHP               | <b>Date</b> 02/06/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ABRAHAM,ERIC,D | <b>Time</b> 16:06:57         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J       | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 2402328118 |
| <b>Expiration</b> 03/28/2025    | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 01/20/2025     |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.084  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.084  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.084  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/06/2024 17:16

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

|                                 |                                          |                                   |                                  |
|---------------------------------|------------------------------------------|-----------------------------------|----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 02/23/2024                   | <b>Type</b> Department Inspection |                                  |
| <b>Inspector</b> NICODEMO,PHIL, | <b>Time</b> 10:33:24                     | <b>Software</b> 8100.27           |                                  |
| <b>0.05 Lot #</b> 202303K       | <b>0.08 Lot #</b> 202303L                | <b>0.20 Lot #</b> 202304C         | <b>0.08 Gas Lot #</b> 06723080A5 |
| <b>Expiration</b> 03/29/2025    | <b>Expiration</b> 03/29/2025             | <b>Expiration</b> 04/05/2025      | <b>Expiration</b> 04/05/2025     |
| <b>Number of Simulators</b> 5   | <b>Average Standard Deviation</b> 0.0013 | <b>Compliance</b> Yes             |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.082  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.083  |
| 0.20 g/210L Test # 6                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.084  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.052  |
| 0.08 g/210L Test # 8                   | 0.084  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.053  |
| 0.08 g/210L Test # 9                   | 0.085  |
| 0.20 g/210L Test # 9                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.054  |
| 0.08 g/210L Test # 10                  | 0.085  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.082  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0019 |
| 0.08 g/210L Test (Std Dev)             | 0.0021 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0010 |

#### Remarks:

Data Download Date/Time 02/23/2024 11:42

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000785

**Agency** FL HIGHWAY PATROL  
**Inspector** RUBEL, NEIL, F

**Date** 03/04/2024  
**Time** 15:21:32

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402328118  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 06/15/2024      **Expiration** 01/20/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 03/04/2024 16:31

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

**Agency** GULF BREEZE PD  
**Inspector** NEFF,STEPHEN,

**Date** 01/25/2024  
**Time** 16:05:18

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402367487  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 04/28/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/25/2024 18:06

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

**Agency** GULF BREEZE PD  
**Inspector** HART,DYLAN,A

**Date** 02/17/2024  
**Time** 19:44:22

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402367487  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 04/28/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/17/2024 21:08

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

**Agency** GULF BREEZE PD  
**Inspector** HART,DYLAN,A

**Date** 03/08/2024  
**Time** 19:43:37

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402367487  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 04/28/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 03/08/2024 21:00

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | ST LUCIE COUNTY SO | <b>Date</b>       | 01/25/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | JONES,CORTLAND,T   | <b>Time</b>       | 16:29:15   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 22722080A1 |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 10/05/2024 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/25/2024 16:38

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

|                             |                    |                                   |            |                   |                       |                       |            |
|-----------------------------|--------------------|-----------------------------------|------------|-------------------|-----------------------|-----------------------|------------|
| <b>Agency</b>               | ST LUCIE COUNTY SO | <b>Date</b>                       | 02/06/2024 | <b>Type</b>       | Department Inspection |                       |            |
| <b>Inspector</b>            | GUTSCHOW,TAYLOR,D  | <b>Time</b>                       | 10:58:20   | <b>Software</b>   | 8100.27               |                       |            |
| <b>0.05 Lot #</b>           | 202303K            | <b>0.08 Lot #</b>                 | 202303L    | <b>0.20 Lot #</b> | 202304C               | <b>0.08 Gas Lot #</b> | 01923080A3 |
| <b>Expiration</b>           | 03/29/2025         | <b>Expiration</b>                 | 03/29/2025 | <b>Expiration</b> | 04/05/2025            | <b>Expiration</b>     | 02/05/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Average Standard Deviation</b> | 0.0004     | <b>Compliance</b> | Yes                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.052  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.082  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.052  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.082  |
|                                             |        | Interferent Detect Test # 5      | INT    |

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.052  |
| 0.08 g/210L Test # 6                   | 0.082  |
| 0.20 g/210L Test # 6                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.052  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.082  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.052  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.052  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.052  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.082  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 02/06/2024 11:07

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> ST LUCIE COUNTY SO | <b>Date</b> 02/26/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LIMB,HUNTER,    | <b>Time</b> 13:59:23         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 22722080A1 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 10/05/2024     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/26/2024 14:08

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | ST LUCIE COUNTY SO | <b>Date</b>       | 03/26/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LIMB,HUNTER,       | <b>Time</b>       | 18:10:54   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 22722080A1 |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 10/05/2024 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 03/26/2024 18:20

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

|                                    |                                          |                                   |                                  |
|------------------------------------|------------------------------------------|-----------------------------------|----------------------------------|
| <b>Agency</b> LARGO PD             | <b>Date</b> 03/05/2024                   | <b>Type</b> Department Inspection |                                  |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:49:38                     | <b>Software</b> 8100.27           |                                  |
| <b>0.05 Lot #</b> 202303K          | <b>0.08 Lot #</b> 202303L                | <b>0.20 Lot #</b> 202304C         | <b>0.08 Gas Lot #</b> 01923080A3 |
| <b>Expiration</b> 03/29/2025       | <b>Expiration</b> 03/29/2025             | <b>Expiration</b> 04/05/2025      | <b>Expiration</b> 02/05/2025     |
| <b>Number of Simulators</b> 5      | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes             |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 03/05/2024 13:59

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

**Agency** LARGO PD      **Date** 03/06/2024      **Type** Agency Inspection  
**Inspector** THORNSBROUGH,DREW,A      **Time** 14:50:45      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 15922080A4  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 08/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/06/2024 15:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

**Agency** ST LUCIE COUNTY SO      **Date** 01/25/2024      **Type** Agency Inspection  
**Inspector** JONES,CORTLAND,T      **Time** 15:54:37      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 08223080A4  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 06/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/25/2024 16:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> ST LUCIE COUNTY SO | <b>Date</b> 02/26/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LIMB,HUNTER,    | <b>Time</b> 13:25:16         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 08223080A4 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 06/05/2025     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 02/26/2024 13:32

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> ST LUCIE COUNTY SO | <b>Date</b> 03/26/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LIMB,HUNTER,    | <b>Time</b> 17:59:32         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 08223080A4 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 06/05/2025     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 03/26/2024 18:07

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | ST LUCIE COUNTY SO | <b>Date</b>       | 01/25/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | JONES,CORTLAND,T   | <b>Time</b>       | 16:07:01   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 08223080A4 |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 06/05/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/25/2024 16:17

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> ST LUCIE COUNTY SO | <b>Date</b> 02/26/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LIMB,HUNTER,    | <b>Time</b> 13:29:51         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 08223080A4 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 06/05/2025     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/26/2024 13:39

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

|                             |                    |                                   |            |                   |                       |                       |            |
|-----------------------------|--------------------|-----------------------------------|------------|-------------------|-----------------------|-----------------------|------------|
| <b>Agency</b>               | ST LUCIE COUNTY SO | <b>Date</b>                       | 03/05/2024 | <b>Type</b>       | Department Inspection |                       |            |
| <b>Inspector</b>            | GUTSCHOW,TAYLOR,D  | <b>Time</b>                       | 14:09:07   | <b>Software</b>   | 8100.27               |                       |            |
| <b>0.05 Lot #</b>           | 202303K            | <b>0.08 Lot #</b>                 | 202303L    | <b>0.20 Lot #</b> | 202304C               | <b>0.08 Gas Lot #</b> | 01923080A3 |
| <b>Expiration</b>           | 03/29/2025         | <b>Expiration</b>                 | 03/29/2025 | <b>Expiration</b> | 04/05/2025            | <b>Expiration</b>     | 02/05/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Average Standard Deviation</b> | 0.0001     | <b>Compliance</b> | Yes                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.082  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 03/05/2024 14:18

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> ST LUCIE COUNTY SO | <b>Date</b> 03/26/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LIMB,HUNTER,    | <b>Time</b> 18:05:07         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J        | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 08223080A4 |
| <b>Expiration</b> 03/28/2025     | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 06/05/2025     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 03/26/2024 18:13

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | GULF COUNTY SO   | <b>Date</b>       | 01/25/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | BRAMMER,DALTON,J | <b>Time</b>       | 22:53:17   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J          | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 402607186  |
| <b>Expiration</b>           | 03/28/2025       | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 12/02/2025 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: Range Exceeded.

**Data Download Date/Time** 01/25/2024 23:12

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | GULF COUNTY SO   | <b>Date</b>       | 02/23/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | BRAMMER,DALTON,J | <b>Time</b>       | 13:37:15   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J          | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 402607186  |
| <b>Expiration</b>           | 03/28/2025       | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 12/02/2025 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/29/2024 12:30

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

**Agency** GULF COUNTY SO **Date** 02/29/2024 **Type** Agency Inspection  
**Inspector** LOWRY, ALFONSO, L **Time** 12:21:21 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 0 **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASS AI FOR INSTRUMENT OPERATION COMPLIANCE NOT DETERMINE

**Data Download Date/Time** 02/29/2024 12:30

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

|                             |                |                                   |            |                   |                       |                       |            |
|-----------------------------|----------------|-----------------------------------|------------|-------------------|-----------------------|-----------------------|------------|
| <b>Agency</b>               | GULF COUNTY SO | <b>Date</b>                       | 03/05/2024 | <b>Type</b>       | Department Inspection |                       |            |
| <b>Inspector</b>            | NICODEMO,PHIL, | <b>Time</b>                       | 13:40:02   | <b>Software</b>   | 8100.27               |                       |            |
| <b>0.05 Lot #</b>           | 202303K        | <b>0.08 Lot #</b>                 | 202303L    | <b>0.20 Lot #</b> | 202304C               | <b>0.08 Gas Lot #</b> | 06723080A5 |
| <b>Expiration</b>           | 03/29/2025     | <b>Expiration</b>                 | 03/29/2025 | <b>Expiration</b> | 04/05/2025            | <b>Expiration</b>     | 04/05/2025 |
| <b>Number of Simulators</b> | 5              | <b>Average Standard Deviation</b> | 0.0003     | <b>Compliance</b> | Yes                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

Remarks:

Data Download Date/Time 03/05/2024 13:50

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

**Agency** GULF COUNTY SO  
**Inspector** BRAMMER,DALTON,J

**Date** 03/20/2024  
**Time** 16:22:25

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402607186  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 12/02/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 03/20/2024 16:33

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | HAMILTON COUNTY SO | <b>Date</b>       | 01/31/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | CURRY,WILLIAM,C    | <b>Time</b>       | 01:27:26   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 29122080A1 |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 12/05/2024 |
| <b>Number of Simulators</b> | 4                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.195  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.083  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.050  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.082  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3               | INT    |

#### Remarks:

FORGOT TO HOOK UP DRY GAS.

Data Download Date/Time 01/31/2024 01:37

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | HAMILTON COUNTY SO | <b>Date</b>       | 02/27/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | CURRY,WILLIAM,C    | <b>Time</b>       | 17:14:13   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 29122080A1 |
| <b>Expiration</b>           | 03/28/2025         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 12/05/2024 |
| <b>Number of Simulators</b> | 4                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.192  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 02/27/2024 17:26

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

**Agency** DIXIE COUNTY SO **Date** 01/31/2024 **Type** Agency Inspection  
**Inspector** MANDOLA, MATTHEW, **Time** 13:14:04 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1          | 0.000  |
| Minimum Sample Volume Check: OK             |        | Interferent Detect Test # (Repeat) 1 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 2          | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # (Repeat) 2 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     | Alcohol Free Test # 3                | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 3          | INT    |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # (Repeat) 3 | 0.000  |
| 0.05 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Test: OK                        | No     |                                      |        |
| 0.20 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                      |        |

#### Remarks:

Int Det: WRONG SOLUTION.

**Data Download Date/Time** 01/31/2024 13:22

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | DIXIE COUNTY SO  | <b>Date</b>       | 01/31/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | MANDOLA,MATTHEW, | <b>Time</b>       | 13:53:22   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J          | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 14323080A3 |
| <b>Expiration</b>           | 03/28/2025       | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 09/05/2025 |
| <b>Number of Simulators</b> | 4                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 01/31/2024 14:00

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

**Agency** DIXIE COUNTY SO      **Date** 02/21/2024      **Type** Agency Inspection  
**Inspector** MANDOLA, MATTHEW,      **Time** 11:14:58      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 14323080A3  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 09/05/2025

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.191  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/21/2024 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> DIXIE COUNTY SO      | <b>Date</b> 03/07/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> MANDOLA, MATTHEW, | <b>Time</b> 09:57:08         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J          | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 14323080A3 |
| <b>Expiration</b> 03/28/2025       | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 09/05/2025     |
| <b>Number of Simulators</b> 4      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 03/07/2024 10:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

|                                  |                                          |                                   |
|----------------------------------|------------------------------------------|-----------------------------------|
| <b>Agency</b> DIXIE COUNTY SO    | <b>Date</b> 03/21/2024                   | <b>Type</b> Department Inspection |
| <b>Inspector</b> NICODEMO, PHIL, | <b>Time</b> 11:16:50                     | <b>Software</b> 8100.27           |
| <b>0.05 Lot #</b> 202303K        | <b>0.08 Lot #</b> 202303L                | <b>0.20 Lot #</b> 202304C         |
| <b>0.08 Gas Lot #</b> 06723080A5 |                                          |                                   |
| <b>Expiration</b> 03/29/2025     | <b>Expiration</b> 03/29/2025             | <b>Expiration</b> 04/05/2025      |
|                                  |                                          | <b>Expiration</b> 04/05/2025      |
| <b>Number of Simulators</b> 5    | <b>Average Standard Deviation</b> 0.0003 | <b>Compliance</b> Yes             |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000809

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 03/21/2024 11:25

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> DIXIE COUNTY SO      | <b>Date</b> 01/31/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> MANDOLA, MATTHEW, | <b>Time</b> 12:02:31         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J          | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 14323080A3 |
| <b>Expiration</b> 03/28/2025       | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 09/05/2025     |
| <b>Number of Simulators</b> 4      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.048  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.079  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.187  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.199  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Interferent Detect Test # 1      | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 2            | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.048  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

05: Purge Fail. 20: Control Outside TEMP LOW.

**Data Download Date/Time** 01/31/2024 12:13

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

|                                 |                                          |                                   |                                  |
|---------------------------------|------------------------------------------|-----------------------------------|----------------------------------|
| <b>Agency</b> DIXIE COUNTY SO   | <b>Date</b> 02/16/2024                   | <b>Type</b> Department Inspection |                                  |
| <b>Inspector</b> NICODEMO,PHIL, | <b>Time</b> 14:55:21                     | <b>Software</b> 8100.27           |                                  |
| <b>0.05 Lot #</b> 202303K       | <b>0.08 Lot #</b> 202303L                | <b>0.20 Lot #</b> 202304C         | <b>0.08 Gas Lot #</b> 06723080A5 |
| <b>Expiration</b> 03/29/2025    | <b>Expiration</b> 03/29/2025             | <b>Expiration</b> 04/05/2025      | <b>Expiration</b> 04/05/2025     |
| <b>Number of Simulators</b> 5   | <b>Average Standard Deviation</b> 0.0003 | <b>Compliance</b> Yes             |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 02/16/2024 15:06

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> DIXIE COUNTY SO      | <b>Date</b> 03/07/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> MANDOLA, MATTHEW, | <b>Time</b> 11:12:58         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J          | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 14323080A3 |
| <b>Expiration</b> 03/28/2025       | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 09/05/2025     |
| <b>Number of Simulators</b> 4      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.191  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 03/07/2024 11:40

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

**Agency** PERRY POLICE DEPT  
**Inspector** HENING, FRED, C

**Date** 01/23/2024  
**Time** 11:19:27

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 08223080A2  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 05/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 01/23/2024 11:36

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

**Agency** PERRY PD **Date** 01/30/2024 **Type** Department Inspection  
**Inspector** NICODEMO,PHIL, **Time** 14:12:27 **Software** 8100.27

**0.05 Lot #** 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG310901  
**Expiration** 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/19/2025

**Number of Simulators** 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

**Remarks:**

**Data Download Date/Time** 01/30/2024 14:25

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

**Agency** HERNANDO COUNTY S.O.      **Date** 01/26/2024      **Type** Agency Inspection  
**Inspector** BAKER,PATRICK,M      **Time** 19:00:56      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 15922080A5  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 07/12/2025      **Expiration** 08/05/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.085  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.085  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.085  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/26/2024 19:12



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

**Agency** HERNANDO COUNTY S.O.      **Date** 02/28/2024      **Type** Agency Inspection  
**Inspector** BAKER,PATRICK,M      **Time** 19:34:22      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 28423080A3  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 07/12/2025      **Expiration** 11/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/28/2024 19:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

**Agency** MARION COUNTY SO      **Date** 01/26/2024      **Type** Department Inspection  
**Inspector** SIDDOWAY,BENJAMIN,W      **Time** 15:39:09      **Software** 8100.27

**0.05 Lot #** 202303K      **0.08 Lot #** 202303L      **0.20 Lot #** 202304C      **0.08 Gas Lot #** AG229803  
**Expiration** 03/29/2025      **Expiration** 03/29/2025      **Expiration** 04/05/2025      **Expiration** 10/25/2024

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1                      | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1                      | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.074  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.197  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.072  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.078  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.071  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3               | INT    |
|                                             |        | Alcohol Free Test # 4                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 4                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 4                      | 0.078  |
|                                             |        | 0.20 g/210L Test # 4                      | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4          | 0.062  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4               | INT    |
|                                             |        | Alcohol Free Test # 5                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 5                      | 0.049  |

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

|                                            |       |
|--------------------------------------------|-------|
| 0.08 g/210L Test # 5                       | 0.079 |
| 0.20 g/210L Test # 5                       | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5           | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 5  | 0.079 |
| Interferent Detect Test # 5                | INT   |
| Alcohol Free Test # 6                      | 0.000 |
| 0.05 g/210L Test # 6                       | 0.050 |
| 0.08 g/210L Test # 6                       | 0.079 |
| 0.20 g/210L Test # 6                       | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 6           | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6  | 0.079 |
| Interferent Detect Test # 6                | INT   |
| Alcohol Free Test # 7                      | 0.000 |
| 0.05 g/210L Test # 7                       | 0.049 |
| 0.08 g/210L Test # 7                       | 0.079 |
| 0.20 g/210L Test # 7                       | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7           | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7  | 0.079 |
| Interferent Detect Test # 7                | INT   |
| Alcohol Free Test # 8                      | 0.000 |
| 0.05 g/210L Test # 8                       | 0.049 |
| 0.08 g/210L Test # 8                       | 0.078 |
| 0.20 g/210L Test # 8                       | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8           | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8  | 0.079 |
| Interferent Detect Test # 8                | INT   |
| Alcohol Free Test # 9                      | 0.000 |
| 0.05 g/210L Test # 9                       | 0.050 |
| 0.08 g/210L Test # 9                       | 0.078 |
| 0.20 g/210L Test # 9                       | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9           | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9  | 0.079 |
| Interferent Detect Test # 9                | INT   |
| Alcohol Free Test # 10                     | 0.000 |
| 0.05 g/210L Test # 10                      | 0.050 |
| 0.08 g/210L Test # 10                      | 0.079 |
| 0.20 g/210L Test # 10                      | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10          | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10 | 0.079 |
| Interferent Detect Test # 10               | INT   |

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

|                                                    |        |
|----------------------------------------------------|--------|
| 0.05 g/210L Test (Std Dev)                         | 0.0004 |
| 0.08 g/210L Test (Std Dev)                         | 0.0005 |
| 0.20 g/210L Test (Std Dev)                         | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev)             | 0.0361 |
| 0.08 g/210L Dry Gas Std Test (Std Dev)<br>(Repeat) | 0.0000 |

**Remarks:**

08: Control Outside Tolerance, DGS TANK EMPTIED, REPLACED .

---

**Data Download Date/Time** 01/26/2024 15:46

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

**Agency** MARION COUNTY SO      **Date** 02/16/2024      **Type** Agency Inspection  
**Inspector** YOUNG,ROBERT,A      **Time** 17:39:35      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test                           | Result |
|---------------------------------------------|--------|--------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1          | 0.000  |
| Date and/or Time Adjusted                   | No     | Alcohol Free Test # (Repeat) 1 | 0.000  |
| Minimum Sample Volume Check: OK             |        |                                |        |
| Barometric Pressure Sensor Check: OK        |        |                                |        |
| Alcohol Free Subject Test: 0.000            | Yes    |                                |        |
| Mouth Alcohol Test: Slope Not Met           | Yes    |                                |        |
| Interferent Detect Test: Interferent Detect | No     |                                |        |
| Diagnostic Check (Post-Inspection): OK      | No     |                                |        |
| Alcohol Free Test: OK                       | No     |                                |        |
| 0.05 g/210L Test: OK                        | No     |                                |        |
| 0.08 g/210L Test: OK                        | No     |                                |        |
| 0.20 g/210L Test: OK                        | No     |                                |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                |        |

#### Remarks:

00: Ambient Fail, Ambient Fail. Non-compliance: .

**Data Download Date/Time** 02/16/2024 19:00

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

**Agency** MARION COUNTY SO      **Date** 02/16/2024      **Type** Agency Inspection  
**Inspector** YOUNG,ROBERT,A      **Time** 17:50:18      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Ambient FailSequence AbortedSequence Aborted. Non-compliance: .

**Data Download Date/Time** 02/16/2024 19:00

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

**Agency** MARION COUNTY SO  
**Inspector** YOUNG,ROBERT,A

**Date** 02/16/2024  
**Time** 18:50:33

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402872586  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 06/15/2024      **Expiration** 10/12/2026

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

**Data Download Date/Time** 02/16/2024 19:00

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

**Agency** MARION COUNTY SO  
**Inspector** YOUNG,ROBERT,A

**Date** 03/22/2024  
**Time** 16:42:10

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 402872586  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 06/15/2024      **Expiration** 10/12/2026

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

**Data Download Date/Time** 03/22/2024 16:51

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

**Agency** HERNANDO COUNTY S.O. **Date** 01/26/2024 **Type** Agency Inspection  
**Inspector** BAKER,PATRICK,M **Time** 19:45:21 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202307C **0.08 Gas Lot #** 08223080A3  
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 07/12/2025 **Expiration** 06/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/26/2024 19:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

|                             |                      |                   |            |                   |                   |                       |            |
|-----------------------------|----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | HERNANDO COUNTY S.O. | <b>Date</b>       | 02/28/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | BAKER,PATRICK,M      | <b>Time</b>       | 19:55:29   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J              | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 08223080A3 |
| <b>Expiration</b>           | 03/28/2025           | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 06/05/2025 |
| <b>Number of Simulators</b> | 5                    | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.196  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.080  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.079  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.080  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3               | INT    |

#### Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 02/28/2024 20:08

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

**Agency** HERNANDO COUNTY S.O.      **Date** 03/19/2024      **Type** Agency Inspection  
**Inspector** BAKER,PATRICK,M      **Time** 17:16:27      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 08223080A3  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 07/12/2025      **Expiration** 06/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/19/2024 17:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

**Agency** SUMTER COUNTY SO      **Date** 01/30/2024      **Type** Agency Inspection  
**Inspector** HAYES, CLINTON, R      **Time** 19:50:03      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 221303  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 01/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: Interferent Detect.

**Data Download Date/Time** 01/30/2024 20:04

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SUMTER COUNTY SO | <b>Date</b>       | 02/28/2024 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | HOLLIS,TERENCE,L | <b>Time</b>       | 19:37:42   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202303J          | <b>0.08 Lot #</b> | 202206B    | <b>0.20 Lot #</b> | 202206C           | <b>0.08 Gas Lot #</b> | 221303     |
| <b>Expiration</b>           | 03/28/2025       | <b>Expiration</b> | 06/14/2024 | <b>Expiration</b> | 06/15/2024        | <b>Expiration</b>     | 01/05/2025 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.077  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.194  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.076  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.000  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3               | INT    |

#### Remarks:

08: Control Outside Tolerance FORGOT TO ATTACH DRY GAS R .

**Data Download Date/Time** 02/28/2024 19:50

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

**Agency** SUMTER COUNTY SO      **Date** 03/21/2024      **Type** Agency Inspection  
**Inspector** HAYES, CLINTON, R      **Time** 19:57:32      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 221303  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 01/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/21/2024 20:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER CO SO  
**Inspector** HAYES, CLINTON, R

**Date** 01/30/2024  
**Time** 19:47:16

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 221303  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 01/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/30/2024 19:57

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER CO SO **Date** 02/28/2024 **Type** Agency Inspection  
**Inspector** HOLLIS,TERENCE,L **Time** 18:31:20 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1          | INT    |
| Minimum Sample Volume Check: OK             |        | Interferent Detect Test # (Repeat) 1 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 2          | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # (Repeat) 2 | INT    |
| Interferent Detect Test: Interferent Detect | No     | Alcohol Free Test # 3                | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 3          | 0.000  |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # (Repeat) 3 | INT    |
| 0.05 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Test: OK                        | No     |                                      |        |
| 0.20 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                      |        |

#### Remarks:

THE 3RD SAMPLE WAS.000 ,RESTART INSPECTION CHANGE ACETON SL. DID NOT COMPLETE THE INTERFERENCE TEST. RESTART THE INSPECT.

**Data Download Date/Time** 02/28/2024 18:46

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER CO SO **Date** 02/28/2024 **Type** Agency Inspection  
**Inspector** HOLLIS,TERENCE,L **Time** 19:33:32 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 221303  
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

2ND ATTEMPT INSPECTION PASSED

**Data Download Date/Time** 02/28/2024 19:43

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER CO SO  
**Inspector** HAYES, CLINTON, R

**Date** 03/21/2024  
**Time** 20:01:39

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202206C      **0.08 Gas Lot #** 221303  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 06/15/2024      **Expiration** 01/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 03/21/2024 20:17

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO **Date** 01/28/2024 **Type** Agency Inspection  
**Inspector** BROWN,JEFFERY,P **Time** 19:28:02 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202307C **0.08 Gas Lot #** 223083  
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 07/12/2025 **Expiration** 06/05/2025

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/31/2024 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO  
**Inspector** BROWN,JEFFERY,P

**Date** 02/28/2024  
**Time** 20:07:23

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202206B      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 223083  
**Expiration** 03/28/2025      **Expiration** 06/14/2024      **Expiration** 07/12/2025      **Expiration** 06/05/2025

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/29/2024 08:09

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO      **Date** 03/29/2024      **Type** Agency Inspection  
**Inspector** BROWN,JEFFERY,P      **Time** 02:01:00      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 223083  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 06/05/2025

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/29/2024 08:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

|                                     |                              |                               |                                  |
|-------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> LAKE COUNTY S.O.      | <b>Date</b> 01/23/2024       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> BOWDEN, MATTHEW, C | <b>Time</b> 11:01:08         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J           | <b>0.08 Lot #</b> 202206B    | <b>0.20 Lot #</b> 202206C     | <b>0.08 Gas Lot #</b> 33522080A5 |
| <b>Expiration</b> 03/28/2025        | <b>Expiration</b> 06/14/2024 | <b>Expiration</b> 06/15/2024  | <b>Expiration</b> 01/05/2025     |
| <b>Number of Simulators</b> 5       | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 01/23/2024 11:12

Printed: 11 April 2024

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 02/29/2024 **Type** Agency Inspection  
**Inspector** BOWDEN, MATTHEW, C **Time** 10:54:07 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33522080A5  
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/29/2024 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 03/22/2024 **Type** Agency Inspection  
**Inspector** BOWDEN, MATTHEW, C **Time** 10:35:51 **Software** 8100.27

**0.05 Lot #** 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33522080A5  
**Expiration** 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/22/2024 10:46

**Printed: 11 April 2024**

COBRA ©1997-2014 CMI, Inc.