

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000200

Agency: FDLE TRAINER

Date/Time: 02/27/2025 14:34:23

Data Download Date/Time: 05/01/2025 14:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

Agency: FDLE

Date/Time: 04/11/2025 14:13:26

Data Download Date/Time: 06/13/2025 17:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FDLE

Date/Time: 07/01/2025 09:32:59

Data Download Date/Time: 07/01/2025 10:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

Agency: FDLE

Date/Time: 04/05/2025 08:51:12

Data Download Date/Time: 04/30/2025 22:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

Agency: FDLE

Date/Time: 03/31/2025 17:11:31

Data Download Date/Time: 04/02/2025 05:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

Agency: COLUMBIA COUNTY SO

Date/Time: 02/17/2025 13:52:14

Data Download Date/Time: 02/25/2025 14:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

Agency: SUWANNEE COUNTY SO

Date/Time: 06/21/2025 22:51:36

Data Download Date/Time: 06/23/2025 13:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

Agency: FL HIGHWAY PATROL

Date/Time: 04/04/2025 12:57:22

Data Download Date/Time: 05/12/2025 14:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 04/04/2025 13:06:08

Data Download Date/Time: 05/12/2025 14:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

Agency: ST LUCIE COUNTY SO

Date/Time: 05/25/2025 19:06:46

Data Download Date/Time: 07/14/2025 10:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ST LUCIE COUNTY SO

Date/Time: 06/12/2025 09:41:36

Data Download Date/Time: 07/14/2025 10:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH P.D.

Date/Time: 02/17/2025 09:36:19

Data Download Date/Time: 02/17/2025 16:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: VERO BEACH P.D.

Date/Time: 02/17/2025 09:43:46

Data Download Date/Time: 02/17/2025 16:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

Agency: ST LUCIE COUNTY SO

Date/Time: 05/29/2025 15:17:31

Data Download Date/Time: 05/29/2025 16:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 03/21/2025 23:14:47

Data Download Date/Time: 03/22/2025 00:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

Agency: HAINES CITY PD

Date/Time: 02/09/2025 20:05:11

Data Download Date/Time: 02/09/2025 20:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HAINES CITY PD

Date/Time: 06/12/2025 15:39:48

Data Download Date/Time: 06/13/2025 17:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

Agency: HERNANDO COUNTY SO

Date/Time: 04/16/2025 21:57:13

Data Download Date/Time: 04/29/2025 05:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

Agency: SUMTER COUNTY SO

Date/Time: 04/30/2025 09:47:58

Data Download Date/Time: 04/30/2025 11:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SUMTER COUNTY SO

Date/Time: 07/01/2025 09:33:26

Data Download Date/Time: 07/01/2025 10:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency: PANAMA CITY BEACH PD

Date/Time: 05/30/2025 14:49:28

Data Download Date/Time: 06/12/2025 14:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PANAMA CITY BEACH PD

Date/Time: 07/04/2025 14:16:51

Data Download Date/Time: 07/04/2025 16:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000860

Agency: POLK COUNTY SO

Date/Time: 03/25/2025 08:34:46

Data Download Date/Time: 03/25/2025 09:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

Agency: MIAMI DADE PD

Date/Time: 06/25/2025 13:29:06

Data Download Date/Time: 07/29/2025 13:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 02/13/2025 20:48:04

Data Download Date/Time: 02/13/2025 21:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ST JOE PD

Date/Time: 03/21/2025 22:19:32

Data Download Date/Time: 04/21/2025 11:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 05/13/2025 17:51:59

Data Download Date/Time: 06/23/2025 23:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PORT ST JOE PD

Date/Time: 05/13/2025 17:52:36

Data Download Date/Time: 06/23/2025 23:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 05/13/2025 17:55:28

Data Download Date/Time: 06/23/2025 23:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PORT ST JOE PD

Date/Time: 05/13/2025 19:11:37

Data Download Date/Time: 06/23/2025 23:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 05/13/2025 19:12:22

Data Download Date/Time: 06/23/2025 23:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

Agency: LEON COUNTY SO

Date/Time: 02/17/2025 17:06:13

Data Download Date/Time: 02/25/2025 10:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

Agency: GADSDEN COUNTY SO

Date/Time: 02/26/2025 10:56:06

Data Download Date/Time: 02/26/2025 12:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 01/23/2025 22:48:33

Data Download Date/Time: 01/24/2025 01:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 01/23/2025 22:49:05

Data Download Date/Time: 01/24/2025 01:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 01/23/2025 23:19:36

Data Download Date/Time: 01/24/2025 01:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 01/24/2025 00:17:01

Data Download Date/Time: 01/24/2025 01:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 02/28/2025 20:58:38

Data Download Date/Time: 02/28/2025 23:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 03/31/2025 02:55:10

Data Download Date/Time: 03/31/2025 04:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 04/27/2025 00:15:33

Data Download Date/Time: 04/30/2025 20:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 04/27/2025 00:17:03

Data Download Date/Time: 04/30/2025 20:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 04/27/2025 01:29:23

Data Download Date/Time: 04/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 04/30/2025 19:33:49

Data Download Date/Time: 04/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 04/30/2025 23:36:36

Data Download Date/Time: 05/28/2025 20:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 04/30/2025 23:43:05

Data Download Date/Time: 05/28/2025 20:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 04/30/2025 23:43:49

Data Download Date/Time: 05/28/2025 20:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 05/01/2025 05:55:11

Data Download Date/Time: 05/28/2025 20:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 05/10/2025 20:47:36

Data Download Date/Time: 05/28/2025 20:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 05/11/2025 00:18:48

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 05/24/2025 03:00:58

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 05/24/2025 03:44:47

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 05/25/2025 05:07:37

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 05/25/2025 06:52:07

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 05/26/2025 00:55:34

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 05/28/2025 19:56:29

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 06/04/2025 04:51:00

Data Download Date/Time: 06/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 06/04/2025 04:52:53

Data Download Date/Time: 06/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 06/07/2025 23:57:38

Data Download Date/Time: 06/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 06/22/2025 21:46:31

Data Download Date/Time: 06/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 06/30/2025 19:16:31

Data Download Date/Time: 06/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

Agency: HARDEE COUNTY SO

Date/Time: 01/26/2025 20:06:55

Data Download Date/Time: 01/31/2025 23:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HARDEE COUNTY SO

Date/Time: 03/06/2025 00:23:28

Data Download Date/Time: 03/18/2025 03:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

Agency: HARDEE COUNTY SO

Date/Time: 03/06/2025 00:28:43

Data Download Date/Time: 03/18/2025 03:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HARDEE COUNTY SO

Date/Time: 03/06/2025 00:46:14

Data Download Date/Time: 03/18/2025 03:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

Agency: HAINES CITY PD

Date/Time: 05/21/2025 18:13:55

Data Download Date/Time: 06/14/2025 05:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HAINES CITY PD

Date/Time: 05/21/2025 18:26:07

Data Download Date/Time: 06/14/2025 05:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

Agency: HAINES CITY PD

Date/Time: 05/21/2025 18:41:51

Data Download Date/Time: 06/14/2025 05:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HAINES CITY PD

Date/Time: 06/14/2025 03:53:04

Data Download Date/Time: 06/14/2025 05:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

Agency: WEST MELBOURNE PD

Date/Time: 04/19/2025 12:08:10

Data Download Date/Time: 04/19/2025 15:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WEST MELBOURNE PD

Date/Time: 06/10/2025 12:00:25

Data Download Date/Time: 06/10/2025 13:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

Agency: PALM BAY P.D.

Date/Time: 03/27/2025 11:32:38

Data Download Date/Time: 03/27/2025 13:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PALM BAY P.D.

Date/Time: 04/30/2025 07:57:42

Data Download Date/Time: 04/30/2025 12:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

Agency: PALM BAY P.D.

Date/Time: 06/26/2025 13:55:01

Data Download Date/Time: 06/26/2025 15:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 01/27/2025 06:59:22

Data Download Date/Time: 01/27/2025 07:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 02/27/2025 06:58:12

Data Download Date/Time: 02/28/2025 06:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 03/25/2025 14:46:52

Data Download Date/Time: 03/26/2025 14:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 03/25/2025 14:48:13

Data Download Date/Time: 03/26/2025 14:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 04/24/2025 07:38:04

Data Download Date/Time: 04/24/2025 09:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 05/25/2025 10:56:57

Data Download Date/Time: 05/25/2025 11:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 05/25/2025 11:03:31

Data Download Date/Time: 05/25/2025 11:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 06/28/2025 10:06:45

Data Download Date/Time: 06/28/2025 11:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001048

Agency: LAKE ALFRED PD

Date/Time: 04/11/2025 09:36:12

Data Download Date/Time: 04/11/2025 16:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LAKE ALFRED PD

Date/Time: 04/11/2025 09:36:50

Data Download Date/Time: 04/11/2025 16:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001048

Agency: LAKE ALFRED PD

Date/Time: 04/11/2025 10:02:18

Data Download Date/Time: 04/11/2025 16:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LAKE ALFRED PD

Date/Time: 04/11/2025 10:51:18

Data Download Date/Time: 04/11/2025 16:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 07/25/2025 11:57:02

Data Download Date/Time: 07/25/2025 13:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency: CORAL SPRINGS PD

Date/Time: 02/11/2025 14:06:51

Data Download Date/Time: 02/28/2025 18:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency: LAKELAND PD

Date/Time: 01/15/2025 14:39:31

Data Download Date/Time: 02/05/2025 13:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

Agency: POLK CSO

Date/Time: 05/20/2025 08:19:51

Data Download Date/Time: 05/20/2025 09:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK CSO

Date/Time: 05/22/2025 08:44:20

Data Download Date/Time: 06/26/2025 06:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

Agency: POLK CSO

Date/Time: 07/29/2025 08:08:22

Data Download Date/Time: 07/29/2025 08:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 03/12/2025 17:42:17

Data Download Date/Time: 03/12/2025 18:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CLEARWATER PD

Date/Time: 03/12/2025 17:44:27

Data Download Date/Time: 03/12/2025 18:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 03/12/2025 17:49:22

Data Download Date/Time: 03/12/2025 18:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CLEARWATER PD

Date/Time: 04/29/2025 19:33:56

Data Download Date/Time: 05/03/2025 00:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 05/27/2025 21:11:50

Data Download Date/Time: 05/31/2025 03:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 06/30/2025 05:41:26

Data Download Date/Time: 07/04/2025 00:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 07/17/2025 18:06:24

Data Download Date/Time: 07/17/2025 19:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 01/17/2025 12:19:58

Data Download Date/Time: 01/17/2025 14:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 02/21/2025 10:34:27

Data Download Date/Time: 02/21/2025 11:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 03/26/2025 13:09:32

Data Download Date/Time: 03/28/2025 11:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 04/30/2025 09:37:54

Data Download Date/Time: 04/30/2025 10:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 04/30/2025 09:39:10

Data Download Date/Time: 04/30/2025 10:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency: HARDEE COUNTY SO

Date/Time: 02/17/2025 15:52:03

Data Download Date/Time: 02/24/2025 11:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 01/14/2025 00:51:53

Data Download Date/Time: 01/15/2025 10:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 01/14/2025 01:45:05

Data Download Date/Time: 01/15/2025 10:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 03/31/2025 07:13:00

Data Download Date/Time: 04/22/2025 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 04/22/2025 09:21:52

Data Download Date/Time: 04/22/2025 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 04/22/2025 09:40:46

Data Download Date/Time: 04/22/2025 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 01/17/2025 12:39:01

Data Download Date/Time: 01/17/2025 13:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 02/21/2025 10:33:35

Data Download Date/Time: 02/21/2025 11:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 03/26/2025 13:09:11

Data Download Date/Time: 03/28/2025 10:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 04/30/2025 09:37:51

Data Download Date/Time: 04/30/2025 11:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency: EGLIN AFB

Date/Time: 05/21/2025 10:00:19

Data Download Date/Time: 06/22/2025 15:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency: FL HIGHWAY PATROL

Date/Time: 04/18/2025 07:16:01

Data Download Date/Time: 05/12/2025 09:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency: FHP TROOP D ORLANDO

Date/Time: 04/10/2025 23:10:54

Data Download Date/Time: 06/09/2025 14:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP TROOP D ORLANDO

Date/Time: 07/18/2025 14:03:27

Data Download Date/Time: 07/18/2025 14:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 01/14/2025 09:53:02

Data Download Date/Time: 01/14/2025 10:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

Agency: NEW SMYRNA BEACH PD

Date/Time: 03/06/2025 08:37:54

Data Download Date/Time: 04/16/2025 10:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: NEW SMYRNA BEACH PD

Date/Time: 06/25/2025 13:16:21

Data Download Date/Time: 07/11/2025 12:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

Agency: EGLIN AFB

Date/Time: 03/12/2025 23:34:06

Data Download Date/Time: 03/13/2025 03:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

Agency: ORANGE CITY P.D.

Date/Time: 02/17/2025 18:03:36

Data Download Date/Time: 02/24/2025 11:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 02/09/2025 13:24:47

Data Download Date/Time: 03/03/2025 16:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

Agency: MIAMI PD

Date/Time: 01/31/2025 21:00:36

Data Download Date/Time: 02/22/2025 03:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MIAMI PD

Date/Time: 03/16/2025 19:00:56

Data Download Date/Time: 03/26/2025 22:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

Agency: PANAMA CITY BEACH PD

Date/Time: 04/04/2025 16:20:40

Data Download Date/Time: 04/14/2025 06:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 07/19/2025 13:26:38

Data Download Date/Time: 07/19/2025 15:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency: POLK COUNTY SO

Date/Time: 02/26/2025 13:57:20

Data Download Date/Time: 02/27/2025 08:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

Agency: FL HIGHWAY PATROL

Date/Time: 03/04/2025 16:28:21

Data Download Date/Time: 04/05/2025 19:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 03/04/2025 16:36:21

Data Download Date/Time: 04/05/2025 19:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

Agency: FL HIGHWAY PATROL

Date/Time: 03/04/2025 16:37:33

Data Download Date/Time: 04/05/2025 19:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency: BREVARD COUNTY S.O.

Date/Time: 02/17/2025 10:52:01

Data Download Date/Time: 02/17/2025 18:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency: PALM BAY PD

Date/Time: 04/30/2025 07:57:03

Data Download Date/Time: 04/30/2025 11:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PALM BAY PD

Date/Time: 06/26/2025 13:53:17

Data Download Date/Time: 06/26/2025 15:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency: PALM BAY PD

Date/Time: 06/26/2025 13:54:34

Data Download Date/Time: 06/26/2025 15:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency: PALM BAY P.D.

Date/Time: 02/26/2025 11:06:29

Data Download Date/Time: 02/26/2025 12:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PALM BAY P.D.

Date/Time: 06/26/2025 13:53:48

Data Download Date/Time: 06/26/2025 15:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency: SEMINOLE COUNTY SO

Date/Time: 02/06/2025 00:15:03

Data Download Date/Time: 02/26/2025 23:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

Agency: BAKER COUNTY SO

Date/Time: 06/23/2025 16:20:58

Data Download Date/Time: 07/17/2025 13:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

Agency: OKEECHOBEE COUNTY SO

Date/Time: 03/13/2025 18:16:42

Data Download Date/Time: 03/13/2025 19:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OKEECHOBEE COUNTY SO

Date/Time: 05/07/2025 04:41:51

Data Download Date/Time: 05/07/2025 05:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

Agency: OKEECHOBEE COUNTY SO

Date/Time: 01/22/2025 10:01:41

Data Download Date/Time: 01/22/2025 11:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OKEECHOBEE COUNTY SO

Date/Time: 03/13/2025 18:14:50

Data Download Date/Time: 03/13/2025 19:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

Agency: OKEECHOBEE COUNTY SO

Date/Time: 05/20/2025 18:19:39

Data Download Date/Time: 05/20/2025 19:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: OKEECHOBEE COUNTY SO

Date/Time: 06/26/2025 17:10:08

Data Download Date/Time: 06/26/2025 17:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

Agency: PORT ST LUCIE PD

Date/Time: 02/18/2025 10:20:13

Data Download Date/Time: 02/18/2025 11:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ST LUCIE PD

Date/Time: 07/10/2025 10:42:19

Data Download Date/Time: 07/10/2025 11:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

Agency: LAKELAND PD

Date/Time: 01/15/2025 14:45:33

Data Download Date/Time: 02/03/2025 14:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: LAKELAND PD

Date/Time: 04/14/2025 08:10:39

Data Download Date/Time: 04/28/2025 10:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

Agency: LAKELAND PD

Date/Time: 04/28/2025 13:46:01

Data Download Date/Time: 05/27/2025 10:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: LAKELAND PD

Date/Time: 04/28/2025 13:51:30

Data Download Date/Time: 05/27/2025 10:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

Agency: LAKELAND PD

Date/Time: 04/28/2025 13:57:13

Data Download Date/Time: 05/27/2025 10:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency: BARTOW PD

Date/Time: 01/16/2025 18:59:51

Data Download Date/Time: 02/09/2025 08:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency: GAINESVILLE PD

Date/Time: 03/25/2025 19:58:16

Data Download Date/Time: 03/25/2025 20:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: GAINESVILLE PD

Date/Time: 03/25/2025 20:07:07

Data Download Date/Time: 03/25/2025 20:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency: SUMTER COUNTY SO

Date/Time: 04/30/2025 09:47:49

Data Download Date/Time: 04/30/2025 11:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SUMTER COUNTY SO

Date/Time: 07/01/2025 09:26:17

Data Download Date/Time: 07/01/2025 10:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency: PALM BAY PD

Date/Time: 04/30/2025 07:55:57

Data Download Date/Time: 04/30/2025 10:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PALM BAY PD

Date/Time: 06/26/2025 13:54:29

Data Download Date/Time: 07/10/2025 10:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Agency: SOUTH MIAMI PD

Date/Time: 01/13/2025 11:23:40

Data Download Date/Time: 01/13/2025 16:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 01/29/2025 16:15:45

Data Download Date/Time: 02/03/2025 23:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 01/29/2025 21:32:36

Data Download Date/Time: 02/03/2025 23:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 04/29/2025 19:36:10

Data Download Date/Time: 05/03/2025 00:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 05/27/2025 21:14:07

Data Download Date/Time: 05/31/2025 03:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 05/27/2025 21:37:31

Data Download Date/Time: 05/31/2025 03:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 06/30/2025 05:44:36

Data Download Date/Time: 07/04/2025 00:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 07/17/2025 18:07:29

Data Download Date/Time: 07/17/2025 19:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

Agency: LAKELAND PD

Date/Time: 02/14/2025 14:53:18

Data Download Date/Time: 02/17/2025 12:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency: PORT ST LUCIE PD

Date/Time: 01/24/2025 19:34:36

Data Download Date/Time: 02/18/2025 11:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ST LUCIE PD

Date/Time: 01/26/2025 08:01:57

Data Download Date/Time: 02/18/2025 11:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency: PORT ST LUCIE PD

Date/Time: 04/04/2025 08:10:25

Data Download Date/Time: 04/27/2025 11:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency: FL HIGHWAY PATROL

Date/Time: 06/11/2025 19:24:46

Data Download Date/Time: 06/11/2025 21:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency: OSCEOLA COUNTY SO

Date/Time: 05/14/2025 11:35:09

Data Download Date/Time: 05/14/2025 13:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency: OSCEOLA COUNTY SO

Date/Time: 05/14/2025 13:59:06

Data Download Date/Time: 06/23/2025 14:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency: LAKELAND PD

Date/Time: 01/15/2025 14:31:14

Data Download Date/Time: 01/29/2025 14:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency:

Date/Time: 06/20/2025 11:59:20

Data Download Date/Time: 07/09/2025 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency:

Date/Time: 06/20/2025 12:08:54

Data Download Date/Time: 07/09/2025 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency:

Date/Time: 06/20/2025 12:09:41

Data Download Date/Time: 07/09/2025 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 01/15/2025 13:11:25

Data Download Date/Time: 01/21/2025 09:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 02/20/2025 09:30:46

Data Download Date/Time: 02/20/2025 11:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 03/26/2025 10:54:54

Data Download Date/Time: 04/23/2025 11:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 04/23/2025 09:21:07

Data Download Date/Time: 04/23/2025 11:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 05/16/2025 07:50:24

Data Download Date/Time: 05/16/2025 10:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 06/21/2025 16:33:14

Data Download Date/Time: 06/21/2025 18:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 07/02/2025 15:50:23

Data Download Date/Time: 07/02/2025 17:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CRESTVIEW PD

Date/Time: 07/02/2025 16:04:47

Data Download Date/Time: 07/02/2025 17:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 07/02/2025 16:36:32

Data Download Date/Time: 07/02/2025 18:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 07/10/2025 01:51:31

Data Download Date/Time: 07/23/2025 16:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 07/23/2025 14:09:42

Data Download Date/Time: 07/23/2025 16:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005404

Agency: POLK COUNTY SO

Date/Time: 04/22/2025 08:22:06

Data Download Date/Time: 04/22/2025 09:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency: FERNANDINA BEACH PD

Date/Time: 05/29/2025 17:25:27

Data Download Date/Time: 06/25/2025 17:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FERNANDINA BEACH PD

Date/Time: 05/29/2025 17:31:07

Data Download Date/Time: 06/25/2025 17:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

Agency: HENDRY COUNTY SO

Date/Time: 04/29/2025 11:49:22

Data Download Date/Time: 06/02/2025 13:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 01/27/2025 07:32:02

Data Download Date/Time: 01/27/2025 08:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 01/27/2025 07:33:56

Data Download Date/Time: 01/27/2025 08:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 01/27/2025 07:35:33

Data Download Date/Time: 01/27/2025 08:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 01/27/2025 07:36:26

Data Download Date/Time: 01/27/2025 08:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 01/27/2025 07:37:12

Data Download Date/Time: 01/27/2025 08:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 03/25/2025 15:27:51

Data Download Date/Time: 03/27/2025 10:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 03/25/2025 15:28:59

Data Download Date/Time: 03/27/2025 10:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 04/24/2025 09:06:20

Data Download Date/Time: 06/28/2025 12:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 05/25/2025 10:38:48

Data Download Date/Time: 06/28/2025 12:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

Agency: MIAMI PD

Date/Time: 01/15/2025 08:27:23

Data Download Date/Time: 01/30/2025 20:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 01/22/2025 15:29:29

Data Download Date/Time: 01/22/2025 16:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BOCA RATON PD

Date/Time: 02/20/2025 09:57:30

Data Download Date/Time: 02/20/2025 10:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 03/20/2025 08:50:30

Data Download Date/Time: 03/20/2025 09:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BOCA RATON PD

Date/Time: 05/09/2025 17:09:18

Data Download Date/Time: 05/09/2025 18:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 06/03/2025 09:54:00

Data Download Date/Time: 06/03/2025 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BOCA RATON PD

Date/Time: 07/22/2025 08:23:06

Data Download Date/Time: 07/22/2025 09:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

Agency: FL HIGHWAY PATROL

Date/Time: 02/04/2025 01:39:23

Data Download Date/Time: 02/17/2025 20:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 02/04/2025 01:55:20

Data Download Date/Time: 02/17/2025 20:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

Agency: FL HIGHWAY PATROL

Date/Time: 04/10/2025 17:48:17

Data Download Date/Time: 04/24/2025 16:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

Agency: FL HIGHWAY PATROL

Date/Time: 05/30/2025 10:51:26

Data Download Date/Time: 06/08/2025 21:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 05/30/2025 10:54:01

Data Download Date/Time: 06/08/2025 21:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

Agency: HOLMES BEACH PD

Date/Time: 01/13/2025 08:51:25

Data Download Date/Time: 02/07/2025 11:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HOLMES BEACH PD

Date/Time: 01/13/2025 08:54:44

Data Download Date/Time: 02/07/2025 11:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

Agency: HOLMES BEACH PD

Date/Time: 01/13/2025 08:56:09

Data Download Date/Time: 02/07/2025 11:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HOLMES BEACH PD

Date/Time: 03/06/2025 11:55:01

Data Download Date/Time: 04/22/2025 17:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

Agency: HOLMES BEACH PD

Date/Time: 07/15/2025 08:10:34

Data Download Date/Time: 07/15/2025 09:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

Agency: MANATEE COUNTY SO

Date/Time: 01/31/2025 13:33:41

Data Download Date/Time: 01/31/2025 13:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

Agency: FL HIGHWAY PATROL

Date/Time: 02/01/2025 18:19:27

Data Download Date/Time: 02/07/2025 09:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 02/01/2025 18:20:53

Data Download Date/Time: 02/07/2025 09:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

Agency: KENNEDY SPACE CENTER

Date/Time: 05/22/2025 11:22:34

Data Download Date/Time: 06/25/2025 08:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: KENNEDY SPACE CENTER

Date/Time: 06/14/2025 23:03:41

Data Download Date/Time: 06/25/2025 08:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

Agency: KENNEDY SPACE CENTER

Date/Time: 07/22/2025 06:45:22

Data Download Date/Time: 07/22/2025 07:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

Agency: FL HIGHWAY PATROL

Date/Time: 05/30/2025 11:16:12

Data Download Date/Time: 06/11/2025 17:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

Agency: FL HIGHWAY PATROL

Date/Time: 02/12/2025 11:49:04

Data Download Date/Time: 04/15/2025 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 02/12/2025 11:50:58

Data Download Date/Time: 04/15/2025 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

Agency: PENSACOLA P.D.

Date/Time: 04/30/2025 11:18:47

Data Download Date/Time: 04/30/2025 13:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

Agency: WALTON COUNTY SO

Date/Time: 01/31/2025 13:19:06

Data Download Date/Time: 06/12/2025 09:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WALTON COUNTY SO

Date/Time: 01/31/2025 13:20:34

Data Download Date/Time: 06/12/2025 09:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

Agency: SUWANNEE COUNTY SO

Date/Time: 04/20/2025 15:41:15

Data Download Date/Time: 04/20/2025 15:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

Agency: FL HIGHWAY PATROL

Date/Time: 07/07/2025 15:12:10

Data Download Date/Time: 07/11/2025 10:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

Agency: FL HIGHWAY PATROL

Date/Time: 02/12/2025 11:22:55

Data Download Date/Time: 02/14/2025 13:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 02/12/2025 11:49:36

Data Download Date/Time: 02/14/2025 13:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

Agency: FL HIGHWAY PATROL

Date/Time: 05/30/2025 11:24:01

Data Download Date/Time: 06/08/2025 21:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

Agency: FL HIGHWAY PATROL

Date/Time: 01/20/2025 03:46:54

Data Download Date/Time: 02/03/2025 12:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

Agency: FFWCC

Date/Time: 06/27/2025 21:37:54

Data Download Date/Time: 06/28/2025 16:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

Agency: POLK COUNTY SO

Date/Time: 03/25/2025 08:32:59

Data Download Date/Time: 03/25/2025 09:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 03/25/2025 08:33:57

Data Download Date/Time: 03/25/2025 09:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

Agency: POLK COUNTY SO

Date/Time: 04/22/2025 08:21:50

Data Download Date/Time: 04/22/2025 09:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 07/02/2025 15:35:15

Data Download Date/Time: 07/29/2025 08:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

Agency: POLK COUNTY SO

Date/Time: 07/29/2025 08:10:10

Data Download Date/Time: 07/29/2025 08:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/19/2025 21:01:55

Data Download Date/Time: 01/19/2025 21:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/16/2025 21:25:36

Data Download Date/Time: 02/16/2025 22:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/20/2025 21:55:01

Data Download Date/Time: 03/20/2025 22:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 05/10/2025 19:54:48

Data Download Date/Time: 05/10/2025 20:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

Agency: POLK COUNTY SO

Date/Time: 01/24/2025 09:28:29

Data Download Date/Time: 01/24/2025 09:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	Fail
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 03/25/2025 08:35:31

Data Download Date/Time: 03/25/2025 09:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

Agency: POLK COUNTY SO

Date/Time: 04/22/2025 08:21:39

Data Download Date/Time: 04/22/2025 09:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 05/20/2025 08:13:15

Data Download Date/Time: 05/20/2025 09:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

Agency: POLK COUNTY SO

Date/Time: 07/29/2025 08:08:59

Data Download Date/Time: 07/29/2025 08:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/19/2025 21:03:40

Data Download Date/Time: 01/19/2025 22:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

Agency: MIAMI PD

Date/Time: 01/31/2025 17:14:55

Data Download Date/Time: 02/22/2025 00:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK
