

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

Agency	VOLUSIA COUNTY SO	Date	01/29/2025	Type	Agency Inspection		
Inspector	MALETTO,NICHOLAS,R	Time	16:22:21	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	403135737
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	08/30/2027
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2025 16:26

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

Agency ST JOHNS COUNTY SO
Inspector MILLS,SCOTT,T

Date 01/31/2025
Time 19:32:59

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402816921
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 08/09/2026

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/31/2025 20:13

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

Agency ST JOHNS COUNTY SO
Inspector MILLS,SCOTT,T

Date 02/26/2025
Time 19:42:36

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402816921
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 08/09/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2025 19:46

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

Agency DELAND PD	Date 01/24/2025	Type Department Inspection
Inspector ARMSTRONG, DESTINEE, N	Time 12:51:20	Software 8100.27

0.05 Lot # 202406K	0.08 Lot # 202406L	0.20 Lot # 202406N	0.08 Gas Lot # AG429602
Expiration 06/19/2026	Expiration 06/19/2026	Expiration 06/20/2026	Expiration 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 01/24/2025 12:51

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

Agency DELAND PD
Inspector WALTER,CRAIG,K

Date 02/28/2025
Time 16:20:04

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 02/28/2025 16:20

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

Agency FLAGLER COUNTY SO **Date** 01/14/2025 **Type** Agency Inspection
Inspector FAIRCLOTH, EDWARD, W **Time** 08:54:14 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 07/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 01/14/2025 08:57

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

Agency	FLAGLER COUNTY SO	Date	02/12/2025	Type	Agency Inspection		
Inspector	FAIRCLOTH,EDWARD,W	Time	08:36:46	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202410F	0.08 Gas Lot #	15924080A2
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	10/30/2026	Expiration	07/05/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 02/12/2025 08:38

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

Agency BROWARD COUNTY SO
Inspector FRAZIER, ANAYA, S

Date 01/08/2025
Time 10:35:35

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402623828
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 12/22/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

S. TILLMAN IN TRANING

Data Download Date/Time 01/08/2025 10:42

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

Agency BROWARD COUNTY SO
Inspector FRAZIER,ANAYA,S

Date 02/05/2025
Time 09:42:28

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402623828
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/22/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

S. TILLMAN IN TRAINING.

Data Download Date/Time 02/05/2025 09:44

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

Agency MANATEE COUNTY SO	Date 01/17/2025	Type Agency Inspection	
Inspector COLEMAN, WILLIAM,	Time 09:52:50	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 403141809
Expiration 02/27/2026	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 09/10/2027
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 01/17/2025 09:53

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

Agency MANATEE COUNTY SO	Date 02/07/2025	Type Agency Inspection	
Inspector NOVAK,CHRISTOPHER,	Time 09:16:24	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 403141809
Expiration 02/27/2026	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 09/10/2027
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/07/2025 09:17

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

Agency	ST. PETERSBURG PD	Date	01/15/2025	Type	Agency Inspection		
Inspector	HALL,DUSTIN,L	Time	22:31:19	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	15924080A2
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	07/05/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/15/2025 22:55

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

Agency	ST. PETERSBURG PD	Date	02/11/2025	Type	Agency Inspection		
Inspector	HALL,DUSTIN,L	Time	22:25:43	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202410F	0.08 Gas Lot #	15924080A2
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	10/30/2026	Expiration	07/05/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/12/2025 00:50

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

Agency	NEW SMYRNA BEACH PD	Date	01/31/2025	Type	Agency Inspection		
Inspector	HUNNEFELD,RALPH,A	Time	13:27:53	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	24923080A4
Expiration	03/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	10/05/2025
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/31/2025 14:12

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

Agency	NEW SMYRNA BEACH PD	Date	02/27/2025	Type	Agency Inspection		
Inspector	HELMS,WILLIAM,A	Time	17:01:56	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	24923080A4
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	10/05/2025
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2025 08:22

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

Agency LEON COUNTY SO
Inspector HENING,FRED,C

Date 01/15/2025
Time 12:40:30

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A3
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 01/15/2025 12:47

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

Agency LEON COUNTY SO
Inspector HENING, FRED, C

Date 02/10/2025
Time 15:08:37

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A3
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TOO MUCH MOUTH ALCOHOL.

Data Download Date/Time 02/11/2025 15:02

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

Agency	VOLUSIA COUNTY SO	Date	01/29/2025	Type	Agency Inspection		
Inspector	MALETTO,NICHOLAS,R	Time	16:36:26	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	402419647
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	04/14/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2025 16:38

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

Agency	VOLUSIA COUNTY SO	Date	02/27/2025	Type	Agency Inspection		
Inspector	MALETTO,NICHOLAS,R	Time	19:00:07	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202410F	0.08 Gas Lot #	402419647
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	10/30/2026	Expiration	04/14/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/27/2025 19:06

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

Agency	EGLIN AFB	Date	01/17/2025	Type	Agency Inspection		
Inspector	TUSKES,ARTHUR,A	Time	19:26:32	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	402648278
Expiration	02/27/2025	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	01/27/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/17/2025 20:28

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

Agency EGLIN AFB
Inspector BURNS,ALVIN,G

Date 02/26/2025
Time 19:14:37

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402648278
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 01/27/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2025 20:15

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

Agency ORMOND BEACH PD
Inspector GARNER,MICHAEL,D

Date 01/23/2025
Time 17:39:24

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477283
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/24/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/23/2025 17:41

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001158

Agency	PONCE INLET PD	Date	01/26/2025	Type	Agency Inspection
Inspector	GASTON,SCOTT,B	Time	16:26:58	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	4	Compliance	Yes
-----------------------------	---	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: .

Data Download Date/Time 01/26/2025 17:13

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001158

Agency PONCE INLET PD
Inspector GASTON,SCOTT,B

Date 01/26/2025
Time 17:08:16

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402558895
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/03/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/26/2025 17:13

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001158

Agency PONCE INLET PD
Inspector GASTON,SCOTT,B

Date 02/23/2025
Time 16:59:46

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402558895
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/03/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Volume Not Met (0.%03ld - Breath Sample Not.

Data Download Date/Time 02/23/2025 17:08

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

Agency CLERMONT PD
Inspector WHEELER,CHAD,T

Date 01/22/2025
Time 10:35:22

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402357111
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 02/10/2025

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/22/2025 10:35

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

Agency CLERMONT PD **Date** 02/25/2025 **Type** Agency Inspection
Inspector COONS,SEAN,C **Time** 11:30:16 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 02/27/2026 **Expiration** **Expiration** **Expiration**

Number of Simulators 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.015
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # (Repeat) 2	0.013
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test # 3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test # (Repeat) 3	0.014
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. .05 ISSUE. WILL RESTART WITH NEW SOLUTION.

Data Download Date/Time 02/25/2025 13:38

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

Agency CLERMONT PD
Inspector COONS,SEAN,C

Date 02/25/2025
Time 13:36:54

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403008479
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/28/2027

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/25/2025 13:38

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency	GROVELAND PD	Date	01/21/2025	Type	Agency Inspection		
Inspector	PENA,RUSSEL,S	Time	12:31:55	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	25024080A3
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	10/05/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 01/21/2025 12:34

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

Agency MIAMI PD	Date 01/30/2025	Type Agency Inspection	
Inspector FERNANDEZ,ANGEL,	Time 18:56:22	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 402367487
Expiration 02/27/2026	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 04/28/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2025 18:56

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

Agency MIAMI PD	Date 02/22/2025	Type Agency Inspection	
Inspector FERNANDEZ,ANGEL,	Time 00:56:22	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 402367487
Expiration 02/27/2026	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 04/28/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/22/2025 00:55

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

Agency MIAMI PD	Date 01/30/2025	Type Agency Inspection	
Inspector FERNANDEZ,ANGEL,	Time 19:18:29	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 402367487
Expiration 02/27/2026	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 04/28/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2025 19:14

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

Agency MIAMI PD	Date 02/22/2025	Type Agency Inspection	
Inspector FERNANDEZ,ANGEL,	Time 03:03:31	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 402367487
Expiration 02/27/2026	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 04/28/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/22/2025 03:06

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency	MARTIN COUNTY SO	Date	01/27/2025	Type	Agency Inspection		
Inspector	THOMAS,CHRISTOPHER,L	Time	17:16:06	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	08223080A5
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	06/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/27/2025 17:30

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

Agency MARTIN COUNTY SO **Date** 02/02/2025 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 23:12:52 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 08223080A5
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/02/2025 23:16

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

Agency BOYNTON BEACH PD **Date** 01/17/2025 **Type** Agency Inspection
Inspector ALEXIS,VLADIMIR, **Time** 22:00:39 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 220213
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/17/2025 22:01

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

Agency BOYNTON BEACH PD **Date** 02/19/2025 **Type** Agency Inspection
Inspector ALEXIS,VLADIMIR, **Time** 18:26:08 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 220213
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/19/2025 18:27

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

Agency PANAMA CITY BEACH PD **Date** 01/07/2025 **Type** Agency Inspection
Inspector MOSS,ROBERT,R **Time** 18:35:42 **Software** 8100.27

0.05 Lot # 202402B **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403005538
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/25/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 01/07/2025 19:35

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency COLLIER COUNTY SO **Date** 01/15/2025 **Type** Agency Inspection
Inspector LANDAS, CARTER, **Time** 12:49:30 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 20524080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

TESTED AT NJC

Data Download Date/Time 01/15/2025 14:48

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency COLLIER COUNTY SO
Inspector GONZALES,YVETTE,R

Date 02/05/2025
Time 15:43:32

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 20524080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/05/2025 16:54

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Agency COLLIER COUNTY SO **Date** 01/15/2025 **Type** Agency Inspection
Inspector LANDAS, CARTER, **Time** 10:18:39 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 20524080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

TESTED AT NJC

Data Download Date/Time 01/15/2025 15:13

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Agency COLLIER COUNTY SO **Date** 02/06/2025 **Type** Agency Inspection
Inspector GONZALES,YVETTE,R **Time** 11:22:41 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 20524080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/06/2025 14:13

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Agency COLLIER COUNTY SO	Date 02/21/2025	Type Department Inspection
Inspector GUTSCHOW,TAYLOR,D	Time 13:34:43	Software 8100.27

0.05 Lot # 202303K	0.08 Lot # 202303L	0.20 Lot # 202304C	0.08 Gas Lot # AG429602
Expiration 03/29/2025	Expiration 03/29/2025	Expiration 04/05/2025	Expiration 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/21/2025 13:34

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Agency COLLIER COUNTY SO
Inspector GONZALES,YVETTE,R

Date 02/27/2025
Time 21:12:46

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 20524080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2025 14:07

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency	PANAMA CITY PD	Date	01/11/2025	Type	Agency Inspection		
Inspector	NELSON,GARRETT,R	Time	21:04:03	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	2403056709
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	05/28/2027
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 01/11/2025 22:04

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency	PANAMA CITY PD	Date	02/02/2025	Type	Agency Inspection		
Inspector	TURNER,JONATHAN,R	Time	19:29:05	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	3024030567
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	05/28/2027
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/02/2025 20:29

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency NAS PENSACOLA PD
Inspector KELLEY,KEITH,H

Date 01/03/2025
Time 12:13:17

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 223083
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/05/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.190
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 01/05/2025 12:32

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

Agency NAS PENSACOLA PD
Inspector KELLEY,KEITH,H

Date 02/01/2025
Time 20:06:56

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 223083
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/05/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.192
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.191
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.192
		0.20 g/210L Test # (Repeat) 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 02/01/2025 21:45

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency	HENDRY COUNTY SO	Date	01/10/2025	Type	Agency Inspection		
Inspector	MCNEIL,DARRIN,J	Time	19:30:31	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	402872586
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	10/12/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2025 14:17

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

Agency HENDRY COUNTY SO **Date** 02/28/2025 **Type** Agency Inspection
Inspector NORWOOD,CHRISTOPHER,J **Time** 19:26:56 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402872586
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 10/12/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/03/2025 11:49

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

Agency	LEE COUNTY SO	Date	01/22/2025	Type	Agency Inspection		
Inspector	DUNCAN,NICHOLAS,	Time	11:45:47	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	24923080A3
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	10/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.189
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test # (Repeat) 1	0.194
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test # 1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.079
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test # 2	0.195
		0.20 g/210L Test # (Repeat) 2	0.195
		0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.073
		0.08 g/210L Test # (Repeat) 3	0.079
		0.20 g/210L Test # 3	0.195
		0.20 g/210L Test # (Repeat) 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance. 20: Control Outside Tolerance-WEAK O-RING-TIGHTEN-RETEST.

Data Download Date/Time 01/22/2025 11:46

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

Agency LEE COUNTY SO **Date** 02/19/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 12:30:29 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/19/2025 12:32

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

Agency LEE COUNTY SO **Date** 02/21/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:13:39 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/21/2025 14:15

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency TYNDALL AFB
Inspector LORD,LANNY,A

Date 01/06/2025
Time 12:23:04

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477284
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/27/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/08/2025 09:55

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency TYNDALL AFB
Inspector LORD,LANNY,A

Date 02/04/2025
Time 12:06:51

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477284
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/27/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2025 13:15

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency	WEST PALM BEACH PD	Date	01/29/2025	Type	Agency Inspection		
Inspector	INGRASSIA,JAMES,A	Time	22:34:44	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	402872586
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	10/12/2026
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/29/2025 22:33

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

Agency	WEST PALM BEACH PD	Date	02/23/2025	Type	Agency Inspection		
Inspector	INGRASSIA,JAMES,A	Time	18:22:13	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	402872586
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	10/12/2026
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed. A F / M A: Improper Sample - BEGAN EARLYOON. 00: Ambient Fail. - RETRY

Data Download Date/Time 02/25/2025 13:17

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency LAKE COUNTY SO **Date** 02/05/2025 **Type** Department Inspection
Inspector ARMSTRONG, DESTINEE, N **Time** 12:58:05 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/05/2025 12:57

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency	DAYTONA BEACH SHORES	Date	01/08/2025	Type	Agency Inspection		
Inspector	MEINEKE,TIMOTHY,N	Time	05:12:06	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	402477284
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	06/27/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/08/2025 05:14

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency DAYTONA BEACH SHORES **Date** 02/13/2025 **Type** Agency Inspection
Inspector MEINEKE,TIMOTHY,N **Time** 05:08:17 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 32324080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/13/2025 05:10

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency EDGEWATER PD
Inspector FEGER, JOSHUA, D

Date 01/12/2025
Time 16:04:39

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402843437
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/08/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

Time-Date changed. Int Det: WRONG SIM USED FOR FIRST TEST. .

Data Download Date/Time 01/12/2025 16:04

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency EDGEWATER PD **Date** 01/15/2025 **Type** Department Inspection
Inspector ARMSTRONG, DESTINEE, N **Time** 13:57:13 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 01/15/2025 13:59

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency	EDGEWATER PD	Date	01/22/2025	Type	Agency Inspection		
Inspector	CHECKSFIELD,STEPHEN,F	Time	10:43:30	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202410F	0.08 Gas Lot #	402843437
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	10/30/2026	Expiration	09/08/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/22/2025 10:42

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency EDGEWATER PD
Inspector FEGER, JOSHUA, D

Date 02/20/2025
Time 10:19:48

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402843437
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/08/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/20/2025 10:19

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

Agency	PUTNAM COUNTY SO	Date	01/31/2025	Type	Agency Inspection		
Inspector	BISHOP,TREVOR,C	Time	15:17:09	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202410F	0.08 Gas Lot #	28423080A2
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	10/30/2026	Expiration	11/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/31/2025 15:23

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

Agency	PUTNAM COUNTY SO	Date	02/28/2025	Type	Agency Inspection		
Inspector	BISHOP,TREVOR,C	Time	23:30:17	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202410F	0.08 Gas Lot #	28423080A2
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	10/30/2026	Expiration	11/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2025 23:33

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency SOUTH DAYTONA PD **Date** 01/05/2025 **Type** Agency Inspection
Inspector MENDOZA,JONATHAN,F **Time** 09:11:35 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403096062
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/15/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/05/2025 09:10

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency	SOUTH DAYTONA PD	Date	02/03/2025	Type	Agency Inspection		
Inspector	MENDOZA,JONATHAN,F	Time	03:50:27	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	403096062
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	07/15/2027
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.046
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.044
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.05 g/210L Test # (Repeat) 3	0.047
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.

Data Download Date/Time 02/03/2025 03:50

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

Agency FL HIGHWAY PATROL **Date** 01/25/2025 **Type** Agency Inspection
Inspector KEMP,TREVOR,D **Time** 19:26:21 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI Detect. Non-compliance: .

Data Download Date/Time 02/24/2025 19:21

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

Agency FL HIGHWAY PATROL **Date** 01/25/2025 **Type** Agency Inspection
Inspector KEMP,TREVOR,D **Time** 19:30:14 **Software** 8100.27

0.05 Lot # **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI Detect. Non-compliance: .

Data Download Date/Time 02/24/2025 19:21

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

Agency	FL HIGHWAY PATROL	Date	01/25/2025	Type	Agency Inspection
Inspector	KEMP,TREVOR,D	Time	19:32:24	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	5	Compliance	No
-----------------------------	---	-------------------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

A F / M A: RFI Detect. Non-compliance: .

Data Download Date/Time 02/24/2025 19:21

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

Agency	FL HIGHWAY PATROL	Date	01/25/2025	Type	Agency Inspection		
Inspector	KEMP,TREVOR,D	Time	20:08:09	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	403135737
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	08/30/2027
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 02/24/2025 19:21

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

Agency FL HIGHWAY PATROL
Inspector KEMP,TREVOR,D

Date 02/24/2025
Time 20:28:40

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 403135737
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 08/30/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 02/24/2025 20:41

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

Agency	FORT LAUDERDALE PD	Date	01/28/2025	Type	Agency Inspection		
Inspector	STONER,NATHAN,C	Time	21:45:16	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202410F	0.08 Gas Lot #	402886996
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	10/30/2026	Expiration	11/01/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.085
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.085
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.084
		Interferent Detect Test # 3	INT

Remarks:

OFC CARTER PRESENT

Data Download Date/Time 01/28/2025 21:49

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

Agency	FORT LAUDERDALE PD	Date	02/25/2025	Type	Agency Inspection		
Inspector	STONER,NATHAN,C	Time	21:45:34	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202410F	0.08 Gas Lot #	402886996
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	10/30/2026	Expiration	11/01/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

OFC CARTER PRESENT

Data Download Date/Time 03/01/2025 14:09

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency	BREVARD COUNTY SO	Date	01/22/2025	Type	Agency Inspection		
Inspector	MCLAUGHLIN,BRADLEY,	Time	09:57:27	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	24923080A3
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	10/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

05: C0T SIM NOT HOOKED UP.

Data Download Date/Time 01/22/2025 10:00

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

Agency BREVARD COUNTY SO **Date** 02/24/2025 **Type** Agency Inspection
Inspector WAGNER,ROBERT,F **Time** 09:37:52 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A3
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/24/2025 09:39

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency MONROE COUNTY SO
Inspector GILROY,ZACHARY,I

Date 01/19/2025
Time 19:57:48

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402886996
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/01/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

KV SPARE INTOX

Data Download Date/Time 01/19/2025 19:54

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency MONROE COUNTY SO
Inspector GILROY,ZACHARY,I

Date 02/19/2025
Time 17:16:15

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403207508
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 12/02/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

KV SPARE

Data Download Date/Time 02/19/2025 17:14

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency	ORANGE COUNTY SO	Date	01/14/2025	Type	Agency Inspection		
Inspector	MELVILLE,KELLY,M	Time	10:13:39	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	28423080A4
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	11/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2025 10:49

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

Agency	ORANGE COUNTY SO	Date	02/18/2025	Type	Agency Inspection		
Inspector	MELVILLE,KELLY,M	Time	10:08:14	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	28423080A4
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	11/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2025 10:07

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency ORANGE COUNTY SO **Date** 01/14/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 12:35:42 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2025 13:06

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

Agency	ORANGE COUNTY SO	Date	02/18/2025	Type	Agency Inspection		
Inspector	MELVILLE,KELLY,M	Time	12:19:08	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	15924080A2
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	07/05/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2025 12:35

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency	ORANGE COUNTY S.O.	Date	01/14/2025	Type	Agency Inspection		
Inspector	MELVILLE,KELLY,M	Time	10:16:28	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	28423080A4
Expiration	02/27/2026	Expiration	10/22/2025	Expiration	07/12/2025	Expiration	11/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2025 10:44

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency ORANGE COUNTY S.O. **Date** 02/13/2025 **Type** Department Inspection
Inspector ARMSTRONG, DESTINEE, N **Time** 10:55:37 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.076
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.076
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.076
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/13/2025 11:06

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Agency	PINELLAS COUNTY SO	Date	01/16/2025	Type	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	Time	12:31:14	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance Yes
-----------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Data Download Date/Time 01/24/2025 09:59

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Agency PINELLAS COUNTY SO **Date** 01/17/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 12:52:38 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Data Download Date/Time 01/24/2025 09:59

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency COCOA PD	Date 01/09/2025	Type Agency Inspection	
Inspector FRAZIER,JAMES,H	Time 20:13:20	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 14323080A2
Expiration 02/27/2026	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 06/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/09/2025 20:16

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

Agency COCOA PD	Date 02/14/2025	Type Agency Inspection	
Inspector FRAZIER,JAMES,H	Time 19:46:01	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 14323080A2
Expiration 02/27/2026	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 06/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/14/2025 19:47

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

Agency	ROCKLEDGE PD	Date	02/01/2025	Type	Agency Inspection		
Inspector	KENNEDY,PATRICK,W	Time	00:02:23	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	403141809
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	09/10/2027
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 02/01/2025 00:02

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.