

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency	SARASOTA COUNTY SO	Date	01/20/2025	Type	Agency Inspection		
Inspector	SANDERS,JOSHUA,P	Time	08:03:37	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	10124080A2
Expiration	03/28/2025	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	06/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Time changed. LID WAS NOT SECURE - NO SEAL. PASSED INSPECT

Data Download Date/Time 01/20/2025 08:03

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO **Date** 02/21/2025 **Type** Agency Inspection
Inspector FELTOVIC, DEP MIKE, **Time** 07:37:55 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT JAIL

Data Download Date/Time 02/21/2025 07:39

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD	Date 01/12/2025	Type Agency Inspection	
Inspector SILLS,MARK,A	Time 08:49:48	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307C	0.20 Lot # 202307C	0.08 Gas Lot # 231603
Expiration 02/27/2026	Expiration 07/12/2025	Expiration 07/12/2025	Expiration 07/05/2026
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/12/2025 08:48

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD	Date 01/12/2025	Type Agency Inspection	
Inspector SILLS,MARK,A	Time 09:29:17	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 231603
Expiration 02/27/2026	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 07/05/2026
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail.

Data Download Date/Time 01/12/2025 09:27

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD	Date 02/09/2025	Type Agency Inspection	
Inspector SILLS,MARK,A	Time 08:37:42	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 231603
Expiration 02/27/2026	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 07/05/2026
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/09/2025 08:35

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 02/21/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:46:50 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.076
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/21/2025 13:45

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 02/28/2025 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 14:08:32 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 02/27/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2025 14:07

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency	CHARLOTTE COUNTY SO	Date	01/12/2025	Type	Agency Inspection		
Inspector	CHANDLER,VANESSA,N	Time	20:51:02	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	06723080A5
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	04/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 01/12/2025 20:48

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO **Date** 02/04/2025 **Type** Agency Inspection
Inspector LEONARD,ALEYA,D **Time** 22:19:29 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 06723080A5
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 02/04/2025 22:16

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency	GAINESVILLE PD	Date	01/27/2025	Type	Agency Inspection		
Inspector	STEMEN,AARON,M	Time	20:11:01	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	24923080A3
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	10/05/2025
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/27/2025 20:14

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency	GAINESVILLE PD	Date	02/26/2025	Type	Agency Inspection		
Inspector	STEMEN,AARON,M	Time	19:03:14	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202410F	0.08 Gas Lot #	24923080A3
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	10/30/2026	Expiration	10/05/2025
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/26/2025 19:06

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency LEVY COUNTY SO
Inspector LEVY, JAYSON, I

Date 01/04/2025
Time 21:28:47

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 01524080A4
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A:USED TOO MUCH MA.

Data Download Date/Time 01/10/2025 10:46

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency LEVY COUNTY SO
Inspector LEVY, JAYSON, I

Date 01/29/2025
Time 14:57:53

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 10124080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2025 17:48

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency	LEVY COUNTY SO	Date	02/02/2025	Type	Agency Inspection		
Inspector	LEVY,JAYSON,I	Time	10:58:34	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	10124080A2
Expiration	03/28/2025	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	06/05/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/02/2025 10:58

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT
Inspector POTTER, MICHAEL, R

Date 01/14/2025
Time 14:19:08

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN SERV. NS MAYPORT

Data Download Date/Time 01/15/2025 09:09

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT **Date** 02/25/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 09:45:50 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

EVED USE AT M/P N/S

Data Download Date/Time 02/25/2025 09:47

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency	PINELLAS COUNTY SO	Date	01/16/2025	Type	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	Time	11:58:43	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Data Download Date/Time 01/24/2025 10:03

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency PINELLAS COUNTY SO	Date 01/17/2025	Type Department Inspection
Inspector GUTSCHOW, TAYLOR, D	Time 13:36:28	Software 8100.27

0.05 Lot # 202303K	0.08 Lot # 202303L	0.20 Lot # 202304C	0.08 Gas Lot # AG429602
Expiration 03/29/2025	Expiration 03/29/2025	Expiration 04/05/2025	Expiration 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 01/24/2025 10:03

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency PINELLAS COUNTY SO **Date** 02/19/2025 **Type** Agency Inspection
Inspector DEANE,M, **Time** 15:16:17 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** AG409204
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/01/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/19/2025 15:16

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency	ORANGE COUNTY SO	Date	01/14/2025	Type	Agency Inspection		
Inspector	MELVILLE,KELLY,M	Time	11:37:52	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	28423080A4
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	11/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2025 13:17

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO **Date** 02/05/2025 **Type** Department Inspection
Inspector ARMSTRONG, DESTINEE, N **Time** 13:12:31 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.202
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/05/2025 13:11

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency	ORANGE COUNTY SO	Date	02/18/2025	Type	Agency Inspection		
Inspector	MELVILLE,KELLY,M	Time	10:16:07	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	15924080A2
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	07/05/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2025 10:17

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 01/14/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 11:42:13 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2025 13:20

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

Agency ORANGE COUNTY SO **Date** 02/18/2025 **Type** Agency Inspection
Inspector MELVILLE,KELLY,M **Time** 10:11:00 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2025 10:11

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency	ORANGE COUNTY SO	Date	01/14/2025	Type	Agency Inspection		
Inspector	MELVILLE,KELLY,M	Time	10:21:20	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	15924080A2
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	07/05/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2025 10:41

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency	ORANGE COUNTY SO	Date	02/18/2025	Type	Agency Inspection		
Inspector	MELVILLE,KELLY,M	Time	11:28:42	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202307C	0.08 Gas Lot #	15924080A2
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	07/12/2025	Expiration	07/05/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2025 11:31

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO **Date** 01/27/2025 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 22:13:14 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 08223080A5
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/27/2025 22:14

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO **Date** 02/02/2025 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 23:03:13 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 08223080A5
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/02/2025 23:02

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency	BISCAYNE NAT PARK	Date	01/24/2025	Type	Agency Inspection		
Inspector	SAAVEDRA, OSCAR,	Time	12:43:48	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	402987895
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	03/05/2027
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/24/2025 12:43

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency	BISCAYNE NAT PARK	Date	02/26/2025	Type	Agency Inspection		
Inspector	GORNEY,WILLIAM,M	Time	10:58:05	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	402987895
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	03/05/2027
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2025 10:59

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 01/28/2025 **Type** Agency Inspection
Inspector HAYES, CLINTON, R **Time** 03:07:45 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/28/2025 03:11

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency	SUMTER COUNTY SO	Date	02/27/2025	Type	Agency Inspection
Inspector	HOLLIS,TERENCE,L	Time	19:09:19	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	5	Compliance	No
-----------------------------	---	-------------------	----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Interferent Detect Test # 1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test # (Repeat) 1	0.000
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 2	0.000
Interferent Detect Test: Interferent Detect	No	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 3	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 3	0.000
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

Int Det:RADIO TO CLOSE TO DUI ROOM,. Non-compliance:INT NOT DETECTED RESTART INSPECTION.

Data Download Date/Time 02/27/2025 19:52

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 02/27/2025 **Type** Agency Inspection
Inspector HOLLIS,TERENCE,L **Time** 19:52:31 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 231603
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/27/2025 19:52

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

Agency TYNDALL AFB
Inspector LORD,LANNY,A

Date 01/06/2025
Time 14:08:39

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477284
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/27/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/08/2025 09:49

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

Agency TYNDALL AFB
Inspector LORD,LANNY,A

Date 02/04/2025
Time 13:27:24

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477284
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/27/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2025 16:08

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency	CAPE CORAL PD	Date	01/14/2025	Type	Agency Inspection		
Inspector	BARKLEY,JEFFREY,G	Time	23:02:37	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202410F	0.08 Gas Lot #	28423080A4
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	10/30/2026	Expiration	11/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/18/2025 03:41

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency	CAPE CORAL PD	Date	02/06/2025	Type	Agency Inspection		
Inspector	BARKLEY,JEFFREY,G	Time	03:02:53	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202410F	0.08 Gas Lot #	28423080A4
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	10/30/2026	Expiration	11/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/06/2025 03:13

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 01/16/2025 **Type** Agency Inspection
Inspector LANDAS, CARTER, **Time** 15:58:00 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

TESTED AT NJC

Data Download Date/Time 01/16/2025 16:58

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO
Inspector GONZALES,YVETTE,R

Date 02/05/2025
Time 12:17:12

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/05/2025 13:21

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency ST PETERSBURG PD
Inspector HALL,DUSTIN,L

Date 01/15/2025
Time 22:27:17

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/15/2025 22:50

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency ST PETERSBURG PD
Inspector HALL,DUSTIN,L

Date 02/11/2025
Time 22:20:42

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A1
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 10/30/2026 **Expiration** 12/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/13/2025 09:34

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency MIAMI PD	Date 01/30/2025	Type Agency Inspection	
Inspector FERNANDEZ,ANGEL,	Time 19:57:48	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 402367487
Expiration 02/27/2026	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 04/28/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.191
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2025 20:03

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency MIAMI PD	Date 02/22/2025	Type Agency Inspection	
Inspector FERNANDEZ,ANGEL,	Time 00:21:17	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 402367487
Expiration 02/27/2026	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 04/28/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/22/2025 00:22

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency	ST JOHNS COUNTY SO	Date	01/31/2025	Type	Agency Inspection		
Inspector	ELLIS,ALEXANDER,	Time	19:52:19	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	402816921
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	08/09/2026
Number of Simulators	4	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/31/2025 20:21

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO
Inspector MILLS,SCOTT,T

Date 02/26/2025
Time 21:04:20

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402816921
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 08/09/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2025 21:08

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency	NORTH MIAMI PD	Date	01/20/2025	Type	Agency Inspection		
Inspector	DE JESUS,RAY,	Time	14:04:59	Software	8100.27		
0.05 Lot #	202303J	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	26824080A1
Expiration	03/28/2025	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	11/05/2026
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/20/2025 14:06

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD	Date 02/05/2025	Type Department Inspection
Inspector GUTSCHOW,TAYLOR,D	Time 13:50:05	Software 8100.27

0.05 Lot # 202303K	0.08 Lot # 202303L	0.20 Lot # 202304C	0.08 Gas Lot # AG429602
Expiration 03/29/2025	Expiration 03/29/2025	Expiration 04/05/2025	Expiration 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 02/05/2025 13:53

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD
Inspector DE JESUS, RAY,

Date 02/23/2025
Time 18:52:01

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 26824080A1
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/23/2025 18:54

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO
Inspector LYONS,DANIEL,E

Date 01/20/2025
Time 17:04:47

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403207507
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/26/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/24/2025 14:40

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO **Date** 02/04/2025 **Type** Department Inspection
Inspector ARMSTRONG, DESTINEE, N **Time** 09:52:18 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG429602
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0008 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.082
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0012
0.20 g/210L Test (Std Dev)	0.0011
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/04/2025 09:54

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY SO
Inspector LYONS,DANIEL,E

Date 02/26/2025
Time 15:44:28

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403207507
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/26/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2025 16:06

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD	Date 01/29/2025	Type Agency Inspection	
Inspector LEVY, JAYSON, I	Time 15:01:04	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 10124080A2
Expiration 03/28/2025	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 06/05/2026
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2025 17:40

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD	Date 02/07/2025	Type Agency Inspection	
Inspector LEVY, JAYSON, I	Time 16:08:11	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 10124080A2
Expiration 03/28/2025	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 06/05/2026
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/07/2025 16:49

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD	Date 01/28/2025	Type Agency Inspection	
Inspector LUTZ,EDWIN,J	Time 09:43:36	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 24923080A4
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 10/05/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/28/2025 09:44

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency	PALM BAY PD	Date	02/26/2025	Type	Agency Inspection		
Inspector	HODGDON,TODD,S	Time	11:58:19	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202410E	0.20 Lot #	202410F	0.08 Gas Lot #	24923080A4
Expiration	02/27/2026	Expiration	10/22/2026	Expiration	10/30/2026	Expiration	10/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2025 11:58

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO **Date** 01/22/2025 **Type** Agency Inspection
Inspector DUNCAN,NICHOLAS, **Time** 11:55:36 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/22/2025 11:56

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO	Date 02/19/2025	Type Agency Inspection	
Inspector DUNCAN, NICHOLAS,	Time 12:49:59	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202307C	0.08 Gas Lot # 31124080A2
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 07/12/2025	Expiration 12/05/2026
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/19/2025 12:54

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK PD **Date** 01/31/2025 **Type** Agency Inspection
Inspector ADAMS,STERLING,M **Time** 02:36:48 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28423080A4
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 01/31/2025 02:53

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency	PINELLAS PARK PD	Date	02/01/2025	Type	Agency Inspection		
Inspector	ADAMS,STERLING,M	Time	19:36:30	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	28423080A4
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	11/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 02/01/2025 19:52

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Agency	SOUTH MIAMI PD	Date	02/05/2025	Type	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	Time	11:13:20	Software	8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators	Compliance	Yes
-----------------------------	-------------------	-----

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Data Download Date/Time 02/13/2025 09:22

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Agency SOUTH MIAMI PD **Date** 02/14/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 11:57:25 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/14/2025 11:55

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 01/12/2025 **Type** Agency Inspection
Inspector CHANDLER,VANESSA,N **Time** 19:52:42 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 06723080A5
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 01/12/2025 19:50

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 02/04/2025 **Type** Agency Inspection
Inspector LEONARD,ALEYA,D **Time** 21:12:34 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 06723080A5
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.075
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 02/04/2025 21:15

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency	CLEARWATER PD	Date	01/30/2025	Type	Agency Inspection		
Inspector	HURT,JONATHAN,R	Time	19:20:23	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	14323080A3
Expiration	02/27/2025	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	09/05/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

ACETONE LOT #2023-B, M/A LOT

Data Download Date/Time 02/03/2025 23:06

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH CSO
Inspector KARLECKE,JASON,C

Date 01/17/2025
Time 14:24:51

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 14323080A3
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/05/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/17/2025 14:25

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO	Date 02/14/2025	Type Department Inspection
Inspector GUTSCHOW,TAYLOR,D	Time 12:30:27	Software 8100.27

0.05 Lot # 202303K	0.08 Lot # 202303L	0.20 Lot # 202304C	0.08 Gas Lot # AG429602
Expiration 03/29/2025	Expiration 03/29/2025	Expiration 04/05/2025	Expiration 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.196
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.052
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/14/2025 12:30

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 01/17/2025 **Type** Agency Inspection
Inspector COLEMAN, WILLIAM, **Time** 14:20:15 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403141809
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/10/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/17/2025 14:22

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO **Date** 02/07/2025 **Type** Agency Inspection
Inspector NOVAK,CHRISTOPHER, **Time** 12:44:01 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403141809
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 09/10/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/07/2025 12:45

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 01/09/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 09:39:47 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 15924080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 07/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

IN SERV @ PTDF

Data Download Date/Time 01/09/2025 09:45

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO **Date** 02/24/2025 **Type** Agency Inspection
Inspector LIGHTSEY,ROBERT,D **Time** 13:54:29 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28424080A3
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

EVED USE AT PTDF

Data Download Date/Time 02/24/2025 13:57

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency PORT ST LUCIE PD **Date** 02/18/2025 **Type** Agency Inspection
Inspector APPELBAUM,KEITH,A **Time** 11:19:55 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 403008479
Expiration 03/28/2025 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 03/28/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/18/2025 11:28

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency NAS JACKSONVILLE PD **Date** 01/15/2025 **Type** Agency Inspection
Inspector POTTER, MICHAEL, R **Time** 15:07:41 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 01923080A4
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 02/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN SERV. NAS JAX

Data Download Date/Time 01/15/2025 15:08

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency NAS JACKSONVILLE PD **Date** 02/25/2025 **Type** Agency Inspection
Inspector POTTER, MICHAEL, R **Time** 14:36:50 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 28424080A3
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

REMOVED FROM SERV FDLE D/I

Data Download Date/Time 02/25/2025 14:37

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

Agency NAS KEY WEST	Date 02/14/2025	Type Department Inspection
Inspector GUTSCHOW,TAYLOR,D	Time 13:31:17	Software 8100.27

0.05 Lot # 202303K	0.08 Lot # 202303L	0.20 Lot # 202304C	0.08 Gas Lot # AG429602
Expiration 03/29/2025	Expiration 03/29/2025	Expiration 04/05/2025	Expiration 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

--

Data Download Date/Time 02/14/2025 13:29

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 01/31/2025 **Type** Agency Inspection
Inspector HUNNEFELD,RALPH,A **Time** 13:35:46 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/31/2025 14:10

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 02/27/2025 **Type** Agency Inspection
Inspector HELMS,WILLIAM,A **Time** 17:09:32 **Software** 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 24923080A4
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 10/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2025 08:24

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency MIAMI BEACH PD **Date** 01/16/2025 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 11:26:16 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402477284
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI PRICE #958 PRESENT

Data Download Date/Time 01/16/2025 13:24

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency MIAMI BEACH PD	Date 02/20/2025	Type Agency Inspection	
Inspector SOCARRAS,PEDRO,A	Time 10:22:16	Software 8100.27	
0.05 Lot # 202303J	0.08 Lot # 202307A	0.20 Lot # 202307C	0.08 Gas Lot # 402477284
Expiration 03/28/2025	Expiration 07/11/2025	Expiration 07/12/2025	Expiration 06/27/2025
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 02/20/2025 11:16

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Agency	HIALEAH PD	Date	01/22/2025	Type	Agency Inspection		
Inspector	DELGADO,ARMANDO,	Time	10:51:23	Software	8100.27		
0.05 Lot #	202402E	0.08 Lot #	202307A	0.20 Lot #	202307C	0.08 Gas Lot #	402370890
Expiration	02/27/2026	Expiration	07/11/2025	Expiration	07/12/2025	Expiration	05/09/2025
Number of Simulators	5	Compliance	Yes				

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/22/2025 12:17

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Agency HIALEAH PD **Date** 02/10/2025 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 11:46:38 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG429602
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/22/2026

Number of Simulators 5 **Average Standard Deviation** 0.0009 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.077
		0.20 g/210L Test # 4	0.196
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.196
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.192
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0019
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/10/2025 12:31

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Agency HIALEAH PD	Date 02/14/2025	Type Department Inspection	
Inspector GUTSCHOW,TAYLOR,D	Time 11:53:35	Software 8100.27	

0.05 Lot # 202303K	0.08 Lot # 202303L	0.20 Lot # 202304C	0.08 Gas Lot # AG429602
Expiration 03/29/2025	Expiration 03/29/2025	Expiration 04/05/2025	Expiration 10/22/2026

Number of Simulators 5	Average Standard Deviation 0.0005	Compliance Yes
-------------------------------	--	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.196
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0014
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 02/14/2025 11:54

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD
Inspector PEREZ,VINCENTE,

Date 01/22/2025
Time 10:55:38

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402370890
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 05/09/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/22/2025 11:56

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD
Inspector MONTERO,JOSE,

Date 02/06/2025
Time 09:38:05

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402370890
Expiration 02/27/2026 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 05/09/2025

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/06/2025 09:45

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD
Inspector OMALLEY,BRIAN,P

Date 01/06/2025
Time 11:12:34

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403005538
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/25/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/06/2025 11:15

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD
Inspector OMALLEY,BRIAN,P

Date 02/26/2025
Time 13:58:52

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403005538
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/25/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/26/2025 14:06

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

Agency MIAMI GARDENS PD
Inspector OMALLEY,BRIAN,P

Date 01/06/2025
Time 11:16:06

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403005538
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/25/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/06/2025 11:17

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

Agency MIAMI GARDENS PD
Inspector OMALLEY,BRIAN,P

Date 02/26/2025
Time 11:12:09

Type Agency Inspection
Software 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 403005538
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 03/25/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/26/2025 11:13

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

Agency TALLAHASSEE PD
Inspector ROY,GARY,A

Date 01/02/2025
Time 20:39:47

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 12/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/02/2025 20:40

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

Agency TALLAHASSEE PD
Inspector ROY,GARY,A

Date 02/12/2025
Time 21:30:01

Type Agency Inspection
Software 8100.27

0.05 Lot # 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 31124080A2
Expiration 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 12/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/12/2025 21:30

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.