

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

**Agency** MARTIN COUNTY SO      **Date** 01/27/2025      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 21:57:30      **Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202410E      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 08223080A5  
**Expiration** 02/27/2026      **Expiration** 10/22/2026      **Expiration** 07/12/2025      **Expiration** 06/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/27/2025 21:56

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

|                             |                      |                   |            |                   |                   |                       |            |
|-----------------------------|----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | MARTIN COUNTY SO     | <b>Date</b>       | 02/03/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | THOMAS,CHRISTOPHER,L | <b>Time</b>       | 02:04:47   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E              | <b>0.08 Lot #</b> | 202410E    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 08223080A5 |
| <b>Expiration</b>           | 02/27/2026           | <b>Expiration</b> | 10/22/2026 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 06/05/2025 |
| <b>Number of Simulators</b> | 5                    | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/03/2025 02:04

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MARTIN COUNTY SO     | <b>Date</b> 02/10/2025 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:35:47   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202303K    | <b>0.08 Lot #</b> 202303L    | <b>0.20 Lot #</b> 202304C    | <b>0.08 Gas Lot #</b> AG429602 |
| <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 04/05/2025 | <b>Expiration</b> 10/22/2026   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 02/10/2025 13:34

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

**Agency** MARTIN COUNTY SO      **Date** 01/27/2025      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 16:48:09      **Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202410E      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 08223080A5  
**Expiration** 02/27/2026      **Expiration** 10/22/2026      **Expiration** 07/12/2025      **Expiration** 06/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/27/2025 16:52

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

|                             |                      |                   |            |                   |                   |                       |            |
|-----------------------------|----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | MARTIN COUNTY SO     | <b>Date</b>       | 02/02/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | THOMAS,CHRISTOPHER,L | <b>Time</b>       | 22:33:13   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E              | <b>0.08 Lot #</b> | 202410E    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 08223080A5 |
| <b>Expiration</b>           | 02/27/2026           | <b>Expiration</b> | 10/22/2026 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 06/05/2025 |
| <b>Number of Simulators</b> | 5                    | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/02/2025 22:35

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 01/31/2025  
**Time** 08:37:55

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202410E      **0.20 Lot #** 202410F      **0.08 Gas Lot #** 402806340  
**Expiration** 02/27/2026      **Expiration** 10/22/2026      **Expiration** 10/30/2026      **Expiration** 07/26/2026

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 01/31/2025 08:36

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 02/28/2025  
**Time** 18:14:42

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202410E      **0.20 Lot #** 202410F      **0.08 Gas Lot #** 402806340  
**Expiration** 02/27/2026      **Expiration** 10/22/2026      **Expiration** 10/30/2026      **Expiration** 07/26/2026

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/28/2025 18:13

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | LAKE COUNTY SO     | <b>Date</b>       | 01/16/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | DESANTIS,JEFFERY,S | <b>Time</b>       | 12:03:39   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 28423080A4 |
| <b>Expiration</b>           | 02/27/2026         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 11/05/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 01/16/2025 12:40

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

|                                   |                              |                               |                                  |
|-----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> LAKE COUNTY SO      | <b>Date</b> 02/11/2025       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> BOWDEN,MATTHEW,C | <b>Time</b> 10:56:41         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202402E         | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 28423080A4 |
| <b>Expiration</b> 02/27/2026      | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 11/05/2025     |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 02/11/2025 11:00

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> BREVARD COUNTY SO  | <b>Date</b> 01/22/2025       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> WAGNER,ROBERT,F | <b>Time</b> 09:25:42         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202402E        | <b>0.08 Lot #</b> 202410E    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 24923080A3 |
| <b>Expiration</b> 02/27/2026     | <b>Expiration</b> 10/22/2024 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 10/05/2025     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.074  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.193  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.194  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.094  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

20: COT SIM HEATER OFF.

**Data Download Date/Time** 01/22/2025 09:27

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> BREVARD COUNTY SO  | <b>Date</b> 02/24/2025       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> WAGNER,ROBERT,F | <b>Time</b> 09:48:14         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202402E        | <b>0.08 Lot #</b> 202410E    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 24923080A3 |
| <b>Expiration</b> 02/27/2026     | <b>Expiration</b> 10/22/2026 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 10/05/2025     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/24/2025 09:57

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | HENDRY COUNTY SO | <b>Date</b>       | 01/10/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | MCNEIL,DARRIN,J  | <b>Time</b>       | 20:39:35   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E          | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402872586  |
| <b>Expiration</b>           | 02/27/2026       | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 10/12/2026 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/03/2025 15:00

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

|                  |                   |             |            |                 |                   |
|------------------|-------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | HENDRY COUNTY SO  | <b>Date</b> | 02/03/2025 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | GUTSCHOW,TAYLOR,D | <b>Time</b> | 10:13:23   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

**Data Download Date/Time** 02/03/2025 15:00

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> HENDRY COUNTY SO     | <b>Date</b> 02/03/2025 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 14:59:47   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202303K    | <b>0.08 Lot #</b> 202303L    | <b>0.20 Lot #</b> 202304C    | <b>0.08 Gas Lot #</b> AG429602 |
| <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 04/05/2025 | <b>Expiration</b> 10/22/2026   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0009 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 02/03/2025 15:00

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | MANATEE COUNTY SO | <b>Date</b>       | 01/17/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | COLEMAN, WILLIAM, | <b>Time</b>       | 14:29:48   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E           | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402987895  |
| <b>Expiration</b>           | 02/27/2026        | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 03/05/2027 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/17/2025 14:30

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

|                                     |                              |                               |                                 |
|-------------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> MANATEE COUNTY SO     | <b>Date</b> 02/07/2025       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> NOVAK,CHRISTOPHER, | <b>Time</b> 09:35:51         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202402E           | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 402987895 |
| <b>Expiration</b> 02/27/2026        | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 03/05/2027    |
| <b>Number of Simulators</b> 5       | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 02/07/2025 09:37

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | CLAY COUNTY SO    | <b>Date</b>       | 01/27/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | PATRONE,RICHARD,L | <b>Time</b>       | 08:18:25   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E           | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 28423080A4 |
| <b>Expiration</b>           | 02/27/2026        | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 11/05/2025 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSPECTION

**Data Download Date/Time** 01/27/2025 08:19

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

**Agency** CLAY COUNTY SO      **Date** 02/27/2025      **Type** Agency Inspection  
**Inspector** JOHNSON,MATTHEW,P      **Time** 09:13:59      **Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 28423080A4  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 11/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSPECTION

**Data Download Date/Time** 02/27/2025 09:15

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

**Agency** PALM BEACH CSO  
**Inspector** KARLECKE,JASON,C

**Date** 01/17/2025  
**Time** 16:47:16

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 14323080A3  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 09/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/17/2025 16:48

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PALM BEACH COUNTY SO | <b>Date</b> 02/03/2025 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:34:20   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202303K    | <b>0.08 Lot #</b> 202303L    | <b>0.20 Lot #</b> 202304C    | <b>0.08 Gas Lot #</b> AG429602 |
| <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 04/05/2025 | <b>Expiration</b> 10/22/2026   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 02/03/2025 13:36

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

**Agency** PALM BEACH COUNTY SO      **Date** 02/07/2025      **Type** Agency Inspection  
**Inspector** KARLECKE,JASON,C      **Time** 16:13:47      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 10124080A2  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 06/05/2026

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/07/2025 16:14

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

|                                   |                              |                               |                                  |
|-----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> PALM BEACH CSO      | <b>Date</b> 01/17/2025       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> KARLECKE,JASON,C | <b>Time</b> 16:53:25         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202303J         | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 10124080A2 |
| <b>Expiration</b> 03/28/2025      | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 06/05/2026     |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 01/17/2025 16:55

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PALM BEACH COUNTY SO | <b>Date</b> 02/03/2025 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:25:57   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202303K    | <b>0.08 Lot #</b> 202303L    | <b>0.20 Lot #</b> 202304C    | <b>0.08 Gas Lot #</b> AG429602 |
| <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 04/05/2025 | <b>Expiration</b> 10/22/2026   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 02/03/2025 13:29

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

**Agency** PALM BEACH COUNTY SO      **Date** 02/07/2025      **Type** Agency Inspection  
**Inspector** KARLECKE,JASON,C      **Time** 16:00:35      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 10124080A2  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 06/05/2026

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/07/2025 16:00

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PALM BEACH COUNTY SO | <b>Date</b> 01/13/2025 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 12:54:22   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202303K    | <b>0.08 Lot #</b> 202303L    | <b>0.20 Lot #</b> 202304C    | <b>0.08 Gas Lot #</b> AG429602 |
| <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 04/05/2025 | <b>Expiration</b> 10/22/2026   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 01/13/2025 13:09

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

**Agency** PALM BEACH COUNTY SO      **Date** 01/17/2025      **Type** Agency Inspection  
**Inspector** KARLECKE,JASON,C      **Time** 16:56:34      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 14323080A3  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 09/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/17/2025 16:57

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

**Agency** PALM BEACH COUNTY SO      **Date** 02/07/2025      **Type** Agency Inspection  
**Inspector** KARLECKE,JASON,C      **Time** 14:14:10      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 14323080A3  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 09/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/07/2025 14:14

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 01/25/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | BLAKE,DAVID,A     | <b>Time</b>       | 21:48:03   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E           | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 3024028434 |
| <b>Expiration</b>           | 02/27/2026        | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 09/08/2026 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INTOX FREE OF CRACKS OR DAMAGE

Data Download Date/Time 01/31/2025 10:06

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL    | <b>Date</b> 01/31/2025 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:39:36   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202303K    | <b>0.08 Lot #</b> 202303L    | <b>0.20 Lot #</b> 202304C    | <b>0.08 Gas Lot #</b> AG429602 |
| <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 04/05/2025 | <b>Expiration</b> 10/22/2026   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.077  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

Remarks:

Data Download Date/Time 01/31/2025 13:41

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL         | <b>Date</b> 01/24/2025 | <b>Type</b> Department Inspection |
| <b>Inspector</b> ARMSTRONG, DESTINEE, N | <b>Time</b> 14:55:10   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202406K    | <b>0.08 Lot #</b> 202406L    | <b>0.20 Lot #</b> 202406N    | <b>0.08 Gas Lot #</b> AG429602 |
| <b>Expiration</b> 06/19/2026 | <b>Expiration</b> 06/19/2026 | <b>Expiration</b> 06/20/2026 | <b>Expiration</b> 10/22/2026   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

Remarks:

Data Download Date/Time 01/24/2025 14:55

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

**Agency** FL HIGHWAY PATROL  
**Inspector** CAPELA,JORDAN,A

**Date** 02/18/2025  
**Time** 16:04:24

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 402623828  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 10/22/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1               | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2                      | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2                      | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2               | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.079  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3               | INT    |

#### Remarks:

08: RFI Detect.

**Data Download Date/Time** 02/18/2025 16:10

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | VOLUSIA COUNTY SO  | <b>Date</b>       | 01/29/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | MALETTO,NICHOLAS,R | <b>Time</b>       | 16:27:26   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402498227  |
| <b>Expiration</b>           | 02/27/2024         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 07/22/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed.

**Data Download Date/Time** 01/29/2025 16:31

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | VOLUSIA COUNTY SO  | <b>Date</b>       | 02/27/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | MALETTO,NICHOLAS,R | <b>Time</b>       | 18:41:11   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E            | <b>0.08 Lot #</b> | 202410E    | <b>0.20 Lot #</b> | 202410F           | <b>0.08 Gas Lot #</b> | 402498227  |
| <b>Expiration</b>           | 02/27/2026         | <b>Expiration</b> | 10/22/2026 | <b>Expiration</b> | 10/30/2026        | <b>Expiration</b>     | 07/22/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/27/2025 18:43

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

**Agency** GAINESVILLE PD  
**Inspector** STEMEN, AARON, M

**Date** 01/27/2025  
**Time** 20:09:49

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 24923080A3  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 10/05/2025

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

**Data Download Date/Time** 01/27/2025 20:12

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

**Agency** GAINESVILLE PD  
**Inspector** STEMEN, AARON, M

**Date** 02/26/2025  
**Time** 18:57:45

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202410E      **0.20 Lot #** 202410F      **0.08 Gas Lot #** 24923080A3  
**Expiration** 02/27/2026      **Expiration** 10/22/2026      **Expiration** 10/30/2026      **Expiration** 10/05/2025

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN COMPLIANCE

**Data Download Date/Time** 02/26/2025 19:03

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

**Agency** HILLSBOROUGH CO SO      **Date** 01/23/2025      **Type** Agency Inspection  
**Inspector** BROWN,KEVIN,C      **Time** 08:35:50      **Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 28423080A2  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 11/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI J HAMMOND #264409 PRESENT FOR INSPECTION

**Data Download Date/Time** 01/23/2025 10:47

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> HILLSBOROUGH CSO     | <b>Date</b> 02/10/2025 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 12:17:57   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202303K    | <b>0.08 Lot #</b> 202303L    | <b>0.20 Lot #</b> 202304C    | <b>0.08 Gas Lot #</b> AG429602 |
| <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 04/05/2025 | <b>Expiration</b> 10/22/2026   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.077  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.047  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.047  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.047  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

#### Remarks:

Data Download Date/Time 02/10/2025 12:18

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

**Agency** HILLSBOROUGH CSO  
**Inspector** BROWN,KEVIN,C

**Date** 02/20/2025  
**Time** 09:11:36

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 28423080A2  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 11/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.075  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI JHAMMOND #264409 PRESENT FOR INSPECTION

**Data Download Date/Time** 02/20/2025 09:53

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | HENDRY COUNTY SO | <b>Date</b>       | 01/10/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | MCNEIL,DARRIN,J  | <b>Time</b>       | 20:31:55   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E          | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402872586  |
| <b>Expiration</b>           | 02/27/2026       | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 10/12/2026 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 01/29/2025 12:04

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

**Agency** HENDRY COUNTY SO      **Date** 01/29/2025      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D      **Time** 15:06:12      **Software** 8100.27

**0.05 Lot #** 202303K      **0.08 Lot #** 202303L      **0.20 Lot #** 202304C      **0.08 Gas Lot #** AG429602  
**Expiration** 03/29/2025      **Expiration** 03/29/2025      **Expiration** 04/05/2025      **Expiration** 10/22/2026

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 01/29/2025 15:07

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> ESCAMBIA COUNTY SO        | <b>Date</b> 02/26/2025 | <b>Type</b> Department Inspection |
| <b>Inspector</b> ARMSTRONG, DESTINEE, N | <b>Time</b> 12:34:00   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202406K    | <b>0.08 Lot #</b> 202406L    | <b>0.20 Lot #</b> 202406N    | <b>0.08 Gas Lot #</b> AG429602 |
| <b>Expiration</b> 06/19/2026 | <b>Expiration</b> 06/19/2026 | <b>Expiration</b> 06/20/2026 | <b>Expiration</b> 10/22/2026   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.077  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 02/26/2025 13:34

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | ESCAMBIA COUNTY SO | <b>Date</b>       | 01/15/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SHELLEY,SAMUEL,L   | <b>Time</b>       | 09:03:45   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E            | <b>0.08 Lot #</b> | 202410E    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402690226  |
| <b>Expiration</b>           | 02/27/2026         | <b>Expiration</b> | 10/22/2026 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 03/09/2026 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 01/15/2025 15:03

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | ESCAMBIA COUNTY SO | <b>Date</b>       | 02/05/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SHELLEY,SAMUEL,L   | <b>Time</b>       | 09:33:06   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E            | <b>0.08 Lot #</b> | 202410E    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402690226  |
| <b>Expiration</b>           | 02/27/2026         | <b>Expiration</b> | 10/22/2026 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 03/09/2026 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/05/2025 10:34

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

**Agency** MIAMI PD **Date** 01/30/2025 **Type** Agency Inspection  
**Inspector** FERNANDEZ,ANGEL, **Time** 18:49:31 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1          | INT    |
| Minimum Sample Volume Check: OK             |        | Interferent Detect Test # (Repeat) 1 | INT    |
| Barometric Pressure Sensor Check: OK        |        | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 2          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # (Repeat) 2 | INT    |
| Interferent Detect Test: Interferent Detect | No     | Alcohol Free Test # 3                | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 3          | 0.008  |
| Alcohol Free Test: OK                       | Yes    |                                      |        |
| 0.05 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Test: OK                        | No     |                                      |        |
| 0.20 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                      |        |

#### Remarks:

Int Det:CHANGE SOLUTION, Ambient Fail.REDO TESTNon-compliance:CHANGE SOLUTIONS AND CLEAR ROOM.

**Data Download Date/Time** 01/30/2025 20:39

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

|                                   |                              |                               |                                 |
|-----------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> MIAMI PD            | <b>Date</b> 01/30/2025       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> FERNANDEZ,ANGEL, | <b>Time</b> 20:35:28         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202402E         | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 402367487 |
| <b>Expiration</b> 02/27/2026      | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 04/28/2025    |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.045  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/30/2025 20:39

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

|                                   |                              |                               |                                 |
|-----------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> MIAMI PD            | <b>Date</b> 02/22/2025       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> FERNANDEZ,ANGEL, | <b>Time</b> 00:58:22         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202402E         | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 402367487 |
| <b>Expiration</b> 02/27/2026      | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 04/28/2025    |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 02/22/2025 01:05

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

**Agency** MONROE COUNTY SO  
**Inspector** GILROY,ZACHARY,I

**Date** 01/19/2025  
**Time** 10:51:40

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 402886996  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 11/01/2026

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PK INTOX

**Data Download Date/Time** 01/19/2025 10:53

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

**Agency** MONROE COUNTY SO  
**Inspector** GILROY,ZACHARY,I

**Date** 02/19/2025  
**Time** 08:11:31

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 402886996  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 11/01/2026

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PK INTOX

**Data Download Date/Time** 02/19/2025 08:34

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

**Agency** FL HIGHWAY PATROL **Date** 01/29/2025 **Type** Agency Inspection  
**Inspector** JACOBSON,KARSON,J **Time** 12:49:15 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 402843437  
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 01/30/2025 14:29

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

**Agency** FL HIGHWAY PATROL **Date** 02/28/2025 **Type** Agency Inspection  
**Inspector** JACOBSON,KARSON,J **Time** 15:50:48 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202307C **0.08 Gas Lot #** 402843437  
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 07/12/2025 **Expiration** 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 02/28/2025 17:14

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

**Agency** PINECREST PD  
**Inspector** VEGA,OBEL,

**Date** 01/22/2025  
**Time** 19:27:05

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202410E      **0.20 Lot #** 202410F      **0.08 Gas Lot #** 402310313  
**Expiration** 03/28/2025      **Expiration** 10/22/2026      **Expiration** 10/30/2026      **Expiration** 01/15/2025

**Number of Simulators** 5

**Compliance** No

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

.20 G/210L NOT WITHIN RANGE. EXPIRED GAS CYLINDER..20 G/210L NOT WITHIN RANGE. EXPIRED GAS CYLINDER..

**Data Download Date/Time** 02/25/2025 12:06

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

**Agency** PINECREST PD      **Date** 02/13/2025      **Type** Agency Inspection  
**Inspector** WADDELL,NATHANIEL,J      **Time** 00:46:11      **Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202410E      **0.20 Lot #** 202410F      **0.08 Gas Lot #** 403207507  
**Expiration** 02/27/2026      **Expiration** 10/22/2026      **Expiration** 10/30/2026      **Expiration** 11/26/2027

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | No     | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/25/2025 12:06

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

**Agency** BROWARD COUNTY SO      **Date** 01/08/2025      **Type** Agency Inspection  
**Inspector** FRAZIER,ANAYA,S      **Time** 11:08:50      **Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 402783528  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 06/27/2026

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

JGUZMAN IN TRAINING

**Data Download Date/Time** 01/08/2025 11:13

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | BROWARD COUNTY SO | <b>Date</b>       | 02/05/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | FRAZIER,ANAYA,S   | <b>Time</b>       | 09:42:56   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E           | <b>0.08 Lot #</b> | 202410E    | <b>0.20 Lot #</b> | 202410F           | <b>0.08 Gas Lot #</b> | 402783528  |
| <b>Expiration</b>           | 02/27/2026        | <b>Expiration</b> | 10/22/2026 | <b>Expiration</b> | 10/30/2026        | <b>Expiration</b>     | 06/27/2026 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

S. TILLMAN IN TRAINING.

**Data Download Date/Time** 02/05/2025 09:57

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PALM BEACH COUNTY SO | <b>Date</b> 01/13/2025 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 12:39:46   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202303K    | <b>0.08 Lot #</b> 202303L    | <b>0.20 Lot #</b> 202304C    | <b>0.08 Gas Lot #</b> AG429602 |
| <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 04/05/2025 | <b>Expiration</b> 10/22/2026   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 01/13/2025 12:44

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

**Agency** PALM BEACH COUNTY SO      **Date** 01/17/2025      **Type** Agency Inspection  
**Inspector** KARLECKE,JASON,C      **Time** 13:06:12      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 14323080A3  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 09/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/17/2025 13:27

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

**Agency** PALM BEACH COUNTY SO      **Date** 02/07/2025      **Type** Agency Inspection  
**Inspector** KARLECKE,JASON,C      **Time** 13:48:56      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 14323080A3  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 09/05/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/07/2025 13:51

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH CSO  
**Inspector** KARLECKE,JASON,C

**Date** 01/17/2025  
**Time** 15:05:54

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 10124080A2  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 06/05/2026

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/17/2025 15:09

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PALM BEACH COUNTY SO | <b>Date</b> 02/03/2025 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:08:24   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202303K    | <b>0.08 Lot #</b> 202303L    | <b>0.20 Lot #</b> 202304C    | <b>0.08 Gas Lot #</b> AG429602 |
| <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 03/29/2025 | <b>Expiration</b> 04/05/2025 | <b>Expiration</b> 10/22/2026   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 02/03/2025 13:38

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH COUNTY SO      **Date** 02/07/2025      **Type** Agency Inspection  
**Inspector** KARLECKE,JASON,C      **Time** 13:10:45      **Software** 8100.27

**0.05 Lot #** 202303J      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 10124080A2  
**Expiration** 03/28/2025      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 06/05/2026

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/07/2025 13:12

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

|                                    |                              |                               |                                 |
|------------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> MANATEE COUNTY SO    | <b>Date</b> 01/17/2025       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> COLEMAN, WILLIAM, | <b>Time</b> 09:23:38         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202402E          | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 403141809 |
| <b>Expiration</b> 02/27/2026       | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 09/10/2027    |
| <b>Number of Simulators</b> 5      | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/17/2025 09:25

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

|                                     |                              |                               |                                 |
|-------------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> MANATEE COUNTY SO     | <b>Date</b> 02/07/2025       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> NOVAK,CHRISTOPHER, | <b>Time</b> 13:01:03         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202402E           | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 403141809 |
| <b>Expiration</b> 02/27/2026        | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 09/10/2027    |
| <b>Number of Simulators</b> 5       | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/07/2025 13:02

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

|                                    |                              |                               |                                 |
|------------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> MANATEE COUNTY SO    | <b>Date</b> 01/17/2025       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> COLEMAN, WILLIAM, | <b>Time</b> 09:57:25         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202402E          | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 402357111 |
| <b>Expiration</b> 02/27/2026       | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 02/10/2025    |
| <b>Number of Simulators</b> 5      | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/17/2025 09:57

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

|                                      |                              |                               |                                 |
|--------------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> MANATEE COUNTY SO      | <b>Date</b> 02/07/2025       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> NOVAK, CHRISTOPHER, | <b>Time</b> 09:22:12         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202402E            | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 402357111 |
| <b>Expiration</b> 02/27/2026         | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 02/10/2025    |
| <b>Number of Simulators</b> 5        | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 02/07/2025 09:25

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

**Agency** HILLSBOROUGH CO SO  
**Inspector** BROWN,KEVIN,C

**Date** 01/23/2025  
**Time** 08:45:20

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 28423080A2  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 11/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI J HAMMOND #264409 PRESENT FOR INSPECTION

**Data Download Date/Time** 01/23/2025 10:37

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

**Agency** HILLSBOROUGH CSO      **Date** 02/10/2025      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D      **Time** 12:18:21      **Software** 8100.27

**0.05 Lot #** 202303K      **0.08 Lot #** 202303L      **0.20 Lot #** 202304C      **0.08 Gas Lot #** AG429602  
**Expiration** 03/29/2025      **Expiration** 03/29/2025      **Expiration** 04/05/2025      **Expiration** 10/22/2026

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.194  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.194  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 02/10/2025 12:32

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

**Agency** HILLSBOROUGH CSO  
**Inspector** BROWN,KEVIN,C

**Date** 02/20/2025  
**Time** 09:03:59

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 28423080A2  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 11/05/2025

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI JHAMMOND #264409 PRESENT FOR INSPECTION

**Data Download Date/Time** 02/20/2025 09:46

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FLAGLER COUNTY SO  | <b>Date</b>       | 01/14/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | FAIRCLOTH,EDWARD,W | <b>Time</b>       | 09:40:31   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E            | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 24923080A2 |
| <b>Expiration</b>           | 02/27/2026         | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 10/05/2025 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

MONTHLY INSPECTION

**Data Download Date/Time** 01/14/2025 09:41

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

**Agency** FLAGLER COUNTY SO      **Date** 01/29/2025      **Type** Department Inspection  
**Inspector** ARMSTRONG, DESTINEE, N      **Time** 11:47:27      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG429602  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 10/22/2026

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 01/29/2025 11:46

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

|                             |                      |                   |            |                   |                   |                       |            |
|-----------------------------|----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | KENNEDY SPACE CENTER | <b>Date</b>       | 01/23/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | MILLER,JEFFREY,A     | <b>Time</b>       | 15:13:28   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E              | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 14323080A2 |
| <b>Expiration</b>           | 02/27/2026           | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 06/05/2025 |
| <b>Number of Simulators</b> | 5                    | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.194  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.195  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

20: RFI Detect.

**Data Download Date/Time** 01/23/2025 15:15

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

|                             |                      |                   |            |                   |                   |                       |            |
|-----------------------------|----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | KENNEDY SPACE CENTER | <b>Date</b>       | 02/27/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SAVARD,CHARLES,M     | <b>Time</b>       | 08:14:34   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E              | <b>0.08 Lot #</b> | 202410E    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 14323080A2 |
| <b>Expiration</b>           | 02/27/2026           | <b>Expiration</b> | 10/22/2026 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 06/05/2025 |
| <b>Number of Simulators</b> | 5                    | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/27/2025 08:14

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | BOCA RATON PD    | <b>Date</b>       | 01/22/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | CRAWFORD,ANDREW, | <b>Time</b>       | 16:05:47   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E          | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 08223080A3 |
| <b>Expiration</b>           | 02/27/2026       | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 06/05/2025 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

PASS

**Data Download Date/Time** 01/22/2025 16:05

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | BOCA RATON PD    | <b>Date</b>       | 02/20/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | CRAWFORD,ANDREW, | <b>Time</b>       | 10:34:01   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E          | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 08223080A3 |
| <b>Expiration</b>           | 02/27/2026       | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 06/05/2025 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

PASS

**Data Download Date/Time** 02/20/2025 10:33

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

|                                 |                              |                               |                                 |
|---------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 01/25/2025       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> KEMP,TREVOR,D  | <b>Time</b> 20:16:42         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202402E       | <b>0.08 Lot #</b> 202410E    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 403141809 |
| <b>Expiration</b> 02/27/2026    | <b>Expiration</b> 10/22/2026 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 09/10/2027    |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NA

**Data Download Date/Time** 01/25/2025 20:31

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 02/24/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | KEMP,TREVOR,D     | <b>Time</b>       | 20:16:41   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E           | <b>0.08 Lot #</b> | 202410E    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 403141809  |
| <b>Expiration</b>           | 02/27/2026        | <b>Expiration</b> | 10/22/2026 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 09/10/2027 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

N/A

**Data Download Date/Time** 02/24/2025 21:14

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

**Agency** FL HIGHWAY PATROL **Date** 01/03/2025 **Type** Agency Inspection  
**Inspector** ARMSTRONG,DESTINEE,N **Time** 12:28:12 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASS AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

**Data Download Date/Time** 01/08/2025 12:36

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

**Agency** FL HIGHWAY PATROL      **Date** 01/08/2025      **Type** Department Inspection  
**Inspector** ARMSTRONG, DESTINEE, N      **Time** 14:11:14      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG429602  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 10/22/2026

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.082  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.082  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.052  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0007 |
| 0.08 g/210L Test (Std Dev)             | 0.0007 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 01/08/2025 14:10

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL | <b>Date</b>       | 02/26/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | DRACH,STEFAN,     | <b>Time</b>       | 09:09:46   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E           | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202410F           | <b>0.08 Gas Lot #</b> | 402397742  |
| <b>Expiration</b>           | 02/27/2026        | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 10/30/2026        | <b>Expiration</b>     | 03/05/2025 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

UNIT CLEAN / FUNCTIONAL AT TIME OF INSPECTION

Data Download Date/Time 02/26/2025 10:37

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

|                                 |                              |                               |                                 |
|---------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> FL HIGHWAY PATROL | <b>Date</b> 01/25/2025       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> KEMP,TREVOR,D  | <b>Time</b> 20:15:57         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202402E       | <b>0.08 Lot #</b> 202410E    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 403135736 |
| <b>Expiration</b> 02/27/2026    | <b>Expiration</b> 10/22/2026 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 08/29/2027    |
| <b>Number of Simulators</b> 5   | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NA

**Data Download Date/Time** 01/25/2025 20:26

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

**Agency** FL HIGHWAY PATROL  
**Inspector** KEMP,TREVOR,D

**Date** 02/24/2025  
**Time** 20:29:38

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202410E      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 403135736  
**Expiration** 02/27/2026      **Expiration** 10/22/2026      **Expiration** 07/12/2025      **Expiration** 08/29/2027

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 02/24/2025 20:43

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

**Agency** MANATEE COUNTY SO      **Date** 01/17/2025      **Type** Agency Inspection  
**Inspector** COLEMAN, WILLIAM,      **Time** 14:24:28      **Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 403056709  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 05/28/2027

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.084  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.084  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.084  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/17/2025 14:25

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

|                                      |                              |                               |                                 |
|--------------------------------------|------------------------------|-------------------------------|---------------------------------|
| <b>Agency</b> MANATEE COUNTY SO      | <b>Date</b> 02/07/2025       | <b>Type</b> Agency Inspection |                                 |
| <b>Inspector</b> NOVAK, CHRISTOPHER, | <b>Time</b> 09:26:46         | <b>Software</b> 8100.27       |                                 |
| <b>0.05 Lot #</b> 202402E            | <b>0.08 Lot #</b> 202307A    | <b>0.20 Lot #</b> 202307C     | <b>0.08 Gas Lot #</b> 403056709 |
| <b>Expiration</b> 02/27/2026         | <b>Expiration</b> 07/11/2025 | <b>Expiration</b> 07/12/2025  | <b>Expiration</b> 05/28/2027    |
| <b>Number of Simulators</b> 5        | <b>Compliance</b> Yes        |                               |                                 |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.084  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.084  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.084  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/07/2025 09:31

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

|                             |                        |                   |            |                   |                   |                       |            |
|-----------------------------|------------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL      | <b>Date</b>       | 01/20/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | GONZALEZ,CHRISTOPHER,M | <b>Time</b>       | 22:08:29   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E                | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402872586  |
| <b>Expiration</b>           | 02/27/2026             | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 10/12/2026 |
| <b>Number of Simulators</b> | 5                      | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CASE CHECKED OK

**Data Download Date/Time** 01/20/2025 22:08

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

|                             |                       |                   |            |                   |                   |                       |            |
|-----------------------------|-----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL     | <b>Date</b>       | 02/17/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | DOBOSIEWICZ,ANTHONY,N | <b>Time</b>       | 20:29:00   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E               | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 402872586  |
| <b>Expiration</b>           | 02/27/2026            | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 10/12/2026 |
| <b>Number of Simulators</b> | 5                     | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CASE CHECKED - NO CRACKS- IN COMPLIANCE

**Data Download Date/Time** 02/17/2025 20:40

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

|                             |                        |                   |            |                   |                   |                       |            |
|-----------------------------|------------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | FL HIGHWAY PATROL      | <b>Date</b>       | 01/20/2025 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | GONZALEZ,CHRISTOPHER,M | <b>Time</b>       | 21:41:32   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202402E                | <b>0.08 Lot #</b> | 202307A    | <b>0.20 Lot #</b> | 202307C           | <b>0.08 Gas Lot #</b> | 2402498227 |
| <b>Expiration</b>           | 02/27/2026             | <b>Expiration</b> | 07/11/2025 | <b>Expiration</b> | 07/12/2025        | <b>Expiration</b>     | 07/22/2025 |
| <b>Number of Simulators</b> | 5                      | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

CASE CHECKED OK

Data Download Date/Time 01/20/2025 22:01

Printed: 11 March 2025

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

**Agency** FL HIGHWAY PATROL      **Date** 02/17/2025      **Type** Agency Inspection  
**Inspector** DOBOSIEWICZ, ANTHONY, N      **Time** 20:21:32      **Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 2402498227  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 07/22/2025

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed.

**Data Download Date/Time** 02/17/2025 20:36

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

**Agency** MANATEE COUNTY SO      **Date** 01/17/2025      **Type** Agency Inspection  
**Inspector** COLEMAN, WILLIAM,      **Time** 09:37:40      **Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 403141809  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 09/10/2027

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/17/2025 09:36

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

**Agency** MANATEE COUNTY SO      **Date** 02/07/2025      **Type** Agency Inspection  
**Inspector** NOVAK,CHRISTOPHER,      **Time** 12:54:51      **Software** 8100.27

**0.05 Lot #** 202402E      **0.08 Lot #** 202307A      **0.20 Lot #** 202307C      **0.08 Gas Lot #** 403141809  
**Expiration** 02/27/2026      **Expiration** 07/11/2025      **Expiration** 07/12/2025      **Expiration** 09/10/2027

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/07/2025 12:55

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

**Agency** FL HIGHWAY PATROL **Date** 01/28/2025 **Type** Agency Inspection  
**Inspector** HITCHCOCK, WILLIAM, D **Time** 01:54:19 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402357111  
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/10/2025

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/02/2025 06:23

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

**Agency** FL HIGHWAY PATROL **Date** 02/28/2025 **Type** Agency Inspection  
**Inspector** HITCHCOCK, WILLIAM, D **Time** 13:40:27 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507  
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/02/2025 06:23

**Printed: 11 March 2025**

COBRA ©1997-2014 CMI, Inc.