

RECEIVED

Florida Department of Law Enforcement FEB 02 2017
Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FT WALTON BEACH PD
Time of Inspection: 02:41

Date of Inspection: 02/01/2017

Serial Number: 80-000738
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK	Yes	No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/25/2018	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.048	0.000 / INT		
0.000	0.048	0.000		
0.000	0.048	0.000		

Number of Simulators Used: 5

Remarks:

INSPECTOR INJURED DURING AGENCY INSPECTION.

UNABLE TO COMPLETE INSPECTION DUE TO
FLOODED SAMPLE CHAMBER.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ROBERT S HULL
Signature and Printed Name

02/01/2017
Date