

RECEIVED

NOV 30 2017

FDLE  
Alcohol Testing Program

Florida Department of Law Enforcement  
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COLUMBIA COUNTY SO  
Time of Inspection: 02:28

Date of Inspection: 11/28/2017

Serial Number: 80-000775  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:09216080A2 Exp: 05/05/2018
0.000	0.049	0.074 / 0.076	0.193	0.079
0.000	0.049	0.075 / 0.076	0.195	0.079
0.000	0.049	0.076 / 0.077	0.195	0.078

Number of Simulators Used: 5


Remarks:

08: Control Outside Tolerance.

*Solution not warm enough*

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 BRETT L SULLIVAN  
Signature and Printed Name

11/28/2017  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

SEP 19 2017

FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Alcohol Testing Program

Agency: COLUMBIA COUNTY SO  
Time of Inspection: 08:26

Date of Inspection: 09/19/2017

Serial Number: 80-000775  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: 5


Remarks:  
A F / M A: Sequence Aborted BYPASS TO OPERATE INSTRUMENT.

*\* Compliance Not determined*

*JSAM*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name

JAKE L SHANAHAN

09/19/2017  
Date

*9/19/17  
JS*