

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 20:52

Date of Inspection: 09/17/2017

Serial Number: 80-000786
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | | No |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34015080A3 Exp: 01/05/2018 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.198 | 0.079 |
| 0.000 | 0.050 | 0.080 | 0.199 | 0.079 |
| 0.000 | 0.050 | 0.080 | 0.200 | 0.079 |


Number of Simulators Used: 5

Remarks:

A F / M A: TOO MUCH MOUTH ALC. FRAZIER OBS/PRESENT, MOUTH ALCOHOL DETECTED ON 2ND ATTEMPT, SECOND INSPECTION CONDUCTED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #16806
 YVETTE D WASHINGTON
 Signature and Printed Name

09/17/2017
 Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 22:29

Date of Inspection: 09/17/2017

Serial Number: 80-000786
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#:34015080A3 Exp: 01/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34015080A3 Exp: 01/05/2018 |
|----------------------------|---|---|--|---|
| 0.000 | 0.050 | 0.079 | 0.000 / 0.198 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.000 / 0.200 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.000 / 0.200 | 0.079 |


Number of Simulators Used: 5

Remarks:

20: ConTRL OUTSIDE TOL, MISTAKENLY USE GAS INSTEAD OF .200 **FRAZER PRESENT & OBSERVED.**

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #14806
Signature and Printed Name

09/17/2017
Date