

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JAN 09 2018
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Vero Beach Police Department

Serial Number: 80-000790

Time of Inspection: 9:00 p.m. Date of Inspection: 11/08/17

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		X
Diagnostic Check (Pre-Inspection): OK		X
Alcohol Free Subject Test: 0.000		X
Mouth Alcohol Test: Slope Not Met		X
Interferent Detect Test: Interferent Detect		X
Diagnostic Check (Post-Inspection): OK		X


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A

Number of Simulators Used: N/A

Remarks: I attempted an agency inspection on the above instrument, when the instrument started smoking and then lost power. It happened during the warm up process and it smelled like wires burning. Unable to complete the monthly agency inspection. The instrument was taken off line and sent out for diagnosis and repair.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Sergeant Philip Huddy #494

Signature and Printed Name

11/08/17

Date