



RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

MAY 23 2017

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO  
Time of Inspection: 15:15

Date of Inspection: 04/27/2017

Serial Number: 80-000795  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: <del>03/08/2013</del> 03/08/2018 QN. 5-277	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 19615080A2 Exp: 09/05/2017
0.000	0.051	0.084	0.204	0.078
0.000	0.051	0.084	0.203	0.078
0.000	0.051	0.083	0.202	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11B-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11B-8, FAC.

James A. Newsome Signature and Printed Name JAMES A. NEWSOME

04/27/2017  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

MAY 08 2017

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

**FDLE**  
Alcohol Testing Program

Agency: GULF COUNTY SO  
Time of Inspection: 15:15

Date of Inspection: 04/27/2017

Serial Number: 80-000795  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: <del>03/08/2013</del> <i>03/08/2018 QN.5-277</i>	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 19615080A2 Exp: 09/05/2017
0.000	0.051	0.084	0.204	0.078
0.000	0.051	0.084	0.203	0.078
0.000	0.051	0.083	0.202	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*James A. Newsome*

Signature and Printed Name JAMES A NEWSOME

04/27/2017  
Date

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

APR 03 2017

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO  
Time of Inspection: 14:44

Date of Inspection: 03/30/2017

Serial Number: 80-000795  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19615080A2 Exp: 09/05/2017
0.000	0.050	0.081	0.203	0.078
0.000	0.051	0.081	0.204	0.077
0.000	0.051	0.082	0.205	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*James A. Newsome*

JAMES A NEWSOME  
Signature and Printed Name

03/30/2017  
Date