

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

DEC 05 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HAMILTON COUNTY SO
Time of Inspection: 01:17

Date of Inspection: 11/28/2017

Serial Number: 80-000797
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:14816080A3 Exp: 07/05/2018
0.000	0.049	0.078	0.195	0.079
0.000	0.049	0.079	0.196	0.079
0.000	0.049	0.079	0.197	0.079

Number of Simulators Used: 4

Remarks:

CORRECTED .08 SOLUTION INFO LOT#201611B EXP DATE 11/15/2018

Sending To FDLE For Yearly Inspection

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

William C Curry

WILLIAM C CURRY

Signature and Printed Name

11/28/2017
Date