

RECEIVED

AUG 17 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 18:40

Date of Inspection: 08/14/2017

Serial Number: 80-000808
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 12716080A7 Exp: 06/05/2018
0.000	0.050	0.081	0.199	0.081
0.000	0.051	0.081	0.199	0.080
0.000	0.051	0.082	0.200	0.080

Number of Simulators Used: 5

Remarks:

AI C. KIRK #135602 present entire machine for SP 246109

(Handwritten initials)

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

(Handwritten signature)

ROGER G SKIPPER

Signature and Printed Name

08/14/2017
Date

*Reviewed
8-16-17
M. St. J. B. [initials]*