

80-000228  
80-000812  
80-000814

**Bell, Danielle**

80-005249

**From:** Collito, Brad <bcollito@hernandosheriff.org>  
**Sent:** Thursday, May 11, 2017 10:00 PM  
**To:** Bell, Danielle; Cameron, John IV; Conrad Juergensmeyer; Hutchinson, Harold; Ivette Miranda; Bell, Danielle; Page, Mike; Paula Tyrell; Sonya Wolthuis; Stevens, Teresa  
**Subject:** May 2017 Hernando County Sheriff's Office Agency Inspections  
**Attachments:** May 2017 Agency Inspections.pdf

Good Evening,

Attached are the May 2017 Agency Inspections for the Hernando County Sheriff's Office.

Instrument 814 and 5249 were not inspected due to both of them being at CMI for servicing. Per CMI, our instruments are scheduled to be looked at on 5/19/2017.

Instrument 228 is an Evidential Loaner instrument that we have received from FDLE while 814 and 5249 are being serviced.

Instrument 812 and 228 were inspected with no issues to report. Both instruments passed inspection.

If I can be of additional assistance, please do not hesitate to contact me.

Respectfully,  
Sgt. B. Collito

**Sergeant Brad Collito**  
**Intake/ Release Unit**  
**Hernando County Detention Center**  
**16425 Spring Hill Dr**  
**Brooksville, FL 34604**  
**T(352)544-2334 F(352)544-2349**  
**bcollito@hernandosheriff.org**

Please note: Florida has a very broad Public Records Law. Most written communications to or from the Hernando County Sheriff's Office are public records available to anyone upon request. Your e-mail, including your e-mail address, may therefore, be subject to public disclosure.

80-000814

RECEIVED

FEB 15 2017

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 12:27

Date of Inspection: 02/15/2017

Serial Number: 80-000814  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2018
0.000	0.048	0.077	0.197	0.081
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Signature* BR 1193

BRAD R COLLITO

Signature and Printed Name

02/15/2017  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
JAN 19 2017  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 08:49

Date of Inspection: 01/19/2017

Serial Number: 80-000814  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201501F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28515080A2 Exp: 11/05/2017
0.000	0.048	0.078	0.196	0.078
0.000	0.048	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Sgt B Collito*

BRAD R COLLITO

Signature and Printed Name

01/19/2017  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
JAN 19 2017

**FDLE**  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.  
Time of Inspection: 08:06

Date of Inspection: 01/19/2017

Serial Number: 80-000814  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
Non compliance: EEPROM FAILURE.

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Sgt B C 192*

BRAD R COLLITO

Signature and Printed Name

01/19/2017  
Date