

80-000815 *Q18*

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

APR 27 2017

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO  
Time of Inspection: 09:50

Date of Inspection: 04/27/2017

Serial Number: 80-000815  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/18/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:200160800A Exp: 09/05/2018
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.199	0.079

Number of Simulators Used: 5

Remarks:

A F / M A: RFI Detect.

*Amended: OFFICER STEPPED INSIDE DUI ROOM AS THE TEST WAS BEING STARTED, WITH RADIO ON.*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gary B Forinash*  
Signature and Printed Name

GARY B FORINASH

04/27/2017  
Date