

80-000825

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

AUG 03 2017

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

FDLE

Agency: MOUNT DORA P.D.  
Time of Inspection: 21:47

Date of Inspection: 07/30/2017

Serial Number: 80-000825  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:583457 Exp: 10/20/2019
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.079	0.199	0.080
0.000	0.049	0.079	0.200	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KEITH A TAYLOR

Signature and Printed Name

07/30/2017  
Date