

8318

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUN 21 2017

FDLE

AGENCY INSPECTION REPORT – INTOXILYZER 8000 Alcohol Testing Program

Agency: MARTIN COUNTY SO
Time of Inspection: 22:28

Date of Inspection: 06/16/2017

Serial Number: 80-000831
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A2 Exp: 10/05/2017
0.000	0.048	0.077	0.195	0.078
0.000	0.048	0.079	0.195	0.077
0.000	0.048	0.078	0.195	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

CHRISTOPHER L THOMAS

06/16/2017
 Date