

RECEIVED
APR 03 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 10:14

Date of Inspection: 03/31/2017

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2018
0.000	0.047	0.078	0.193	0.079
0.000	0.048	0.079	0.194	0.079
0.000	0.048	0.079	0.195	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M. Bowden 1100

SANDI W CHESSHER
Signature and Printed Name

S. Chessher

03/31/2017
Date

80-000836 Qrs

RECEIVED

FEB 27 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 09:10

Date of Inspection: 02/27/2017

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2018
0.000	0.048	0.080	0.196	0.080
0.000	0.048	0.080	0.197	0.080
0.000	0.049	0.080	0.197	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW BOWEN


SANDRA W CHESSHER 
Signature and Printed Name

02/27/2017
Date

80-000836

RECEIVED
JAN 30 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD (LAKE SO)
Time of Inspection: 09:35

Date of Inspection: 01/30/2017

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 12716080A7 Exp: 06/05/2018
0.000	0.048	0.078	0.195	0.079
0.000	0.049	0.079	0.196	0.079
0.000	0.049	0.080	0.196	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW BOWDEN
M/BOWDEN 1100

SIGNATURE AND PRINTED NAME

SANDRA W CHESSHER *[Signature]* 950

01/30/2017
Date