

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

APR 28 2017

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO  
Time of Inspection: 11:46

Date of Inspection: 04/28/2017

Serial Number: 80-000867  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 279870 Exp: 02/04/2019
0.000	0.048	0.078	0.188 / 0.192	0.081
0.000	0.048	0.078	0.189 / 0.192	0.081
0.000	0.048	0.079	0.190 / 0.193	0.081

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance.

HEATING ELEMENT WAS TURNED OFF ACCIDENTALLY. TURNED ON HEATING ELEMENT LET IT RUN, MADE SURE SIMULATOR WAS SEALED AND CONNECTIONS WERE TIGHT. RE-RAN 0.20 TEST AND HAD NO PROBLEM.

*J. Huff* 8568/774

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*J. Huff* 8568/774

JON B HUFF

Signature and Printed Name

04/28/2017  
Date