

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 17:23

Date of Inspection: 07/13/2017

Serial Number: 80-000870
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:14968 Exp: 08/06/2018
/ 0.000	0.048	0.080	0.201	0.082
/ 0.000	0.049	0.080	0.201	0.081
/ 0.000	0.049	0.080	0.202	0.081

Number of Simulators Used: 5


Remarks:

00: Ambient Fail.

ALCOHOL FREE TEST: DGS WAS STILL CONNECTED TO CALIBRATION INLET, CAUSING AMBIENT FAIL. DISCONNECTED THE DGS AND CONNECTED SIMULATOR TO CALIBRATION INLET. RE-RAN AFT AND NO ISSUE. *JLB*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


8568/174
JON B HUFF

Signature and Printed Name

07/13/2017
Date