



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CORAL GABLES PD  
Time of Inspection: 13:28

Date of Inspection: 12/04/2017

Serial Number: 80-000871  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: 4

**Remarks:**

A P / M A: . Non-compliance: Mouth Alcohol Registered AT . 239 and did not read slope not met  
  
Reason is unknown

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_ DANIEL A SMITH  
 Signature and Printed Name

12/04/2017  
 Date

RECEIVED

FEB 14 2017

FDLE

Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CORAL GABLES PD  
Time of Inspection: 11:37

Date of Inspection: 01/26/2017

Serial Number: 80-000871  
Software: 810C.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2016	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201504C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 26316C90A1 Exp: 11/05/2015
/ 0.000	0.050	0.081	0.196	0.081
/ 0.000	0.050	0.080	0.197	0.081
/ 0.000	0.051	0.081	0.197	0.080

Number of Simulators Used: 4

Remarks:

00: Ambient Fail.

ONLY ONE TEST WAS COMPLETED FOR THE ALCOHOL FREE.  
 PRIOR TO TESTING THE INSTRUMENT COMPLETED AN AMBIENT FAIL  
 FOR REASONS UNKNOWN. AFTER CLEARING ITSELF THE ALCOHOL FREE  
 TEST WAS COMPLETED AND PASSED O.K.

The above instrument complies (  ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIEL A SMITH

Signature and Printed Name

01/26/2017  
Date