

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

JAN 10 2017

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

*BK*

Agency: PINELLAS COUNTY SO  
Time of Inspection: 12:24

Date of Inspection: 01/06/2017

Serial Number: 80-000889  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

*N/A Compliance Not Determined*

*SP*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Danielle M Bell*

DANIELLE M BELL

Signature and Printed Name

01/06/2017  
Date