

80-000901 *WB*

RECEIVED  
DEC 23 2017

Florida Department of Law Enforcement  
Alcohol Testing Program

FDLE  
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EFWCC TAMPA

Time of Inspection: 22:54

Date of Inspection: 12/14/2017

Serial Number: 80-000901

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No
		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.05 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____
0.000				
0.000				
0.000				

Number of Analyzers Used: \_\_\_\_\_

Remarks:

The instrument was not able to clear the sample chamber after a breath sample analysis due to an excess of mouth alcohol. The inspection was started and in final inspection was in compliance.

The above instrument complies ( ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Signature]*  
Signature and Printed Name

Date

80-000901 DMS

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

JUN 23 2017

FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FFWCC TAMPA

Time of Inspection: 23:19

Date of Inspection: 06/16/2017

Serial Number: 80-000901

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:20016080A1 Exp: 09/05/2018
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.051	0.080	0.198	0.079

Number of Simulators Used: 5

Remarks:  
IN COMPLIANCE

REMINDER:

DRY GAS SENSOR IS INOPERABLE.

THANK YOU FOR ALL THAT YOU DO, DANIELLE!

*Damon*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DAMON J PULASKI  
Signature and Printed Name

06/16/2017  
Date