

80-000902 *FAS*

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAR 24 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC TAMPA
Time of Inspection: 21:31

Date of Inspection: 03/24/2017

Serial Number: 80-000902
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:
PASSED A/F TOO MUCH M/A. PUR+PURGE FAIL AFTER M/A SUSPECTED ROOM AIR CONTAMINATED BY DRY GAS STANDARD - (LEAK) FDLE NOTIFIED. INSPECTION WAS ABORTED. A SUBSEQUENT INSPECTION WAS COMPLETED AND IN COMPLIANCE.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 DAMON J PULASKI
Signature and Printed Name

03/24/2017
Date

80-000902

RECEIVED

FEB 14 2017

FDLE Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC TAMPA
Time of Inspection: 23:31

Date of Inspection: 02/13/2017

Serial Number: 80-000902
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:20016080A1 Exp: 09/05/2018
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.051	0.080	0.199	0.080

Number of Simulators Used: 5

Remarks:
IN COMPLIANCE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DAMON J PULASKI

Signature and Printed Name

02/13/2017
Date

80-0090 RECEIVED

FEB 14 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC TAMPA
Time of Inspection: 22:36

Date of Inspection: 02/13/2017

Serial Number: 80-000902
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded. EXCEEDED AMT RESIDUAL LEFT OVER. Non-compliance: TOO MUCH ALCOHOL/RESIDUAL ALCOHOL. THE INSTRUMENT PASSED THE ALCOHOL FREE SUBJECT TEST. MOUTH ALCOHOL; TOO MUCH ALCOHOL IN MOUTH; THEREFORE, TEST RESULTED AS RANGE EXCEEDED. THE FOLLOWING TESTS RESULTED ~~AND~~ AS RESIDUAL ALCOHOL IN MOUTH. I ABORTED THE INSPECTION AND RESTARTED. THE FOLLOWING AGENCY INSPECTION ON THE INSTRUMENT WAS IN COMPLIANCE.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DAMON J PULASKI

Signature and Printed Name

02/13/2017
Date