

Florida Department of Law Enforcement Alcohol Testing Program

SP939 RECEIVED

FEB 27 2017

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: LEE COUNTY SO
Time of Inspection: 07:56

Date of Inspection: 02/22/2017

Serial Number: 80-000939
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.049			
0.000	0.050			
0.000	0.050			

Number of Simulators Used: 5

Remarks:

Instrument went into Final Stage After .05 Test Instead of Continuing Inspection. Instrument was not in Compliance, form changed Below.

[Signature]
98049

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] ROMAN SERRANO
Signature and Printed Name

02/22/2017
Date

COPY