

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUL 27 2017

1043
SP

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 15:03

Date of Inspection: 07/24/2017

Serial Number: 80-001043
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 24515080A3 Exp: 10/05/2017
0.000	0.046	0.076	0.193	0.078
0.000	0.046	0.077	0.195	0.078
0.000	0.047	0.077	0.195	0.078

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY _____ D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 Signature and Printed Name

07/24/2017
Date

GREGORY S GLASCO

1043
SP

RECEIVED

Florida Department of Law Enforcement JUN 01 2017

Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 08:51

Date of Inspection: 05/28/2017

Serial Number: 80-001043
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A6 Exp: 06/05/2018
0.000	0.049	0.077	0.196	0.077
0.000	0.048	0.078	0.197	0.077
0.000	0.048	0.077	0.197	0.077

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184
Signature and Printed Name

SHON W ABELN

05/28/2017
Date

1043 SP
RECEIVED
 MAY 04 2017
 FDLE
 Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: **HIGHLANDS COUNTY SO** Date of Inspection: **04/23/2017** Serial Number: **80-001043**
 Time of Inspection: **11:03** Software: **8100.27**

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716090A6 Exp: 06/05/2018
0.000	0.049	0.079	0.200	0.076
0.000	0.049	0.079	0.199	0.076
0.000	0.049	0.078	0.199	0.077

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
 BY Shon W. Abeln **2184 D.S.**

The above instrument complies () does not comply () with Chapter 11D-8, FAC.
 I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln **2184** SHON W ABELN
 Signature and Printed Name
 04/23/2017
 Date