

1044
B

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

MAR 03 2017

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD
Time of Inspection: 11:24

Date of Inspection: 02/21/2017

Serial Number: 80-001044
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1694353 Exp: 06/10/2017
0.000	0.047	0.077	0.196	0.080
0.000	0.048	0.076	0.196	0.080
0.000	0.048	0.077	0.195	0.081

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

KYLE DUFFEY

Signature and Printed Name

02/21/2017
Date

1044
R

RECEIVED

FEB 28 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD
Time of Inspection: 10:38

Date of Inspection: 02/13/2017

Serial Number: 80-001044
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1694353 Exp: 06/10/2017
0.000	0.048	0.074 / 0.075	0.196	0.079
0.000	0.049	0.075 / 0.076	0.196	0.079
0.000	0.048	0.076 / 0.076	0.195	0.079

Number of Simulators Used: 4

Remarks:
08: Control Outside Tolerance. POSS AIR LEAK AT INTAKE TUBE



The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Kyle Duffey
#17979

Signature and Printed Name

KYLE DUFFEY

02/13/2017
Date

1044

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JAN 30 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD
Time of Inspection: 11:02

Date of Inspection: 01/24/2017

Serial Number: 80-001044
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1694353 Exp: 06/10/2017
0.000	0.048	0.075	0.195	0.080
0.000	0.048	0.076	0.194	0.080
0.000	0.049	0.076	0.195	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

KYLE DUFFEY

Signature and Printed Name

01/24/2017
Date