

1046 (7)

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

JAN 26 2017

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO  
Time of Inspection: 14:24

Date of Inspection: 01/21/2017

Serial Number: 80-001046  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:2016080A1 Exp: 09/05/2018
0.000	0.050	0.079	0.200	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.080	0.200	0.081


Number of Simulators Used: 5

Remarks:  
MONTHLY INSPECTION

*RETURN TO FILE.  
LEAKING, 80 GAS*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
7124 RICHARD L PATRONE  
Signature and Printed Name

01/21/2017  
Date