

RECEIVED

OCT 17 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD
Time of Inspection: 11:22

Date of Inspection: 10/17/2017

Serial Number: 80-001048
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 34015080A2 Exp: 01/05/2018
0.000	0.048	0.075	0.201	0.080
0.000	0.048	0.078	0.201	0.080
0.000	0.049	0.078	0.201	0.080

Number of Simulators Used: 3

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Anthony W Gettle

Signature and Printed Name

10/17/2017
Date

RECEIVED

OCT 17 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD

Time of Inspection: 11:57

Date of Inspection: 10/16/2017

Serial Number: 80-001048

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.047	0.076	0.184 / 0.187	
0.000	0.047	0.076	0.186 / 0.188	
0.000	0.047	0.077	0.187 / 0.190	

Number of Simulators Used: 3

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

- NOTIFIED FDLE 10/16/17 11:55AM
- FAILURE WAS DETERMINED TO BE THE SIMULATOR'S HEATING ELEMENT AND WAS SENT TO BE REPAIRED.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ANTHONY W GETTLE

Signature and Printed Name

10/16/2017
Date

80-001048 *WB*

RECEIVED

OCT 17 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD

Time of Inspection: 11:57

Date of Inspection: 10/16/2017

Serial Number: 80-001048

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.047	0.076	0.184 / 0.187	
0.000	0.047	0.076	0.186 / 0.188	
0.000	0.047	0.077	0.187 / 0.190	

Number of Simulators Used: 3

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

- NOTIFIED FDLE 10/16/17 11:55AM

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ANTHONY W GETTLE

Signature and Printed Name

10/16/2017

Date