

**Bell, Danielle**

---

**From:** Michael D. Weiskopf <Michael.Weiskopf@stpete.org>  
**Sent:** Saturday, January 14, 2017 11:51 PM  
**To:** Bell, Danielle  
**Subject:** Intoxilyzer Unit 80-001051

I am sending you Intoxilyzer Unit 80-001051 to you for its Department inspection. I have completed the Agency inspection on it with no issues. We discovered the 12 volt plug is not powering the unit, however, when other instruments are plugged in to the same power supply there are no issues. Please advise on your findings.

Thanks

Officer Michael Weiskopf  
DUI Unit/Agency Inspector  
St Petersburg Police Department  
1300 1<sup>st</sup> Avenue North  
St Petersburg, FL 33705  
Office: 727-893-7099  
Audix: 727-551-3108

Your Sunshine City