

1059
②

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 03 2017

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

FDLE

Agency: FERNANDINA BEACH PD
Time of Inspection: 10:34

Date of Inspection: 04/11/2017

Serial Number: 80-001059
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:15615080A3 Exp: 07/05/2017
0.000	0.051	0.082	0.200	0.077
0.000	0.052	0.082	0.202	0.077
0.000	0.051	0.082	0.202	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jason Smith

JASON E SMITH

Signature and Printed Name

04/11/2017
Date