

RECEIVED

FEB 06 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WINTER HAVEN PD
Time of Inspection: 11:23

Date of Inspection: 01/30/2017

Serial Number: 80-001060
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2018
0.000	0.043 / 0.048	0.082	0.202	0.082
0.000	0.046 / 0.049	0.082	0.204	0.082
0.000	0.047 / 0.049	0.082	0.205	0.081

Number of Simulators Used: 3

Remarks:

05: Control Outside Tolerance.

* WAITED (5) five minutes and restarted TEST when the instrument asked.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Det. Robert E. Bonsall #450 Signature and Printed Name

ROBERT E BONSALL

01/30/2017
Date

80-001060

RECEIVED

FEB 06 2017

FDLE Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WINTER HAVEN PD
Time of Inspection: 09:05

Date of Inspection: 01/30/2017

Serial Number: 80-001060
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: : Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: : Exp:
0.000	0.050	0.081		
0.000	0.050	0.081		
0.000	0.050	0.082		

Number of Simulators Used: 3

Remarks: *Accidentally HIT (N) for next test. Conducted re-inspection.*

28 1/31/17

The above instrument complies (~~X~~) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Robert E. Bonsall #450
Signature and Printed Name ROBERT E BONSTALL

01/30/2017
Date

80-001060

RECEIVED
FEB 03 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WINTER HAVEN PD
Time of Inspection: 09:05

Date of Inspection: 01/30/2017

Serial Number: 80-001060
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.050	0.081		
0.000	0.050	0.081		
0.000	0.050	0.082		

Number of Simulators Used: 3

Remarks:

Accidentally HIT (N) See NEXT TEST. Conductor RE-INSPECTION.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Robert E. Bonsall #450

ROBERT E BONSTALL

Signature and Printed Name

01/30/2017
Date