

RECEIVED

NOV 13 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 14:30

Date of Inspection: 11/06/2017

Serial Number: 80-001067
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK	Yes	No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34015080A3 Exp: 01/05/2018
0.000	0.049	0.079	0.198	0.074 / 0.075
0.000	0.049	0.079	0.198	0.074 / 0.075
0.000	0.049	0.079	0.198	0.075 / 0.075

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded. 08: Control Outside Tolerance.

*A Frazier observed inspections
Too much Alcohol for Range exceeded*

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

YVETTE D WASHINGTON

Signature and Printed Name

11/06/2017
Date