

1069

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 16 2017

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 13:01

Date of Inspection: 06/15/2017

Serial Number: 80-001069
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17316080A3 Exp: 06/05/2018
0.000	0.051	0.083	0.202	0.079
0.000	0.052	0.082	0.202	0.079
0.000	0.051	0.082	0.201	0.079

Number of Simulators Used: 5

Remarks:

AMMENDMENT:

M 1599

Removed from service for annual preventive maintenance and Department Inspection 06/15/2017

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] M 1599 MICHAEL R. POTTER
Signature and Printed Name

06/15/2017
Date