

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
MAY 09 2017  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK COUNTY SO  
Time of Inspection: 14:37

Date of Inspection: 05/09/2017

Serial Number: 80-001080  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:20016080AJ Exp: 09/05/2018
0.000	0.046	0.078	0.195 / 0.200	0.080 / 0.081
0.000	0.047	0.079	0.197 / 0.200	0.081 / 0.081
0.000	0.047	0.080	/ 0.200	RFI / 0.081

Number of Simulators Used: 5

Remarks:  
20: RFI Detect. 08: RFI Detect.DOOR TO TESTING AREA OPENED

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 → 6493

BARBIE R WILSON

---

Signature and Printed Name

05/09/2017  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED  
MAR 06 2017

Alcohol Testing Program  
FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK COUNTY SO  
Time of Inspection: 13:13

Date of Inspection: 03/06/2017

Serial Number: 80-001080  
Software: 8100.27

Check or Test	YES	NO
Data and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.047	0.081	0.198	0.077
0.000	0.047	0.081	0.198	0.077
0.000	0.047	0.081	0.199	0.077

Number of Simulators Used: 5

Remarks:  
A E / N A: Range Exceeded.

used too much mouth alcohol *RW* 3/6/17

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*RW* → 6493

BARBIE R WILSON

Signature and Printed Name

03/06/2017  
Date

80-001080

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

FEB 09 2017

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK COUNTY SO  
Time of Inspection: 10:40

Date of Inspection: 02/09/2017

Serial Number: 80-001080  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.048	0.079	0.194	0.076
0.000	0.049	0.078	0.194	0.076
0.000	0.048	0.079	0.193	0.076

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 10493

BARBIE R WILSON

Signature and Printed Name

02/09/2017  
Date