

1083
②

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

SEP 13 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 14:10

Date of Inspection: 09/12/2017

Serial Number: 80-001083
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:26316080A2 Exp: 11/05/2018
0.000	0.047	0.079	0.198	0.078
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.079	0.198	0.078

Number of Simulators Used: 5

Remarks: * AMENDMENT *

PLACED INTO SERVO. 9.12.2017
W 7599

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Michael R Potter W 7599 MICHAEL R POTTER
Signature and Printed Name

09/12/2017
Date

1083
②

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 10:43

Date of Inspection: 02/24/2017

Serial Number: 80-001083
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17316080A3 Exp: 08/05/2018
0.000	0.047	0.077	0.201	0.079
0.000	0.048	0.078	0.202	0.078
0.000	0.048	0.078	0.202	0.079

Number of Simulators Used: 5

Remarks:
A F / M A: No Sample Provided.

AMENDMENT:

I DID NOT PROVIDE A BREATHER SAMPLE IN TIME AND THE SEQUENCE TIMED OUT. RE-RAN THIS PORTION OF THE INSPECTION. *[Signature]* W 7599

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] W 7599

MICHAEL R POTTER

Signature and Printed Name

02/24/2017
Date