









RECEIVED

Florida Department of Law Enforcement AUG 29 2017  
Alcohol Testing Program

FDLE  
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TYNDALL AFB  
Time of Inspection: 14:20

Date of Inspection: 06/03/2017

Serial Number: 80-001230  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A5 Exp: 06/05/2018
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.081	0.200	0.079

Number of Simulators Used: 5

Remarks:  
N/A

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name

SUSAN M BARGE

06/03/2017  
Date

RECEIVED

Florida Department of Law Enforcement <sup>JUL 26 2017</sup>  
Alcohol Testing Program

FDLE  
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TYNDALL AFB

Time of Inspection: 17:38

Date of Inspection: 07/24/2017

Serial Number: 80-001230

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.050	0.081	0.201	0.079
0.000	0.050	0.081	0.201	0.079
0.000	0.051	0.082	0.201	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR C LANTIS

Signature and Printed Name

07/24/2017  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TYNDALL AFB

Serial Number: 80-001230

Time of Inspection: 14:20

Date of Inspection: 06/03/2017

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A5 Exp: 06/05/2018
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.081	0.200	0.079

Number of Simulators Used: 5

Remarks:

N/A

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 Signature and Printed Name

SUSAN M BARGE

06/03/2017  
 Date

RECEIVED

APR 10 2017

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TYNDALL AFB

Time of Inspection: 21:44

Date of Inspection: 04/03/2017

Serial Number: 80-001230

Software: 8100.27

heck or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A5 Exp: 06/05/2018
0.000	0.048	0.079	0.202	0.080
0.000	0.049	0.079	0.202	0.080
0.000	0.049	0.079	0.202	0.080

Number of Simulators Used: 5

Remarks:  
NA

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name

04/03/2017  
Date