

RECEIVED

APR 03 2017

Florida Department of Law Enforcement
Alcohol Testing Program

1243
Revised 3/28/17
B JH 643

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EDGEWATER P.D.
Time of Inspection: 02:26

Date of Inspection: 03/29/2017

Serial Number: 80-001243
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1789170 Exp: 01/14/2018
0.000	0.048	0.076	0.196	0.074 / 0.074
0.000	0.049	0.077	0.197	0.075 / 0.075
0.000	0.049	0.078	0.197	0.075 / 0.075

Number of Simulators Used: 2

Remarks:
08: Control Outside Tolerance, Control Outside Tolerance.

The above instrument complies does not comply with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brian W Gintz

BRIAN W GINTZ

Signature and Printed Name

03/29/2017
Date