

Florida Department of Law Enforcement Alcohol Testing Program

COPY

RECEIVED

JUN 06 2017

AGENCY INSPECTION REPORT - INTOXILYZER 8000

FDLE

Agency: ROCKLEDGE P.D.

Time of Inspection: 11:17

Date of Inspection: 05/29/2017

Serial Number: 80-001261
Alcohol Testing Program

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 572920 Exp: 09/13/2019
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.080	0.198	0.079


Number of Simulators Used: 4

Remarks:

Time-Date changed.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 Signature and Printed Name PATRICK W KENNEDY

05/29/2017
 Date

1261

RECEIVED

Florida Department of Law Enforcement MAR 03 2017

Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.

Time of Inspection: 08:04

Date of Inspection: 02/27/2017

Serial Number: 80-001261

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.049	0.079	0.189 / 0.194	0.080
0.000	0.049	0.079	0.195 / 0.198	0.079
0.000	0.050	0.080	0.192 / 0.195	0.079

Number of Simulators Used: 4

Remarks:

20: Control Outside Tolerance: TIGHTENED SEAL ON SIMULATOR

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name IVETTE GOMEZ

02/27/2017
Date

RECEIVED

JAN 10 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 19:09

Date of Inspection: 01/03/2017

Serial Number: 80-001261
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.080	0.199	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

IVETTE GOMEZ #180
Signature and Printed Name

01/03/2017
Date