

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
File SP

JAN 12 2018

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 11:29

Date of Inspection: 12/26/2017

Serial Number: 80-001261
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201703A Exp: 03/07/2019 | 0.08g/210L Test (g/210L) Lot#: 201709A Exp: 09/12/2019 | 0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 572920 Exp: 09/13/2019 |
|----------------------------|--|--|--|--|
| 0.000 | 0.049 | 0.080 | 0.196 | 0.079 |
| 0.000 | 0.049 | 0.079 | 0.196 | 0.079 |
| 0.000 | 0.049 | 0.079 | 0.196 | 0.079 |

Number of Simulators Used: 4

Remarks:

USED NEW SOLUTIONS, CHECKED SIMULATORS FOR AIR LEAKS. COMPLIANCE GAINED DURING THIS INSPECTION.
PWK #185 12/26/2017

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SG I. Patrick W. Kennedy PATRICK W KENNEDY
Signature and Printed Name

12/26/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED *201*

JAN 12 2018

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 10:23

Date of Inspection: 12/26/2017

Serial Number: 80-001261
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |
| | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019 | 0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____ |
|----------------------------|---|---|---|--|
| 0.000 | 0.050 | 0.077 | 0.188 / 0.178 | |
| 0.000 | 0.050 | 0.078 | 0.201 / 0.185 | |
| 0.000 | 0.049 | 0.078 | 0.189 / 0.188 | |

Number of Simulators Used: 4

Remarks:
20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick W Kennedy
Signature and Printed Name

12/26/2017
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

DEC 04 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 11:24

Date of Inspection: 11/28/2017

Serial Number: 80-001261
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | Yes | |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019 | 0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.078 | 0.194 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.195 | 0.080 |
| 0.000 | 0.049 | 0.077 | 0.195 | 0.080 |

Number of Simulators Used: 4

Remarks:
Time-Date changed.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 PATRICK W KENNEDY
Signature and Printed Name

11/28/2017
Date

1201 SP

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

OCT 31 2017

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

At Alcohol Testing Program

Agency: ROCKLEDGE P.D.

Time of Inspection: 02:05

Date of Inspection: 10/26/2017

Serial Number: 80-001261

Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | |
| Diagnostic Check (Pre-Inspection): OK | | No |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |
| | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019 | 0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.080 | 0.198 | 0.080 |
| 0.000 | 0.050 | 0.080 | 0.199 | 0.080 |
| 0.000 | 0.051 | 0.080 | 0.199 | 0.080 |

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick W Kennedy #185 PATRICK W KENNEDY
Signature and Printed Name

10/26/2017
Date

1261 SF RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

OCT 20 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.

Time of Inspection: 07:43

Date of Inspection: 09/30/2017

Serial Number: 80-001261

Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019 | 0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.080 | 0.199 | 0.079 |
| 0.000 | 0.049 | 0.080 | 0.200 | 0.079 |
| 0.000 | 0.050 | 0.080 | 0.200 | 0.079 |

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature #170

IVETTE GOMEZ

Signature and Printed Name

09/30/2017
Date

RECEIVED

AUG 04 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 03:56

Date of Inspection: 07/28/2017

Serial Number: 80-001261
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019 | 0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019 |
|----------------------------|---|---|---|---|
| 0.000 | 0.050 | 0.079 | 0.199 | 0.078 |
| 0.000 | 0.050 | 0.080 | 0.200 | 0.079 |
| 0.000 | 0.050 | 0.081 | 0.200 | 0.079 |

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt Ivette Gomez #180

IVETTE GOMEZ

Signature and Printed Name

07/28/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUL 07 2017

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 08:35

Date of Inspection: 06/26/2017

Serial Number: 80-001261
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.197 | 0.079 |
| 0.000 | 0.049 | 0.080 | 0.198 | 0.079 |
| 0.000 | 0.049 | 0.080 | 0.199 | 0.080 |

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SGT Patrick W. Kennedy #185 PATRICK W KENNEDY
Signature and Printed Name

06/26/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

COPY

RECEIVED

JUN 06 2017

AGENCY INSPECTION REPORT - INTOXILYZER 8000

FDLE

Agency: ROCKLEDGE P.D.

Time of Inspection: 11:17

Date of Inspection: 05/29/2017

Serial Number: 80-001261
Software: 8100.27

Alcohol Testing Program

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | Yes | |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 572920 Exp: 09/13/2019 |
|----------------------------|--|--|--|--|
| 0.000 | 0.050 | 0.080 | 0.197 | 0.080 |
| 0.000 | 0.050 | 0.080 | 0.197 | 0.080 |
| 0.000 | 0.050 | 0.080 | 0.198 | 0.079 |

Number of Simulators Used: 4

Remarks:
Time-Date changed.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 Signature and Printed Name PATRICK W KENNEDY

05/29/2017
 Date

1261

RECEIVED

Florida Department of Law Enforcement MAR 03 2017

Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.

Time of Inspection: 08:04

Date of Inspection: 02/27/2017

Serial Number: 80-001261

Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.189 / 0.194 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.195 / 0.198 | 0.079 |
| 0.000 | 0.050 | 0.080 | 0.192 / 0.195 | 0.079 |

Number of Simulators Used: 4

Remarks:

20: Control Outside Tolerance: TIGHTENED SEAL ON SIMULATOR

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name IVETTE GOMEZ

02/27/2017
Date

RECEIVED

JAN 10 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ROCKLEDGE P.D.
Time of Inspection: 19:09

Date of Inspection: 01/03/2017

Serial Number: 80-001261
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:572920 Exp: 09/13/2019 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.198 | 0.079 |
| 0.000 | 0.050 | 0.080 | 0.199 | 0.079 |
| 0.000 | 0.050 | 0.080 | 0.199 | 0.079 |

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #180 IVETTE GOMEZ
Signature and Printed Name

01/03/2017
Date