

1266 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.
Time of Inspection: 02:20

Date of Inspection: 04/23/2017

Serial Number: 80-001266
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 24515080A2 Exp: 10/05/2017
0.000	0.046	0.078	0.196	0.080
0.000	0.046	0.078	0.197	0.079
0.000	0.047	0.079	0.197	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ

Signature and Printed Name

04/23/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

1266
SP

RECEIVED

MAR 28 2017

AGENCY INSPECTION REPORT - INTOXILYZER 8000

FDLE
Alcohol Testing Program

Agency: PALM BAY P.D.
Time of Inspection: 19:39

Date of Inspection: 03/27/2017

Serial Number: 80-001266
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A2 Exp: 10/05/2017
0.000	0.048	0.078	0.194	0.079
0.000	0.048	0.078	0.193	0.079
0.000	0.048	0.078	0.194	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



EDWIN J LUTZ

Signature and Printed Name

03/27/2017
Date

SP 1264

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.
Time of Inspection: 12:18

Date of Inspection: 02/26/2017

Serial Number: 80-001266
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 24515080A2 Exp: 10/05/2017
0.000	0.048	0.078	0.193	0.079
0.000	0.048	0.079	0.195	0.079
0.000	0.049	0.079	0.195	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


EDWIN J LUTZ
Signature and Printed Name

02/26/2017
Date

D. Morgan

PALM BAY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001009
02/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:16
Control Test	0.190	11:17
Air Blank	0.000	11:17
Control Test	0.191	11:18
Air Blank	0.000	11:18
Control Test Stats		
Average	0.1905	
Std Dev	0.0007	
Rel Std Dev(%)	0.3712	




Operator's Signature

SP1266

PALM BAY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001266
02/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:47
Control Test	0.047	10:47
Air Blank	0.000	10:48
Control Test	0.047	10:49
Air Blank	0.000	10:49
Control Test Stats		
Average	0.0470	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

PALM BAY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001721
02/26/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:41
Control Test	0.076	10:42
Air Blank	0.000	10:42
Control Test	0.077	10:43
Air Blank	0.000	10:43
Control Test Stats		
Average	0.0765	
Std Dev	0.0007	
Rel Std Dev(%)	0.9243	



Operator's Signature