

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

FEB 14 2017

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 10:41

Date of Inspection: ^{12/06/2016}~~10/06/2016~~ (PE) Serial Number: 80-001276
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22315080A2 Exp: 09/05/2017
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.201	0.078
0.000	0.049	0.078	0.200	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PABLO ESPINOZA

Signature and Printed Name

^{12/6/2016}~~10/06/2016~~ (PE)
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

FEB 14 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 06:00

Date of Inspection: ~~10/06/2016~~ ^{11/24/2016} PE

Serial Number: 80-001276
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22315080A2 Exp: 09/05/2017
0.000	0.049	0.077	0.198	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.048	0.078	0.199	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name PABLO ESPINOZA

Date ~~10/06/2016~~ ^{11/24/2016} PE