

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

FEB 10 2017

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Alcohol Testing Program

Agency: BRADFORD COUNTY SO
Time of Inspection: 08:48

Date of Inspection: 02/08/2017

Serial Number: 80-001290
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
BYPASS TO OPERATE INSTRUMENT


BK

[Handwritten initials]

* Compliance Not Determined

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name JAKE L SHANAHAN

02/08/2017
Date