

1340
SR

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY SO
Time of Inspection: 21:09

Date of Inspection: 09/22/2017

Serial Number: 80-001340
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.043 / 0.045	0.074 / 0.075	0.189 / 0.188	
0.000	0.043 / 0.045	0.074 / 0.075	0.189 / 0.189	
0.000	0.043 / 0.045	0.074 / 0.076	0.189 / 0.189	

Number of Simulators Used: 4

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance.20: CONTROL OUTSIDE TOLERANCENon-compliance:CONTROL TEST .05 .08 & .20 OUTSIDE TOLERANCE.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JOSHUA E PITTS
Signature and Printed Name

09/22/2017
Date