

80-001367

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

FEB 07 2017

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: PINELLAS COUNTY SO
Time of Inspection: 10:26

Date of Inspection: 02/03/2017

Serial Number: 80-001367
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG622905 Exp: 08/16/2018
0.000	0.048	0.077	0.191	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.050	0.080	0.201	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M DEANE

Signature and Printed Name

02/03/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

FEB 07 2017

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: PINELLAS COUNTY SO
Time of Inspection: 00:48

Date of Inspection: 02/03/2017

Serial Number: 80-001367
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG622905 Exp: 08/16/2018
0.000	0.050	0.078	0.194	0.081
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.197	0.081

Number of Simulators Used: 5

Remarks:

A F / M A: TOO MUCH MOUTH ALC. IN MOUTH.

The above instrument complies (^{MD}) does not comply (^{MD}) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M. Deane
M DEANE
Signature and Printed Name

02/03/2017
Date