

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

OCT 26 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: Orange County Sheriff's Office

Serial Number: 80-001418

Time of Inspection: N/A

Date of Inspection: 10/19/2016

Software:

CHECK OR TEST	YES	NO
Date and/or Time Adjusted	N/A	N/A
Diagnostic Check (Pre-Inspection): OK	N/A	N/A
Alcohol Free Subject Test: 0.000	N/A	N/A
Mouth Alcohol Test: Slope Not Met	N/A	N/A
Interferent Detect Test: Interferent Detect	N/A	N/A
Diagnostic Check (Post-Inspection): OK	N/A	N/A

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A


Number of Simulators Used: N/A

Remarks: Unable to perform an Agency Inspection due to instrument unable to clear (DISABLED MODE). I notified Department Inspector Jake Shanahan via telephone. I am sending the instrument to FDLE per his request.

(The information entered on this Agency Inspection Report was hand-typed by Agency Inspector K. Melville)

The above instrument complies (N/A) does not comply (N/A) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit ~~and that I performed this inspection~~ in accordance with the provisions of Chapter 11D-8, FAC.

OK


 Signature and Printed Name Kelly M. Melville

10/19/2016

Date