

RECEIVED

JUL 05 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BISCAYNE NAT PARK
Time of Inspection: 14:55

Date of Inspection: 06/22/2017

Serial Number: 80-001442
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:20016080A2 Exp: 09/05/2018
0.000	0.049	0.080	0.196	0.077
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.078

Number of Simulators Used: 5

Remarks:

Int Det: RECONNECT SIMULATOR.

The simulator to Intoxilyzer connection was problematic. I changed simulators, reheated solution, and ran test.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Kevin Millan MBPD #141

KEVIN MILLAN

Signature and Printed Name

06/22/2017
Date