



80-001865

RECEIVED  
MAR 07 2017

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 09:04

Date of Inspection: 03/07/2017

Serial Number: 80-001865  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.044 / 0.046	0.075	0.183 / 0.170	
0.000	0.045 / 0.046	0.075	0.187 / 0.179	
0.000	0.046 / 0.047	0.076	RFI / 0.181	

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance. 20: RFI Detect, Control Outside Tolerance. Non-compliance: .

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L. FRANZ  
Signature and Printed Name

03/07/2017  
Date

80-001865 RAS

RECEIVED

FEB 28 2017

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 11:33

Date of Inspection: 02/27/2017


Serial Number: 80-001865  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No *
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22315080A2 Exp: 09/05/2017
0.000	0.047	0.078	0.195	0.077
0.000	0.048	0.078	0.198	0.077
0.000	0.049	0.079	0.197	0.077

Number of Simulators Used: 5

Remarks:  
A F / M A :

\* Residual Mouth Alcohol PRESENT, TEST REPEATED! 

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GORDON L FRANZ

Signature and Printed Name

02/27/2017  
Date

80-001865

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED  
JAN 25 2017  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 10:47

Date of Inspection: 01/24/2017

Serial Number: 80-001865  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22315080A2 Exp: 09/05/2017
0.000	0.047	0.076	0.193	0.078
0.000	0.047	0.076	0.195	0.078
0.000	0.047	0.077	0.196	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L FRANZ  
Signature and Printed Name

01/24/2017  
Date